


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200112-126
Instrument name:	Phadia-B
Instrument type:	Phadia 250
Status:	Ready
Started:	12-Jan-20 7:05:51 PM
Ended:	12-Jan-20 7:33:58 PM
Approved:	Approved (RJB)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-B
Instrument type:	Phadia 250
Measured:	23-Dec-19 3:51:37 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	18	0.00			
		3:51:37 PM	19	0.00	OK		
		3:52:36 PM	18	0.00	OK		
		%CV	4.5	>99			
CAL-0.35	0.35	Mean	147	0.35			
		3:53:37 PM	146	0.35	OK		
		3:54:36 PM	148	0.35	OK		
		%CV	0.6	0.7			
CAL-0.70	0.70	Mean	271	0.69			
		3:55:37 PM	274	0.70	OK		
		3:56:36 PM	268	0.68	OK		
		%CV	1.6	1.7			
CAL-3.50	3.50	Mean	1279	3.51			
		3:57:37 PM	1283	3.52	OK		
		3:58:36 PM	1275	3.50	OK		
		%CV	0.5	0.5			
CAL-17.5	17.5	Mean	5583	17.5			
		3:59:37 PM	5591	17.5	OK		
		4:00:36 PM	5575	17.5	OK		
		%CV	0.2	0.2			
CAL-100	100	Mean	17182	100			
		4:01:37 PM	17224	101	OK		
		4:02:36 PM	17140	99.4	OK		
		%CV	0.3	0.9			

Curve data	Value	Logged mean	Difference
ED-20	10.1	9.67	+3.97
ED-50	30.1	29.6	+1.76
ED-80	62.3	62.2	+0.20
Slope	211	216	-2.31

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

No quality controls in this run

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000336	phad	5	7:05:51 PM	36	0.25 PAU/I	---	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000335	fx2	5	7:12:51 PM	21	0.04 kUA/I	---	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000335	fx22	5	7:13:51 PM	30	0.16 kUA/l	---	Negative	
011990000335	wx1	5	7:14:51 PM	184	2.27 kUA/l	---	Positive	
011990000335	t20	5	7:15:51 PM	24	0.08 kUA/l	---		
011990000336	fx5	5	7:06:51 PM	25	0.09 kUA/l	---	Negative	
011990000336	fx1	5	7:07:51 PM	20	0.03 kUA/l	---	Negative	
011990000336	fx2	5	7:08:51 PM	14	---	---	Negative	
011990000336	fx22	5	7:09:51 PM	28	0.14 kUA/l	---	Negative	
011990000336	wx1	5	7:10:51 PM	33	0.21 kUA/l	---	Negative	
011990000336	t20	5	7:11:51 PM	18	0.01 kUA/l	---		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	3:51:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	3:52:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:53:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:54:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:55:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:56:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:57:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:58:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	3:59:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	4:00:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:01:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:02:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000335	fx2	7:12:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000335	fx22	7:13:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000335	wx1	7:14:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000335	t20	7:15:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000336	phad	7:05:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000336	fx5	7:06:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000336	fx1	7:07:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000336	fx2	7:08:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000336	fx22	7:09:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000336	wx1	7:10:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000336	t20	7:11:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	