


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200121-383  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 21-Jan-20 2:06:42 PM  
Ended: 21-Jan-20 5:58:36 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

### Curve Controls

No curve controls in this run

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	3:16:39 PM	6164	17.0 kUA/l	<lab>	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000151	phad	5	3:02:40 PM	37	0.24 PAU/l	0	Negative	
011990000153	phad	5	3:19:39 PM	30	0.15 PAU/l	0	Negative	
011990000154	phad	5	2:54:40 PM	1397	16.3 PAU/l	3	Positive	
011990000158	phad	5	4:38:36 PM	126	1.28 PAU/l	2	Positive	
011990000161	phad	5	2:06:42 PM	87	0.82 PAU/l	2	Positive	
011990000164	phad	5	2:38:41 PM	100	0.97 PAU/l	2	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000167	phad	5	5:10:35 PM	32	0.18 PAU/I	0	Negative	
011990000168	phad	5	4:46:36 PM	33	0.19 PAU/I	0	Negative	
011990000170	phad	5	3:43:38 PM	47	0.36 PAU/I	1	Positive	
011990000171	phad	5	2:46:41 PM	34	0.20 PAU/I	0	Negative	
011990000175	phad	5	2:22:41 PM	45	0.33 PAU/I	0	Negative	
011990000176	phad	5	2:30:41 PM	45	0.33 PAU/I	0	Negative	
011990000179	phad	5	4:11:37 PM	57	0.47 PAU/I	1	Positive	
011990000181	phad	5	2:14:42 PM	34	0.21 PAU/I	0	Negative	
011990000182	phad	5	3:10:39 PM	259	2.82 PAU/I	2	Positive	
011990000185	phad	5	3:27:39 PM	131	1.33 PAU/I	2	Positive	
011990000186	phad	5	4:54:35 PM	539	6.07 PAU/I	3	Positive	
011990000188	phad	5	5:02:35 PM	40	0.28 PAU/I	0	Negative	
011990000189	phad	5	4:30:36 PM	28	0.13 PAU/I	0	Negative	
011990000193	phad	5	3:35:38 PM	42	0.30 PAU/I	0	Negative	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000151	fx5	5	3:03:40 PM	25	0.10 kUA/I	0	Negative	
011990000151	fx1	5	3:04:40 PM	30	0.16 kUA/I	0	Negative	
011990000151	fx2	5	3:05:40 PM	43	0.31 kUA/I	0	Negative	
011990000151	fx22	5	3:06:40 PM	32	0.18 kUA/I	0	Negative	
011990000151	gx2	5	3:07:40 PM	27	0.12 kUA/I	0	Negative	
011990000151	wx1	5	3:08:40 PM	27	0.13 kUA/I	0	Negative	
011990000151	t20	5	3:09:40 PM	18	0.01 kUA/I	0		
011990000153	fx5	5	3:20:39 PM	43	0.31 kUA/I	0	Negative	
011990000153	fx1	5	3:21:39 PM	19	0.02 kUA/I	0	Negative	
011990000153	fx2	5	3:22:39 PM	15	Low RU	0	Negative	
011990000153	fx22	5	3:23:39 PM	29	0.15 kUA/I	0	Negative	
011990000153	gx2	5	3:24:39 PM	20	0.04 kUA/I	0	Negative	
011990000153	wx1	5	3:25:39 PM	27	0.12 kUA/I	0	Negative	
011990000153	t20	5	3:26:39 PM	16	Low RU	0		
011990000154	fx5	5	2:55:40 PM	458	5.13 kUA/I	3	Positive	
011990000154	fx1	5	2:56:40 PM	661	7.50 kUA/I	3	Positive	
011990000154	fx2	5	2:57:40 PM	1890	22.5 kUA/I	4	Positive	
011990000154	fx22	5	2:58:40 PM	761	8.68 kUA/I	3	Positive	
011990000154	gx2	5	2:59:40 PM	354	3.92 kUA/I	3	Positive	
011990000154	wx1	5	3:00:40 PM	631	7.15 kUA/I	3	Positive	
011990000154	t20	5	3:01:40 PM	650	7.37 kUA/I	3		
011990000158	fx5	5	4:39:36 PM	21	0.05 kUA/I	0	Negative	
011990000158	fx1	5	4:40:36 PM	21	0.05 kUA/I	0	Negative	
011990000158	fx2	5	4:41:36 PM	78	0.71 kUA/I	2	Positive	
011990000158	fx22	5	4:42:36 PM	34	0.20 kUA/I	0	Negative	
011990000158	gx2	5	4:43:36 PM	26	0.11 kUA/I	0	Negative	
011990000158	wx1	5	4:44:36 PM	44	0.32 kUA/I	0	Negative	
011990000158	t20	5	4:45:36 PM	25	0.10 kUA/I	0		
011990000161	fx5	5	2:07:42 PM	26	0.11 kUA/I	0	Negative	
011990000161	fx1	5	2:08:42 PM	17	Low RU	0	Negative	
011990000161	fx2	5	2:09:42 PM	46	0.34 kUA/I	0	Negative	
011990000161	fx22	5	2:10:42 PM	32	0.18 kUA/I	0	Negative	
011990000161	gx2	5	2:11:42 PM	25	0.10 kUA/I	0	Negative	
011990000161	wx1	5	2:12:42 PM	39	0.26 kUA/I	0	Negative	
011990000161	t20	5	2:13:42 PM	38	0.25 kUA/I	0		

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000164	fx5	5	2:39:41 PM	21	0.05 kUA/l	0	Negative	
011990000164	fx1	5	2:40:41 PM	17	Low RU	0	Negative	
011990000164	fx2	5	2:41:41 PM	17	Low RU	0	Negative	
011990000164	fx22	5	2:42:41 PM	33	0.19 kUA/l	0	Negative	
011990000164	gx2	5	2:43:40 PM	20	0.03 kUA/l	0	Negative	
011990000164	wx1	5	2:44:41 PM	30	0.15 kUA/l	0	Negative	
011990000164	t20	5	2:45:40 PM	17	Low RU	0		
011990000167	fx5	5	5:11:35 PM	19	0.03 kUA/l	0	Negative	
011990000167	fx1	5	5:12:35 PM	15	Low RU	0	Negative	
011990000167	fx2	5	5:13:35 PM	15	Low RU	0	Negative	
011990000167	fx22	5	5:14:35 PM	34	0.20 kUA/l	0	Negative	
011990000167	gx2	5	5:15:35 PM	20	0.04 kUA/l	0	Negative	
011990000167	wx1	5	5:16:35 PM	39	0.26 kUA/l	0	Negative	
011990000167	t20	5	5:17:35 PM	16	Low RU	0		
011990000168	fx5	5	4:47:36 PM	22	0.06 kUA/l	0	Negative	
011990000168	fx1	5	4:48:36 PM	16	Low RU	0	Negative	
011990000168	fx2	5	4:49:36 PM	13	Low RU	0	Negative	
011990000168	fx22	5	4:50:36 PM	25	0.10 kUA/l	0	Negative	
011990000168	gx2	5	4:51:36 PM	36	0.23 kUA/l	0	Negative	
011990000168	wx1	5	4:52:36 PM	30	0.16 kUA/l	0	Negative	
011990000168	t20	5	4:53:35 PM	15	Low RU	0		
011990000170	fx5	5	3:44:38 PM	29	0.14 kUA/l	0	Negative	
011990000170	fx1	5	3:45:38 PM	27	0.12 kUA/l	0	Negative	
011990000170	fx2	5	3:46:38 PM	13	Low RU	0	Negative	
011990000170	fx22	5	3:47:38 PM	51	0.41 kUA/l	1	Positive	
011990000170	gx2	5	3:48:38 PM	25	0.09 kUA/l	0	Negative	
011990000170	wx1	5	3:49:38 PM	192	2.04 kUA/l	2	Positive	
011990000170	t20	5	3:50:38 PM	79	0.73 kUA/l	2		
011990000171	fx5	5	2:47:41 PM	21	0.05 kUA/l	0	Negative	
011990000171	fx1	5	2:48:40 PM	21	0.05 kUA/l	0	Negative	
011990000171	fx2	5	2:49:40 PM	14	Low RU	0	Negative	
011990000171	fx22	5	2:50:40 PM	34	0.20 kUA/l	0	Negative	
011990000171	gx2	5	2:51:40 PM	25	0.10 kUA/l	0	Negative	
011990000171	wx1	5	2:52:40 PM	30	0.16 kUA/l	0	Negative	
011990000171	t20	5	2:53:40 PM	18	0.02 kUA/l	0		
011990000175	fx5	5	2:23:41 PM	21	0.05 kUA/l	0	Negative	
011990000175	fx1	5	2:24:41 PM	19	0.03 kUA/l	0	Negative	
011990000175	fx2	5	2:25:41 PM	36	0.22 kUA/l	0	Negative	
011990000175	fx22	5	2:26:42 PM	34	0.21 kUA/l	0	Negative	
011990000175	gx2	5	2:27:41 PM	29	0.14 kUA/l	0	Negative	
011990000175	wx1	5	2:28:41 PM	32	0.18 kUA/l	0	Negative	
011990000175	t20	5	2:29:41 PM	19	0.03 kUA/l	0		
011990000176	fx5	5	2:31:41 PM	25	0.10 kUA/l	0	Negative	
011990000176	fx1	5	2:32:41 PM	20	0.04 kUA/l	0	Negative	
011990000176	fx2	5	2:33:41 PM	23	0.08 kUA/l	0	Negative	
011990000176	fx22	5	2:34:41 PM	35	0.21 kUA/l	0	Negative	
011990000176	gx2	5	2:35:41 PM	29	0.15 kUA/l	0	Negative	
011990000176	wx1	5	2:36:41 PM	130	1.33 kUA/l	2	Positive	
011990000176	t20	5	2:37:41 PM	19	0.03 kUA/l	0		
011990000179	fx5	5	4:12:37 PM	52	0.41 kUA/l	1	Positive	
011990000179	fx1	5	4:13:37 PM	32	0.18 kUA/l	0	Negative	
011990000179	fx2	5	4:14:37 PM	29	0.15 kUA/l	0	Negative	
011990000179	fx22	5	4:26:37 PM	40	0.27 kUA/l	0	Negative	
011990000179	gx2	5	4:27:37 PM	25	0.09 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000179	wx1	5	4:28:37 PM	34	0.20 kUA/I	0	Negative	
011990000179	t20	5	4:29:36 PM	24	0.08 kUA/I	0		
011990000181	fx5	5	2:15:42 PM	21	0.04 kUA/I	0	Negative	
011990000181	fx1	5	2:16:42 PM	15	Low RU	0	Negative	
011990000181	fx2	5	2:17:42 PM	16	Low RU	0	Negative	
011990000181	fx22	5	2:18:42 PM	31	0.17 kUA/I	0	Negative	
011990000181	gx2	5	2:19:42 PM	26	0.10 kUA/I	0	Negative	
011990000181	wx1	5	2:20:42 PM	29	0.15 kUA/I	0	Negative	
011990000181	t20	5	2:21:42 PM	17	Low RU	0		
011990000182	fx5	5	3:11:39 PM	75	0.69 kUA/I	1	Positive	
011990000182	fx1	5	3:12:39 PM	63	0.54 kUA/I	1	Positive	
011990000182	fx2	5	3:13:39 PM	73	0.66 kUA/I	1	Positive	
011990000182	fx22	5	3:14:40 PM	71	0.64 kUA/I	1	Positive	
011990000182	gx2	5	3:15:39 PM	87	0.82 kUA/I	2	Positive	
011990000182	wx1	5	3:17:39 PM	72	0.65 kUA/I	1	Positive	
011990000182	t20	5	3:18:39 PM	48	0.37 kUA/I	1		
011990000185	fx5	5	3:28:39 PM	47	0.36 kUA/I	1	Positive	
011990000185	fx1	5	3:29:39 PM	45	0.33 kUA/I	0	Negative	
011990000185	fx2	5	3:30:39 PM	172	1.80 kUA/I	2	Positive	
011990000185	fx22	5	3:31:38 PM	61	0.53 kUA/I	1	Positive	
011990000185	gx2	5	3:32:39 PM	45	0.34 kUA/I	0	Negative	
011990000185	wx1	5	3:33:39 PM	57	0.47 kUA/I	1	Positive	
011990000185	t20	5	3:34:39 PM	37	0.24 kUA/I	0		
011990000186	fx5	5	4:55:35 PM	309	3.39 kUA/I	2	Positive	
011990000186	fx1	5	4:56:36 PM	24	0.08 kUA/I	0	Negative	
011990000186	fx2	5	4:57:36 PM	497	5.58 kUA/I	3	Positive	
011990000186	fx22	5	4:58:35 PM	46	0.34 kUA/I	0	Negative	
011990000186	gx2	5	4:59:35 PM	40	0.27 kUA/I	0	Negative	
011990000186	wx1	5	5:00:35 PM	122	1.22 kUA/I	2	Positive	
011990000186	t20	5	5:01:35 PM	57	0.47 kUA/I	1		
011990000188	fx5	5	5:03:35 PM	35	0.21 kUA/I	0	Negative	
011990000188	fx1	5	5:04:35 PM	28	0.13 kUA/I	0	Negative	
011990000188	fx2	5	5:05:35 PM	16	Low RU	0	Negative	
011990000188	fx22	5	5:06:35 PM	32	0.18 kUA/I	0	Negative	
011990000188	gx2	5	5:07:35 PM	26	0.11 kUA/I	0	Negative	
011990000188	wx1	5	5:08:35 PM	41	0.28 kUA/I	0	Negative	
011990000188	t20	5	5:09:35 PM	30	0.15 kUA/I	0		
011990000189	fx5	5	4:31:36 PM	19	0.02 kUA/I	0	Negative	
011990000189	fx1	5	4:32:36 PM	15	Low RU	0	Negative	
011990000189	fx2	5	4:33:36 PM	11	Low RU	0	Negative	
011990000189	fx22	5	4:34:36 PM	32	0.18 kUA/I	0	Negative	
011990000189	gx2	5	4:35:36 PM	18	0.01 kUA/I	0	Negative	
011990000189	wx1	5	4:36:36 PM	26	0.11 kUA/I	0	Negative	
011990000189	t20	5	4:37:36 PM	16	Low RU	0		
011990000193	fx5	5	3:36:38 PM	25	0.10 kUA/I	0	Negative	
011990000193	fx1	5	3:37:38 PM	23	0.07 kUA/I	0	Negative	
011990000193	fx2	5	3:38:38 PM	14	Low RU	0	Negative	
011990000193	fx22	5	3:39:38 PM	35	0.22 kUA/I	0	Negative	
011990000193	gx2	5	3:40:38 PM	18	0.01 kUA/I	0	Negative	
011990000193	wx1	5	3:41:38 PM	164	1.71 kUA/I	2	Positive	
011990000193	t20	5	3:42:38 PM	45	0.33 kUA/I	0		
012001900678A	f27	1	5:28:34 PM	32	0.04 kUA/I			
012001900678A	f83	1	5:29:34 PM	15	0.00 kUA/I			
012001900678A	d2	1	5:30:34 PM	25	0.02 kUA/I			

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
012001900678A	f245	1	5:31:34 PM	11	0.00 kUA/I			
012001900678A	d1	1	5:32:34 PM	86	0.16 kUA/I			
012001900678A	f2	1	5:33:34 PM	30	0.03 kUA/I			
012001900678A	f41	1	5:34:34 PM	11	0.00 kUA/I			
012001900678A	f24	1	5:35:34 PM	28	0.03 kUA/I			
012001900678A	f4	1	5:36:34 PM	30	0.03 kUA/I			
012001900678A	mx2	1	5:37:34 PM	36			Negative	
012001900678A	f13	1	5:38:34 PM	21	0.01 kUA/I			
012001900678A	f422	1	5:39:34 PM	13	0.00 kUA/I			
012001900678A	f423	1	5:40:34 PM	14	0.00 kUA/I			
012001900678A	f424	1	5:41:34 PM	15	0.00 kUA/I			
012001900678A	f352	1	5:42:34 PM	15	0.00 kUA/I			
012001900698A	f422	1	5:45:34 PM	11189	38.17 kUA/I			
012001900698A	f424	1	5:46:33 PM	38	0.05 kUA/I			
012001900698A	f352	1	5:47:33 PM	22	0.01 kUA/I			
012001900900A	f18	1	5:48:33 PM	14	0.00 kUA/I			
012001900900A	f202	1	5:49:33 PM	497	1.12 kUA/I			
012001900900A	t3	1	5:50:33 PM	14	0.00 kUA/I			
012001900900A	f17	1	5:51:33 PM	58	0.10 kUA/I			
012001900900A	f345	1	5:52:33 PM	76	0.14 kUA/I			
012001900900A	f201	1	5:53:33 PM	372	0.82 kUA/I			
012001900900A	f203	1	5:54:33 PM	442	0.99 kUA/I			
012001900900A	f256	1	5:55:33 PM	638	1.45 kUA/I			
012001900900A	f13	1	5:56:33 PM	17	0.00 kUA/I			
012001900900A	f425	1	5:57:33 PM	14	0.00 kUA/I			
012001900954A	f1	1	5:43:33 PM	4115	10.5 kUA/I			
012001900954A	f13	1	5:44:34 PM	329	0.72 kUA/I			
012002000645A	ex72	1	5:18:34 PM	13			Negative	
012002000645A	mx2	1	5:19:34 PM	29			Negative	
012002000645A	gx2	1	5:20:34 PM	16	0.00 kUA/I	0	Negative	
012002000645A	e1	1	5:21:34 PM	9	0.00 kUA/I			
012002000645A	g2	1	5:22:34 PM	16	0.00 kUA/I			
012002000645A	e5	1	5:23:34 PM	11	0.00 kUA/I			
012002000645A	d1	1	5:24:35 PM	20	0.01 kUA/I			
012002000645A	m3	1	5:25:34 PM	13	0.00 kUA/I			
012002000645A	e215	1	5:26:34 PM	10	0.00 kUA/I			
012002000645A	i206	1	5:27:34 PM	15	0.00 kUA/I			

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
-----------------	------	----------	----------	-----------	---------	----------	----------	-----------	---------	---------



# Laboratory Report

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	3:16:39 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012001900900A	f18	5:48:33 PM	421BJ 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			Exp
012001900900A	f202	5:49:33 PM	997KA 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	t3	5:50:33 PM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	f17	5:51:33 PM	D2FBX 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	f345	5:52:33 PM	BTXGL 20/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	f201	5:53:33 PM	906J3 21/03	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	f203	5:54:33 PM	998HG 20/11	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	f256	5:55:33 PM	CMEBH 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	f13	5:56:33 PM	375MS 20/04	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f27	5:28:34 PM	450CC 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f83	5:29:34 PM	428C8 20/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	d2	5:30:34 PM	358E2 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f245	5:31:34 PM	925HU 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	d1	5:32:34 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f2	5:33:34 PM	362H0 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f41	5:34:34 PM	456CN 20/03	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f24	5:35:34 PM	424EF 20/11	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f4	5:36:34 PM	363F1 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	mx2	5:37:34 PM	851J1 21/03	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f13	5:38:34 PM	375MS 20/04	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f422	5:39:34 PM	CDFH4 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f423	5:40:34 PM	CDGHN 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f424	5:41:34 PM	CDHH0 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900678A	f352	5:42:34 PM	CEZH5 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900900A	f425	5:57:33 PM	CDPGE 20/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900954A	f1	5:43:33 PM	361EN 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900954A	f13	5:44:34 PM	375MS 20/04	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	ex72	5:18:34 PM	521G9 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	mx2	5:19:34 PM	851J1 21/03	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	gx2	5:20:34 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	e1	5:21:34 PM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	g2	5:22:34 PM	379CG 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	e5	5:23:34 PM	360F3 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	d1	5:24:35 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	m3	5:25:34 PM	368K6 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	e215	5:26:34 PM	BP5GE 20/04	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012002000645A	i206	5:27:34 PM	C07GX 20/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900698A	f422	5:45:34 PM	CDFH4 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900698A	f424	5:46:33 PM	CDHH0 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900698A	f352	5:47:33 PM	CEZH5 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000185	phad	3:27:39 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000185	fx5	3:28:39 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000185	fx1	3:29:39 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000185	fx2	3:30:39 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000185	fx22	3:31:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000185	gx2	3:32:39 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000185	wx1	3:33:39 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000185	t20	3:34:39 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	phad	3:35:38 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	fx5	3:36:38 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	fx1	3:37:38 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	fx2	3:38:38 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	fx22	3:39:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	gx2	3:40:38 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	wx1	3:41:38 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000193	t20	3:42:38 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000170	phad	3:43:38 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000170	fx5	3:44:38 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000170	fx1	3:45:38 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000170	fx2	3:46:38 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000170	fx22	3:47:38 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000170	gx2	3:48:38 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000170	wx1	3:49:38 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000170	t20	3:50:38 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	phad	4:11:37 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	fx5	4:12:37 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	fx1	4:13:37 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	fx2	4:14:37 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	fx22	4:26:37 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	gx2	4:27:37 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	wx1	4:28:37 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000179	t20	4:29:36 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	phad	4:30:36 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	fx5	4:31:36 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	fx1	4:32:36 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	fx2	4:33:36 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	fx22	4:34:36 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	gx2	4:35:36 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	wx1	4:36:36 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000189	t20	4:37:36 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	phad	4:38:36 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	fx5	4:39:36 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	fx1	4:40:36 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	fx2	4:41:36 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	fx22	4:42:36 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	gx2	4:43:36 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	wx1	4:44:36 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000158	t20	4:45:36 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	phad	4:46:36 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	fx5	4:47:36 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	fx1	4:48:36 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	fx2	4:49:36 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	fx22	4:50:36 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	gx2	4:51:36 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	wx1	4:52:36 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000168	t20	4:53:35 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	phad	4:54:35 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	fx5	4:55:35 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	fx1	4:56:36 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	fx2	4:57:36 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	fx22	4:58:35 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	gx2	4:59:35 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	wx1	5:00:35 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000186	t20	5:01:35 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	phad	5:02:35 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	fx5	5:03:35 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	fx1	5:04:35 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	fx2	5:05:35 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	fx22	5:06:35 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	gx2	5:07:35 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	wx1	5:08:35 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000188	t20	5:09:35 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	phad	5:10:35 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	fx5	5:11:35 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	fx1	5:12:35 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	fx2	5:13:35 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	fx22	5:14:35 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	gx2	5:15:35 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	wx1	5:16:35 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000167	t20	5:17:35 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBS 21/07	
011990000161	phad	2:06:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000161	fx5	2:07:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000161	fx1	2:08:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000161	fx2	2:09:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000161	fx22	2:10:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000161	gx2	2:11:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000161	wx1	2:12:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000161	t20	2:13:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000181	phad	2:14:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000181	fx5	2:15:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000181	fx1	2:16:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000181	fx2	2:17:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000181	fx22	2:18:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000181	gx2	2:19:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000181	wx1	2:20:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000181	t20	2:21:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	phad	2:22:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	fx5	2:23:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	fx1	2:24:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	fx2	2:25:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	fx22	2:26:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	gx2	2:27:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	wx1	2:28:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000175	t20	2:29:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	phad	2:30:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	fx5	2:31:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	fx1	2:32:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	fx2	2:33:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	fx22	2:34:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	gx2	2:35:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	wx1	2:36:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000176	t20	2:37:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	phad	2:38:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	fx5	2:39:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	fx1	2:40:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	fx2	2:41:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	fx22	2:42:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	gx2	2:43:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	wx1	2:44:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000164	t20	2:45:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	phad	2:46:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	fx5	2:47:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	fx1	2:48:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	fx2	2:49:40 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	fx22	2:50:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	gx2	2:51:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	wx1	2:52:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000171	t20	2:53:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	phad	2:54:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	fx5	2:55:40 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	fx1	2:56:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	fx2	2:57:40 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	fx22	2:58:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	gx2	2:59:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	wx1	3:00:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000154	t20	3:01:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	phad	3:02:40 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	fx5	3:03:40 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	fx1	3:04:40 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	fx2	3:05:40 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	fx22	3:06:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	gx2	3:07:40 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	wx1	3:08:40 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000151	t20	3:09:40 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	phad	3:10:39 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	fx5	3:11:39 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	fx1	3:12:39 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	fx2	3:13:39 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	fx22	3:14:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	gx2	3:15:39 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	wx1	3:17:39 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000182	t20	3:18:39 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000153	phad	3:19:39 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000153	fx5	3:20:39 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000153	fx1	3:21:39 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000153	fx2	3:22:39 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000153	fx22	3:23:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000153	gx2	3:24:39 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000153	wx1	3:25:39 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	
011990000153	t20	3:26:39 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BS 21/07	

### Errors

#### Expired lot numbers used for Sample

352D0 ImmunoCAP