


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number:	200130-390
Instrument name:	Phadia-A
Instrument type:	Phadia 250
Status:	Ready
Started:	30-Jan-20 11:09:51 AM
Ended:	30-Jan-20 2:42:46 PM
Approved:	Approved (faziz)
Calibrators in run:	No

### Calibration Information

Instrument name:	Phadia-A
Instrument type:	Phadia 250
Measured:	05-Jan-20 1:26:01 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	11:09:51 AM	334	0.74	OK	
CC-2	17.5	11:10:51 AM	6533	18.3	OK	

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	11:12:51 AM	3515	8.82 kUA/I	CZ1AR	OK	
slgE Neg	d1	1	11:13:51 AM	24	0.02 kUA/I	BL1CP	OK	
slgE L	e1	1	11:14:51 AM	1416	3.31 kUA/I	CZ4AS	OK	
IQC271119	d1	1	11:41:50 AM	6423	17.9 kUA/I	<lab>	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000651	phad	5	11:54:49 AM	29	0.14 PAU/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000652	phad	5	11:46:49 AM	33	0.19 PAU/I	0	Negative	
011990000653	phad	5	11:37:50 AM	68	0.60 PAU/I	1	Positive	
011990000655	phad	5	11:22:50 AM	41	0.28 PAU/I	0	Negative	
011990000656	phad	5	12:34:48 PM	32	0.18 PAU/I	0	Negative	
011990000661	phad	5	2:02:44 PM	677	7.69 PAU/I	3	Positive	
011990000662	phad	5	2:10:44 PM	32	0.18 PAU/I	0	Negative	
011990000663	phad	5	2:18:44 PM	42	0.30 PAU/I	0	Negative	
011990000664	phad	5	2:26:43 PM	35	0.21 PAU/I	0	Negative	
011990000665	phad	5	2:34:43 PM	37	0.23 PAU/I	0	Negative	
011990000666	phad	5	12:50:47 PM	2113	25.3 PAU/I	4	Positive	
011990000667	phad	5	1:46:45 PM	44	0.32 PAU/I	0	Negative	
011990000668	phad	5	1:38:45 PM	35	0.21 PAU/I	0	Negative	
011990000669	phad	5	1:30:46 PM	43	0.31 PAU/I	0	Negative	
011990000670	phad	5	1:22:46 PM	692	7.87 PAU/I	3	Positive	
011990000671	phad	5	12:42:47 PM	68	0.60 PAU/I	1	Positive	
011990000672	phad	5	1:06:46 PM	37	0.24 PAU/I	0	Negative	
011990000673	phad	5	1:14:46 PM	35	0.21 PAU/I	0	Negative	
011990000674	phad	5	12:58:47 PM	41	0.29 PAU/I	0	Negative	
011990000675	phad	5	1:54:45 PM	34	0.20 PAU/I	0	Negative	
011990000876	phad	5	11:11:51 AM	41	0.28 PAU/I	0	Negative	
011990000877	phad	5	12:02:49 PM	205	2.18 PAU/I	2	Positive	
011990000878	phad	5	12:10:48 PM	55	0.45 PAU/I	1	Positive	
011990000879	phad	5	12:18:48 PM	37	0.24 PAU/I	0	Negative	
011990000880	phad	5	12:26:48 PM	36	0.22 PAU/I	0	Negative	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000651	fx5	5	11:55:49 AM	21	0.05 kUA/I	0	Negative	
011990000651	fx1	5	11:56:49 AM	15	Low RU	0	Negative	
011990000651	fx2	5	11:57:49 AM	12	Low RU	0	Negative	
011990000651	fx22	5	11:58:49 AM	28	0.13 kUA/I	0	Negative	
011990000651	gx2	5	11:59:49 AM	28	0.13 kUA/I	0	Negative	
011990000651	wx1	5	12:00:49 PM	29	0.15 kUA/I	0	Negative	
011990000651	t20	5	12:01:49 PM	17	Low RU	0		
011990000652	fx5	5	11:47:50 AM	21	0.05 kUA/I	0	Negative	
011990000652	fx1	5	11:48:49 AM	14	Low RU	0	Negative	
011990000652	fx2	5	11:49:49 AM	14	Low RU	0	Negative	
011990000652	fx22	5	11:50:49 AM	28	0.14 kUA/I	0	Negative	
011990000652	gx2	5	11:51:49 AM	22	0.07 kUA/I	0	Negative	
011990000652	wx1	5	11:52:49 AM	34	0.20 kUA/I	0	Negative	
011990000652	t20	5	11:53:49 AM	18	0.01 kUA/I	0		
011990000653	fx5	5	11:38:50 AM	19	0.02 kUA/I	0	Negative	
011990000653	fx1	5	11:39:50 AM	20	0.04 kUA/I	0	Negative	
011990000653	fx2	5	11:40:50 AM	156	1.62 kUA/I	2	Positive	
011990000653	fx22	5	11:42:49 AM	31	0.17 kUA/I	0	Negative	
011990000653	gx2	5	11:43:49 AM	24	0.09 kUA/I	0	Negative	
011990000653	wx1	5	11:44:49 AM	34	0.20 kUA/I	0	Negative	
011990000653	t20	5	11:45:49 AM	18	0.02 kUA/I	0		
011990000655	fx5	5	11:23:50 AM	20	0.04 kUA/I	0	Negative	
011990000655	fx1	5	11:24:50 AM	20	0.03 kUA/I	0	Negative	
011990000655	fx2	5	11:32:50 AM	16	Low RU	0	Negative	
011990000655	fx22	5	11:33:50 AM	31	0.17 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000655	gx2	5	11:34:50 AM	21	0.05 kUA/l	0	Negative	
011990000655	wx1	5	11:35:50 AM	31	0.17 kUA/l	0	Negative	
011990000655	t20	5	11:36:50 AM	18	0.02 kUA/l	0		
011990000656	fx5	5	12:35:48 PM	21	0.05 kUA/l	0	Negative	
011990000656	fx1	5	12:36:48 PM	15	Low RU	0	Negative	
011990000656	fx2	5	12:37:47 PM	15	Low RU	0	Negative	
011990000656	fx22	5	12:38:47 PM	28	0.13 kUA/l	0	Negative	
011990000656	gx2	5	12:39:47 PM	21	0.05 kUA/l	0	Negative	
011990000656	wx1	5	12:40:47 PM	30	0.15 kUA/l	0	Negative	
011990000656	t20	5	12:41:47 PM	17	Low RU	0		
011990000661	fx5	5	2:03:44 PM	34	0.20 kUA/l	0	Negative	
011990000661	fx1	5	2:04:44 PM	24	0.09 kUA/l	0	Negative	
011990000661	fx2	5	2:05:44 PM	368	4.08 kUA/l	3	Positive	
011990000661	fx22	5	2:06:44 PM	38	0.25 kUA/l	0	Negative	
011990000661	gx2	5	2:07:44 PM	38	0.25 kUA/l	0	Negative	
011990000661	wx1	5	2:08:44 PM	110	1.09 kUA/l	2	Positive	
011990000661	t20	5	2:09:44 PM	56	0.46 kUA/l	1		
011990000662	fx5	5	2:11:44 PM	26	0.10 kUA/l	0	Negative	
011990000662	fx1	5	2:12:44 PM	16	Low RU	0	Negative	
011990000662	fx2	5	2:13:44 PM	13	Low RU	0	Negative	
011990000662	fx22	5	2:14:44 PM	35	0.21 kUA/l	0	Negative	
011990000662	gx2	5	2:15:44 PM	30	0.16 kUA/l	0	Negative	
011990000662	wx1	5	2:16:44 PM	31	0.17 kUA/l	0	Negative	
011990000662	t20	5	2:17:44 PM	16	Low RU	0		
011990000663	fx5	5	2:19:44 PM	29	0.14 kUA/l	0	Negative	
011990000663	fx1	5	2:20:44 PM	19	0.03 kUA/l	0	Negative	
011990000663	fx2	5	2:21:44 PM	15	Low RU	0	Negative	
011990000663	fx22	5	2:22:44 PM	28	0.13 kUA/l	0	Negative	
011990000663	gx2	5	2:23:43 PM	25	0.10 kUA/l	0	Negative	
011990000663	wx1	5	2:24:44 PM	31	0.17 kUA/l	0	Negative	
011990000663	t20	5	2:25:43 PM	18	0.02 kUA/l	0		
011990000664	fx5	5	2:27:44 PM	22	0.07 kUA/l	0	Negative	
011990000664	fx1	5	2:28:44 PM	17	Low RU	0	Negative	
011990000664	fx2	5	2:29:43 PM	18	0.01 kUA/l	0	Negative	
011990000664	fx22	5	2:30:43 PM	30	0.16 kUA/l	0	Negative	
011990000664	gx2	5	2:31:43 PM	24	0.08 kUA/l	0	Negative	
011990000664	wx1	5	2:32:43 PM	28	0.13 kUA/l	0	Negative	
011990000664	t20	5	2:33:43 PM	15	Low RU	0		
011990000665	fx5	5	2:35:43 PM	22	0.06 kUA/l	0	Negative	
011990000665	fx1	5	2:36:43 PM	24	0.09 kUA/l	0	Negative	
011990000665	fx2	5	2:37:43 PM	18	0.01 kUA/l	0	Negative	
011990000665	fx22	5	2:38:43 PM	36	0.22 kUA/l	0	Negative	
011990000665	gx2	5	2:39:43 PM	23	0.08 kUA/l	0	Negative	
011990000665	wx1	5	2:40:43 PM	33	0.19 kUA/l	0	Negative	
011990000665	t20	5	2:41:43 PM	18	0.02 kUA/l	0		
011990000666	fx5	5	12:51:47 PM	29	0.15 kUA/l	0	Negative	
011990000666	fx1	5	12:52:47 PM	20	0.04 kUA/l	0	Negative	
011990000666	fx2	5	12:53:47 PM	14	Low RU	0	Negative	
011990000666	fx22	5	12:54:47 PM	38	0.25 kUA/l	0	Negative	
011990000666	gx2	5	12:55:47 PM	30	0.16 kUA/l	0	Negative	
011990000666	wx1	5	12:56:47 PM	34	0.20 kUA/l	0	Negative	
011990000666	t20	5	12:57:47 PM	21	0.05 kUA/l	0		
011990000667	fx5	5	1:47:45 PM	21	0.05 kUA/l	0	Negative	
011990000667	fx1	5	1:48:45 PM	16	Low RU	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000667	fx2	5	1:49:45 PM	17	Low RU	0	Negative	
011990000667	fx22	5	1:50:45 PM	31	0.17 kUA/I	0	Negative	
011990000667	gx2	5	1:51:45 PM	53	0.42 kUA/I	1	Positive	
011990000667	wx1	5	1:52:45 PM	33	0.20 kUA/I	0	Negative	
011990000667	t20	5	1:53:45 PM	17	Low RU	0		
011990000668	fx5	5	1:39:45 PM	24	0.08 kUA/I	0	Negative	
011990000668	fx1	5	1:40:45 PM	19	0.02 kUA/I	0	Negative	
011990000668	fx2	5	1:41:45 PM	15	Low RU	0	Negative	
011990000668	fx22	5	1:42:45 PM	29	0.15 kUA/I	0	Negative	
011990000668	gx2	5	1:43:45 PM	24	0.08 kUA/I	0	Negative	
011990000668	wx1	5	1:44:45 PM	28	0.13 kUA/I	0	Negative	
011990000668	t20	5	1:45:45 PM	17	Low RU	0		
011990000669	fx5	5	1:31:46 PM	25	0.10 kUA/I	0	Negative	
011990000669	fx1	5	1:32:46 PM	17	0.01 kUA/I	0	Negative	
011990000669	fx2	5	1:33:45 PM	15	Low RU	0	Negative	
011990000669	fx22	5	1:34:45 PM	32	0.19 kUA/I	0	Negative	
011990000669	gx2	5	1:35:45 PM	22	0.06 kUA/I	0	Negative	
011990000669	wx1	5	1:36:45 PM	33	0.19 kUA/I	0	Negative	
011990000669	t20	5	1:37:45 PM	17	Low RU	0		
011990000670	fx5	5	1:23:46 PM	42	0.29 kUA/I	0	Negative	
011990000670	fx1	5	1:24:46 PM	22	0.06 kUA/I	0	Negative	
011990000670	fx2	5	1:25:46 PM	942	10.8 kUA/I	3	Positive	
011990000670	fx22	5	1:26:46 PM	205	2.18 kUA/I	2	Positive	
011990000670	gx2	5	1:27:46 PM	63	0.55 kUA/I	1	Positive	
011990000670	wx1	5	1:28:46 PM	84	0.78 kUA/I	2	Positive	
011990000670	t20	5	1:29:45 PM	68	0.60 kUA/I	1		
011990000671	fx5	5	12:43:47 PM	25	0.10 kUA/I	0	Negative	
011990000671	fx1	5	12:44:48 PM	20	0.04 kUA/I	0	Negative	
011990000671	fx2	5	12:45:47 PM	19	0.02 kUA/I	0	Negative	
011990000671	fx22	5	12:46:47 PM	31	0.17 kUA/I	0	Negative	
011990000671	gx2	5	12:47:47 PM	22	0.06 kUA/I	0	Negative	
011990000671	wx1	5	12:48:47 PM	62	0.53 kUA/I	1	Positive	
011990000671	t20	5	12:49:47 PM	19	0.03 kUA/I	0		
011990000672	fx5	5	1:07:46 PM	33	0.19 kUA/I	0	Negative	
011990000672	fx1	5	1:08:46 PM	16	Low RU	0	Negative	
011990000672	fx2	5	1:09:46 PM	12	Low RU	0	Negative	
011990000672	fx22	5	1:10:46 PM	30	0.16 kUA/I	0	Negative	
011990000672	gx2	5	1:11:46 PM	24	0.08 kUA/I	0	Negative	
011990000672	wx1	5	1:12:47 PM	28	0.14 kUA/I	0	Negative	
011990000672	t20	5	1:13:46 PM	18	0.01 kUA/I	0		
011990000673	fx5	5	1:15:46 PM	22	0.06 kUA/I	0	Negative	
011990000673	fx1	5	1:16:46 PM	14	Low RU	0	Negative	
011990000673	fx2	5	1:17:46 PM	14	Low RU	0	Negative	
011990000673	fx22	5	1:18:46 PM	30	0.15 kUA/I	0	Negative	
011990000673	gx2	5	1:19:46 PM	20	0.04 kUA/I	0	Negative	
011990000673	wx1	5	1:20:46 PM	30	0.16 kUA/I	0	Negative	
011990000673	t20	5	1:21:46 PM	16	Low RU	0		
011990000674	fx5	5	12:59:47 PM	29	0.14 kUA/I	0	Negative	
011990000674	fx1	5	1:00:46 PM	21	0.05 kUA/I	0	Negative	
011990000674	fx2	5	1:01:46 PM	18	0.01 kUA/I	0	Negative	
011990000674	fx22	5	1:02:46 PM	33	0.20 kUA/I	0	Negative	
011990000674	gx2	5	1:03:47 PM	22	0.06 kUA/I	0	Negative	
011990000674	wx1	5	1:04:47 PM	36	0.23 kUA/I	0	Negative	
011990000674	t20	5	1:05:46 PM	19	0.02 kUA/I	0		

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000675	fx5	5	1:55:45 PM	19	0.03 kUA/l	0	Negative	
011990000675	fx1	5	1:56:44 PM	15	Low RU	0	Negative	
011990000675	fx2	5	1:57:44 PM	15	Low RU	0	Negative	
011990000675	fx22	5	1:58:44 PM	30	0.16 kUA/l	0	Negative	
011990000675	gx2	5	1:59:45 PM	22	0.06 kUA/l	0	Negative	
011990000675	wx1	5	2:00:44 PM	28	0.13 kUA/l	0	Negative	
011990000675	t20	5	2:01:44 PM	16	Low RU	0		
011990000876	fx5	5	11:15:51 AM	32	0.19 kUA/l	0	Negative	
011990000876	fx1	5	11:16:50 AM	18	0.02 kUA/l	0	Negative	
011990000876	fx2	5	11:17:50 AM	13	Low RU	0	Negative	
011990000876	fx22	5	11:18:50 AM	31	0.17 kUA/l	0	Negative	
011990000876	gx2	5	11:19:50 AM	25	0.10 kUA/l	0	Negative	
011990000876	wx1	5	11:20:51 AM	32	0.18 kUA/l	0	Negative	
011990000876	t20	5	11:21:50 AM	20	0.04 kUA/l	0		
011990000877	fx5	5	12:03:49 PM	30	0.16 kUA/l	0	Negative	
011990000877	fx1	5	12:04:48 PM	22	0.06 kUA/l	0	Negative	
011990000877	fx2	5	12:05:49 PM	18	0.02 kUA/l	0	Negative	
011990000877	fx22	5	12:06:49 PM	37	0.24 kUA/l	0	Negative	
011990000877	gx2	5	12:07:49 PM	564	6.37 kUA/l	3	Positive	
011990000877	wx1	5	12:08:49 PM	63	0.54 kUA/l	1	Positive	
011990000877	t20	5	12:09:49 PM	25	0.10 kUA/l	0		
011990000878	fx5	5	12:11:48 PM	22	0.06 kUA/l	0	Negative	
011990000878	fx1	5	12:12:48 PM	17	0.01 kUA/l	0	Negative	
011990000878	fx2	5	12:13:48 PM	21	0.05 kUA/l	0	Negative	
011990000878	fx22	5	12:14:48 PM	33	0.19 kUA/l	0	Negative	
011990000878	gx2	5	12:15:48 PM	24	0.08 kUA/l	0	Negative	
011990000878	wx1	5	12:16:48 PM	33	0.20 kUA/l	0	Negative	
011990000878	t20	5	12:17:48 PM	20	0.04 kUA/l	0		
011990000879	fx5	5	12:19:48 PM	21	0.05 kUA/l	0	Negative	
011990000879	fx1	5	12:20:48 PM	16	Low RU	0	Negative	
011990000879	fx2	5	12:21:48 PM	13	Low RU	0	Negative	
011990000879	fx22	5	12:22:48 PM	31	0.16 kUA/l	0	Negative	
011990000879	gx2	5	12:23:48 PM	20	0.04 kUA/l	0	Negative	
011990000879	wx1	5	12:24:48 PM	30	0.16 kUA/l	0	Negative	
011990000879	t20	5	12:25:48 PM	16	Low RU	0		
011990000880	fx5	5	12:27:48 PM	22	0.06 kUA/l	0	Negative	
011990000880	fx1	5	12:28:48 PM	18	0.01 kUA/l	0	Negative	
011990000880	fx2	5	12:29:48 PM	13	Low RU	0	Negative	
011990000880	fx22	5	12:30:48 PM	31	0.17 kUA/l	0	Negative	
011990000880	gx2	5	12:31:48 PM	23	0.07 kUA/l	0	Negative	
011990000880	wx1	5	12:32:48 PM	29	0.15 kUA/l	0	Negative	
011990000880	t20	5	12:33:47 PM	21	0.05 kUA/l	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	



# Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	11:09:51 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	11:10:51 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	11:12:51 AM	352D0 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ1AR 21/08		Exp
slgE Neg	d1	11:13:51 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		BL1CP 19/12!		Exp
slgE L	e1	11:14:51 AM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ4AS 20/03		
IQC271119	d1	11:41:50 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000876	phad	11:11:51 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000876	fx5	11:15:51 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000876	fx1	11:16:50 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000876	fx2	11:17:50 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000876	fx22	11:18:50 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000876	gx2	11:19:50 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000876	wx1	11:20:51 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000876	t20	11:21:50 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	phad	12:02:49 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	fx5	12:03:49 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	fx1	12:04:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	fx2	12:05:49 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	fx22	12:06:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	gx2	12:07:49 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	wx1	12:08:49 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000877	t20	12:09:49 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	phad	12:10:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	fx5	12:11:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	fx1	12:12:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	fx2	12:13:48 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	fx22	12:14:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	gx2	12:15:48 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	wx1	12:16:48 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000878	t20	12:17:48 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	phad	12:18:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	fx5	12:19:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	fx1	12:20:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	fx2	12:21:48 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	fx22	12:22:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	gx2	12:23:48 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	wx1	12:24:48 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000879	t20	12:25:48 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	phad	12:26:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	fx5	12:27:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	fx1	12:28:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	fx2	12:29:48 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	fx22	12:30:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	gx2	12:31:48 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	wx1	12:32:48 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000880	t20	12:33:47 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	phad	11:54:49 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	fx5	11:55:49 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	fx1	11:56:49 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	fx2	11:57:49 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	fx22	11:58:49 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	gx2	11:59:49 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	wx1	12:00:49 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000651	t20	12:01:49 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000656	phad	12:34:48 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000656	fx5	12:35:48 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000656	fx1	12:36:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000656	fx2	12:37:47 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000656	fx22	12:38:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000656	gx2	12:39:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000656	wx1	12:40:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000656	t20	12:41:47 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	phad	2:02:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	fx5	2:03:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	fx1	2:04:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	fx2	2:05:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	fx22	2:06:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	gx2	2:07:44 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	wx1	2:08:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000661	t20	2:09:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	phad	12:50:47 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	fx5	12:51:47 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	fx1	12:52:47 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	fx2	12:53:47 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	fx22	12:54:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	gx2	12:55:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	wx1	12:56:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000666	t20	12:57:47 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	phad	12:42:47 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	fx5	12:43:47 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	fx1	12:44:48 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	fx2	12:45:47 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	fx22	12:46:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	gx2	12:47:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	wx1	12:48:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000671	t20	12:49:47 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	phad	11:46:49 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	fx5	11:47:50 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	fx1	11:48:49 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	fx2	11:49:49 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	fx22	11:50:49 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	gx2	11:51:49 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	wx1	11:52:49 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000652	t20	11:53:49 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	phad	2:10:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	fx5	2:11:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	fx1	2:12:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	fx2	2:13:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	fx22	2:14:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	gx2	2:15:44 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	wx1	2:16:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000662	t20	2:17:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	phad	1:46:45 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	fx5	1:47:45 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	fx1	1:48:45 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	fx2	1:49:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	fx22	1:50:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	gx2	1:51:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	wx1	1:52:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000667	t20	1:53:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	phad	1:06:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	fx5	1:07:46 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	fx1	1:08:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	fx2	1:09:46 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	fx22	1:10:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	gx2	1:11:46 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	wx1	1:12:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000672	t20	1:13:46 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000673	phad	1:14:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000673	fx5	1:15:46 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000673	fx1	1:16:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000673	fx2	1:17:46 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000673	fx22	1:18:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000673	gx2	1:19:46 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000673	wx1	1:20:46 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000673	t20	1:21:46 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	phad	12:58:47 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	fx5	12:59:47 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	fx1	1:00:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	fx2	1:01:46 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	fx22	1:02:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	gx2	1:03:47 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	wx1	1:04:47 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000674	t20	1:05:46 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	phad	1:54:45 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	fx5	1:55:45 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	fx1	1:56:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	fx2	1:57:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	fx22	1:58:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	gx2	1:59:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	wx1	2:00:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000675	t20	2:01:44 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	phad	1:38:45 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	fx5	1:39:45 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	fx1	1:40:45 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	fx2	1:41:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	fx22	1:42:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	gx2	1:43:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	wx1	1:44:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000668	t20	1:45:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	phad	1:30:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	fx5	1:31:46 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	fx1	1:32:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	fx2	1:33:45 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	fx22	1:34:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	gx2	1:35:45 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	wx1	1:36:45 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000669	t20	1:37:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	phad	1:22:46 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	fx5	1:23:46 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	fx1	1:24:46 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	fx2	1:25:46 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	fx22	1:26:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	gx2	1:27:46 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	wx1	1:28:46 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000670	t20	1:29:45 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	phad	2:18:44 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	fx5	2:19:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	fx1	2:20:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	fx2	2:21:44 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	fx22	2:22:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	gx2	2:23:43 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	wx1	2:24:44 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000663	t20	2:25:43 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	phad	2:26:43 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	fx5	2:27:44 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	fx1	2:28:44 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	fx2	2:29:43 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	fx22	2:30:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	gx2	2:31:43 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	wx1	2:32:43 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000664	t20	2:33:43 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000665	phad	2:34:43 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000665	fx5	2:35:43 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000665	fx1	2:36:43 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000665	fx2	2:37:43 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000665	fx22	2:38:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000665	gx2	2:39:43 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000665	wx1	2:40:43 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000665	t20	2:41:43 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	phad	11:37:50 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	fx5	11:38:50 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	fx1	11:39:50 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	fx2	11:40:50 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	fx22	11:42:49 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	gx2	11:43:49 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	wx1	11:44:49 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000653	t20	11:45:49 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	phad	11:22:50 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	fx5	11:23:50 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	fx1	11:24:50 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	fx2	11:32:50 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	fx22	11:33:50 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	gx2	11:34:50 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	wx1	11:35:50 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000655	t20	11:36:50 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

### Errors

#### Expired lot numbers used for Quality Control

352D0	ImmunoCAP
BL1CP	Quality control vial


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200130-390  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 30-Jan-20 2:42:43 PM  
Ended: 30-Jan-20 4:00:02 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

No curve controls in this run

### Quality Controls

No quality controls in this run

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000654	phad	5	2:50:42 PM	34	0.21 PAU/l	0	Negative	
011990000657	phad	5	3:14:42 PM	32	0.18 PAU/l	0	Negative	
011990000658	phad	5	3:06:42 PM	1593	18.8 PAU/l	4	Positive	
011990000659	phad	5	2:58:42 PM	82	0.76 PAU/l	2	Positive	
011990000660	phad	5	2:42:43 PM	39	0.26 PAU/l	0	Negative	

## Laboratory Report

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000654	fx5	5	2:51:42 PM	22	0.06kUA/l	0	Negative	
011990000654	fx1	5	2:52:42 PM	18	0.01kUA/l	0	Negative	
011990000654	fx2	5	2:53:42 PM	15	Low RU	0	Negative	
011990000654	fx22	5	2:54:42 PM	32	0.18kUA/l	0	Negative	
011990000654	gx2	5	2:55:42 PM	27	0.12kUA/l	0	Negative	
011990000654	wx1	5	2:56:42 PM	33	0.19kUA/l	0	Negative	
011990000654	t20	5	2:57:42 PM	19	0.02kUA/l	0		
011990000657	fx5	5	3:15:42 PM	23	0.07kUA/l	0	Negative	
011990000657	fx1	5	3:16:41 PM	21	0.05kUA/l	0	Negative	
011990000657	fx2	5	3:17:41 PM	13	Low RU	0	Negative	
011990000657	fx22	5	3:18:41 PM	28	0.13kUA/l	0	Negative	
011990000657	gx2	5	3:19:41 PM	23	0.07kUA/l	0	Negative	
011990000657	wx1	5	3:19:59 PM	---	---	---	---	Err
011990000657	t20	5	3:21:41 PM	17	Low RU	0		
011990000658	fx5	5	3:07:42 PM	23	0.08kUA/l	0	Negative	
011990000658	fx1	5	3:08:42 PM	19	0.02kUA/l	0	Negative	
011990000658	fx2	5	3:09:42 PM	20	0.04kUA/l	0	Negative	
011990000658	fx22	5	3:10:42 PM	33	0.19kUA/l	0	Negative	
011990000658	gx2	5	3:11:41 PM	28	0.13kUA/l	0	Negative	
011990000658	wx1	5	3:12:41 PM	31	0.17kUA/l	0	Negative	
011990000658	t20	5	3:13:41 PM	18	0.01kUA/l	0		
011990000659	fx5	5	2:59:42 PM	28	0.13kUA/l	0	Negative	
011990000659	fx1	5	3:00:42 PM	23	0.07kUA/l	0	Negative	
011990000659	fx2	5	3:01:42 PM	16	Low RU	0	Negative	
011990000659	fx22	5	3:02:42 PM	36	0.23kUA/l	0	Negative	
011990000659	gx2	5	3:03:42 PM	25	0.10kUA/l	0	Negative	
011990000659	wx1	5	3:04:42 PM	56	0.46kUA/l	1	Positive	
011990000659	t20	5	3:05:42 PM	30	0.16kUA/l	0		
011990000660	fx5	5	2:43:43 PM	43	0.30kUA/l	0	Negative	
011990000660	fx1	5	2:44:43 PM	17	0.01kUA/l	0	Negative	
011990000660	fx2	5	2:45:42 PM	15	Low RU	0	Negative	
011990000660	fx22	5	2:46:43 PM	31	0.17kUA/l	0	Negative	
011990000660	gx2	5	2:47:43 PM	26	0.11kUA/l	0	Negative	
011990000660	wx1	5	2:48:43 PM	36	0.22kUA/l	0	Negative	
011990000660	t20	5	2:49:42 PM	19	0.02kUA/l	0		
011990000881	fx22	5	3:21:59 PM	---	---	---	---	Err
011990000881	gx2	5	3:23:41 PM	34	0.20kUA/l	0	Negative	
011990000881	wx1	5	3:23:59 PM	---	---	---	---	Err
011990000881	t20	5	3:25:41 PM	19	0.03kUA/l	0		
011990000886	wx1	5	3:26:41 PM	225	2.42kUA/l	2	Positive	
012002300410A	e1	2	3:27:41 PM	9	Low RU			
012002300410A	d2	2	3:28:41 PM	15	Low RU			
012002300410A	t214	2	3:29:41 PM	12	---			
012002300410A	e5	2	3:30:41 PM	12	Low RU			
012002300410A	f245	2	3:31:41 PM	10	---			
012002300410A	d1	2	3:32:40 PM	22	0.02kUA/l			
012002300410A	f2	2	3:33:41 PM	17	0.00kUA/l			
012002300410A	f4	2	3:34:41 PM	20	0.02kUA/l			
012002300410A	f13	2	3:35:41 PM	15	<0.10			
012002300410A	f425	2	3:36:41 PM	11	---			
012002300410A	f428	2	3:37:40 PM	12	---			
012002300410A	f439	2	3:38:40 PM	11	---			
012002300410A	f440	2	3:39:40 PM	13	---			
012002300410A	ex72	2	3:40:40 PM	16			Negative	

# Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
012002300410A	mx2	2	3:41:40 PM	29			Negative	
012002300410A	gx3	2	3:42:40 PM	16			Negative	
012002700111A	f2	1	3:52:40 PM	98	0.19kUA/I			
012002701016A	f425	1	3:43:40 PM	1237	2.88kUA/I			
012002701016A	f428	1	3:44:40 PM	34	0.04kUA/I			
012002701016A	f439	1	3:45:40 PM	30	0.03kUA/I			
012002701016A	f440	1	3:46:40 PM	1484	3.48kUA/I			
012002701016A	f422	1	3:47:40 PM	11923	42.11kUA/I			
012002701016A	f423	1	3:48:40 PM	21717	141.26kUA/I			
012002701016A	f424	1	3:49:40 PM	9782	31.3 kUA/I			
012002701016A	f427	1	3:50:40 PM	35	0.04kUA/I			
012002701016A	f352	1	3:51:40 PM	38	0.05kUA/I			

## Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012002300410A	e1	3:27:41 PM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	d2	3:28:41 PM	358E2 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	t214	3:29:41 PM	C2PGA 20/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	e5	3:30:41 PM	360F3 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f245	3:31:41 PM	925HU 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	d1	3:32:40 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f2	3:33:41 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f4	3:34:41 PM	363F1 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f13	3:35:41 PM	375MS 20/04	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f425	3:36:41 PM	CDPGE 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f428	3:37:40 PM	CFBGD 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f439	3:38:40 PM	CZPGE 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	f440	3:39:40 PM	D0MGF 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	ex72	3:40:40 PM	521G9 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	mx2	3:41:40 PM	851J1 21/03	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002300410A	gx3	3:42:40 PM	434B9 20/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f425	3:43:40 PM	CDPGE 20/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f428	3:44:40 PM	CFBGD 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f439	3:45:40 PM	CZPGE 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f440	3:46:40 PM	D0MGF 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f422	3:47:40 PM	CDFH4 20/10	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f423	3:48:40 PM	CDGHN 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f424	3:49:40 PM	CDHH0 20/09	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f427	3:50:40 PM	CFCHO 20/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002701016A	f352	3:51:40 PM	CEZH5 20/12	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	fx22	3:21:59 PM	BPHGE 21/06	A1HGR 21/06					
011990000881	gx2	3:23:41 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	wx1	3:23:59 PM	436BB 21/05	A1HGR 21/06					
011990000881	t20	3:25:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
012002700111A	f2	3:52:40 PM	362H0 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02			



## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000657	phad	3:14:42 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000657	fx5	3:15:42 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000657	fx1	3:16:41 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000657	fx2	3:17:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000657	fx22	3:18:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000657	gx2	3:19:41 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000657	wx1	3:19:59 PM	436BB 21/05	A1HGR 21/06					
011990000657	t20	3:21:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	phad	3:06:42 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	fx5	3:07:42 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	fx1	3:08:42 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	fx2	3:09:42 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	fx22	3:10:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	gx2	3:11:41 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	wx1	3:12:41 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000658	t20	3:13:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	phad	2:58:42 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	fx5	2:59:42 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	fx1	3:00:42 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	fx2	3:01:42 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	fx22	3:02:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	gx2	3:03:42 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	wx1	3:04:42 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000659	t20	3:05:42 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	phad	2:42:43 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	fx5	2:43:43 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	fx1	2:44:43 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	fx2	2:45:42 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	fx22	2:46:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	gx2	2:47:43 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	wx1	2:48:43 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000660	t20	2:49:42 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	phad	2:50:42 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	fx5	2:51:42 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	fx1	2:52:42 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	fx2	2:53:42 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	fx22	2:54:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	gx2	2:55:42 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	wx1	2:56:42 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000654	t20	2:57:42 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	wx1	3:26:41 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

### Errors

Identity	Test	Measured	Error
011990000881	fx22	3:21:59 PM	306-Shortage of stop solution
011990000881	wx1	3:23:59 PM	306-Shortage of stop solution
011990000657	wx1	3:19:59 PM	306-Shortage of stop solution