


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200116-130
Instrument name:	Phadia-B
Instrument type:	Phadia 250
Status:	Ready
Started:	16-Jan-20 10:35:50 AM
Ended:	16-Jan-20 2:52:39 PM
Approved:	Approved (faziz)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-B
Instrument type:	Phadia 250
Measured:	23-Dec-19 3:51:37 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	18	0.00			
		3:51:37 PM	19	0.00	OK		
		3:52:36 PM	18	0.00	OK		
		%CV	4.5	>99			
CAL-0.35	0.35	Mean	147	0.35			
		3:53:37 PM	146	0.35	OK		
		3:54:36 PM	148	0.35	OK		
		%CV	0.6	0.7			
CAL-0.70	0.70	Mean	271	0.69			
		3:55:37 PM	274	0.70	OK		
		3:56:36 PM	268	0.68	OK		
		%CV	1.6	1.7			
CAL-3.50	3.50	Mean	1279	3.51			
		3:57:37 PM	1283	3.52	OK		
		3:58:36 PM	1275	3.50	OK		
		%CV	0.5	0.5			
CAL-17.5	17.5	Mean	5583	17.5			
		3:59:37 PM	5591	17.5	OK		
		4:00:36 PM	5575	17.5	OK		
		%CV	0.2	0.2			
CAL-100	100	Mean	17182	100			
		4:01:37 PM	17224	101	OK		
		4:02:36 PM	17140	99.4	OK		
		%CV	0.3	0.9			

Curve data	Value	Logged mean	Difference
ED-20	10.1	9.67	+3.97
ED-50	30.1	29.6	+1.76
ED-80	62.3	62.2	+0.20
Slope	211	216	-2.31

Number of curves included in mean: 52

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	10:35:50 AM	292	0.75	OK	Not
CC-2	17.5	10:36:50 AM	5779	18.2	OK	Not

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	10:37:50 AM	4003	11.9 kUA/l	CZ1AR	OK	Not
slgE L	e1	1	10:38:50 AM	1073	2.92 kUA/l	CZ4AN	OK	Not
slgE Neg	d1	1	10:39:08 AM	---	---	BL1CP	----	Not, Err
			Expected value:	---				
			Expected SD:	---				
			SD limit:	---				
			Range min:	---				
			Range max:	0.30				
			Limit CV%:	---				
IQC271119	d1	1	10:47:50 AM	5672	17.8 kUA/l	<lab>	----	Not

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
			<i>Expected value:</i>	---				
			<i>Expected SD:</i>	---				
			<i>SD limit:</i>	---				
			<i>Range min:</i>	---				
			<i>Range max:</i>	---				
			<i>Limit CV%:</i>	---				
slgE Neg	d1	1	10:51:49 AM	19	0.00 kUA/I	BL1CP	OK	Not

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000353	phad	5	11:53:47 AM	93	1.03 PAU/I	2	Positive	
011990000354	phad	5	12:08:47 PM	40	0.30 PAU/I	0	Negative	
011990000357	phad	5	1:49:43 PM	184	2.27 PAU/I	2	Positive	
011990000358	phad	5	1:41:43 PM	132	1.56 PAU/I	2	Positive	
011990000359	phad	5	2:30:41 PM	29	0.15 PAU/I	0	Negative	
011990000360	phad	5	2:38:41 PM	3242	47.2 PAU/I	4	Positive	
011990000363	phad	5	11:47:47 AM	30	0.17 PAU/I	0	Negative	
011990000364	phad	5	12:00:47 PM	224	2.81 PAU/I	2	Positive	
011990000367	phad	5	1:58:43 PM	33	0.21 PAU/I	0	Negative	
011990000368	phad	5	1:33:44 PM	50	0.44 PAU/I	1	Positive	
011990000374	phad	5	12:31:46 PM	35	0.23 PAU/I	0	Negative	
011990000378	phad	5	1:11:44 PM	36	0.26 PAU/I	0	Negative	
011990000379	phad	5	1:25:44 PM	136	1.61 PAU/I	2	Positive	
011990000383	phad	5	11:41:47 AM	38	0.27 PAU/I	0	Negative	
011990000384	phad	5	12:16:46 PM	34	0.22 PAU/I	0	Negative	
011990000387	phad	5	2:06:43 PM	256	3.26 PAU/I	2	Positive	
011990000388	phad	5	12:55:45 PM	33	0.21 PAU/I	0	Negative	
011990000389	phad	5	1:03:44 PM	83	0.89 PAU/I	2	Positive	
011990000393	phad	5	11:35:48 AM	32	0.19 PAU/I	0	Negative	
011990000394	phad	5	12:47:46 PM	1463	20.2 PAU/I	4	Positive	
011990000395	phad	5	12:39:46 PM	30	0.16 PAU/I	0	Negative	
011990000510	phad	5	2:22:42 PM	83	0.89 PAU/I	2	Positive	
011990000919	phad	5	2:14:42 PM	31	0.19 PAU/I	0	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000353	fx5	5	11:54:47 AM	52	0.46 kUA/I	1	Positive	
011990000353	fx1	5	11:55:47 AM	28	0.14 kUA/I	0	Negative	
011990000353	fx2	5	11:56:47 AM	254	3.22 kUA/I	2	Positive	
011990000353	fx22	5	11:57:47 AM	57	0.54 kUA/I	1	Positive	
011990000353	gx2	5	11:23:48 AM	50	0.44 kUA/I	1	Positive	
011990000353	wx1	5	11:58:47 AM	47	0.40 kUA/I	1	Positive	
011990000353	t20	5	11:59:47 AM	29	0.16 kUA/I	0		
011990000354	fx5	5	12:09:46 PM	22	0.06 kUA/I	0	Negative	
011990000354	fx1	5	12:10:46 PM	16	Low RU	0	Negative	
011990000354	fx2	5	12:11:46 PM	105	1.19 kUA/I	2	Positive	
011990000354	fx22	5	12:12:47 PM	32	0.20 kUA/I	0	Negative	
011990000354	gx2	5	12:13:47 PM	24	0.09 kUA/I	0	Negative	
011990000354	wx1	5	12:14:46 PM	31	0.18 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000354	t20	5	12:15:46 PM	18	Low RU	0		
011990000357	fx5	5	1:50:43 PM	41	0.32 kUA/l	0	Negative	
011990000357	fx1	5	1:51:43 PM	16	Low RU	0	Negative	
011990000357	fx2	5	1:52:43 PM	67	0.67 kUA/l	1	Positive	
011990000357	fx22	5	1:53:43 PM	29	0.15 kUA/l	0	Negative	
011990000357	gx2	5	1:54:43 PM	25	0.09 kUA/l	0	Negative	
011990000357	wx1	5	1:55:43 PM	245	3.10 kUA/l	2	Positive	
011990000357	t20	5	1:56:43 PM	32	0.19 kUA/l	0		
011990000358	fx5	5	1:42:43 PM	37	0.26 kUA/l	0	Negative	
011990000358	fx1	5	1:43:43 PM	28	0.15 kUA/l	0	Negative	
011990000358	fx2	5	1:44:43 PM	32	0.19 kUA/l	0	Negative	
011990000358	fx22	5	1:45:43 PM	28	0.13 kUA/l	0	Negative	
011990000358	gx2	5	1:46:43 PM	21	0.04 kUA/l	0	Negative	
011990000358	wx1	5	1:47:43 PM	36	0.24 kUA/l	0	Negative	
011990000358	t20	5	1:48:43 PM	19	0.02 kUA/l	0		
011990000359	fx5	5	2:31:42 PM	21	0.04 kUA/l	0	Negative	
011990000359	fx1	5	2:32:42 PM	19	0.02 kUA/l	0	Negative	
011990000359	fx2	5	2:33:41 PM	11	Low RU	0	Negative	
011990000359	fx22	5	2:34:41 PM	27	0.12 kUA/l	0	Negative	
011990000359	gx2	5	2:35:41 PM	19	0.01 kUA/l	0	Negative	
011990000359	wx1	5	2:36:41 PM	26	0.12 kUA/l	0	Negative	
011990000359	t20	5	2:37:41 PM	15	Low RU	0		
011990000360	fx5	5	2:39:41 PM	79	0.83 kUA/l	2	Positive	
011990000360	fx1	5	2:40:41 PM	118	1.37 kUA/l	2	Positive	
011990000363	fx5	5	11:48:47 AM	20	0.03 kUA/l	0	Negative	
011990000363	fx1	5	11:49:47 AM	16	Low RU	0	Negative	
011990000363	fx2	5	11:50:47 AM	11	Low RU	0	Negative	
011990000363	fx22	5	11:51:47 AM	29	0.16 kUA/l	0	Negative	
011990000363	gx2	5	11:21:49 AM	22	0.06 kUA/l	0	Negative	
011990000363	wx1	5	11:52:47 AM	27	0.13 kUA/l	0	Negative	
011990000363	t20	5	11:22:48 AM	17	Low RU	0		
011990000364	fx5	5	12:01:47 PM	27	0.13 kUA/l	0	Negative	
011990000364	fx1	5	12:02:47 PM	31	0.18 kUA/l	0	Negative	
011990000364	fx2	5	12:03:47 PM	102	1.14 kUA/l	2	Positive	
011990000364	fx22	5	12:04:47 PM	39	0.29 kUA/l	0	Negative	
011990000364	gx2	5	12:05:47 PM	25	0.10 kUA/l	0	Negative	
011990000364	wx1	5	12:06:47 PM	505	6.68 kUA/l	3	Positive	
011990000364	t20	5	12:07:47 PM	54	0.50 kUA/l	1		
011990000367	fx5	5	1:59:43 PM	28	0.14 kUA/l	0	Negative	
011990000367	fx1	5	2:00:43 PM	17	Low RU	0	Negative	
011990000367	fx2	5	2:01:42 PM	18	0.01 kUA/l	0	Negative	
011990000367	fx22	5	2:02:42 PM	28	0.14 kUA/l	0	Negative	
011990000367	gx2	5	2:03:42 PM	21	0.05 kUA/l	0	Negative	
011990000367	wx1	5	2:04:42 PM	26	0.11 kUA/l	0	Negative	
011990000367	t20	5	2:05:43 PM	16	Low RU	0		
011990000368	fx5	5	1:34:43 PM	25	0.10 kUA/l	0	Negative	
011990000368	fx1	5	1:35:43 PM	23	0.07 kUA/l	0	Negative	
011990000368	fx2	5	1:36:43 PM	24	0.08 kUA/l	0	Negative	
011990000368	fx22	5	1:37:43 PM	37	0.26 kUA/l	0	Negative	
011990000368	gx2	5	1:38:43 PM	32	0.20 kUA/l	0	Negative	
011990000368	wx1	5	1:39:44 PM	32	0.20 kUA/l	0	Negative	
011990000368	t20	5	1:40:43 PM	18	0.01 kUA/l	0		
011990000374	fx5	5	12:32:46 PM	18	Low RU	0	Negative	
011990000374	fx1	5	12:33:46 PM	17	Low RU	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000374	fx2	5	12:34:46 PM	19	0.02 kUA/l	0	Negative	
011990000374	fx22	5	12:35:46 PM	30	0.17 kUA/l	0	Negative	
011990000374	gx2	5	12:36:45 PM	21	0.05 kUA/l	0	Negative	
011990000374	wx1	5	12:37:45 PM	31	0.17 kUA/l	0	Negative	
011990000374	t20	5	12:38:46 PM	17	Low RU	0		
011990000378	fx5	5	1:12:44 PM	21	0.04 kUA/l	0	Negative	
011990000378	fx1	5	1:13:44 PM	16	Low RU	0	Negative	
011990000378	fx2	5	1:14:44 PM	13	Low RU	0	Negative	
011990000378	fx22	5	1:15:44 PM	28	0.14 kUA/l	0	Negative	
011990000378	gx2	5	1:16:44 PM	20	0.02 kUA/l	0	Negative	
011990000378	wx1	5	1:17:44 PM	27	0.13 kUA/l	0	Negative	
011990000378	t20	5	1:18:44 PM	16	Low RU	0		
011990000379	fx5	5	1:26:44 PM	80	0.85 kUA/l	2	Positive	
011990000379	fx1	5	1:27:44 PM	16	Low RU	0	Negative	
011990000379	fx2	5	1:28:44 PM	951	12.9 kUA/l	3	Positive	
011990000379	fx22	5	1:29:43 PM	30	0.17 kUA/l	0	Negative	
011990000379	gx2	5	1:30:44 PM	54	0.49 kUA/l	1	Positive	
011990000379	wx1	5	1:31:44 PM	32	0.19 kUA/l	0	Negative	
011990000379	t20	5	1:32:44 PM	16	Low RU	0		
011990000383	fx5	5	11:42:47 AM	21	0.04 kUA/l	0	Negative	
011990000383	fx1	5	11:43:47 AM	14	Low RU	0	Negative	
011990000383	fx2	5	11:44:47 AM	13	Low RU	0	Negative	
011990000383	fx22	5	11:45:48 AM	26	0.11 kUA/l	0	Negative	
011990000383	gx2	5	11:19:48 AM	22	0.05 kUA/l	0	Negative	
011990000383	wx1	5	11:46:48 AM	29	0.15 kUA/l	0	Negative	
011990000383	t20	5	11:20:49 AM	15	Low RU	0		
011990000384	fx5	5	12:17:46 PM	22	0.06 kUA/l	0	Negative	
011990000384	fx1	5	12:18:46 PM	14	Low RU	0	Negative	
011990000384	fx2	5	12:19:46 PM	12	Low RU	0	Negative	
011990000384	fx22	5	12:20:46 PM	29	0.16 kUA/l	0	Negative	
011990000384	gx2	5	12:28:46 PM	20	0.03 kUA/l	0	Negative	
011990000384	wx1	5	12:29:46 PM	31	0.19 kUA/l	0	Negative	
011990000384	t20	5	12:30:46 PM	16	Low RU	0		
011990000387	fx5	5	2:07:42 PM	18	Low RU	0	Negative	
011990000387	fx1	5	2:08:42 PM	15	Low RU	0	Negative	
011990000387	fx2	5	2:09:42 PM	514	6.80 kUA/l	3	Positive	
011990000387	fx22	5	2:10:42 PM	27	0.12 kUA/l	0	Negative	
011990000387	gx2	5	2:11:42 PM	21	0.05 kUA/l	0	Negative	
011990000387	wx1	5	2:12:42 PM	29	0.15 kUA/l	0	Negative	
011990000387	t20	5	2:13:42 PM	15	Low RU	0		
011990000388	fx5	5	12:56:45 PM	17	Low RU	0	Negative	
011990000388	fx1	5	12:57:45 PM	20	0.03 kUA/l	0	Negative	
011990000388	fx2	5	12:58:45 PM	14	Low RU	0	Negative	
011990000388	fx22	5	12:59:45 PM	26	0.11 kUA/l	0	Negative	
011990000388	gx2	5	1:00:45 PM	21	0.04 kUA/l	0	Negative	
011990000388	wx1	5	1:01:45 PM	27	0.13 kUA/l	0	Negative	
011990000388	t20	5	1:02:45 PM	16	Low RU	0		
011990000389	fx5	5	1:04:45 PM	17	Low RU	0	Negative	
011990000389	fx1	5	1:05:45 PM	16	Low RU	0	Negative	
011990000389	fx2	5	1:06:45 PM	83	0.89 kUA/l	2	Positive	
011990000389	fx22	5	1:07:45 PM	26	0.12 kUA/l	0	Negative	
011990000389	gx2	5	1:08:44 PM	21	0.04 kUA/l	0	Negative	
011990000389	wx1	5	1:09:44 PM	40	0.31 kUA/l	0	Negative	
011990000389	t20	5	1:10:44 PM	22	0.05 kUA/l	0		

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000393	fx5	5	11:36:48 AM	19	0.02 kUA/l	0	Negative	
011990000393	fx1	5	11:37:48 AM	13	Low RU	0	Negative	
011990000393	fx2	5	11:38:48 AM	11	Low RU	0	Negative	
011990000393	fx22	5	11:39:48 AM	26	0.11 kUA/l	0	Negative	
011990000393	gx2	5	11:17:48 AM	23	0.07 kUA/l	0	Negative	
011990000393	wx1	5	11:40:48 AM	27	0.13 kUA/l	0	Negative	
011990000393	t20	5	11:18:48 AM	16	Low RU	0		
011990000394	fx5	5	12:48:45 PM	25	0.10 kUA/l	0	Negative	
011990000394	fx1	5	12:49:45 PM	18	Low RU	0	Negative	
011990000394	fx2	5	12:50:45 PM	312	4.02 kUA/l	3	Positive	
011990000394	fx22	5	12:51:45 PM	32	0.19 kUA/l	0	Negative	
011990000394	gx2	5	12:52:45 PM	24	0.08 kUA/l	0	Negative	
011990000394	wx1	5	12:53:45 PM	29	0.16 kUA/l	0	Negative	
011990000394	t20	5	12:54:45 PM	18	Low RU	0		
011990000395	fx5	5	12:40:45 PM	18	Low RU	0	Negative	
011990000395	fx1	5	12:41:45 PM	15	Low RU	0	Negative	
011990000395	fx2	5	12:42:45 PM	14	Low RU	0	Negative	
011990000395	fx22	5	12:43:45 PM	30	0.17 kUA/l	0	Negative	
011990000395	gx2	5	12:44:45 PM	17	---	---	Negative	
011990000395	wx1	5	12:45:45 PM	26	0.11 kUA/l	0	Negative	
011990000395	t20	5	12:46:45 PM	15	Low RU	0		
011990000510	fx5	5	2:23:42 PM	20	0.03 kUA/l	0	Negative	
011990000510	fx1	5	2:24:42 PM	18	Low RU	0	Negative	
011990000510	fx2	5	2:25:42 PM	200	2.49 kUA/l	2	Positive	
011990000510	fx22	5	2:26:41 PM	33	0.21 kUA/l	0	Negative	
011990000510	gx2	5	2:27:42 PM	24	0.09 kUA/l	0	Negative	
011990000510	wx1	5	2:28:42 PM	39	0.28 kUA/l	0	Negative	
011990000510	t20	5	2:29:41 PM	27	0.13 kUA/l	0		
011990000919	fx5	5	2:15:42 PM	20	0.02 kUA/l	0	Negative	
011990000919	fx1	5	2:16:42 PM	27	0.12 kUA/l	0	Negative	
011990000919	fx2	5	2:17:42 PM	10	Low RU	0	Negative	
011990000919	fx22	5	2:18:42 PM	25	0.10 kUA/l	0	Negative	
011990000919	gx2	5	2:19:42 PM	20	0.02 kUA/l	0	Negative	
011990000919	wx1	5	2:20:42 PM	26	0.11 kUA/l	0	Negative	
011990000919	t20	5	2:21:42 PM	14	Low RU	0		
012001401046A	f425	1	11:00:49 AM	1020	2.77 kUA/l			
012001401046A	f428	1	11:01:49 AM	1056	2.87 kUA/l			
012001401046A	f439	1	11:02:49 AM	930	2.52 kUA/l			
012001401046A	f440	1	11:03:49 AM	1616	4.48 kUA/l			
012001401046A	f422	1	11:04:49 AM	1440	3.97 kUA/l			
012001401046A	f423	1	11:05:49 AM	919	2.49 kUA/l			
012001401046A	f424	1	11:06:49 AM	1104	3.01 kUA/l			
012001401046A	f427	1	11:07:49 AM	902	2.44 kUA/l			
012001401046A	f352	1	11:08:49 AM	1168	3.19 kUA/l			
012001500838A	f20	1	10:52:49 AM	22	0.01 kUA/l			
012001500838A	f79	1	10:53:49 AM	48	0.08 kUA/l			
012001500838A	f17	1	10:54:50 AM	19	0.00 kUA/l			
012001500838A	d1	1	10:55:50 AM	20	0.00 kUA/l			
012001500838A	f2	1	10:56:49 AM	525	1.39 kUA/l			
012001500838A	f40	1	10:57:49 AM	15	0.00 kUA/l			
012001500875A	f245	1	10:40:50 AM	10	0.00 kUA/l			
012001500875A	f2	1	10:41:50 AM	14	0.00 kUA/l			
012001500875A	f24	1	10:42:50 AM	15	0.00 kUA/l			
012001500875A	f14	1	10:48:50 AM	11	0.00 kUA/l			

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
012001500875A	f4	1	10:49:49 AM	17	0.00 kUA/l			
012001500875A	f13	1	10:50:50 AM	13	0.00 kUA/l			

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	3:51:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	3:52:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:53:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	3:54:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:55:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	3:56:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:57:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	3:58:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	3:59:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	4:00:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:01:37 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	4:02:36 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	10:35:50 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	10:36:50 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	10:37:50 AM	352D0 19/12/1	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
slgE L	e1	10:38:50 AM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06/1		Exp
slgE Neg	d1	10:39:08 AM	357J6 21/01	A1HGR 21/06				BL1CP 19/12/1		Exp
IQC271119	d1	10:47:50 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		
slgE Neg	d1	10:51:49 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12/1		Exp

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012001500838A	f20	10:52:49 AM	422CS 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500838A	f79	10:53:49 AM	525JR 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500838A	f17	10:54:50 AM	D2FBX 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500838A	d1	10:55:50 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500838A	f2	10:56:49 AM	362H0 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500838A	f40	10:57:49 AM	455DT 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500875A	f245	10:40:50 AM	925HU 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500875A	f2	10:41:50 AM	362H0 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500875A	f24	10:42:50 AM	424EF 20/11	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500875A	f14	10:48:50 AM	364DN 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500875A	f4	10:49:49 AM	363F1 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001500875A	f13	10:50:50 AM	375MS 20/04	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f425	11:00:49 AM	CDPGE 20/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f428	11:01:49 AM	CFBGD 20/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f439	11:02:49 AM	CZPGE 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f440	11:03:49 AM	D0MGF 20/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f422	11:04:49 AM	CDFH3 20/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f423	11:05:49 AM	CDGHN 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f424	11:06:49 AM	CDHH0 20/09	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f427	11:07:49 AM	CFCHO 20/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001401046A	f352	11:08:49 AM	CEZH5 20/12	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
011990000394	phad	12:47:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000394	fx5	12:48:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000394	fx1	12:49:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000394	fx2	12:50:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000394	fx22	12:51:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000394	gx2	12:52:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000394	wx1	12:53:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000394	t20	12:54:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000395	phad	12:39:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000395	fx5	12:40:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000395	fx1	12:41:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000395	fx2	12:42:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000395	fx22	12:43:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000395	gx2	12:44:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000395	wx1	12:45:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000395	t20	12:46:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	phad	12:31:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	fx5	12:32:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	fx1	12:33:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	fx2	12:34:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	fx22	12:35:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	gx2	12:36:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	wx1	12:37:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000374	t20	12:38:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	phad	12:16:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	fx5	12:17:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	fx1	12:18:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	fx2	12:19:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	fx22	12:20:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	gx2	12:28:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	wx1	12:29:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000384	t20	12:30:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	phad	12:08:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	fx5	12:09:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	fx1	12:10:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	fx2	12:11:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	fx22	12:12:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	gx2	12:13:47 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	wx1	12:14:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000354	t20	12:15:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	phad	12:00:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	fx5	12:01:47 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	fx1	12:02:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	fx2	12:03:47 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	fx22	12:04:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	gx2	12:05:47 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	wx1	12:06:47 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000364	t20	12:07:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	phad	11:53:47 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	fx5	11:54:47 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	fx1	11:55:47 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	fx2	11:56:47 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	fx22	11:57:47 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	gx2	11:23:48 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	wx1	11:58:47 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000353	t20	11:59:47 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	phad	11:47:47 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	fx5	11:48:47 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	fx1	11:49:47 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	fx2	11:50:47 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	fx22	11:51:47 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	gx2	11:21:49 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	wx1	11:52:47 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000363	t20	11:22:48 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	phad	11:41:47 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	fx5	11:42:47 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	fx1	11:43:47 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	fx2	11:44:47 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	fx22	11:45:48 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	gx2	11:19:48 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	wx1	11:46:48 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000383	t20	11:20:49 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000393	phad	11:35:48 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000393	fx5	11:36:48 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000393	fx1	11:37:48 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000393	fx2	11:38:48 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000393	fx22	11:39:48 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000393	gx2	11:17:48 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000393	wx1	11:40:48 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000393	t20	11:18:48 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	phad	12:55:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	fx5	12:56:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	fx1	12:57:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	fx2	12:58:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	fx22	12:59:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	gx2	1:00:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	wx1	1:01:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000388	t20	1:02:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	phad	1:03:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	fx5	1:04:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	fx1	1:05:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	fx2	1:06:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	fx22	1:07:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	gx2	1:08:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	wx1	1:09:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000389	t20	1:10:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	phad	1:11:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	fx5	1:12:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	fx1	1:13:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	fx2	1:14:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	fx22	1:15:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	gx2	1:16:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	wx1	1:17:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000378	t20	1:18:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	phad	1:25:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	fx5	1:26:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	fx1	1:27:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	fx2	1:28:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	fx22	1:29:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	gx2	1:30:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	wx1	1:31:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000379	t20	1:32:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	phad	1:33:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	fx5	1:34:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	fx1	1:35:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	fx2	1:36:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	fx22	1:37:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	gx2	1:38:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	wx1	1:39:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000368	t20	1:40:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	phad	1:41:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	fx5	1:42:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	fx1	1:43:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	fx2	1:44:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	fx22	1:45:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	gx2	1:46:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	wx1	1:47:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000358	t20	1:48:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	phad	1:49:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	fx5	1:50:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	fx1	1:51:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	fx2	1:52:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	fx22	1:53:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	gx2	1:54:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	wx1	1:55:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000357	t20	1:56:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000367	phad	1:58:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000367	fx5	1:59:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000367	fx1	2:00:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000367	fx2	2:01:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000367	fx22	2:02:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000367	gx2	2:03:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000367	wx1	2:04:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000367	t20	2:05:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	phad	2:06:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	fx5	2:07:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	fx1	2:08:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	fx2	2:09:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	fx22	2:10:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	gx2	2:11:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	wx1	2:12:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000387	t20	2:13:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	phad	2:14:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	fx5	2:15:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	fx1	2:16:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	fx2	2:17:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	fx22	2:18:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	gx2	2:19:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	wx1	2:20:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000919	t20	2:21:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	phad	2:22:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	fx5	2:23:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	fx1	2:24:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	fx2	2:25:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	fx22	2:26:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	gx2	2:27:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	wx1	2:28:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000510	t20	2:29:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	phad	2:30:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	fx5	2:31:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	fx1	2:32:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	fx2	2:33:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	fx22	2:34:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	gx2	2:35:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	wx1	2:36:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000359	t20	2:37:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000360	phad	2:38:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000360	fx5	2:39:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000360	fx1	2:40:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Errors

Identity	Test	Measured	Error
slgE Neg	d1	10:39:08 AM	402-Error in sample pipetting module

Expired lot numbers used for Quality Control

352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial

Notes

Identity	Test	Measured	Comment
CC-1	a_IgE	10:35:50 AM	Ok - RJB
CC-2	a_IgE	10:36:50 AM	Ok - RJB
slgE M	t3	10:37:50 AM	Ok - RJB
slgE L	e1	10:38:50 AM	Ok - RJB
slgE Neg	d1	10:39:08 AM	repeated - RJB
IQC271119	d1	10:47:50 AM	Ok - RJB



Laboratory Report

Identity	Test	Measured	Comment
slgE Neg	d1	10:51:49 AM	Ok - RJB