


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200127-387  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 27-Jan-20 11:16:24 AM  
Ended: 27-Jan-20 2:55:18 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	11:16:24 AM	326	0.72	OK	
CC-2	17.5	11:17:24 AM	6537	18.3	OK	

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	11:18:24 AM	3476	8.71 kUA/l	CZ1AR	OK	
slgE Neg	d1	1	11:19:24 AM	23	0.01 kUA/l	BL1CP	OK	
slgE L	e1	1	11:20:24 AM	1400	3.28 kUA/l	CZ4AS	OK	
IQC271119	d1	1	11:21:24 AM	6293	17.5 kUA/l	<lab>	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000401	phad	5	1:02:20 PM	30	0.16 PAU/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000402	phad	5	12:03:23 PM	36	0.22 PAU/I	0	Negative	
011990000403	phad	5	11:22:24 AM	54	0.44 PAU/I	1	Positive	
011990000413	phad	5	12:31:22 PM	48	0.37 PAU/I	1	Positive	
011990000414	phad	5	11:56:23 AM	33	0.19 PAU/I	0	Negative	
011990000415	phad	5	11:41:23 AM	379	4.21 PAU/I	3	Positive	
011990000425	phad	5	12:24:22 PM	268	2.91 PAU/I	2	Positive	
011990000426	phad	5	11:25:24 AM	55	0.44 PAU/I	1	Positive	
011990000427	phad	5	11:49:23 AM	84	0.79 PAU/I	2	Positive	
011990000437	phad	5	12:17:22 PM	32	0.18 PAU/I	0	Negative	
011990000449	phad	5	12:10:22 PM	34	0.21 PAU/I	0	Negative	
011990000776	phad	5	1:42:19 PM	35	0.21 PAU/I	0	Negative	
011990000777	phad	5	1:34:20 PM	712	8.11 PAU/I	3	Positive	
011990000778	phad	5	1:26:20 PM	44	0.33 PAU/I	0	Negative	
011990000779	phad	5	1:18:20 PM	101	0.98 PAU/I	2	Positive	
011990000780	phad	5	1:10:20 PM	1570	18.5 PAU/I	4	Positive	
011990000781	phad	5	2:14:18 PM	34	0.20 PAU/I	0	Negative	
011990000782	phad	5	2:06:18 PM	370	4.10 PAU/I	3	Positive	
011990000783	phad	5	1:58:18 PM	68	0.60 PAU/I	1	Positive	
011990000785	phad	5	1:50:19 PM	45	0.33 PAU/I	0	Negative	
011990000786	phad	5	2:38:17 PM	34	0.20 PAU/I	0	Negative	
011990000788	phad	5	2:54:16 PM	290	3.18 PAU/I	2	Positive	
011990000789	phad	5	2:46:16 PM	31	0.16 PAU/I	0	Negative	
011990000790	phad	5	2:22:18 PM	2776	34.0 PAU/I	4	Positive	
011990000795	phad	5	2:30:17 PM	59	0.50 PAU/I	1	Positive	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000401	fx5	5	1:03:20 PM	25	0.09 kUA/I	0	Negative	
011990000401	fx1	5	1:04:20 PM	21	0.05 kUA/I	0	Negative	
011990000401	fx2	5	1:05:20 PM	12	Low RU	0	Negative	
011990000401	fx22	5	1:06:20 PM	31	0.17 kUA/I	0	Negative	
011990000401	gx2	5	1:07:20 PM	25	0.10 kUA/I	0	Negative	
011990000401	wx1	5	1:08:20 PM	29	0.14 kUA/I	0	Negative	
011990000401	t20	5	1:09:20 PM	18	0.01 kUA/I	0		
011990000402	fx5	5	12:04:23 PM	20	0.04 kUA/I	0	Negative	
011990000402	fx1	5	12:05:22 PM	21	0.05 kUA/I	0	Negative	
011990000402	fx2	5	12:06:22 PM	22	0.06 kUA/I	0	Negative	
011990000402	fx22	5	12:52:21 PM	30	0.16 kUA/I	0	Negative	
011990000402	gx2	5	12:07:22 PM	22	0.06 kUA/I	0	Negative	
011990000402	wx1	5	12:08:22 PM	30	0.16 kUA/I	0	Negative	
011990000402	t20	5	12:09:22 PM	17	Low RU	0		
011990000403	fx5	5	11:23:24 AM	21	0.05 kUA/I	0	Negative	
011990000403	fx1	5	11:24:24 AM	20	0.03 kUA/I	0	Negative	
011990000403	fx2	5	11:36:23 AM	13	Low RU	0	Negative	
011990000403	fx22	5	11:37:23 AM	30	0.15 kUA/I	0	Negative	
011990000403	gx2	5	11:38:23 AM	19	0.02 kUA/I	0	Negative	
011990000403	wx1	5	11:39:23 AM	27	0.12 kUA/I	0	Negative	
011990000403	t20	5	11:40:23 AM	17	Low RU	0		
011990000413	fx5	5	12:32:21 PM	26	0.11 kUA/I	0	Negative	
011990000413	fx1	5	12:56:21 PM	18	0.01 kUA/I	0	Negative	
011990000413	fx2	5	12:57:21 PM	21	0.05 kUA/I	0	Negative	
011990000413	fx22	5	12:58:21 PM	31	0.17 kUA/I	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000413	gx2	5	12:59:20 PM	21	0.05 kUA/l	0	Negative	
011990000413	wx1	5	1:00:20 PM	272	2.97 kUA/l	2	Positive	
011990000413	t20	5	1:01:20 PM	18	0.01 kUA/l	0		
011990000414	fx5	5	11:57:23 AM	21	0.05 kUA/l	0	Negative	
011990000414	fx1	5	11:58:23 AM	21	0.05 kUA/l	0	Negative	
011990000414	fx2	5	11:59:22 AM	14	Low RU	0	Negative	
011990000414	fx22	5	12:51:21 PM	33	0.19 kUA/l	0	Negative	
011990000414	gx2	5	12:00:22 PM	23	0.07 kUA/l	0	Negative	
011990000414	wx1	5	12:01:23 PM	29	0.14 kUA/l	0	Negative	
011990000414	t20	5	12:02:23 PM	16	Low RU	0		
011990000415	fx5	5	11:42:24 AM	60	0.51 kUA/l	1	Positive	
011990000415	fx1	5	11:43:23 AM	36	0.23 kUA/l	0	Negative	
011990000415	fx2	5	11:44:23 AM	227	2.45 kUA/l	2	Positive	
011990000415	fx22	5	11:45:23 AM	42	0.30 kUA/l	0	Negative	
011990000415	gx2	5	11:46:23 AM	38	0.25 kUA/l	0	Negative	
011990000415	wx1	5	11:47:23 AM	84	0.79 kUA/l	2	Positive	
011990000415	t20	5	11:48:23 AM	42	0.30 kUA/l	0		
011990000425	fx5	5	12:25:21 PM	37	0.24 kUA/l	0	Negative	
011990000425	fx1	5	12:26:22 PM	31	0.17 kUA/l	0	Negative	
011990000425	fx2	5	12:27:21 PM	20	0.04 kUA/l	0	Negative	
011990000425	fx22	5	12:55:21 PM	42	0.30 kUA/l	0	Negative	
011990000425	gx2	5	12:28:21 PM	265	2.88 kUA/l	2	Positive	
011990000425	wx1	5	12:29:22 PM	73	0.66 kUA/l	1	Positive	
011990000425	t20	5	12:30:22 PM	63	0.54 kUA/l	1		
011990000426	fx5	5	11:26:24 AM	34	0.21 kUA/l	0	Negative	
011990000426	fx1	5	11:27:24 AM	39	0.26 kUA/l	0	Negative	
011990000426	fx2	5	11:28:24 AM	71	0.64 kUA/l	1	Positive	
011990000426	fx22	5	12:46:21 PM	38	0.25 kUA/l	0	Negative	
011990000426	gx2	5	11:29:24 AM	107	1.05 kUA/l	2	Positive	
011990000426	wx1	5	11:30:24 AM	77	0.70 kUA/l	2	Positive	
011990000426	t20	5	12:47:21 PM	23	0.08 kUA/l	0		
011990000427	fx5	5	11:50:23 AM	20	0.04 kUA/l	0	Negative	
011990000427	fx1	5	11:51:23 AM	18	0.02 kUA/l	0	Negative	
011990000427	fx2	5	11:52:23 AM	13	Low RU	0	Negative	
011990000427	fx22	5	12:50:21 PM	31	0.17 kUA/l	0	Negative	
011990000427	gx2	5	11:53:23 AM	225	2.42 kUA/l	2	Positive	
011990000427	wx1	5	11:54:23 AM	27	0.12 kUA/l	0	Negative	
011990000427	t20	5	11:55:23 AM	15	Low RU	0		
011990000437	fx5	5	12:18:22 PM	21	0.05 kUA/l	0	Negative	
011990000437	fx1	5	12:19:22 PM	17	0.01 kUA/l	0	Negative	
011990000437	fx2	5	12:20:22 PM	12	Low RU	0	Negative	
011990000437	fx22	5	12:54:21 PM	30	0.16 kUA/l	0	Negative	
011990000437	gx2	5	12:21:22 PM	22	0.06 kUA/l	0	Negative	
011990000437	wx1	5	12:22:22 PM	29	0.14 kUA/l	0	Negative	
011990000437	t20	5	12:23:22 PM	16	Low RU	0		
011990000438	fx2	5	11:33:24 AM	13	Low RU	0	Negative	
011990000438	fx22	5	12:48:21 PM	32	0.18 kUA/l	0	Negative	
011990000438	gx2	5	11:34:24 AM	22	0.06 kUA/l	0	Negative	
011990000438	wx1	5	11:35:23 AM	30	0.16 kUA/l	0	Negative	
011990000438	t20	5	12:49:21 PM	16	Low RU	0		
011990000439	fx2	5	11:32:24 AM	30	0.16 kUA/l	0	Negative	
011990000449	fx5	5	12:11:22 PM	25	0.10 kUA/l	0	Negative	
011990000449	fx1	5	12:12:22 PM	14	Low RU	0	Negative	
011990000449	fx2	5	12:13:22 PM	19	0.03 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000449	fx22	5	12:53:21 PM	32	0.18 kUA/l	0	Negative	
011990000449	gx2	5	12:14:22 PM	21	0.05 kUA/l	0	Negative	
011990000449	wx1	5	12:15:22 PM	27	0.13 kUA/l	0	Negative	
011990000449	t20	5	12:16:22 PM	16	Low RU	0		
011990000450	fx2	5	11:31:24 AM	22	0.06 kUA/l	0	Negative	
011990000776	fx5	5	1:43:19 PM	24	0.08 kUA/l	0	Negative	
011990000776	fx1	5	1:44:19 PM	17	Low RU	0	Negative	
011990000776	fx2	5	1:45:19 PM	14	Low RU	0	Negative	
011990000776	fx22	5	1:46:19 PM	33	0.20 kUA/l	0	Negative	
011990000776	gx2	5	1:47:19 PM	24	0.08 kUA/l	0	Negative	
011990000776	wx1	5	1:48:19 PM	31	0.17 kUA/l	0	Negative	
011990000776	t20	5	1:49:19 PM	17	0.01 kUA/l	0		
011990000777	fx5	5	1:35:19 PM	28	0.13 kUA/l	0	Negative	
011990000777	fx1	5	1:36:19 PM	19	0.03 kUA/l	0	Negative	
011990000777	fx2	5	1:37:19 PM	658	7.47 kUA/l	3	Positive	
011990000777	fx22	5	1:38:19 PM	32	0.18 kUA/l	0	Negative	
011990000777	gx2	5	1:39:19 PM	50	0.39 kUA/l	1	Positive	
011990000777	wx1	5	1:40:19 PM	143	1.47 kUA/l	2	Positive	
011990000777	t20	5	1:41:19 PM	117	1.16 kUA/l	2		
011990000778	fx5	5	1:27:20 PM	48	0.37 kUA/l	1	Positive	
011990000778	fx1	5	1:28:19 PM	31	0.17 kUA/l	0	Negative	
011990000778	fx2	5	1:29:19 PM	17	Low RU	0	Negative	
011990000778	fx22	5	1:30:19 PM	36	0.23 kUA/l	0	Negative	
011990000778	gx2	5	1:31:19 PM	30	0.15 kUA/l	0	Negative	
011990000778	wx1	5	1:32:19 PM	39	0.26 kUA/l	0	Negative	
011990000778	t20	5	1:33:19 PM	24	0.08 kUA/l	0		
011990000779	fx5	5	1:19:20 PM	22	0.06 kUA/l	0	Negative	
011990000779	fx1	5	1:20:20 PM	36	0.23 kUA/l	0	Negative	
011990000779	fx2	5	1:21:20 PM	55	0.45 kUA/l	1	Positive	
011990000779	fx22	5	1:22:20 PM	29	0.15 kUA/l	0	Negative	
011990000779	gx2	5	1:23:20 PM	27	0.12 kUA/l	0	Negative	
011990000779	wx1	5	1:24:20 PM	56	0.46 kUA/l	1	Positive	
011990000779	t20	5	1:25:20 PM	24	0.09 kUA/l	0		
011990000780	fx5	5	1:11:20 PM	21	0.05 kUA/l	0	Negative	
011990000780	fx1	5	1:12:20 PM	25	0.10 kUA/l	0	Negative	
011990000780	fx2	5	1:13:20 PM	18	0.01 kUA/l	0	Negative	
011990000780	fx22	5	1:14:20 PM	35	0.22 kUA/l	0	Negative	
011990000780	gx2	5	1:15:20 PM	96	0.93 kUA/l	2	Positive	
011990000780	wx1	5	1:16:20 PM	52	0.42 kUA/l	1	Positive	
011990000780	t20	5	1:17:20 PM	20	0.04 kUA/l	0		
011990000781	fx5	5	2:15:18 PM	24	0.09 kUA/l	0	Negative	
011990000781	fx1	5	2:16:18 PM	18	0.02 kUA/l	0	Negative	
011990000781	fx2	5	2:17:18 PM	14	Low RU	0	Negative	
011990000781	fx22	5	2:18:17 PM	29	0.14 kUA/l	0	Negative	
011990000781	gx2	5	2:19:17 PM	18	0.01 kUA/l	0	Negative	
011990000781	wx1	5	2:20:17 PM	28	0.14 kUA/l	0	Negative	
011990000781	t20	5	2:21:18 PM	18	0.02 kUA/l	0		
011990000782	fx5	5	2:07:18 PM	22	0.06 kUA/l	0	Negative	
011990000782	fx1	5	2:08:18 PM	18	0.01 kUA/l	0	Negative	
011990000782	fx2	5	2:09:18 PM	20	0.04 kUA/l	0	Negative	
011990000782	fx22	5	2:10:18 PM	32	0.17 kUA/l	0	Negative	
011990000782	gx2	5	2:11:18 PM	25	0.10 kUA/l	0	Negative	
011990000782	wx1	5	2:12:18 PM	29	0.15 kUA/l	0	Negative	
011990000782	t20	5	2:13:18 PM	18	0.02 kUA/l	0		

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000783	fx5	5	1:59:18 PM	21	0.05 kUA/l	0	Negative	
011990000783	fx1	5	2:00:18 PM	17	Low RU	0	Negative	
011990000783	fx2	5	2:01:18 PM	33	0.19 kUA/l	0	Negative	
011990000783	fx22	5	2:02:18 PM	29	0.15 kUA/l	0	Negative	
011990000783	gx2	5	2:03:18 PM	25	0.10 kUA/l	0	Negative	
011990000783	wx1	5	2:04:18 PM	34	0.20 kUA/l	0	Negative	
011990000783	t20	5	2:05:18 PM	20	0.04 kUA/l	0		
011990000785	fx5	5	1:51:18 PM	23	0.07 kUA/l	0	Negative	
011990000785	fx1	5	1:52:18 PM	19	0.02 kUA/l	0	Negative	
011990000785	fx2	5	1:53:19 PM	28	0.13 kUA/l	0	Negative	
011990000785	fx22	5	1:54:19 PM	33	0.20 kUA/l	0	Negative	
011990000785	gx2	5	1:55:18 PM	19	0.03 kUA/l	0	Negative	
011990000785	wx1	5	1:56:18 PM	29	0.15 kUA/l	0	Negative	
011990000785	t20	5	1:57:18 PM	18	0.02 kUA/l	0		
011990000786	fx5	5	2:39:17 PM	21	0.05 kUA/l	0	Negative	
011990000786	fx1	5	2:40:17 PM	14	Low RU	0	Negative	
011990000786	fx2	5	2:41:17 PM	15	Low RU	0	Negative	
011990000786	fx22	5	2:42:17 PM	27	0.12 kUA/l	0	Negative	
011990000786	gx2	5	2:43:17 PM	21	0.05 kUA/l	0	Negative	
011990000786	wx1	5	2:44:17 PM	26	0.11 kUA/l	0	Negative	
011990000786	t20	5	2:45:16 PM	16	Low RU	0		
011990000789	fx5	5	2:47:16 PM	29	0.15 kUA/l	0	Negative	
011990000789	fx1	5	2:48:16 PM	15	Low RU	0	Negative	
011990000789	fx2	5	2:49:17 PM	21	0.05 kUA/l	0	Negative	
011990000789	fx22	5	2:50:17 PM	27	0.12 kUA/l	0	Negative	
011990000789	gx2	5	2:51:16 PM	20	0.03 kUA/l	0	Negative	
011990000789	wx1	5	2:52:16 PM	26	0.11 kUA/l	0	Negative	
011990000789	t20	5	2:53:16 PM	17	Low RU	0		
011990000790	fx5	5	2:23:18 PM	64	0.55 kUA/l	1	Positive	
011990000790	fx1	5	2:24:17 PM	19	0.03 kUA/l	0	Negative	
011990000790	fx2	5	2:25:17 PM	48	0.36 kUA/l	1	Positive	
011990000790	fx22	5	2:26:17 PM	32	0.18 kUA/l	0	Negative	
011990000790	gx2	5	2:27:17 PM	29	0.14 kUA/l	0	Negative	
011990000790	wx1	5	2:28:17 PM	47	0.35 kUA/l	1	Positive	
011990000790	t20	5	2:29:17 PM	20	0.03 kUA/l	0		
011990000795	fx5	5	2:31:17 PM	19	0.03 kUA/l	0	Negative	
011990000795	fx1	5	2:32:17 PM	17	0.01 kUA/l	0	Negative	
011990000795	fx2	5	2:33:17 PM	70	0.62 kUA/l	1	Positive	
011990000795	fx22	5	2:34:17 PM	28	0.14 kUA/l	0	Negative	
011990000795	gx2	5	2:35:17 PM	20	0.03 kUA/l	0	Negative	
011990000795	wx1	5	2:36:17 PM	28	0.13 kUA/l	0	Negative	
011990000795	t20	5	2:37:17 PM	18	0.01 kUA/l	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	



# Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_lgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_lgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_lgE	11:16:24 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	
CC-2	a_lgE	11:17:24 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	11:18:24 AM	352D0 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ1AR 21/08		Exp
slgE Neg	d1	11:19:24 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		BL1CP 19/12!		Exp
slgE L	e1	11:20:24 AM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ4AS 20/03		
IQC271119	d1	11:21:24 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000439	fx2	11:32:24 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	phad	11:49:23 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	fx5	11:50:23 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	fx1	11:51:23 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	fx2	11:52:23 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	fx22	12:50:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	gx2	11:53:23 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	wx1	11:54:23 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000427	t20	11:55:23 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	phad	11:41:23 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	fx5	11:42:24 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	fx1	11:43:23 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	fx2	11:44:23 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	fx22	11:45:23 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	gx2	11:46:23 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	wx1	11:47:23 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000415	t20	11:48:23 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	phad	11:22:24 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	fx5	11:23:24 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	fx1	11:24:24 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	fx2	11:36:23 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	fx22	11:37:23 AM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	gx2	11:38:23 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	wx1	11:39:23 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000403	t20	11:40:23 AM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000450	fx2	11:31:24 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000438	fx2	11:33:24 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000438	fx22	12:48:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000438	gx2	11:34:24 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000438	wx1	11:35:23 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000438	t20	12:49:21 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	phad	11:25:24 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	fx5	11:26:24 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	fx1	11:27:24 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	fx2	11:28:24 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	fx22	12:46:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	gx2	11:29:24 AM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	wx1	11:30:24 AM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000426	t20	12:47:21 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	phad	11:56:23 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	fx5	11:57:23 AM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	fx1	11:58:23 AM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	fx2	11:59:22 AM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	fx22	12:51:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	gx2	12:00:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	wx1	12:01:23 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000414	t20	12:02:23 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000402	phad	12:03:23 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000402	fx5	12:04:23 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000402	fx1	12:05:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000402	fx2	12:06:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000402	fx22	12:52:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000402	gx2	12:07:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000402	wx1	12:08:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000402	t20	12:09:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	phad	12:10:22 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	fx5	12:11:22 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	fx1	12:12:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	fx2	12:13:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	fx22	12:53:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	gx2	12:14:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	wx1	12:15:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000449	t20	12:16:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	phad	12:17:22 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	fx5	12:18:22 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	fx1	12:19:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	fx2	12:20:22 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	fx22	12:54:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	gx2	12:21:22 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	wx1	12:22:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000437	t20	12:23:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	phad	12:24:22 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	fx5	12:25:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	fx1	12:26:22 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	fx2	12:27:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	fx22	12:55:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	gx2	12:28:21 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	wx1	12:29:22 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000425	t20	12:30:22 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	phad	12:31:22 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	fx5	12:32:21 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	fx1	12:56:21 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	fx2	12:57:21 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	fx22	12:58:21 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	gx2	12:59:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	wx1	1:00:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000413	t20	1:01:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	phad	1:02:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	fx5	1:03:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	fx1	1:04:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	fx2	1:05:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	fx22	1:06:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	gx2	1:07:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	wx1	1:08:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000401	t20	1:09:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	phad	1:10:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	fx5	1:11:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	fx1	1:12:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	fx2	1:13:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	fx22	1:14:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	gx2	1:15:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	wx1	1:16:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000780	t20	1:17:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	phad	1:18:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	fx5	1:19:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	fx1	1:20:20 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	fx2	1:21:20 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	fx22	1:22:20 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	gx2	1:23:20 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	wx1	1:24:20 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000779	t20	1:25:20 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000778	phad	1:26:20 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000778	fx5	1:27:20 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	



30-Jan-20 9:16:16 AM

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000789	fx2	2:49:17 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000789	fx22	2:50:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000789	gx2	2:51:16 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000789	wx1	2:52:16 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000789	t20	2:53:16 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000788	phad	2:54:16 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	phad	2:38:17 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	fx5	2:39:17 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	fx1	2:40:17 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	fx2	2:41:17 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	fx22	2:42:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	gx2	2:43:17 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	wx1	2:44:17 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000786	t20	2:45:16 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	phad	2:30:17 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	fx5	2:31:17 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	fx1	2:32:17 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	fx2	2:33:17 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	fx22	2:34:17 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	gx2	2:35:17 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	wx1	2:36:17 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000795	t20	2:37:17 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

### Errors

#### Expired lot numbers used for Quality Control

352D0	ImmunoCAP
BL1CP	Quality control vial


## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200127-387  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 27-Jan-20 2:55:16 PM  
Ended: 27-Jan-20 7:05:09 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

No curve controls in this run

### Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	5:55:10 PM	5960	16.3 kUA/l	<lab>	OK	

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000630	phad	5	3:56:14 PM	37	0.24 PAU/l	0	Negative	
011990000632	phad	5	4:20:13 PM	57	0.47 PAU/l	1	Positive	
011990000633	phad	5	4:12:13 PM	42	0.30 PAU/l	0	Negative	
011990000634	phad	5	4:04:13 PM	33	0.19 PAU/l	0	Negative	
011990000787	phad	5	3:16:15 PM	46	0.35 PAU/l	0	Negative	
011990000791	phad	5	3:48:14 PM	97	0.94 PAU/l	2	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000792	phad	5	3:40:14 PM	35	0.21 PAU/I	0	Negative	
011990000793	phad	5	3:32:14 PM	33	0.19 PAU/I	0	Negative	
011990000794	phad	5	3:24:15 PM	43	0.30 PAU/I	0	Negative	
011990000796	phad	5	4:47:12 PM	76	0.69 PAU/I	1	Positive	
011990000827	phad	5	7:04:07 PM	35	0.21 PAU/I	0	Negative	
011990000828	phad	5	6:56:07 PM	34	0.20 PAU/I	0	Negative	
011990000829	phad	5	6:48:07 PM	33	0.20 PAU/I	0	Negative	
011990000830	phad	5	6:40:07 PM	32	0.18 PAU/I	0	Negative	
011990000831	phad	5	5:11:11 PM	39	0.26 PAU/I	0	Negative	
011990000833	phad	5	6:32:08 PM	131	1.33 PAU/I	2	Positive	
011990000836	phad	5	5:51:09 PM	94	0.90 PAU/I	2	Positive	
011990000837	phad	5	5:43:10 PM	56	0.46 PAU/I	1	Positive	
011990000838	phad	5	5:35:10 PM	231	2.49 PAU/I	2	Positive	
011990000839	phad	5	5:27:11 PM	1685	19.9 PAU/I	4	Positive	
011990000840	phad	5	5:19:11 PM	34	0.21 PAU/I	0	Negative	
011990000842	phad	5	6:24:08 PM	84	0.78 PAU/I	2	Positive	
011990000843	phad	5	6:16:09 PM	1889	22.5 PAU/I	4	Positive	
011990000844	phad	5	6:08:09 PM	611	6.91 PAU/I	3	Positive	
011990000845	phad	5	6:00:09 PM	192	2.04 PAU/I	2	Positive	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000630	fx5	5	3:57:14 PM	21	0.05 kUA/I	0	Negative	
011990000630	fx1	5	3:58:13 PM	16	Low RU	0	Negative	
011990000630	fx2	5	3:59:13 PM	14	Low RU	0	Negative	
011990000630	fx22	5	4:00:13 PM	29	0.14 kUA/I	0	Negative	
011990000630	gx2	5	4:01:13 PM	21	0.05 kUA/I	0	Negative	
011990000630	wx1	5	4:02:13 PM	32	0.18 kUA/I	0	Negative	
011990000630	t20	5	4:03:14 PM	17	0.01 kUA/I	0		
011990000632	fx5	5	4:21:12 PM	23	0.08 kUA/I	0	Negative	
011990000632	fx1	5	4:22:12 PM	18	0.02 kUA/I	0	Negative	
011990000632	fx2	5	4:23:13 PM	14	Low RU	0	Negative	
011990000632	fx22	5	4:24:13 PM	31	0.16 kUA/I	0	Negative	
011990000632	gx2	5	4:25:13 PM	25	0.10 kUA/I	0	Negative	
011990000632	wx1	5	4:26:12 PM	34	0.20 kUA/I	0	Negative	
011990000632	t20	5	4:27:12 PM	17	Low RU	0		
011990000633	fx5	5	4:13:13 PM	27	0.12 kUA/I	0	Negative	
011990000633	fx1	5	4:14:13 PM	21	0.05 kUA/I	0	Negative	
011990000633	fx2	5	4:15:13 PM	15	Low RU	0	Negative	
011990000633	fx22	5	4:16:13 PM	30	0.16 kUA/I	0	Negative	
011990000633	gx2	5	4:17:13 PM	112	1.11 kUA/I	2	Positive	
011990000633	wx1	5	4:18:13 PM	32	0.17 kUA/I	0	Negative	
011990000633	t20	5	4:19:13 PM	18	0.02 kUA/I	0		
011990000634	fx5	5	4:05:13 PM	23	0.07 kUA/I	0	Negative	
011990000634	fx1	5	4:06:13 PM	16	Low RU	0	Negative	
011990000634	fx2	5	4:07:13 PM	13	Low RU	0	Negative	
011990000634	fx22	5	4:08:13 PM	32	0.18 kUA/I	0	Negative	
011990000634	gx2	5	4:09:13 PM	27	0.12 kUA/I	0	Negative	
011990000634	wx1	5	4:10:13 PM	32	0.18 kUA/I	0	Negative	
011990000634	t20	5	4:11:13 PM	16	Low RU	0		
011990000787	fx5	5	3:17:15 PM	20	0.04 kUA/I	0	Negative	
011990000787	fx1	5	3:18:15 PM	15	Low RU	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000787	fx2	5	3:19:15 PM	20	0.04 kUA/l	0	Negative	
011990000787	fx22	5	3:20:15 PM	29	0.15 kUA/l	0	Negative	
011990000787	gx2	5	3:21:15 PM	23	0.08 kUA/l	0	Negative	
011990000787	wx1	5	3:22:15 PM	30	0.16 kUA/l	0	Negative	
011990000787	t20	5	3:23:15 PM	20	0.04 kUA/l	0		
011990000788	fx5	5	2:55:16 PM	23	0.07 kUA/l	0	Negative	
011990000788	fx1	5	2:56:16 PM	18	0.02 kUA/l	0	Negative	
011990000788	fx2	5	2:57:16 PM	134	1.37 kUA/l	2	Positive	
011990000788	fx22	5	2:58:16 PM	32	0.18 kUA/l	0	Negative	
011990000788	gx2	5	2:59:16 PM	29	0.15 kUA/l	0	Negative	
011990000788	wx1	5	3:14:15 PM	46	0.34 kUA/l	0	Negative	
011990000788	t20	5	3:15:15 PM	26	0.10 kUA/l	0		
011990000791	fx5	5	3:49:14 PM	22	0.07 kUA/l	0	Negative	
011990000791	fx1	5	3:50:14 PM	19	0.02 kUA/l	0	Negative	
011990000791	fx2	5	3:51:14 PM	33	0.19 kUA/l	0	Negative	
011990000791	fx22	5	3:52:14 PM	38	0.25 kUA/l	0	Negative	
011990000791	gx2	5	3:53:14 PM	30	0.16 kUA/l	0	Negative	
011990000791	wx1	5	3:54:14 PM	57	0.47 kUA/l	1	Positive	
011990000791	t20	5	3:55:14 PM	37	0.24 kUA/l	0		
011990000792	fx5	5	3:41:14 PM	24	0.08 kUA/l	0	Negative	
011990000792	fx1	5	3:42:14 PM	18	0.01 kUA/l	0	Negative	
011990000792	fx2	5	3:43:14 PM	17	Low RU	0	Negative	
011990000792	fx22	5	3:44:14 PM	31	0.16 kUA/l	0	Negative	
011990000792	gx2	5	3:45:14 PM	30	0.16 kUA/l	0	Negative	
011990000792	wx1	5	3:46:14 PM	33	0.19 kUA/l	0	Negative	
011990000792	t20	5	3:47:14 PM	18	0.01 kUA/l	0		
011990000793	fx5	5	3:33:14 PM	23	0.08 kUA/l	0	Negative	
011990000793	fx1	5	3:34:14 PM	18	0.01 kUA/l	0	Negative	
011990000793	fx2	5	3:35:15 PM	15	Low RU	0	Negative	
011990000793	fx22	5	3:36:15 PM	27	0.12 kUA/l	0	Negative	
011990000793	gx2	5	3:37:15 PM	22	0.06 kUA/l	0	Negative	
011990000793	wx1	5	3:38:14 PM	27	0.13 kUA/l	0	Negative	
011990000793	t20	5	3:39:14 PM	17	Low RU	0		
011990000794	fx5	5	3:25:15 PM	25	0.10 kUA/l	0	Negative	
011990000794	fx1	5	3:26:15 PM	22	0.06 kUA/l	0	Negative	
011990000794	fx2	5	3:27:15 PM	15	Low RU	0	Negative	
011990000794	fx22	5	3:28:15 PM	34	0.20 kUA/l	0	Negative	
011990000794	gx2	5	3:29:15 PM	28	0.13 kUA/l	0	Negative	
011990000794	wx1	5	3:30:15 PM	33	0.19 kUA/l	0	Negative	
011990000794	t20	5	3:31:15 PM	23	0.07 kUA/l	0		
011990000796	fx5	5	4:48:12 PM	47	0.36 kUA/l	1	Positive	
011990000796	fx1	5	4:49:11 PM	31	0.17 kUA/l	0	Negative	
011990000796	fx2	5	4:50:12 PM	37	0.24 kUA/l	0	Negative	
011990000796	fx22	5	4:51:12 PM	42	0.30 kUA/l	0	Negative	
011990000796	gx2	5	5:08:11 PM	29	0.14 kUA/l	0	Negative	
011990000796	wx1	5	5:09:11 PM	33	0.19 kUA/l	0	Negative	
011990000796	t20	5	5:10:11 PM	22	0.06 kUA/l	0		
011990000828	fx5	5	6:57:07 PM	27	0.12 kUA/l	0	Negative	
011990000828	fx1	5	6:58:07 PM	20	0.04 kUA/l	0	Negative	
011990000828	fx2	5	6:59:07 PM	14	Low RU	0	Negative	
011990000828	fx22	5	7:00:07 PM	28	0.13 kUA/l	0	Negative	
011990000828	gx2	5	7:01:07 PM	22	0.06 kUA/l	0	Negative	
011990000828	wx1	5	7:02:07 PM	28	0.14 kUA/l	0	Negative	
011990000828	t20	5	7:03:07 PM	17	Low RU	0		



## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000829	fx5	5	6:49:07 PM	57	0.48 kUA/l	1	Positive	
011990000829	fx1	5	6:50:07 PM	15	Low RU	0	Negative	
011990000829	fx2	5	6:51:07 PM	14	Low RU	0	Negative	
011990000829	fx22	5	6:52:07 PM	29	0.14 kUA/l	0	Negative	
011990000829	gx2	5	6:53:07 PM	21	0.05 kUA/l	0	Negative	
011990000829	wx1	5	6:54:07 PM	30	0.15 kUA/l	0	Negative	
011990000829	t20	5	6:55:07 PM	17	Low RU	0		
011990000830	fx5	5	6:41:08 PM	23	0.07 kUA/l	0	Negative	
011990000830	fx1	5	6:42:08 PM	18	0.01 kUA/l	0	Negative	
011990000830	fx2	5	6:43:08 PM	12	Low RU	0	Negative	
011990000830	fx22	5	6:44:08 PM	27	0.13 kUA/l	0	Negative	
011990000830	gx2	5	6:45:08 PM	21	0.05 kUA/l	0	Negative	
011990000830	wx1	5	6:46:07 PM	29	0.15 kUA/l	0	Negative	
011990000830	t20	5	6:47:07 PM	17	0.01 kUA/l	0		
011990000831	fx5	5	5:12:11 PM	23	0.07 kUA/l	0	Negative	
011990000831	fx1	5	5:13:11 PM	28	0.13 kUA/l	0	Negative	
011990000831	fx2	5	5:14:11 PM	14	Low RU	0	Negative	
011990000831	fx22	5	5:15:11 PM	29	0.14 kUA/l	0	Negative	
011990000831	gx2	5	5:16:11 PM	31	0.17 kUA/l	0	Negative	
011990000831	wx1	5	5:17:11 PM	29	0.15 kUA/l	0	Negative	
011990000831	t20	5	5:18:11 PM	16	Low RU	0		
011990000833	fx5	5	6:33:08 PM	31	0.17 kUA/l	0	Negative	
011990000833	fx1	5	6:34:08 PM	28	0.14 kUA/l	0	Negative	
011990000833	fx2	5	6:35:08 PM	28	0.14 kUA/l	0	Negative	
011990000833	fx22	5	6:36:08 PM	49	0.38 kUA/l	1	Positive	
011990000833	gx2	5	6:37:08 PM	87	0.82 kUA/l	2	Positive	
011990000833	wx1	5	6:38:08 PM	62	0.53 kUA/l	1	Positive	
011990000833	t20	5	6:39:07 PM	49	0.38 kUA/l	1		
011990000836	fx5	5	5:52:09 PM	40	0.27 kUA/l	0	Negative	
011990000836	fx1	5	5:53:09 PM	17	Low RU	0	Negative	
011990000836	fx2	5	5:54:09 PM	15	Low RU	0	Negative	
011990000836	fx22	5	5:56:09 PM	35	0.22 kUA/l	0	Negative	
011990000836	gx2	5	5:57:09 PM	23	0.08 kUA/l	0	Negative	
011990000836	wx1	5	5:58:09 PM	36	0.23 kUA/l	0	Negative	
011990000836	t20	5	5:59:09 PM	19	0.03 kUA/l	0		
011990000837	fx5	5	5:44:10 PM	22	0.06 kUA/l	0	Negative	
011990000837	fx1	5	5:45:09 PM	25	0.10 kUA/l	0	Negative	
011990000837	fx2	5	5:46:10 PM	33	0.19 kUA/l	0	Negative	
011990000837	fx22	5	5:47:10 PM	31	0.17 kUA/l	0	Negative	
011990000837	gx2	5	5:48:10 PM	22	0.06 kUA/l	0	Negative	
011990000837	wx1	5	5:49:09 PM	75	0.69 kUA/l	1	Positive	
011990000837	t20	5	5:50:09 PM	22	0.06 kUA/l	0		
011990000838	fx5	5	5:36:10 PM	74	0.67 kUA/l	1	Positive	
011990000838	fx1	5	5:37:10 PM	22	0.06 kUA/l	0	Negative	
011990000838	fx2	5	5:38:10 PM	151	1.57 kUA/l	2	Positive	
011990000838	fx22	5	5:39:10 PM	43	0.32 kUA/l	0	Negative	
011990000838	gx2	5	5:40:10 PM	76	0.69 kUA/l	1	Positive	
011990000838	wx1	5	5:41:10 PM	53	0.42 kUA/l	1	Positive	
011990000838	t20	5	5:42:10 PM	28	0.14 kUA/l	0		
011990000839	fx5	5	5:28:10 PM	28	0.13 kUA/l	0	Negative	
011990000839	fx1	5	5:29:10 PM	17	Low RU	0	Negative	
011990000839	fx2	5	5:30:10 PM	128	1.29 kUA/l	2	Positive	
011990000839	fx22	5	5:31:10 PM	32	0.17 kUA/l	0	Negative	
011990000839	gx2	5	5:32:10 PM	25	0.10 kUA/l	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000839	wx1	5	5:33:10 PM	34	0.20 kUA/l	0	Negative	
011990000839	t20	5	5:34:10 PM	17	Low RU	0		
011990000840	fx5	5	5:20:11 PM	22	0.06 kUA/l	0	Negative	
011990000840	fx1	5	5:21:11 PM	34	0.20 kUA/l	0	Negative	
011990000840	fx2	5	5:22:10 PM	15	Low RU	0	Negative	
011990000840	fx22	5	5:23:10 PM	31	0.17 kUA/l	0	Negative	
011990000840	gx2	5	5:24:10 PM	21	0.04 kUA/l	0	Negative	
011990000840	wx1	5	5:25:10 PM	29	0.15 kUA/l	0	Negative	
011990000840	t20	5	5:26:10 PM	17	0.01 kUA/l	0		
011990000842	fx5	5	6:25:08 PM	24	0.08 kUA/l	0	Negative	
011990000842	fx1	5	6:26:08 PM	24	0.08 kUA/l	0	Negative	
011990000842	fx2	5	6:27:08 PM	101	0.98 kUA/l	2	Positive	
011990000842	fx22	5	6:28:08 PM	30	0.16 kUA/l	0	Negative	
011990000842	gx2	5	6:29:08 PM	30	0.15 kUA/l	0	Negative	
011990000842	wx1	5	6:30:08 PM	39	0.27 kUA/l	0	Negative	
011990000842	t20	5	6:31:08 PM	23	0.08 kUA/l	0		
011990000843	fx5	5	6:17:08 PM	38	0.25 kUA/l	0	Negative	
011990000843	fx1	5	6:18:08 PM	36	0.23 kUA/l	0	Negative	
011990000843	fx2	5	6:19:08 PM	17	Low RU	0	Negative	
011990000843	fx22	5	6:20:08 PM	37	0.24 kUA/l	0	Negative	
011990000843	gx2	5	6:21:08 PM	186	1.97 kUA/l	2	Positive	
011990000843	wx1	5	6:22:08 PM	137	1.40 kUA/l	2	Positive	
011990000843	t20	5	6:23:09 PM	22	0.07 kUA/l	0		
011990000844	fx5	5	6:09:09 PM	38	0.25 kUA/l	0	Negative	
011990000844	fx1	5	6:10:09 PM	22	0.06 kUA/l	0	Negative	
011990000844	fx2	5	6:11:09 PM	55	0.45 kUA/l	1	Positive	
011990000844	fx22	5	6:12:08 PM	34	0.20 kUA/l	0	Negative	
011990000844	gx2	5	6:13:08 PM	27	0.12 kUA/l	0	Negative	
011990000844	wx1	5	6:14:09 PM	39	0.26 kUA/l	0	Negative	
011990000844	t20	5	6:15:09 PM	18	0.01 kUA/l	0		
011990000845	fx5	5	6:01:09 PM	29	0.15 kUA/l	0	Negative	
011990000845	fx1	5	6:02:09 PM	20	0.03 kUA/l	0	Negative	
011990000845	fx2	5	6:03:09 PM	50	0.39 kUA/l	1	Positive	
011990000845	fx22	5	6:04:09 PM	31	0.16 kUA/l	0	Negative	
011990000845	gx2	5	6:05:09 PM	29	0.14 kUA/l	0	Negative	
011990000845	wx1	5	6:06:09 PM	43	0.31 kUA/l	0	Negative	
011990000845	t20	5	6:07:09 PM	23	0.07 kUA/l	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
-----------------	------	----------	----------	-----------	---------	----------	----------	-----------	---------	---------

## Laboratory Report

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	5:55:10 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000788	fx5	2:55:16 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000788	fx1	2:56:16 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000788	fx2	2:57:16 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000788	fx22	2:58:16 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000788	gx2	2:59:16 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000788	wx1	3:14:15 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000788	t20	3:15:15 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	phad	3:16:15 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	fx5	3:17:15 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	fx1	3:18:15 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	fx2	3:19:15 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	fx22	3:20:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	gx2	3:21:15 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	wx1	3:22:15 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000787	t20	3:23:15 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	phad	3:24:15 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	fx5	3:25:15 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	fx1	3:26:15 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	fx2	3:27:15 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	fx22	3:28:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	gx2	3:29:15 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	wx1	3:30:15 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000794	t20	3:31:15 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	phad	3:32:14 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	fx5	3:33:14 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	fx1	3:34:14 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	fx2	3:35:15 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	fx22	3:36:15 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	gx2	3:37:15 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	wx1	3:38:14 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000793	t20	3:39:14 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	phad	3:40:14 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	fx5	3:41:14 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	fx1	3:42:14 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	fx2	3:43:14 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	fx22	3:44:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	gx2	3:45:14 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	wx1	3:46:14 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000792	t20	3:47:14 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	phad	3:48:14 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	fx5	3:49:14 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	fx1	3:50:14 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	fx2	3:51:14 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	fx22	3:52:14 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	gx2	3:53:14 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	wx1	3:54:14 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000791	t20	3:55:14 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	phad	3:56:14 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	fx5	3:57:14 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	fx1	3:58:13 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	fx2	3:59:13 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	fx22	4:00:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	gx2	4:01:13 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	wx1	4:02:13 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000630	t20	4:03:14 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000634	phad	4:04:13 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000634	fx5	4:05:13 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000634	fx1	4:06:13 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000634	fx2	4:07:13 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000634	fx22	4:08:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000634	gx2	4:09:13 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000634	wx1	4:10:13 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

# Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000634	t20	4:11:13 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	phad	4:12:13 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	fx5	4:13:13 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	fx1	4:14:13 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	fx2	4:15:13 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	fx22	4:16:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	gx2	4:17:13 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	wx1	4:18:13 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000633	t20	4:19:13 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	phad	4:20:13 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	fx5	4:21:12 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	fx1	4:22:12 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	fx2	4:23:13 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	fx22	4:24:13 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	gx2	4:25:13 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	wx1	4:26:12 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000632	t20	4:27:12 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	phad	4:47:12 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	fx5	4:48:12 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	fx1	4:49:11 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	fx2	4:50:12 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	fx22	4:51:12 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	gx2	5:08:11 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	wx1	5:09:11 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000796	t20	5:10:11 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	phad	6:40:07 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	fx5	6:41:08 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	fx1	6:42:08 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	fx2	6:43:08 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	fx22	6:44:08 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	gx2	6:45:08 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	wx1	6:46:07 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000830	t20	6:47:07 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	phad	6:48:07 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	fx5	6:49:07 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	fx1	6:50:07 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	fx2	6:51:07 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	fx22	6:52:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	gx2	6:53:07 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	wx1	6:54:07 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000829	t20	6:55:07 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	phad	6:56:07 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	fx5	6:57:07 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	fx1	6:58:07 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	fx2	6:59:07 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	fx22	7:00:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	gx2	7:01:07 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	wx1	7:02:07 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000828	t20	7:03:07 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000827	phad	7:04:07 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	phad	6:32:08 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	fx5	6:33:08 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	fx1	6:34:08 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	fx2	6:35:08 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	fx22	6:36:08 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	gx2	6:37:08 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	wx1	6:38:08 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000833	t20	6:39:07 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000831	phad	5:11:11 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000831	fx5	5:12:11 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000831	fx1	5:13:11 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000831	fx2	5:14:11 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000831	fx22	5:15:11 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000831	gx2	5:16:11 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000831	wx1	5:17:11 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	



30-Jan-20 9:16:16 AM

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000842	phad	6:24:08 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000842	fx5	6:25:08 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000842	fx1	6:26:08 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000842	fx2	6:27:08 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000842	fx22	6:28:08 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000842	gx2	6:29:08 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000842	wx1	6:30:08 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000842	t20	6:31:08 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	




## Laboratory Report

### ImmunoCAP Specific IgE 0-100

Run number: 200127-387  
Instrument name: Phadia-A  
Instrument type: Phadia 250  
Status: Ready  
Started: 27-Jan-20 7:05:07 PM  
Ended: 27-Jan-20 8:46:24 PM  
Approved: Approved (faziz)  
Calibrators in run: No

### Calibration Information

Instrument name: Phadia-A  
Instrument type: Phadia 250  
Measured: 05-Jan-20 1:26:01 PM  
Conjugate lot: A1HGR  
Acceptance: OK  
Active: Yes  
Calculation type: Rodbard 5-par



## Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

### Curve Controls

No curve controls in this run

### Quality Controls

No quality controls in this run

### Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000640	phad	5	7:36:05 PM	38	0.25 PAU/l	0	Negative	
011990000826	phad	5	7:12:07 PM	160	1.67 PAU/l	2	Positive	
011990000832	phad	5	7:44:05 PM	37	0.24 PAU/l	0	Negative	
011990000834	phad	5	7:28:06 PM	469	5.25 PAU/l	3	Positive	
011990000835	phad	5	7:20:06 PM	98	0.95 PAU/l	2	Positive	
011990000841	phad	5	8:08:04 PM	33	0.19 PAU/l	0	Negative	
011990000846	phad	5	8:32:03 PM	380	4.22 PAU/l	3	Positive	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000847	phad	5	8:24:03 PM	484	5.43 PAU/I	3	Positive	
011990000848	phad	5	8:16:04 PM	366	4.06 PAU/I	3	Positive	
011990000849	phad	5	8:00:05 PM	54	0.43 PAU/I	1	Positive	
011990000850	phad	5	7:52:05 PM	37	0.24 PAU/I	0	Negative	

### Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000640	fx5	5	7:37:05 PM	23	0.08 kUA/I	0	Negative	
011990000640	fx1	5	7:38:05 PM	18	0.01 kUA/I	0	Negative	
011990000640	fx2	5	7:39:06 PM	18	0.01 kUA/I	0	Negative	
011990000640	fx22	5	7:40:05 PM	33	0.19 kUA/I	0	Negative	
011990000640	gx2	5	7:41:05 PM	22	0.06 kUA/I	0	Negative	
011990000640	wx1	5	7:42:05 PM	30	0.15 kUA/I	0	Negative	
011990000640	t20	5	7:43:05 PM	19	0.03 kUA/I	0		
011990000826	fx5	5	7:13:07 PM	23	0.07 kUA/I	0	Negative	
011990000826	fx1	5	7:14:07 PM	17	Low RU	0	Negative	
011990000826	fx2	5	7:15:06 PM	169	1.77 kUA/I	2	Positive	
011990000826	fx22	5	7:16:06 PM	34	0.20 kUA/I	0	Negative	
011990000826	gx2	5	7:17:06 PM	25	0.10 kUA/I	0	Negative	
011990000826	wx1	5	7:18:06 PM	159	1.66 kUA/I	2	Positive	
011990000826	t20	5	7:19:06 PM	22	0.06 kUA/I	0		
011990000827	fx5	5	7:05:07 PM	21	0.04 kUA/I	0	Negative	
011990000827	fx1	5	7:06:07 PM	25	0.10 kUA/I	0	Negative	
011990000827	fx2	5	7:07:07 PM	13	Low RU	0	Negative	
011990000827	fx22	5	7:08:07 PM	31	0.16 kUA/I	0	Negative	
011990000827	gx2	5	7:09:07 PM	21	0.05 kUA/I	0	Negative	
011990000827	wx1	5	7:10:06 PM	32	0.18 kUA/I	0	Negative	
011990000827	t20	5	7:11:07 PM	17	Low RU	0		
011990000832	fx5	5	7:45:05 PM	24	0.08 kUA/I	0	Negative	
011990000832	fx1	5	7:46:05 PM	17	Low RU	0	Negative	
011990000832	fx2	5	7:47:05 PM	23	0.08 kUA/I	0	Negative	
011990000832	fx22	5	7:48:05 PM	28	0.13 kUA/I	0	Negative	
011990000832	gx2	5	7:49:05 PM	29	0.15 kUA/I	0	Negative	
011990000832	wx1	5	7:50:05 PM	32	0.18 kUA/I	0	Negative	
011990000832	t20	5	7:51:05 PM	17	Low RU	0		
011990000834	fx5	5	7:29:06 PM	31	0.17 kUA/I	0	Negative	
011990000834	fx1	5	7:30:06 PM	25	0.10 kUA/I	0	Negative	
011990000834	fx2	5	7:31:06 PM	223	2.40 kUA/I	2	Positive	
011990000834	fx22	5	7:32:06 PM	41	0.29 kUA/I	0	Negative	
011990000834	gx2	5	7:33:06 PM	43	0.30 kUA/I	0	Negative	
011990000834	wx1	5	7:34:06 PM	189	2.00 kUA/I	2	Positive	
011990000834	t20	5	7:35:06 PM	50	0.40 kUA/I	1		
011990000835	fx5	5	7:21:06 PM	21	0.05 kUA/I	0	Negative	
011990000835	fx1	5	7:22:06 PM	17	Low RU	0	Negative	
011990000835	fx2	5	7:23:06 PM	73	0.66 kUA/I	1	Positive	
011990000835	fx22	5	7:24:06 PM	31	0.17 kUA/I	0	Negative	
011990000835	gx2	5	7:25:06 PM	28	0.14 kUA/I	0	Negative	
011990000835	wx1	5	7:26:06 PM	46	0.35 kUA/I	0	Negative	
011990000835	t20	5	7:27:06 PM	21	0.05 kUA/I	0		
011990000841	fx5	5	8:09:04 PM	20	0.03 kUA/I	0	Negative	
011990000841	fx1	5	8:10:04 PM	16	Low RU	0	Negative	
011990000841	fx2	5	8:11:04 PM	13	Low RU	0	Negative	

## Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000841	fx22	5	8:12:04 PM	32	0.18 kUA/I	0	Negative	
011990000841	gx2	5	8:13:04 PM	23	0.07 kUA/I	0	Negative	
011990000841	wx1	5	8:14:04 PM	29	0.14 kUA/I	0	Negative	
011990000841	t20	5	8:15:04 PM	17	Low RU	0		
011990000846	fx5	5	8:33:03 PM	25	0.10 kUA/I	0	Negative	
011990000846	fx1	5	8:34:03 PM	33	0.19 kUA/I	0	Negative	
011990000846	fx2	5	8:35:03 PM	24	0.08 kUA/I	0	Negative	
011990000846	fx22	5	8:36:03 PM	32	0.18 kUA/I	0	Negative	
011990000846	gx2	5	8:37:03 PM	26	0.11 kUA/I	0	Negative	
011990000846	wx1	5	8:38:03 PM	38	0.25 kUA/I	0	Negative	
011990000846	t20	5	8:39:03 PM	25	0.09 kUA/I	0		
011990000847	fx5	5	8:25:03 PM	24	0.09 kUA/I	0	Negative	
011990000847	fx1	5	8:26:04 PM	16	Low RU	0	Negative	
011990000847	fx2	5	8:27:04 PM	25	0.10 kUA/I	0	Negative	
011990000847	fx22	5	8:28:03 PM	32	0.17 kUA/I	0	Negative	
011990000847	gx2	5	8:29:03 PM	26	0.11 kUA/I	0	Negative	
011990000847	wx1	5	8:30:03 PM	632	7.16 kUA/I	3	Positive	
011990000847	t20	5	8:31:03 PM	40	0.27 kUA/I	0		
011990000848	fx5	5	8:17:04 PM	38	0.25 kUA/I	0	Negative	
011990000848	fx1	5	8:18:04 PM	23	0.07 kUA/I	0	Negative	
011990000848	fx2	5	8:19:04 PM	242	2.62 kUA/I	2	Positive	
011990000848	fx22	5	8:20:04 PM	49	0.38 kUA/I	1	Positive	
011990000848	gx2	5	8:21:03 PM	41	0.29 kUA/I	0	Negative	
011990000848	wx1	5	8:22:04 PM	67	0.59 kUA/I	1	Positive	
011990000848	t20	5	8:23:03 PM	31	0.17 kUA/I	0		
011990000849	fx5	5	8:01:04 PM	68	0.60 kUA/I	1	Positive	
011990000849	fx1	5	8:02:04 PM	22	0.06 kUA/I	0	Negative	
011990000849	fx2	5	8:03:04 PM	44	0.32 kUA/I	0	Negative	
011990000849	fx22	5	8:04:04 PM	40	0.27 kUA/I	0	Negative	
011990000849	gx2	5	8:05:04 PM	36	0.23 kUA/I	0	Negative	
011990000849	wx1	5	8:06:04 PM	37	0.24 kUA/I	0	Negative	
011990000849	t20	5	8:07:04 PM	20	0.04 kUA/I	0		
011990000850	fx5	5	7:53:05 PM	64	0.56 kUA/I	1	Positive	
011990000850	fx1	5	7:54:05 PM	18	0.01 kUA/I	0	Negative	
011990000850	fx2	5	7:55:05 PM	20	0.03 kUA/I	0	Negative	
011990000850	fx22	5	7:56:04 PM	36	0.22 kUA/I	0	Negative	
011990000850	gx2	5	7:57:05 PM	21	0.05 kUA/I	0	Negative	
011990000850	wx1	5	7:58:05 PM	30	0.16 kUA/I	0	Negative	
011990000850	t20	5	7:59:05 PM	18	0.01 kUA/I	0		

### Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000827	fx5	7:05:07 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000827	fx1	7:06:07 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000827	fx2	7:07:07 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000827	fx22	7:08:07 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000827	gx2	7:09:07 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000827	wx1	7:10:06 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000827	t20	7:11:07 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	phad	7:12:07 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	fx5	7:13:07 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	fx1	7:14:07 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	fx2	7:15:06 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	fx22	7:16:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	gx2	7:17:06 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	wx1	7:18:06 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000826	t20	7:19:06 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	phad	7:20:06 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	fx5	7:21:06 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	fx1	7:22:06 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	fx2	7:23:06 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	fx22	7:24:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	gx2	7:25:06 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	wx1	7:26:06 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000835	t20	7:27:06 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	phad	7:28:06 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	fx5	7:29:06 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	fx1	7:30:06 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	fx2	7:31:06 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	fx22	7:32:06 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	gx2	7:33:06 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	wx1	7:34:06 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000834	t20	7:35:06 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	phad	7:36:05 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	fx5	7:37:05 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	fx1	7:38:05 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	fx2	7:39:06 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	fx22	7:40:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	gx2	7:41:05 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	wx1	7:42:05 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000640	t20	7:43:05 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	phad	7:44:05 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	fx5	7:45:05 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	fx1	7:46:05 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	fx2	7:47:05 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	fx22	7:48:05 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	gx2	7:49:05 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	wx1	7:50:05 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000832	t20	7:51:05 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	phad	8:08:04 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	fx5	8:09:04 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	fx1	8:10:04 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	fx2	8:11:04 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	fx22	8:12:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	gx2	8:13:04 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	wx1	8:14:04 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000841	t20	8:15:04 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	phad	7:52:05 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	fx5	7:53:05 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	fx1	7:54:05 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	fx2	7:55:05 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	fx22	7:56:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	gx2	7:57:05 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	wx1	7:58:05 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000850	t20	7:59:05 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000849	phad	8:00:05 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000849	fx5	8:01:04 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000849	fx1	8:02:04 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000849	fx2	8:03:04 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

## Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000849	fx22	8:04:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000849	gx2	8:05:04 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000849	wx1	8:06:04 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000849	t20	8:07:04 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	phad	8:16:04 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	fx5	8:17:04 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	fx1	8:18:04 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	fx2	8:19:04 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	fx22	8:20:04 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	gx2	8:21:03 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	wx1	8:22:04 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000848	t20	8:23:03 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	phad	8:24:03 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	fx5	8:25:03 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	fx1	8:26:04 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	fx2	8:27:04 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	fx22	8:28:03 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	gx2	8:29:03 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	wx1	8:30:03 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000847	t20	8:31:03 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	phad	8:32:03 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	fx5	8:33:03 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	fx1	8:34:03 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	fx2	8:35:03 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	fx22	8:36:03 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	gx2	8:37:03 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	wx1	8:38:03 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000846	t20	8:39:03 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	