


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200120-382
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 20-Jan-20 11:40:54 AM
Ended: 20-Jan-20 3:19:48 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	11:40:54 AM	303	0.66	OK	
CC-2	17.5	11:41:54 AM	6225	17.2	OK	

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	11:43:53 AM	3390	8.47 kUA/l	CZ1AR	OK	
slgE L	e1	1	11:44:53 AM	1091	2.53 kUA/l	CZ4AN	OK	
slgE Neg	d1	1	11:45:53 AM	21	0.01 kUA/l	BL1CP	OK	
IQC271119	d1	1	11:46:54 AM	6323	17.6 kUA/l	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000551	phad	5	2:15:49 PM	30	0.15 PAU/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000552	phad	5	2:07:49 PM	32	0.18 PAU/I	0	Negative	
011990000555	phad	5	12:29:52 PM	1212	14.1 PAU/I	3	Positive	
011990000556	phad	5	1:11:50 PM	28	0.13 PAU/I	0	Negative	
011990000557	phad	5	2:31:48 PM	27	0.12 PAU/I	0	Negative	
011990000562	phad	5	1:51:49 PM	31	0.17 PAU/I	0	Negative	
011990000563	phad	5	12:37:52 PM	447	4.99 PAU/I	3	Positive	
011990000564	phad	5	1:03:51 PM	26	0.10 PAU/I	0	Negative	
011990000565	phad	5	2:23:48 PM	32	0.18 PAU/I	0	Negative	
011990000567	phad	5	12:45:51 PM	367	4.07 PAU/I	3	Positive	
011990000573	phad	5	1:27:50 PM	1262	14.7 PAU/I	3	Positive	
011990000574	phad	5	1:19:51 PM	30	0.15 PAU/I	0	Negative	
011990000575	phad	5	1:35:50 PM	36	0.23 PAU/I	0	Negative	
011990000576	phad	5	1:59:49 PM	501	5.63 PAU/I	3	Positive	
011990000577	phad	5	12:05:53 PM	33	0.20 PAU/I	0	Negative	
011990000578	phad	5	1:43:49 PM	251	2.73 PAU/I	2	Positive	
011990000579	phad	5	11:49:53 AM	31	0.17 PAU/I	0	Negative	
011990000580	phad	5	11:57:53 AM	2963	36.5 PAU/I	4	Positive	
011990000584	phad	5	2:39:48 PM	3308	41.2 PAU/I	4	Positive	
011990000585	phad	5	2:47:47 PM	42	0.30 PAU/I	0	Negative	
011990000586	phad	5	2:55:47 PM	27	0.12 PAU/I	0	Negative	
011990000587	phad	5	3:03:47 PM	58	0.49 PAU/I	1	Positive	
011990000588	phad	5	3:11:47 PM	27	0.12 PAU/I	0	Negative	
011990000589	phad	5	12:21:52 PM	27	0.13 PAU/I	0	Negative	
011990000590	phad	5	12:13:52 PM	31	0.17 PAU/I	0	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000551	fx5	5	2:16:48 PM	19	0.03 kUA/I	0	Negative	
011990000551	fx1	5	2:17:48 PM	16	Low RU	0	Negative	
011990000551	fx2	5	2:18:48 PM	12	Low RU	0	Negative	
011990000551	fx22	5	2:19:48 PM	28	0.13 kUA/I	0	Negative	
011990000551	gx2	5	2:20:48 PM	19	0.02 kUA/I	0	Negative	
011990000551	wx1	5	2:21:48 PM	30	0.16 kUA/I	0	Negative	
011990000551	t20	5	2:22:48 PM	17	Low RU	0		
011990000552	fx5	5	2:08:49 PM	21	0.05 kUA/I	0	Negative	
011990000552	fx1	5	2:09:48 PM	18	0.01 kUA/I	0	Negative	
011990000552	fx2	5	2:10:48 PM	15	Low RU	0	Negative	
011990000552	fx22	5	2:11:48 PM	32	0.19 kUA/I	0	Negative	
011990000552	gx2	5	2:12:48 PM	29	0.14 kUA/I	0	Negative	
011990000552	wx1	5	2:13:48 PM	33	0.20 kUA/I	0	Negative	
011990000552	t20	5	2:14:48 PM	17	0.01 kUA/I	0		
011990000555	fx5	5	12:30:52 PM	56	0.46 kUA/I	1	Positive	
011990000555	fx1	5	12:31:52 PM	52	0.41 kUA/I	1	Positive	
011990000555	fx2	5	12:32:52 PM	33	0.19 kUA/I	0	Negative	
011990000555	fx22	5	12:33:52 PM	61	0.52 kUA/I	1	Positive	
011990000555	gx2	5	12:34:52 PM	1077	12.5 kUA/I	3	Positive	
011990000555	wx1	5	12:35:52 PM	283	3.09 kUA/I	2	Positive	
011990000555	t20	5	12:36:51 PM	55	0.45 kUA/I	1		
011990000556	fx5	5	1:12:51 PM	16	Low RU	0	Negative	
011990000556	fx1	5	1:13:50 PM	21	0.05 kUA/I	0	Negative	
011990000556	fx2	5	1:14:50 PM	16	Low RU	0	Negative	
011990000556	fx22	5	1:15:50 PM	28	0.14 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000556	gx2	5	1:16:50 PM	19	0.03 kUA/l	0	Negative	
011990000556	wx1	5	1:17:50 PM	30	0.16 kUA/l	0	Negative	
011990000556	t20	5	1:18:51 PM	17	Low RU	0		
011990000557	fx5	5	2:32:48 PM	19	0.02 kUA/l	0	Negative	
011990000557	fx1	5	2:33:47 PM	23	0.07 kUA/l	0	Negative	
011990000557	fx2	5	2:34:48 PM	11	Low RU	0	Negative	
011990000557	fx22	5	2:35:48 PM	28	0.13 kUA/l	0	Negative	
011990000557	gx2	5	2:36:48 PM	21	0.05 kUA/l	0	Negative	
011990000557	wx1	5	2:37:48 PM	32	0.18 kUA/l	0	Negative	
011990000557	t20	5	2:38:47 PM	16	Low RU	0		
011990000562	fx5	5	1:52:49 PM	19	0.03 kUA/l	0	Negative	
011990000562	fx1	5	1:53:49 PM	17	Low RU	0	Negative	
011990000562	fx2	5	1:54:49 PM	16	Low RU	0	Negative	
011990000562	fx22	5	1:55:49 PM	27	0.12 kUA/l	0	Negative	
011990000562	gx2	5	1:56:49 PM	19	0.03 kUA/l	0	Negative	
011990000562	wx1	5	1:57:49 PM	30	0.15 kUA/l	0	Negative	
011990000562	t20	5	1:58:49 PM	17	Low RU	0		
011990000563	fx5	5	12:38:52 PM	19	0.03 kUA/l	0	Negative	
011990000563	fx1	5	12:39:51 PM	16	Low RU	0	Negative	
011990000563	fx2	5	12:40:51 PM	36	0.23 kUA/l	0	Negative	
011990000563	fx22	5	12:41:51 PM	33	0.20 kUA/l	0	Negative	
011990000563	gx2	5	12:42:52 PM	22	0.06 kUA/l	0	Negative	
011990000563	wx1	5	12:43:52 PM	314	3.46 kUA/l	2	Positive	
011990000563	t20	5	12:44:51 PM	42	0.30 kUA/l	0		
011990000564	fx5	5	1:04:51 PM	22	0.07 kUA/l	0	Negative	
011990000564	fx1	5	1:05:51 PM	17	Low RU	0	Negative	
011990000564	fx2	5	1:06:50 PM	16	Low RU	0	Negative	
011990000564	fx22	5	1:07:51 PM	28	0.13 kUA/l	0	Negative	
011990000564	gx2	5	1:08:50 PM	21	0.05 kUA/l	0	Negative	
011990000564	wx1	5	1:09:51 PM	32	0.19 kUA/l	0	Negative	
011990000564	t20	5	1:10:51 PM	17	Low RU	0		
011990000565	fx5	5	2:24:48 PM	18	0.02 kUA/l	0	Negative	
011990000565	fx1	5	2:25:48 PM	16	Low RU	0	Negative	
011990000565	fx2	5	2:26:48 PM	11	Low RU	0	Negative	
011990000565	fx22	5	2:27:48 PM	30	0.16 kUA/l	0	Negative	
011990000565	gx2	5	2:28:48 PM	24	0.09 kUA/l	0	Negative	
011990000565	wx1	5	2:29:48 PM	32	0.18 kUA/l	0	Negative	
011990000565	t20	5	2:30:48 PM	15	Low RU	0		
011990000567	fx5	5	12:46:51 PM	27	0.12 kUA/l	0	Negative	
011990000567	fx1	5	12:47:51 PM	23	0.08 kUA/l	0	Negative	
011990000567	fx2	5	12:58:51 PM	34	0.20 kUA/l	0	Negative	
011990000567	fx22	5	12:59:51 PM	31	0.17 kUA/l	0	Negative	
011990000567	gx2	5	1:00:51 PM	20	0.04 kUA/l	0	Negative	
011990000567	wx1	5	1:01:51 PM	30	0.16 kUA/l	0	Negative	
011990000567	t20	5	1:02:51 PM	18	0.01 kUA/l	0		
011990000573	fx5	5	1:28:50 PM	21	0.05 kUA/l	0	Negative	
011990000573	fx1	5	1:29:50 PM	15	Low RU	0	Negative	
011990000573	fx2	5	1:30:50 PM	31	0.17 kUA/l	0	Negative	
011990000573	fx22	5	1:31:50 PM	29	0.14 kUA/l	0	Negative	
011990000573	gx2	5	1:32:50 PM	24	0.08 kUA/l	0	Negative	
011990000573	wx1	5	1:33:50 PM	39	0.26 kUA/l	0	Negative	
011990000573	t20	5	1:34:50 PM	18	0.02 kUA/l	0		
011990000574	fx5	5	1:20:50 PM	22	0.06 kUA/l	0	Negative	
011990000574	fx1	5	1:21:50 PM	24	0.09 kUA/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000574	fx2	5	1:22:50 PM	13	Low RU	0	Negative	
011990000574	fx22	5	1:23:50 PM	29	0.15 kUA/l	0	Negative	
011990000574	gx2	5	1:24:50 PM	24	0.08 kUA/l	0	Negative	
011990000574	wx1	5	1:25:50 PM	32	0.18 kUA/l	0	Negative	
011990000574	t20	5	1:26:50 PM	17	0.01 kUA/l	0		
011990000575	fx5	5	1:36:49 PM	46	0.35 kUA/l	0	Negative	
011990000575	fx1	5	1:37:50 PM	24	0.09 kUA/l	0	Negative	
011990000575	fx2	5	1:38:50 PM	15	Low RU	0	Negative	
011990000575	fx22	5	1:39:50 PM	34	0.21 kUA/l	0	Negative	
011990000575	gx2	5	1:40:50 PM	23	0.08 kUA/l	0	Negative	
011990000575	wx1	5	1:41:49 PM	37	0.24 kUA/l	0	Negative	
011990000575	t20	5	1:42:49 PM	26	0.10 kUA/l	0		
011990000576	fx5	5	2:00:49 PM	34	0.20 kUA/l	0	Negative	
011990000576	fx1	5	2:01:49 PM	21	0.05 kUA/l	0	Negative	
011990000576	fx2	5	2:02:49 PM	120	1.20 kUA/l	2	Positive	
011990000576	fx22	5	2:03:49 PM	37	0.24 kUA/l	0	Negative	
011990000576	gx2	5	2:04:48 PM	49	0.38 kUA/l	1	Positive	
011990000576	wx1	5	2:05:49 PM	231	2.49 kUA/l	2	Positive	
011990000576	t20	5	2:06:49 PM	23	0.07 kUA/l	0		
011990000577	fx5	5	12:06:53 PM	21	0.05 kUA/l	0	Negative	
011990000577	fx1	5	12:07:53 PM	16	Low RU	0	Negative	
011990000577	fx2	5	12:08:53 PM	20	0.04 kUA/l	0	Negative	
011990000577	fx22	5	12:09:53 PM	27	0.13 kUA/l	0	Negative	
011990000577	gx2	5	12:10:52 PM	20	0.04 kUA/l	0	Negative	
011990000577	wx1	5	12:11:52 PM	29	0.15 kUA/l	0	Negative	
011990000577	t20	5	12:12:52 PM	16	Low RU	0		
011990000578	fx5	5	1:44:49 PM	43	0.31 kUA/l	0	Negative	
011990000578	fx1	5	1:45:49 PM	19	0.02 kUA/l	0	Negative	
011990000578	fx2	5	1:46:49 PM	158	1.65 kUA/l	2	Positive	
011990000578	fx22	5	1:47:50 PM	31	0.16 kUA/l	0	Negative	
011990000578	gx2	5	1:48:49 PM	30	0.15 kUA/l	0	Negative	
011990000578	wx1	5	1:49:49 PM	52	0.41 kUA/l	1	Positive	
011990000578	t20	5	1:50:49 PM	27	0.13 kUA/l	0		
011990000579	fx5	5	11:50:53 AM	20	0.04 kUA/l	0	Negative	
011990000579	fx1	5	11:51:53 AM	17	Low RU	0	Negative	
011990000579	fx2	5	11:52:53 AM	13	Low RU	0	Negative	
011990000579	fx22	5	11:53:53 AM	29	0.14 kUA/l	0	Negative	
011990000579	gx2	5	11:54:53 AM	22	0.06 kUA/l	0	Negative	
011990000579	wx1	5	11:55:53 AM	29	0.14 kUA/l	0	Negative	
011990000579	t20	5	11:56:53 AM	18	0.01 kUA/l	0		
011990000580	fx5	5	11:58:53 AM	24	0.08 kUA/l	0	Negative	
011990000580	fx1	5	11:59:53 AM	20	0.04 kUA/l	0	Negative	
011990000580	fx2	5	12:00:53 PM	35	0.22 kUA/l	0	Negative	
011990000580	fx22	5	12:01:53 PM	33	0.20 kUA/l	0	Negative	
011990000580	gx2	5	12:02:53 PM	23	0.07 kUA/l	0	Negative	
011990000580	wx1	5	12:03:53 PM	34	0.21 kUA/l	0	Negative	
011990000580	t20	5	12:04:53 PM	22	0.06 kUA/l	0		
011990000584	fx5	5	2:40:47 PM	997	11.5 kUA/l	3	Positive	
011990000584	fx1	5	2:41:47 PM	373	4.14 kUA/l	3	Positive	
011990000584	fx2	5	2:42:47 PM	2273	27.4 kUA/l	4	Positive	
011990000584	fx22	5	2:43:48 PM	453	5.07 kUA/l	3	Positive	
011990000584	gx2	5	2:44:48 PM	1050	12.1 kUA/l	3	Positive	
011990000584	wx1	5	2:45:48 PM	513	5.77 kUA/l	3	Positive	
011990000584	t20	5	2:46:47 PM	515	5.79 kUA/l	3		

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000585	fx5	5	2:48:47 PM	20	0.04 kUA/l	0	Negative	
011990000585	fx1	5	2:49:47 PM	16	Low RU	0	Negative	
011990000585	fx2	5	2:50:47 PM	13	Low RU	0	Negative	
011990000585	fx22	5	2:51:47 PM	30	0.15 kUA/l	0	Negative	
011990000585	gx2	5	2:52:47 PM	22	0.06 kUA/l	0	Negative	
011990000585	wx1	5	2:53:47 PM	49	0.38 kUA/l	1	Positive	
011990000585	t20	5	2:54:47 PM	22	0.06 kUA/l	0		
011990000586	fx5	5	2:56:47 PM	19	0.02 kUA/l	0	Negative	
011990000586	fx1	5	2:57:47 PM	16	Low RU	0	Negative	
011990000586	fx2	5	2:58:47 PM	12	Low RU	0	Negative	
011990000586	fx22	5	2:59:47 PM	27	0.13 kUA/l	0	Negative	
011990000586	gx2	5	3:00:47 PM	20	0.03 kUA/l	0	Negative	
011990000586	wx1	5	3:01:47 PM	31	0.17 kUA/l	0	Negative	
011990000586	t20	5	3:02:47 PM	15	Low RU	0		
011990000587	fx5	5	3:04:47 PM	20	0.04 kUA/l	0	Negative	
011990000587	fx1	5	3:05:47 PM	19	0.03 kUA/l	0	Negative	
011990000587	fx2	5	3:06:47 PM	25	0.10 kUA/l	0	Negative	
011990000587	fx22	5	3:07:46 PM	34	0.20 kUA/l	0	Negative	
011990000587	gx2	5	3:08:46 PM	22	0.06 kUA/l	0	Negative	
011990000587	wx1	5	3:09:46 PM	33	0.19 kUA/l	0	Negative	
011990000587	t20	5	3:10:46 PM	18	0.01 kUA/l	0		
011990000588	fx5	5	3:12:46 PM	18	0.01 kUA/l	0	Negative	
011990000588	fx1	5	3:13:46 PM	19	0.02 kUA/l	0	Negative	
011990000588	fx2	5	3:14:46 PM	12	Low RU	0	Negative	
011990000588	fx22	5	3:15:46 PM	29	0.14 kUA/l	0	Negative	
011990000588	gx2	5	3:16:46 PM	18	0.02 kUA/l	0	Negative	
011990000588	wx1	5	3:17:46 PM	30	0.16 kUA/l	0	Negative	
011990000588	t20	5	3:18:46 PM	16	Low RU	0		
011990000589	fx5	5	12:22:52 PM	27	0.12 kUA/l	0	Negative	
011990000589	fx1	5	12:23:53 PM	18	0.01 kUA/l	0	Negative	
011990000589	fx2	5	12:24:52 PM	13	Low RU	0	Negative	
011990000589	fx22	5	12:25:52 PM	30	0.16 kUA/l	0	Negative	
011990000589	gx2	5	12:26:52 PM	21	0.06 kUA/l	0	Negative	
011990000589	wx1	5	12:27:52 PM	30	0.16 kUA/l	0	Negative	
011990000589	t20	5	12:28:52 PM	17	Low RU	0		
011990000590	fx5	5	12:14:53 PM	18	0.01 kUA/l	0	Negative	
011990000590	fx1	5	12:15:53 PM	19	0.02 kUA/l	0	Negative	
011990000590	fx2	5	12:16:52 PM	13	Low RU	0	Negative	
011990000590	fx22	5	12:17:52 PM	29	0.15 kUA/l	0	Negative	
011990000590	gx2	5	12:18:52 PM	23	0.07 kUA/l	0	Negative	
011990000590	wx1	5	12:19:52 PM	27	0.13 kUA/l	0	Negative	
011990000590	t20	5	12:20:52 PM	16	Low RU	0		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_lgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_lgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_lgE	11:40:54 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_lgE	11:41:54 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	11:43:53 AM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
slgE L	e1	11:44:53 AM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06!		Exp
slgE Neg	d1	11:45:53 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12!		Exp
IQC271119	d1	11:46:54 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000579	phad	11:49:53 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000579	fx5	11:50:53 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000579	fx1	11:51:53 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000579	fx2	11:52:53 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000579	fx22	11:53:53 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000579	gx2	11:54:53 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000579	wx1	11:55:53 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000579	t20	11:56:53 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	phad	11:57:53 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	fx5	11:58:53 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	fx1	11:59:53 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	fx2	12:00:53 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	fx22	12:01:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	gx2	12:02:53 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	wx1	12:03:53 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000580	t20	12:04:53 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	phad	12:05:53 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	fx5	12:06:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	fx1	12:07:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	fx2	12:08:53 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	fx22	12:09:53 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	gx2	12:10:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	wx1	12:11:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000577	t20	12:12:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	phad	12:21:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	fx5	12:22:52 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	fx1	12:23:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	fx2	12:24:52 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	fx22	12:25:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	gx2	12:26:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	wx1	12:27:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000589	t20	12:28:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	phad	12:13:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	fx5	12:14:53 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	fx1	12:15:53 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	fx2	12:16:52 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	fx22	12:17:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	gx2	12:18:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	wx1	12:19:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000590	t20	12:20:52 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	phad	12:37:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	fx5	12:38:52 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	fx1	12:39:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	fx2	12:40:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	fx22	12:41:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	gx2	12:42:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	wx1	12:43:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000563	t20	12:44:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000555	phad	12:29:52 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000555	fx5	12:30:52 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000555	fx1	12:31:52 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000555	fx2	12:32:52 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000555	fx22	12:33:52 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000555	gx2	12:34:52 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000555	wx1	12:35:52 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000555	t20	12:36:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	phad	1:03:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	fx5	1:04:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	fx1	1:05:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	fx2	1:06:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	fx22	1:07:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	gx2	1:08:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	wx1	1:09:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000564	t20	1:10:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	phad	12:45:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	fx5	12:46:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	fx1	12:47:51 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	fx2	12:58:51 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	fx22	12:59:51 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	gx2	1:00:51 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	wx1	1:01:51 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000567	t20	1:02:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	phad	1:11:50 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	fx5	1:12:51 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	fx1	1:13:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	fx2	1:14:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	fx22	1:15:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	gx2	1:16:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	wx1	1:17:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000556	t20	1:18:51 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	phad	1:27:50 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	fx5	1:28:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	fx1	1:29:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	fx2	1:30:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	fx22	1:31:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	gx2	1:32:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	wx1	1:33:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000573	t20	1:34:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	phad	1:19:51 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	fx5	1:20:50 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	fx1	1:21:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	fx2	1:22:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	fx22	1:23:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	gx2	1:24:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	wx1	1:25:50 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000574	t20	1:26:50 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	phad	1:43:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	fx5	1:44:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	fx1	1:45:49 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	fx2	1:46:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	fx22	1:47:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	gx2	1:48:49 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	wx1	1:49:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000578	t20	1:50:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	phad	1:35:50 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	fx5	1:36:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	fx1	1:37:50 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	fx2	1:38:50 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	fx22	1:39:50 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	gx2	1:40:50 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	wx1	1:41:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000575	t20	1:42:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000576	phad	1:59:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000576	fx5	2:00:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000576	fx1	2:01:49 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000576	fx2	2:02:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000576	fx22	2:03:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000576	gx2	2:04:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000576	wx1	2:05:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000576	t20	2:06:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	phad	1:51:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	fx5	1:52:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	fx1	1:53:49 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	fx2	1:54:49 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	fx22	1:55:49 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	gx2	1:56:49 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	wx1	1:57:49 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000562	t20	1:58:49 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	phad	2:15:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	fx5	2:16:48 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	fx1	2:17:48 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	fx2	2:18:48 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	fx22	2:19:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	gx2	2:20:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	wx1	2:21:48 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000551	t20	2:22:48 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	phad	2:07:49 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	fx5	2:08:49 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	fx1	2:09:48 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	fx2	2:10:48 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	fx22	2:11:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	gx2	2:12:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	wx1	2:13:48 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000552	t20	2:14:48 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	phad	2:31:48 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	fx5	2:32:48 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	fx1	2:33:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	fx2	2:34:48 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	fx22	2:35:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	gx2	2:36:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	wx1	2:37:48 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000557	t20	2:38:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	phad	2:23:48 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	fx5	2:24:48 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	fx1	2:25:48 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	fx2	2:26:48 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	fx22	2:27:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	gx2	2:28:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	wx1	2:29:48 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000565	t20	2:30:48 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	phad	2:39:48 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	fx5	2:40:47 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	fx1	2:41:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	fx2	2:42:47 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	fx22	2:43:48 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	gx2	2:44:48 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	wx1	2:45:48 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000584	t20	2:46:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	phad	2:47:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	fx5	2:48:47 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	fx1	2:49:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	fx2	2:50:47 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	fx22	2:51:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	gx2	2:52:47 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	wx1	2:53:47 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000585	t20	2:54:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000586	phad	2:55:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000586	fx5	2:56:47 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000586	fx1	2:57:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000586	fx2	2:58:47 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000586	fx22	2:59:47 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000586	gx2	3:00:47 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000586	wx1	3:01:47 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000586	t20	3:02:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	phad	3:03:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	fx5	3:04:47 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	fx1	3:05:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	fx2	3:06:47 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	fx22	3:07:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	gx2	3:08:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	wx1	3:09:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000587	t20	3:10:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	phad	3:11:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	fx5	3:12:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	fx1	3:13:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	fx2	3:14:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	fx22	3:15:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	gx2	3:16:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	wx1	3:17:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000588	t20	3:18:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Errors

Expired lot numbers used for Quality Control

352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200120-382
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 20-Jan-20 3:19:46 PM
Ended: 20-Jan-20 5:35:01 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

No curve controls in this run

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	5:25:41 PM	6349	17.7 kUA/l	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000152	phad	5	4:05:45 PM	324	3.56 PAU/l	3	Positive	
011990000561	phad	5	3:57:45 PM	192	2.04 PAU/l	2	Positive	
011990000569	phad	5	3:19:46 PM	29	0.15 PAU/l	0	Negative	
011990000570	phad	5	3:27:46 PM	984	11.3 PAU/l	3	Positive	
011990000571	phad	5	3:41:46 PM	29	0.14 PAU/l	0	Negative	
011990000572	phad	5	3:49:45 PM	34	0.21 PAU/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000581	phad	5	5:01:42 PM	74	0.67 PAU/I	1	Positive	
011990000582	phad	5	5:09:42 PM	68	0.61 PAU/I	1	Positive	
011990000583	phad	5	5:17:41 PM	135	1.37 PAU/I	2	Positive	
011990000591	phad	5	4:45:43 PM	33	0.19 PAU/I	0	Negative	
011990000592	phad	5	4:53:43 PM	35	0.21 PAU/I	0	Negative	
011990000593	phad	5	4:29:44 PM	237	2.56 PAU/I	2	Positive	
011990000594	phad	5	4:37:43 PM	1861	22.1 PAU/I	4	Positive	
011990000595	phad	5	4:13:45 PM	160	1.67 PAU/I	2	Positive	
011990000596	phad	5	4:21:45 PM	55	0.45 PAU/I	1	Positive	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000152	fx5	5	4:06:44 PM	23	0.07 kUA/I	0	Negative	
011990000152	fx1	5	4:07:45 PM	21	0.05 kUA/I	0	Negative	
011990000152	fx2	5	4:08:44 PM	17	Low RU	0	Negative	
011990000152	fx22	5	4:09:45 PM	48	0.37 kUA/I	1	Positive	
011990000152	gx2	5	4:10:44 PM	33	0.20 kUA/I	0	Negative	
011990000152	wx1	5	4:11:45 PM	42	0.30 kUA/I	0	Negative	
011990000152	t20	5	4:12:44 PM	28	0.13 kUA/I	0		
011990000561	fx5	5	3:58:44 PM	22	0.06 kUA/I	0	Negative	
011990000561	fx1	5	3:59:45 PM	21	0.05 kUA/I	0	Negative	
011990000561	fx2	5	4:00:45 PM	27	0.12 kUA/I	0	Negative	
011990000561	fx22	5	4:01:45 PM	31	0.16 kUA/I	0	Negative	
011990000561	gx2	5	4:02:44 PM	22	0.06 kUA/I	0	Negative	
011990000561	wx1	5	4:03:45 PM	35	0.22 kUA/I	0	Negative	
011990000561	t20	5	4:04:44 PM	19	0.02 kUA/I	0		
011990000569	fx5	5	3:20:46 PM	21	0.05 kUA/I	0	Negative	
011990000569	fx1	5	3:21:46 PM	17	Low RU	0	Negative	
011990000569	fx2	5	3:22:46 PM	14	Low RU	0	Negative	
011990000569	fx22	5	3:23:46 PM	31	0.17 kUA/I	0	Negative	
011990000569	gx2	5	3:24:46 PM	29	0.15 kUA/I	0	Negative	
011990000569	wx1	5	3:25:46 PM	31	0.17 kUA/I	0	Negative	
011990000569	t20	5	3:26:46 PM	17	Low RU	0		
011990000570	fx5	5	3:28:46 PM	26	0.10 kUA/I	0	Negative	
011990000570	fx1	5	3:29:46 PM	20	0.04 kUA/I	0	Negative	
011990000570	fx2	5	3:30:46 PM	23	0.07 kUA/I	0	Negative	
011990000570	fx22	5	3:31:46 PM	43	0.30 kUA/I	0	Negative	
011990000570	gx2	5	3:38:45 PM	26	0.11 kUA/I	0	Negative	
011990000570	wx1	5	3:39:46 PM	32	0.18 kUA/I	0	Negative	
011990000570	t20	5	3:40:46 PM	21	0.05 kUA/I	0		
011990000571	fx5	5	3:42:45 PM	19	0.03 kUA/I	0	Negative	
011990000571	fx1	5	3:43:45 PM	19	0.02 kUA/I	0	Negative	
011990000571	fx2	5	3:44:45 PM	11	Low RU	0	Negative	
011990000571	fx22	5	3:45:46 PM	31	0.16 kUA/I	0	Negative	
011990000571	gx2	5	3:46:45 PM	21	0.05 kUA/I	0	Negative	
011990000571	wx1	5	3:47:45 PM	27	0.13 kUA/I	0	Negative	
011990000571	t20	5	3:48:45 PM	17	0.01 kUA/I	0		
011990000572	fx5	5	3:50:45 PM	26	0.11 kUA/I	0	Negative	
011990000572	fx1	5	3:51:45 PM	25	0.10 kUA/I	0	Negative	
011990000572	fx2	5	3:52:45 PM	20	0.04 kUA/I	0	Negative	
011990000572	fx22	5	3:53:45 PM	31	0.17 kUA/I	0	Negative	
011990000572	gx2	5	3:54:45 PM	34	0.20 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000572	wx1	5	3:55:45 PM	36	0.22 kUA/l	0	Negative	
011990000572	t20	5	3:56:44 PM	19	0.03 kUA/l	0		
011990000581	fx5	5	5:02:42 PM	25	0.10 kUA/l	0	Negative	
011990000581	fx1	5	5:03:42 PM	18	0.02 kUA/l	0	Negative	
011990000581	fx2	5	5:04:41 PM	18	0.02 kUA/l	0	Negative	
011990000581	fx22	5	5:05:42 PM	32	0.18 kUA/l	0	Negative	
011990000581	gx2	5	5:06:42 PM	22	0.06 kUA/l	0	Negative	
011990000581	wx1	5	5:07:42 PM	30	0.16 kUA/l	0	Negative	
011990000581	t20	5	5:08:42 PM	17	Low RU	0		
011990000582	fx5	5	5:10:41 PM	31	0.16 kUA/l	0	Negative	
011990000582	fx1	5	5:11:42 PM	23	0.07 kUA/l	0	Negative	
011990000582	fx2	5	5:12:41 PM	33	0.20 kUA/l	0	Negative	
011990000582	fx22	5	5:13:42 PM	42	0.30 kUA/l	0	Negative	
011990000582	gx2	5	5:14:42 PM	37	0.24 kUA/l	0	Negative	
011990000582	wx1	5	5:15:41 PM	40	0.27 kUA/l	0	Negative	
011990000582	t20	5	5:16:41 PM	21	0.05 kUA/l	0		
011990000583	fx5	5	5:18:41 PM	136	1.40 kUA/l	2	Positive	
011990000583	fx1	5	5:19:41 PM	73	0.66 kUA/l	1	Positive	
011990000583	fx2	5	5:20:41 PM	46	0.34 kUA/l	0	Negative	
011990000583	fx22	5	5:21:41 PM	41	0.28 kUA/l	0	Negative	
011990000583	gx2	5	5:22:41 PM	41	0.29 kUA/l	0	Negative	
011990000583	wx1	5	5:23:41 PM	48	0.36 kUA/l	1	Positive	
011990000583	t20	5	5:24:41 PM	24	0.09 kUA/l	0		
011990000591	fx5	5	4:46:43 PM	23	0.07 kUA/l	0	Negative	
011990000591	fx1	5	4:47:43 PM	17	Low RU	0	Negative	
011990000591	fx2	5	4:48:42 PM	18	0.02 kUA/l	0	Negative	
011990000591	fx22	5	4:49:43 PM	30	0.16 kUA/l	0	Negative	
011990000591	gx2	5	4:50:42 PM	27	0.12 kUA/l	0	Negative	
011990000591	wx1	5	4:51:43 PM	28	0.13 kUA/l	0	Negative	
011990000591	t20	5	4:52:42 PM	19	0.02 kUA/l	0		
011990000592	fx5	5	4:54:42 PM	27	0.12 kUA/l	0	Negative	
011990000592	fx1	5	4:55:43 PM	19	0.03 kUA/l	0	Negative	
011990000592	fx2	5	4:56:42 PM	13	Low RU	0	Negative	
011990000592	fx22	5	4:57:43 PM	34	0.20 kUA/l	0	Negative	
011990000592	gx2	5	4:58:42 PM	22	0.06 kUA/l	0	Negative	
011990000592	wx1	5	4:59:43 PM	33	0.19 kUA/l	0	Negative	
011990000592	t20	5	5:00:42 PM	18	0.02 kUA/l	0		
011990000593	fx5	5	4:30:43 PM	26	0.11 kUA/l	0	Negative	
011990000593	fx1	5	4:31:44 PM	20	0.03 kUA/l	0	Negative	
011990000593	fx2	5	4:32:43 PM	275	2.99 kUA/l	2	Positive	
011990000593	fx22	5	4:33:44 PM	32	0.18 kUA/l	0	Negative	
011990000593	gx2	5	4:34:43 PM	23	0.07 kUA/l	0	Negative	
011990000593	wx1	5	4:35:44 PM	29	0.15 kUA/l	0	Negative	
011990000593	t20	5	4:36:43 PM	17	Low RU	0		
011990000594	fx5	5	4:38:43 PM	308	3.39 kUA/l	2	Positive	
011990000594	fx1	5	4:39:44 PM	312	3.43 kUA/l	2	Positive	
011990000594	fx2	5	4:40:43 PM	30	0.15 kUA/l	0	Negative	
011990000594	fx22	5	4:41:44 PM	300	3.29 kUA/l	2	Positive	
011990000594	gx2	5	4:42:43 PM	961	11.1 kUA/l	3	Positive	
011990000594	wx1	5	4:43:43 PM	1275	14.8 kUA/l	3	Positive	
011990000594	t20	5	4:44:42 PM	1800	21.3 kUA/l	4		
011990000595	fx5	5	4:14:44 PM	20	0.04 kUA/l	0	Negative	
011990000595	fx1	5	4:15:45 PM	17	0.01 kUA/l	0	Negative	
011990000595	fx2	5	4:16:44 PM	62	0.53 kUA/l	1	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000595	fx22	5	4:17:45 PM	31	0.16 kUA/l	0	Negative	
011990000595	gx2	5	4:18:44 PM	138	1.42 kUA/l	2	Positive	
011990000595	wx1	5	4:19:44 PM	32	0.18 kUA/l	0	Negative	
011990000595	t20	5	4:20:44 PM	17	Low RU	0		
011990000596	fx5	5	4:22:44 PM	21	0.05 kUA/l	0	Negative	
011990000596	fx1	5	4:23:44 PM	18	0.01 kUA/l	0	Negative	
011990000596	fx2	5	4:24:43 PM	30	0.16 kUA/l	0	Negative	
011990000596	fx22	5	4:25:44 PM	31	0.16 kUA/l	0	Negative	
011990000596	gx2	5	4:26:43 PM	24	0.08 kUA/l	0	Negative	
011990000596	wx1	5	4:27:44 PM	34	0.20 kUA/l	0	Negative	
011990000596	t20	5	4:28:44 PM	18	0.01 kUA/l	0		
012001600647A	t20	1	5:26:41 PM	39	0.05 kUA/l	0		
012001900698A	i70	1	5:27:41 PM	21	0.01 kUA/l			

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	5:25:41 PM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
012001600647A	t20	5:26:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			
012001900698A	i70	5:27:41 PM	568GG 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02			Exp
011990000569	phad	3:19:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000569	fx5	3:20:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000569	fx1	3:21:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000569	fx2	3:22:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000569	fx22	3:23:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000569	gx2	3:24:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000569	wx1	3:25:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000569	t20	3:26:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	phad	3:27:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	fx5	3:28:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	fx1	3:29:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	fx2	3:30:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	fx22	3:31:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	gx2	3:38:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	wx1	3:39:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000570	t20	3:40:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000571	phad	3:41:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000571	fx5	3:42:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000571	fx1	3:43:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000571	fx2	3:44:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000571	fx22	3:45:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000571	gx2	3:46:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000571	wx1	3:47:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000571	t20	3:48:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	phad	3:49:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	fx5	3:50:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	fx1	3:51:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	fx2	3:52:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	fx22	3:53:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	gx2	3:54:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	wx1	3:55:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000572	t20	3:56:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	phad	3:57:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	fx5	3:58:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	fx1	3:59:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	fx2	4:00:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	fx22	4:01:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	gx2	4:02:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	wx1	4:03:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000561	t20	4:04:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	phad	4:05:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	fx5	4:06:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	fx1	4:07:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	fx2	4:08:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	fx22	4:09:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	gx2	4:10:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	wx1	4:11:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000152	t20	4:12:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	phad	4:13:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	fx5	4:14:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	fx1	4:15:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	fx2	4:16:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	fx22	4:17:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	gx2	4:18:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	wx1	4:19:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000595	t20	4:20:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	phad	4:21:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	fx5	4:22:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	fx1	4:23:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	fx2	4:24:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	fx22	4:25:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	gx2	4:26:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	wx1	4:27:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000596	t20	4:28:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	phad	4:29:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	fx5	4:30:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	fx1	4:31:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	fx2	4:32:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	fx22	4:33:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	gx2	4:34:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	wx1	4:35:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000593	t20	4:36:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	phad	4:37:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	fx5	4:38:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	fx1	4:39:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	fx2	4:40:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	fx22	4:41:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	gx2	4:42:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	wx1	4:43:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000594	t20	4:44:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	phad	4:45:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	fx5	4:46:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	fx1	4:47:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	fx2	4:48:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	fx22	4:49:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	gx2	4:50:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	wx1	4:51:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000591	t20	4:52:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000592	phad	4:53:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000592	fx5	4:54:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000592	fx1	4:55:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000592	fx2	4:56:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000592	fx22	4:57:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000592	gx2	4:58:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000592	wx1	4:59:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000592	t20	5:00:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	phad	5:01:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	fx5	5:02:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	fx1	5:03:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	fx2	5:04:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	fx22	5:05:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	gx2	5:06:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	wx1	5:07:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000581	t20	5:08:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	phad	5:09:42 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	fx5	5:10:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	fx1	5:11:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	fx2	5:12:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	fx22	5:13:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	gx2	5:14:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	wx1	5:15:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000582	t20	5:16:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	phad	5:17:41 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	fx5	5:18:41 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	fx1	5:19:41 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	fx2	5:20:41 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	fx22	5:21:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	gx2	5:22:41 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	wx1	5:23:41 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000583	t20	5:24:41 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Errors

Expired lot numbers used for Sample

568GG ImmunoCAP