


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number:	200129-389
Instrument name:	Phadia-A
Instrument type:	Phadia 250
Status:	Ready
Started:	29-Jan-20 11:29:33 AM
Ended:	29-Jan-20 3:40:43 PM
Approved:	Approved (faziz)
Calibrators in run:	No

Calibration Information

Instrument name:	Phadia-A
Instrument type:	Phadia 250
Measured:	05-Jan-20 1:26:01 PM
Conjugate lot:	A1HGR
Acceptance:	OK
Active:	Yes
Calculation type:	Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	11:29:33 AM	309	0.68	OK	
CC-2	17.5	11:30:34 AM	6319	17.6	OK	

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	11:31:34 AM	3920	9.98 kUA/l	CZ1AR	OK	
slgE Neg	d1	1	11:32:33 AM	23	0.02 kUA/l	BL1CP	OK	
slgE L	e1	1	11:33:33 AM	1291	3.01 kUA/l	CZ4AS	OK	
IQC271119	d1	1	11:34:33 AM	6401	17.8 kUA/l	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000631	phad	5	2:53:26 PM	44	0.32 PAU/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000635	phad	5	2:37:26 PM	197	2.10 PAU/I	2	Positive	
011990000636	phad	5	3:08:25 PM	39	0.26 PAU/I	0	Negative	
011990000637	phad	5	2:45:26 PM	34	0.21 PAU/I	0	Negative	
011990000858	phad	5	12:13:32 PM	141	1.45 PAU/I	2	Positive	
011990000860	phad	5	11:59:32 AM	33	0.19 PAU/I	0	Negative	
011990000861	phad	5	12:53:30 PM	36	0.22 PAU/I	0	Negative	
011990000862	phad	5	12:45:30 PM	32	0.18 PAU/I	0	Negative	
011990000863	phad	5	12:37:31 PM	34	0.20 PAU/I	0	Negative	
011990000864	phad	5	12:29:31 PM	32	0.18 PAU/I	0	Negative	
011990000865	phad	5	12:21:31 PM	853	9.77 PAU/I	3	Positive	
011990000866	phad	5	1:01:30 PM	70	0.63 PAU/I	1	Positive	
011990000867	phad	5	1:09:29 PM	30	0.16 PAU/I	0	Negative	
011990000868	phad	5	1:17:29 PM	48	0.36 PAU/I	1	Positive	
011990000869	phad	5	1:25:29 PM	321	3.54 PAU/I	3	Positive	
011990000870	phad	5	1:33:29 PM	35	0.22 PAU/I	0	Negative	
011990000871	phad	5	1:41:29 PM	706	8.04 PAU/I	3	Positive	
011990000872	phad	5	1:49:28 PM	37	0.23 PAU/I	0	Negative	
011990000873	phad	5	1:57:28 PM	45	0.33 PAU/I	0	Negative	
011990000874	phad	5	2:13:27 PM	33	0.19 PAU/I	0	Negative	
011990000875	phad	5	2:21:27 PM	166	1.73 PAU/I	2	Positive	
011990000893	phad	5	3:32:25 PM	67	0.60 PAU/I	1	Positive	
011990000894	phad	5	3:24:25 PM	34	0.20 PAU/I	0	Negative	
011990000895	phad	5	3:16:25 PM	1659	19.6 PAU/I	4	Positive	
011990000896	phad	5	2:29:27 PM	772	8.81 PAU/I	3	Positive	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000631	fx5	5	3:01:25 PM	20	0.04 kUA/I	0	Negative	
011990000631	fx1	5	3:02:26 PM	20	0.04 kUA/I	0	Negative	
011990000631	fx2	5	3:03:25 PM	31	0.16 kUA/I	0	Negative	
011990000631	fx22	5	3:04:26 PM	32	0.18 kUA/I	0	Negative	
011990000631	gx2	5	3:05:25 PM	21	0.04 kUA/I	0	Negative	
011990000631	wx1	5	3:06:25 PM	29	0.14 kUA/I	0	Negative	
011990000631	t20	5	3:07:25 PM	17	0.01 kUA/I	0		
011990000635	fx5	5	2:38:26 PM	23	0.07 kUA/I	0	Negative	
011990000635	fx1	5	2:39:26 PM	30	0.16 kUA/I	0	Negative	
011990000635	fx2	5	2:40:26 PM	16	Low RU	0	Negative	
011990000635	fx22	5	2:41:26 PM	31	0.17 kUA/I	0	Negative	
011990000635	gx2	5	2:42:26 PM	25	0.10 kUA/I	0	Negative	
011990000635	wx1	5	2:43:26 PM	678	7.70 kUA/I	3	Positive	
011990000635	t20	5	2:44:26 PM	25	0.10 kUA/I	0		
011990000636	fx5	5	3:09:25 PM	29	0.15 kUA/I	0	Negative	
011990000636	fx1	5	3:10:25 PM	16	Low RU	0	Negative	
011990000636	fx2	5	3:11:25 PM	20	0.04 kUA/I	0	Negative	
011990000636	fx22	5	3:12:26 PM	32	0.18 kUA/I	0	Negative	
011990000636	gx2	5	3:13:25 PM	29	0.14 kUA/I	0	Negative	
011990000636	wx1	5	3:14:25 PM	32	0.18 kUA/I	0	Negative	
011990000636	t20	5	3:15:25 PM	19	0.02 kUA/I	0		
011990000637	fx5	5	2:46:26 PM	23	0.07 kUA/I	0	Negative	
011990000637	fx1	5	2:47:26 PM	16	Low RU	0	Negative	
011990000637	fx2	5	2:48:26 PM	11	Low RU	0	Negative	
011990000637	fx22	5	2:49:26 PM	29	0.14 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000637	gx2	5	2:50:26 PM	22	0.06 kUA/l	0	Negative	
011990000637	wx1	5	2:51:26 PM	29	0.14 kUA/l	0	Negative	
011990000637	t20	5	2:52:26 PM	19	0.02 kUA/l	0		
011990000858	fx5	5	12:14:32 PM	27	0.12 kUA/l	0	Negative	
011990000858	fx1	5	12:15:32 PM	19	0.02 kUA/l	0	Negative	
011990000858	fx2	5	12:16:32 PM	31	0.17 kUA/l	0	Negative	
011990000858	fx22	5	12:17:31 PM	32	0.18 kUA/l	0	Negative	
011990000858	gx2	5	12:18:31 PM	74	0.67 kUA/l	1	Positive	
011990000858	wx1	5	12:19:31 PM	39	0.26 kUA/l	0	Negative	
011990000858	t20	5	12:20:31 PM	31	0.17 kUA/l	0		
011990000860	fx5	5	12:00:32 PM	23	0.07 kUA/l	0	Negative	
011990000860	fx1	5	12:01:32 PM	15	Low RU	0	Negative	
011990000860	fx2	5	12:02:32 PM	14	Low RU	0	Negative	
011990000860	fx22	5	12:03:32 PM	31	0.17 kUA/l	0	Negative	
011990000860	gx2	5	12:10:32 PM	20	0.04 kUA/l	0	Negative	
011990000860	wx1	5	12:11:32 PM	27	0.13 kUA/l	0	Negative	
011990000860	t20	5	12:12:32 PM	17	Low RU	0		
011990000861	fx5	5	12:54:30 PM	20	0.04 kUA/l	0	Negative	
011990000861	fx1	5	12:55:30 PM	17	Low RU	0	Negative	
011990000861	fx2	5	12:56:31 PM	15	Low RU	0	Negative	
011990000861	fx22	5	12:57:30 PM	32	0.18 kUA/l	0	Negative	
011990000861	gx2	5	12:58:31 PM	22	0.06 kUA/l	0	Negative	
011990000861	wx1	5	12:59:30 PM	30	0.15 kUA/l	0	Negative	
011990000861	t20	5	1:00:30 PM	17	0.01 kUA/l	0		
011990000862	fx5	5	12:46:31 PM	26	0.11 kUA/l	0	Negative	
011990000862	fx1	5	12:47:30 PM	21	0.05 kUA/l	0	Negative	
011990000862	fx2	5	12:48:31 PM	13	Low RU	0	Negative	
011990000862	fx22	5	12:49:30 PM	29	0.14 kUA/l	0	Negative	
011990000862	gx2	5	12:50:31 PM	22	0.06 kUA/l	0	Negative	
011990000862	wx1	5	12:51:30 PM	28	0.13 kUA/l	0	Negative	
011990000862	t20	5	12:52:31 PM	17	0.01 kUA/l	0		
011990000863	fx5	5	12:38:31 PM	23	0.07 kUA/l	0	Negative	
011990000863	fx1	5	12:39:31 PM	16	Low RU	0	Negative	
011990000863	fx2	5	12:40:31 PM	17	Low RU	0	Negative	
011990000863	fx22	5	12:41:31 PM	30	0.16 kUA/l	0	Negative	
011990000863	gx2	5	12:42:31 PM	22	0.06 kUA/l	0	Negative	
011990000863	wx1	5	12:43:31 PM	27	0.12 kUA/l	0	Negative	
011990000863	t20	5	12:44:31 PM	16	Low RU	0		
011990000864	fx5	5	12:30:31 PM	22	0.06 kUA/l	0	Negative	
011990000864	fx1	5	12:31:31 PM	25	0.10 kUA/l	0	Negative	
011990000864	fx2	5	12:32:31 PM	13	Low RU	0	Negative	
011990000864	fx22	5	12:33:31 PM	28	0.14 kUA/l	0	Negative	
011990000864	gx2	5	12:34:31 PM	20	0.03 kUA/l	0	Negative	
011990000864	wx1	5	12:35:31 PM	27	0.12 kUA/l	0	Negative	
011990000864	t20	5	12:36:31 PM	17	Low RU	0		
011990000865	fx5	5	12:22:31 PM	19	0.03 kUA/l	0	Negative	
011990000865	fx1	5	12:23:31 PM	18	0.01 kUA/l	0	Negative	
011990000865	fx2	5	12:24:32 PM	129	1.31 kUA/l	2	Positive	
011990000865	fx22	5	12:25:31 PM	28	0.14 kUA/l	0	Negative	
011990000865	gx2	5	12:26:31 PM	25	0.09 kUA/l	0	Negative	
011990000865	wx1	5	12:27:31 PM	38	0.25 kUA/l	0	Negative	
011990000865	t20	5	12:28:31 PM	21	0.05 kUA/l	0		
011990000866	fx5	5	1:02:30 PM	25	0.09 kUA/l	0	Negative	
011990000866	fx1	5	1:03:30 PM	24	0.09 kUA/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000866	fx2	5	1:04:30 PM	30	0.15 kUA/I	0	Negative	
011990000866	fx22	5	1:05:30 PM	36	0.23 kUA/I	0	Negative	
011990000866	gx2	5	1:06:30 PM	29	0.14 kUA/I	0	Negative	
011990000866	wx1	5	1:07:30 PM	36	0.23 kUA/I	0	Negative	
011990000866	t20	5	1:08:30 PM	21	0.05 kUA/I	0		
011990000867	fx5	5	1:10:30 PM	22	0.07 kUA/I	0	Negative	
011990000867	fx1	5	1:11:29 PM	26	0.11 kUA/I	0	Negative	
011990000867	fx2	5	1:12:30 PM	12	Low RU	0	Negative	
011990000867	fx22	5	1:13:30 PM	28	0.13 kUA/I	0	Negative	
011990000867	gx2	5	1:14:30 PM	22	0.06 kUA/I	0	Negative	
011990000867	wx1	5	1:15:29 PM	27	0.12 kUA/I	0	Negative	
011990000867	t20	5	1:16:30 PM	27	0.12 kUA/I	0		
011990000868	fx5	5	1:18:30 PM	23	0.07 kUA/I	0	Negative	
011990000868	fx1	5	1:19:29 PM	16	Low RU	0	Negative	
011990000868	fx2	5	1:20:30 PM	15	Low RU	0	Negative	
011990000868	fx22	5	1:21:29 PM	30	0.15 kUA/I	0	Negative	
011990000868	gx2	5	1:22:30 PM	22	0.07 kUA/I	0	Negative	
011990000868	wx1	5	1:23:29 PM	30	0.15 kUA/I	0	Negative	
011990000868	t20	5	1:24:30 PM	20	0.03 kUA/I	0		
011990000869	fx5	5	1:26:30 PM	24	0.08 kUA/I	0	Negative	
011990000869	fx1	5	1:27:29 PM	27	0.13 kUA/I	0	Negative	
011990000869	fx2	5	1:28:30 PM	143	1.47 kUA/I	2	Positive	
011990000869	fx22	5	1:29:29 PM	29	0.15 kUA/I	0	Negative	
011990000869	gx2	5	1:30:30 PM	29	0.14 kUA/I	0	Negative	
011990000869	wx1	5	1:31:29 PM	47	0.36 kUA/I	1	Positive	
011990000869	t20	5	1:32:29 PM	26	0.11 kUA/I	0		
011990000870	fx5	5	1:34:29 PM	148	1.53 kUA/I	2	Positive	
011990000870	fx1	5	1:35:28 PM	15	Low RU	0	Negative	
011990000870	fx2	5	1:36:29 PM	13	Low RU	0	Negative	
011990000870	fx22	5	1:37:28 PM	32	0.18 kUA/I	0	Negative	
011990000870	gx2	5	1:38:29 PM	23	0.08 kUA/I	0	Negative	
011990000870	wx1	5	1:39:28 PM	28	0.14 kUA/I	0	Negative	
011990000870	t20	5	1:40:29 PM	18	0.01 kUA/I	0		
011990000871	fx5	5	1:42:29 PM	22	0.06 kUA/I	0	Negative	
011990000871	fx1	5	1:43:28 PM	23	0.07 kUA/I	0	Negative	
011990000871	fx2	5	1:44:29 PM	14	Low RU	0	Negative	
011990000871	fx22	5	1:45:28 PM	28	0.14 kUA/I	0	Negative	
011990000871	gx2	5	1:46:29 PM	21	0.06 kUA/I	0	Negative	
011990000871	wx1	5	1:47:28 PM	27	0.12 kUA/I	0	Negative	
011990000871	t20	5	1:48:29 PM	20	0.04 kUA/I	0		
011990000872	fx5	5	1:50:29 PM	22	0.06 kUA/I	0	Negative	
011990000872	fx1	5	1:51:28 PM	20	0.03 kUA/I	0	Negative	
011990000872	fx2	5	1:52:29 PM	13	Low RU	0	Negative	
011990000872	fx22	5	1:53:28 PM	30	0.16 kUA/I	0	Negative	
011990000872	gx2	5	1:54:29 PM	24	0.08 kUA/I	0	Negative	
011990000872	wx1	5	1:55:28 PM	28	0.13 kUA/I	0	Negative	
011990000872	t20	5	1:56:29 PM	17	0.01 kUA/I	0		
011990000873	fx5	5	1:58:29 PM	21	0.05 kUA/I	0	Negative	
011990000873	fx1	5	1:59:28 PM	19	0.03 kUA/I	0	Negative	
011990000873	fx2	5	2:00:28 PM	15	Low RU	0	Negative	
011990000873	fx22	5	2:01:28 PM	32	0.18 kUA/I	0	Negative	
011990000873	gx2	5	2:02:28 PM	25	0.10 kUA/I	0	Negative	
011990000873	wx1	5	2:03:27 PM	28	0.14 kUA/I	0	Negative	
011990000873	t20	5	2:04:28 PM	20	0.04 kUA/I	0		

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000874	fx5	5	2:14:28 PM	27	0.12 kUA/l	0	Negative	
011990000874	fx1	5	2:15:27 PM	18	0.01 kUA/l	0	Negative	
011990000874	fx2	5	2:16:28 PM	51	0.41 kUA/l	1	Positive	
011990000874	fx22	5	2:17:27 PM	29	0.14 kUA/l	0	Negative	
011990000874	gx2	5	2:18:28 PM	35	0.22 kUA/l	0	Negative	
011990000874	wx1	5	2:19:27 PM	27	0.12 kUA/l	0	Negative	
011990000874	t20	5	2:20:28 PM	17	Low RU	0		
011990000875	fx5	5	2:22:28 PM	27	0.12 kUA/l	0	Negative	
011990000875	fx1	5	2:23:27 PM	25	0.10 kUA/l	0	Negative	
011990000875	fx2	5	2:24:27 PM	49	0.37 kUA/l	1	Positive	
011990000875	fx22	5	2:25:27 PM	30	0.15 kUA/l	0	Negative	
011990000875	gx2	5	2:26:28 PM	24	0.08 kUA/l	0	Negative	
011990000875	wx1	5	2:27:27 PM	34	0.21 kUA/l	0	Negative	
011990000875	t20	5	2:28:27 PM	20	0.04 kUA/l	0		
011990000893	fx5	5	3:33:25 PM	23	0.08 kUA/l	0	Negative	
011990000893	fx1	5	3:34:25 PM	16	Low RU	0	Negative	
011990000893	fx2	5	3:35:24 PM	41	0.28 kUA/l	0	Negative	
011990000893	fx22	5	3:36:25 PM	32	0.18 kUA/l	0	Negative	
011990000893	gx2	5	3:37:24 PM	24	0.09 kUA/l	0	Negative	
011990000893	wx1	5	3:38:25 PM	36	0.23 kUA/l	0	Negative	
011990000893	t20	5	3:39:36 PM	20	0.03 kUA/l	0		
011990000894	fx5	5	3:25:25 PM	21	0.05 kUA/l	0	Negative	
011990000894	fx1	5	3:26:25 PM	17	Low RU	0	Negative	
011990000894	fx2	5	3:27:25 PM	18	0.01 kUA/l	0	Negative	
011990000894	fx22	5	3:28:25 PM	30	0.16 kUA/l	0	Negative	
011990000894	gx2	5	3:29:25 PM	24	0.09 kUA/l	0	Negative	
011990000894	wx1	5	3:30:25 PM	31	0.17 kUA/l	0	Negative	
011990000894	t20	5	3:31:24 PM	17	Low RU	0		
011990000895	fx5	5	3:17:25 PM	42	0.30 kUA/l	0	Negative	
011990000895	fx1	5	3:18:25 PM	22	0.06 kUA/l	0	Negative	
011990000895	fx2	5	3:19:25 PM	297	3.26 kUA/l	2	Positive	
011990000895	fx22	5	3:20:25 PM	40	0.27 kUA/l	0	Negative	
011990000895	gx2	5	3:21:25 PM	49	0.39 kUA/l	1	Positive	
011990000895	wx1	5	3:22:25 PM	125	1.26 kUA/l	2	Positive	
011990000895	t20	5	3:23:25 PM	93	0.89 kUA/l	2		
011990000896	fx5	5	2:30:27 PM	27	0.12 kUA/l	0	Negative	
011990000896	fx1	5	2:31:26 PM	20	0.04 kUA/l	0	Negative	
011990000896	fx2	5	2:32:27 PM	1409	16.5 kUA/l	3	Positive	
011990000896	fx22	5	2:33:26 PM	32	0.18 kUA/l	0	Negative	
011990000896	gx2	5	2:34:27 PM	50	0.40 kUA/l	1	Positive	
011990000896	wx1	5	2:35:26 PM	72	0.64 kUA/l	1	Positive	
011990000896	t20	5	2:36:27 PM	33	0.19 kUA/l	0		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_IgE	11:29:33 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	
CC-2	a_IgE	11:30:34 AM	865FX 20/02	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	11:31:34 AM	352D0 19/12!	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ1AR 21/08		Exp
slgE Neg	d1	11:32:33 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		BL1CP 19/12!		Exp
slgE L	e1	11:33:33 AM	359EP 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		CZ4AS 20/03		
IQC271119	d1	11:34:33 AM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000860	phad	11:59:32 AM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000860	fx5	12:00:32 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000860	fx1	12:01:32 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000860	fx2	12:02:32 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000860	fx22	12:03:32 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000860	gx2	12:10:32 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000860	wx1	12:11:32 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000860	t20	12:12:32 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	phad	12:13:32 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	fx5	12:14:32 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	fx1	12:15:32 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	fx2	12:16:32 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	fx22	12:17:31 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	gx2	12:18:31 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	wx1	12:19:31 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000858	t20	12:20:31 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	phad	12:21:31 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	fx5	12:22:31 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	fx1	12:23:31 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	fx2	12:24:32 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	fx22	12:25:31 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	gx2	12:26:31 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	wx1	12:27:31 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000865	t20	12:28:31 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	phad	12:29:31 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	fx5	12:30:31 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	fx1	12:31:31 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	fx2	12:32:31 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	fx22	12:33:31 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	gx2	12:34:31 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	wx1	12:35:31 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000864	t20	12:36:31 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	phad	12:37:31 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	fx5	12:38:31 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	fx1	12:39:31 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	fx2	12:40:31 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	fx22	12:41:31 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	gx2	12:42:31 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	wx1	12:43:31 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000863	t20	12:44:31 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	phad	12:45:30 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	fx5	12:46:31 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	fx1	12:47:30 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	fx2	12:48:31 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	fx22	12:49:30 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	gx2	12:50:31 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	wx1	12:51:30 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000862	t20	12:52:31 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000861	phad	12:53:30 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000861	fx5	12:54:30 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000861	fx1	12:55:30 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000861	fx2	12:56:31 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000861	fx22	12:57:30 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000861	gx2	12:58:31 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000861	wx1	12:59:30 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000861	t20	1:00:30 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	phad	1:33:29 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	fx5	1:34:29 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	fx1	1:35:28 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	fx2	1:36:29 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	fx22	1:37:28 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	gx2	1:38:29 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	wx1	1:39:28 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000870	t20	1:40:29 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	phad	1:25:29 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	fx5	1:26:30 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	fx1	1:27:29 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	fx2	1:28:30 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	fx22	1:29:29 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	gx2	1:30:30 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	wx1	1:31:29 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000869	t20	1:32:29 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	phad	1:17:29 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	fx5	1:18:30 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	fx1	1:19:29 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	fx2	1:20:30 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	fx22	1:21:29 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	gx2	1:22:30 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	wx1	1:23:29 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000868	t20	1:24:30 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	phad	1:09:29 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	fx5	1:10:30 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	fx1	1:11:29 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	fx2	1:12:30 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	fx22	1:13:30 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	gx2	1:14:30 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	wx1	1:15:29 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000867	t20	1:16:30 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	phad	1:01:30 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	fx5	1:02:30 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	fx1	1:03:30 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	fx2	1:04:30 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	fx22	1:05:30 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	gx2	1:06:30 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	wx1	1:07:30 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000866	t20	1:08:30 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	phad	2:21:27 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	fx5	2:22:28 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	fx1	2:23:27 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	fx2	2:24:27 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	fx22	2:25:27 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	gx2	2:26:28 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	wx1	2:27:27 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000875	t20	2:28:27 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	phad	2:13:27 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	fx5	2:14:28 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	fx1	2:15:27 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	fx2	2:16:28 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	fx22	2:17:27 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	gx2	2:18:28 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	wx1	2:19:27 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000874	t20	2:20:28 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000873	phad	1:57:28 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000873	fx5	1:58:29 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000873	fx1	1:59:28 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000873	fx2	2:00:28 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000873	fx22	2:01:28 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000873	gx2	2:02:28 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000873	wx1	2:03:27 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000873	t20	2:04:28 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	phad	1:49:28 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	fx5	1:50:29 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	fx1	1:51:28 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	fx2	1:52:29 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	fx22	1:53:28 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	gx2	1:54:29 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	wx1	1:55:28 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000872	t20	1:56:29 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	phad	1:41:29 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	fx5	1:42:29 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	fx1	1:43:28 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	fx2	1:44:29 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	fx22	1:45:28 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	gx2	1:46:29 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	wx1	1:47:28 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000871	t20	1:48:29 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	phad	2:29:27 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	fx5	2:30:27 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	fx1	2:31:26 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	fx2	2:32:27 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	fx22	2:33:26 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	gx2	2:34:27 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	wx1	2:35:26 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000896	t20	2:36:27 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	phad	2:37:26 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	fx5	2:38:26 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	fx1	2:39:26 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	fx2	2:40:26 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	fx22	2:41:26 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	gx2	2:42:26 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	wx1	2:43:26 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000635	t20	2:44:26 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	phad	2:45:26 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	fx5	2:46:26 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	fx1	2:47:26 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	fx2	2:48:26 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	fx22	2:49:26 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	gx2	2:50:26 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	wx1	2:51:26 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000637	t20	2:52:26 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	phad	2:53:26 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	fx5	3:01:25 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	fx1	3:02:26 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	fx2	3:03:25 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	fx22	3:04:26 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	gx2	3:05:25 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	wx1	3:06:25 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000631	t20	3:07:25 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	phad	3:08:25 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	fx5	3:09:25 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	fx1	3:10:25 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	fx2	3:11:25 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	fx22	3:12:26 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	gx2	3:13:25 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	wx1	3:14:25 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000636	t20	3:15:25 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000893	phad	3:32:25 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	
011990000893	fx5	3:33:25 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJOBR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000893	fx1	3:34:25 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000893	fx2	3:35:24 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000893	fx22	3:36:25 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000893	gx2	3:37:24 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000893	wx1	3:38:25 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000893	t20	3:39:36 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	phad	3:24:25 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	fx5	3:25:25 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	fx1	3:26:25 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	fx2	3:27:25 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	fx22	3:28:25 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	gx2	3:29:25 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	wx1	3:30:25 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000894	t20	3:31:24 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	phad	3:16:25 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	fx5	3:17:25 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	fx1	3:18:25 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	fx2	3:19:25 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	fx22	3:20:25 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	gx2	3:21:25 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	wx1	3:22:25 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000895	t20	3:23:25 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

Errors

Expired lot numbers used for Quality Control

352D0	ImmunoCAP
BL1CP	Quality control vial


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200129-389
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 29-Jan-20 3:40:42 PM
Ended: 29-Jan-20 5:32:58 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.69
Slope	225	216	+4.00

Number of curves included in mean: 54

Curve Controls

No curve controls in this run

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
IQC271119	d1	1	5:11:38 PM	6420	17.9 kUA/l	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000881	phad	5	5:09:38 PM	35	0.22 PAU/l	0	Negative	
011990000882	phad	5	5:01:39 PM	33	0.20 PAU/l	0	Negative	
011990000883	phad	5	4:53:39 PM	34	0.20 PAU/l	0	Negative	
011990000884	phad	5	4:45:39 PM	778	8.88 PAU/l	3	Positive	
011990000885	phad	5	4:37:40 PM	34	0.20 PAU/l	0	Negative	
011990000886	phad	5	4:29:41 PM	726	8.27 PAU/l	3	Positive	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000887	phad	5	4:21:41 PM	37	0.23 PAU/I	0	Negative	
011990000888	phad	5	4:13:41 PM	1571	18.5 PAU/I	4	Positive	
011990000889	phad	5	4:05:41 PM	41	0.29 PAU/I	0	Negative	
011990000890	phad	5	3:56:42 PM	33	0.19 PAU/I	0	Negative	
011990000891	phad	5	3:48:42 PM	33	0.19 PAU/I	0	Negative	
011990000892	phad	5	3:40:42 PM	37	0.24 PAU/I	0	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000881	fx5	5	5:10:38 PM	25	0.09 kUA/I	0	Negative	
011990000881	fx1	5	5:12:38 PM	18	0.02 kUA/I	0	Negative	
011990000881	fx2	5	5:25:38 PM	19	0.02 kUA/I	0	Negative	
011990000882	fx5	5	5:02:39 PM	24	0.09 kUA/I	0	Negative	
011990000882	fx1	5	5:03:39 PM	20	0.04 kUA/I	0	Negative	
011990000882	fx2	5	5:04:39 PM	13	Low RU	0	Negative	
011990000882	fx22	5	5:05:39 PM	29	0.15 kUA/I	0	Negative	
011990000882	gx2	5	5:06:38 PM	22	0.07 kUA/I	0	Negative	
011990000882	wx1	5	5:07:39 PM	32	0.18 kUA/I	0	Negative	
011990000882	t20	5	5:08:39 PM	19	0.03 kUA/I	0		
011990000883	fx5	5	4:54:39 PM	25	0.09 kUA/I	0	Negative	
011990000883	fx1	5	4:55:39 PM	18	0.02 kUA/I	0	Negative	
011990000883	fx2	5	4:56:39 PM	16	Low RU	0	Negative	
011990000883	fx22	5	4:57:39 PM	32	0.18 kUA/I	0	Negative	
011990000883	gx2	5	4:58:39 PM	24	0.08 kUA/I	0	Negative	
011990000883	wx1	5	4:59:39 PM	29	0.14 kUA/I	0	Negative	
011990000883	t20	5	5:00:39 PM	19	0.03 kUA/I	0		
011990000884	fx5	5	4:46:39 PM	86	0.81 kUA/I	2	Positive	
011990000884	fx1	5	4:47:39 PM	69	0.61 kUA/I	1	Positive	
011990000884	fx2	5	4:48:40 PM	366	4.05 kUA/I	3	Positive	
011990000884	fx22	5	4:49:40 PM	67	0.59 kUA/I	1	Positive	
011990000884	gx2	5	4:50:39 PM	83	0.77 kUA/I	2	Positive	
011990000884	wx1	5	4:51:39 PM	218	2.34 kUA/I	2	Positive	
011990000884	t20	5	4:52:39 PM	159	1.65 kUA/I	2		
011990000885	fx5	5	4:38:40 PM	20	0.04 kUA/I	0	Negative	
011990000885	fx1	5	4:39:40 PM	19	0.03 kUA/I	0	Negative	
011990000885	fx2	5	4:40:40 PM	13	Low RU	0	Negative	
011990000885	fx22	5	4:41:40 PM	30	0.15 kUA/I	0	Negative	
011990000885	gx2	5	4:42:40 PM	22	0.06 kUA/I	0	Negative	
011990000885	wx1	5	4:43:39 PM	29	0.15 kUA/I	0	Negative	
011990000885	t20	5	4:44:40 PM	18	0.01 kUA/I	0		
011990000886	fx5	5	4:30:40 PM	25	0.10 kUA/I	0	Negative	
011990000886	fx1	5	4:31:40 PM	17	Low RU	0	Negative	
011990000886	fx2	5	4:32:40 PM	57	0.47 kUA/I	1	Positive	
011990000886	fx22	5	4:33:40 PM	28	0.13 kUA/I	0	Negative	
011990000886	gx2	5	4:34:40 PM	50	0.39 kUA/I	1	Positive	
011990000886	wx1	5	4:34:58 PM	---	---	---	---	Err
011990000886	t20	5	4:36:40 PM	22	0.06 kUA/I	0		
011990000887	fx5	5	4:22:41 PM	23	0.07 kUA/I	0	Negative	
011990000887	fx1	5	4:23:41 PM	17	Low RU	0	Negative	
011990000887	fx2	5	4:24:41 PM	13	Low RU	0	Negative	
011990000887	fx22	5	4:25:40 PM	28	0.13 kUA/I	0	Negative	
011990000887	gx2	5	4:26:40 PM	20	0.03 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000887	wx1	5	4:27:40 PM	32	0.18 kUA/l	0	Negative	
011990000887	t20	5	4:28:40 PM	19	0.02 kUA/l	0		
011990000888	fx5	5	4:14:41 PM	62	0.53 kUA/l	1	Positive	
011990000888	fx1	5	4:15:41 PM	39	0.27 kUA/l	0	Negative	
011990000888	fx2	5	4:16:41 PM	1200	13.9 kUA/l	3	Positive	
011990000888	fx22	5	4:17:41 PM	89	0.84 kUA/l	2	Positive	
011990000888	gx2	5	4:18:41 PM	1658	19.6 kUA/l	4	Positive	
011990000888	wx1	5	4:19:41 PM	271	2.96 kUA/l	2	Positive	
011990000888	t20	5	4:20:41 PM	196	2.08 kUA/l	2		
011990000889	fx5	5	4:06:41 PM	20	0.04 kUA/l	0	Negative	
011990000889	fx1	5	4:07:41 PM	15	Low RU	0	Negative	
011990000889	fx2	5	4:08:41 PM	16	Low RU	0	Negative	
011990000889	fx22	5	4:09:41 PM	31	0.16 kUA/l	0	Negative	
011990000889	gx2	5	4:10:41 PM	147	1.52 kUA/l	2	Positive	
011990000889	wx1	5	4:11:41 PM	31	0.17 kUA/l	0	Negative	
011990000889	t20	5	4:12:41 PM	21	0.05 kUA/l	0		
011990000890	fx5	5	3:58:42 PM	21	0.05 kUA/l	0	Negative	
011990000890	fx1	5	3:59:41 PM	15	Low RU	0	Negative	
011990000890	fx2	5	4:00:41 PM	13	Low RU	0	Negative	
011990000890	fx22	5	4:01:41 PM	34	0.20 kUA/l	0	Negative	
011990000890	gx2	5	4:02:42 PM	20	0.04 kUA/l	0	Negative	
011990000890	wx1	5	4:03:41 PM	30	0.15 kUA/l	0	Negative	
011990000890	t20	5	4:04:41 PM	17	Low RU	0		
011990000891	fx5	5	3:49:42 PM	36	0.23 kUA/l	0	Negative	
011990000891	fx1	5	3:50:42 PM	16	Low RU	0	Negative	
011990000891	fx2	5	3:51:42 PM	13	Low RU	0	Negative	
011990000891	fx22	5	3:52:42 PM	30	0.15 kUA/l	0	Negative	
011990000891	gx2	5	3:53:42 PM	28	0.13 kUA/l	0	Negative	
011990000891	wx1	5	3:54:42 PM	31	0.17 kUA/l	0	Negative	
011990000891	t20	5	3:55:41 PM	18	0.01 kUA/l	0		
011990000892	fx5	5	3:41:42 PM	27	0.12 kUA/l	0	Negative	
011990000892	fx1	5	3:42:43 PM	17	Low RU	0	Negative	
011990000892	fx2	5	3:43:42 PM	18	0.01 kUA/l	0	Negative	
011990000892	fx22	5	3:44:42 PM	33	0.19 kUA/l	0	Negative	
011990000892	gx2	5	3:45:42 PM	23	0.08 kUA/l	0	Negative	
011990000892	wx1	5	3:46:42 PM	33	0.19 kUA/l	0	Negative	
011990000892	t20	5	3:47:42 PM	19	0.02 kUA/l	0		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_IgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_IgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
-----------------	------	----------	----------	-----------	---------	----------	----------	-----------	---------	---------

Laboratory Report

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
IQC271119	d1	5:11:38 PM	357J6 21/01	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000891	phad	3:48:42 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000891	fx5	3:49:42 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000891	fx1	3:50:42 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000891	fx2	3:51:42 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000891	fx22	3:52:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000891	gx2	3:53:42 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000891	wx1	3:54:42 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000891	t20	3:55:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	phad	3:40:42 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	fx5	3:41:42 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	fx1	3:42:43 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	fx2	3:43:42 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	fx22	3:44:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	gx2	3:45:42 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	wx1	3:46:42 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000892	t20	3:47:42 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	phad	4:29:41 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	fx5	4:30:40 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	fx1	4:31:40 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	fx2	4:32:40 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	fx22	4:33:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	gx2	4:34:40 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000886	wx1	4:34:58 PM	436BB 21/05	A1HGR 21/06					
011990000886	t20	4:36:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	phad	4:21:41 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	fx5	4:22:41 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	fx1	4:23:41 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	fx2	4:24:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	fx22	4:25:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	gx2	4:26:40 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	wx1	4:27:40 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000887	t20	4:28:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	phad	4:13:41 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	fx5	4:14:41 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	fx1	4:15:41 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	fx2	4:16:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	fx22	4:17:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	gx2	4:18:41 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	wx1	4:19:41 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000888	t20	4:20:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	phad	4:05:41 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	fx5	4:06:41 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	fx1	4:07:41 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	fx2	4:08:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	fx22	4:09:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	gx2	4:10:41 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	wx1	4:11:41 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000889	t20	4:12:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	phad	3:56:42 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	fx5	3:58:42 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	fx1	3:59:41 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	fx2	4:00:41 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	fx22	4:01:41 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	gx2	4:02:42 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	wx1	4:03:41 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000890	t20	4:04:41 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	phad	5:09:38 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	fx5	5:10:38 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	fx1	5:12:38 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000881	fx2	5:25:38 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000882	phad	5:01:39 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000882	fx5	5:02:39 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000882	fx1	5:03:39 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000882	fx2	5:04:39 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000882	fx22	5:05:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000882	gx2	5:06:38 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000882	wx1	5:07:39 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000882	t20	5:08:39 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	phad	4:53:39 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	fx5	4:54:39 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	fx1	4:55:39 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	fx2	4:56:39 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	fx22	4:57:39 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	gx2	4:58:39 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	wx1	4:59:39 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000883	t20	5:00:39 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	phad	4:45:39 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	fx5	4:46:39 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	fx1	4:47:39 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	fx2	4:48:40 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	fx22	4:49:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	gx2	4:50:39 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	wx1	4:51:39 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000884	t20	4:52:39 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	phad	4:37:40 PM	610EL 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	fx5	4:38:40 PM	513EX 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	fx1	4:39:40 PM	447C1 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	fx2	4:40:40 PM	448BT 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	fx22	4:41:40 PM	BPHGE 21/06	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	gx2	4:42:40 PM	433BL 21/08	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	wx1	4:43:39 PM	436BB 21/05	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	
011990000885	t20	4:44:40 PM	475GG 21/07	A1HGR 21/06	BW1SZ 20/03	BNDMB 21/02		AJ0BR 21/05	

Errors

Identity	Test	Measured	Error
011990000886	wx1	4:34:58 PM	303-Shortage of conjugate