


Laboratory Report

ImmunoCAP Specific IgE 0-100

Run number: 200121-383
Instrument name: Phadia-A
Instrument type: Phadia 250
Status: Ready
Started: 21-Jan-20 10:34:50 AM
Ended: 21-Jan-20 2:06:44 PM
Approved: Approved (faziz)
Calibrators in run: No

Calibration Information

Instrument name: Phadia-A
Instrument type: Phadia 250
Measured: 05-Jan-20 1:26:01 PM
Conjugate lot: A1HGR
Acceptance: OK
Active: Yes
Calculation type: Rodbard 5-par



Laboratory Report

Calibrator	Conc	Measured	Resp	Calc	Judgement	Auto	Remarks
CAL-0.00	0.001	Mean	17	0.00			
		1:26:01 PM	19	0.01	OK		
		1:27:01 PM	16	0.00	OK		
		%CV	12.1	>99			
CAL-0.35	0.35	Mean	168	0.35			
		1:28:02 PM	164	0.34	OK		
		1:29:01 PM	172	0.36	OK		
		%CV	3.2	3.5			
CAL-0.70	0.70	Mean	316	0.69			
		1:30:01 PM	318	0.70	OK		
		1:31:01 PM	313	0.69	OK		
		%CV	0.9	0.9			
CAL-3.50	3.50	Mean	1494	3.51			
		1:32:01 PM	1487	3.49	OK		
		1:33:01 PM	1501	3.52	OK		
		%CV	0.7	0.7			
CAL-17.5	17.5	Mean	6298	17.5			
		1:34:01 PM	6324	17.6	OK		
		1:35:01 PM	6272	17.4	OK		
		%CV	0.6	0.7			
CAL-100	100	Mean	18931	100.0			
		1:36:01 PM	18832	98.8	OK		
		1:37:01 PM	19029	101	OK		
		%CV	0.7	1.7			

Curve data	Value	Logged mean	Difference
ED-20	9.59	9.68	-0.92
ED-50	29.9	29.6	+0.97
ED-80	63.2	62.2	+1.70
Slope	225	216	+4.13

Number of curves included in mean: 52

Curve Controls

Identity	Expected	Measured	Resp	Conc	Judgement	Remarks
CC-1	0.70	10:34:50 AM	306	0.67	OK	
CC-2	17.5	10:35:50 AM	6345	17.6	OK	

Quality Controls

Identity	Test	Dil	Measured	Resp	Conc/Quot	Lot	Judgement	Remarks
slgE M	t3	1	10:36:49 AM	3498	8.78 kUA/l	CZ1AR	OK	
slgE L	e1	1	10:37:49 AM	1235	2.87 kUA/l	CZ4AN	OK	
slgE Neg	d1	1	10:38:49 AM	23	0.01 kUA/l	BL1CP	OK	
IQC271119	d1	1	10:39:49 AM	6744	19.0 kUA/l	<lab>	OK	

Samples - Phadiatop

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000155	phad	5	12:46:45 PM	43	0.31 PAU/l	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000156	phad	5	11:26:48 AM	70	0.62 PAU/I	1	Positive	
011990000157	phad	5	10:54:49 AM	31	0.17 PAU/I	0	Negative	
011990000159	phad	5	11:34:47 AM	28	0.13 PAU/I	0	Negative	
011990000160	phad	5	12:06:47 PM	30	0.15 PAU/I	0	Negative	
011990000162	phad	5	1:10:44 PM	167	1.75 PAU/I	2	Positive	
011990000163	phad	5	1:18:44 PM	43	0.31 PAU/I	0	Negative	
011990000165	phad	5	1:26:43 PM	64	0.56 PAU/I	1	Positive	
011990000166	phad	5	1:50:43 PM	49	0.38 PAU/I	1	Positive	
011990000169	phad	5	11:42:47 AM	44	0.32 PAU/I	0	Negative	
011990000172	phad	5	1:02:45 PM	51	0.41 PAU/I	1	Positive	
011990000173	phad	5	1:42:43 PM	33	0.19 PAU/I	0	Negative	
011990000174	phad	5	1:34:43 PM	30	0.16 PAU/I	0	Negative	
011990000177	phad	5	12:54:45 PM	35	0.21 PAU/I	0	Negative	
011990000178	phad	5	10:46:49 AM	34	0.21 PAU/I	0	Negative	
011990000180	phad	5	11:50:47 AM	74	0.68 PAU/I	1	Positive	
011990000183	phad	5	1:58:43 PM	402	4.48 PAU/I	3	Positive	
011990000184	phad	5	12:22:46 PM	35	0.21 PAU/I	0	Negative	
011990000187	phad	5	12:38:45 PM	30	0.15 PAU/I	0	Negative	
011990000190	phad	5	11:58:47 AM	32	0.17 PAU/I	0	Negative	
011990000191	phad	5	12:30:46 PM	65	0.56 PAU/I	1	Positive	
011990000192	phad	5	12:14:46 PM	1031	11.9 PAU/I	3	Positive	
011990000194	phad	5	11:18:48 AM	64	0.56 PAU/I	1	Positive	
011990000195	phad	5	11:10:48 AM	30	0.16 PAU/I	0	Negative	
011990000196	phad	5	11:02:49 AM	29	0.15 PAU/I	0	Negative	

Samples - ImmunoCAP Specific IgE 0-100

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000155	fx5	5	12:47:45 PM	25	0.10 kUA/I	0	Negative	
011990000155	fx1	5	12:48:45 PM	19	0.03 kUA/I	0	Negative	
011990000155	fx2	5	12:49:45 PM	37	0.24 kUA/I	0	Negative	
011990000155	fx22	5	12:50:45 PM	29	0.14 kUA/I	0	Negative	
011990000155	gx2	5	12:51:45 PM	23	0.07 kUA/I	0	Negative	
011990000155	wx1	5	12:52:45 PM	30	0.16 kUA/I	0	Negative	
011990000155	t20	5	12:53:45 PM	15	Low RU	0		
011990000156	fx5	5	11:27:48 AM	20	0.03 kUA/I	0	Negative	
011990000156	fx1	5	11:28:48 AM	19	0.03 kUA/I	0	Negative	
011990000156	fx2	5	11:29:48 AM	71	0.64 kUA/I	1	Positive	
011990000156	fx22	5	11:30:48 AM	30	0.16 kUA/I	0	Negative	
011990000156	gx2	5	11:31:47 AM	22	0.07 kUA/I	0	Negative	
011990000156	wx1	5	11:32:48 AM	28	0.14 kUA/I	0	Negative	
011990000156	t20	5	11:33:47 AM	17	Low RU	0		
011990000157	fx5	5	10:55:49 AM	19	0.02 kUA/I	0	Negative	
011990000157	fx1	5	10:56:49 AM	15	Low RU	0	Negative	
011990000157	fx2	5	10:57:49 AM	11	Low RU	0	Negative	
011990000157	fx22	5	10:58:49 AM	28	0.13 kUA/I	0	Negative	
011990000157	gx2	5	10:59:49 AM	18	0.01 kUA/I	0	Negative	
011990000157	wx1	5	11:00:49 AM	29	0.14 kUA/I	0	Negative	
011990000157	t20	5	11:01:49 AM	16	Low RU	0		
011990000159	fx5	5	11:35:47 AM	21	0.06 kUA/I	0	Negative	
011990000159	fx1	5	11:36:47 AM	16	Low RU	0	Negative	
011990000159	fx2	5	11:37:48 AM	12	Low RU	0	Negative	
011990000159	fx22	5	11:38:47 AM	28	0.13 kUA/I	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000159	gx2	5	11:39:47 AM	23	0.08 kUA/l	0	Negative	
011990000159	wx1	5	11:40:47 AM	27	0.12 kUA/l	0	Negative	
011990000159	t20	5	11:41:47 AM	16	Low RU	0		
011990000160	fx5	5	12:07:46 PM	20	0.03 kUA/l	0	Negative	
011990000160	fx1	5	12:08:46 PM	17	Low RU	0	Negative	
011990000160	fx2	5	12:09:46 PM	14	Low RU	0	Negative	
011990000160	fx22	5	12:10:46 PM	30	0.15 kUA/l	0	Negative	
011990000160	gx2	5	12:11:46 PM	22	0.06 kUA/l	0	Negative	
011990000160	wx1	5	12:12:46 PM	37	0.24 kUA/l	0	Negative	
011990000160	t20	5	12:13:46 PM	16	Low RU	0		
011990000162	fx5	5	1:11:44 PM	26	0.11 kUA/l	0	Negative	
011990000162	fx1	5	1:12:44 PM	15	Low RU	0	Negative	
011990000162	fx2	5	1:13:44 PM	16	Low RU	0	Negative	
011990000162	fx22	5	1:14:44 PM	33	0.19 kUA/l	0	Negative	
011990000162	gx2	5	1:15:44 PM	26	0.11 kUA/l	0	Negative	
011990000162	wx1	5	1:16:44 PM	50	0.39 kUA/l	1	Positive	
011990000162	t20	5	1:17:44 PM	19	0.02 kUA/l	0		
011990000163	fx5	5	1:19:44 PM	21	0.05 kUA/l	0	Negative	
011990000163	fx1	5	1:20:43 PM	13	Low RU	0	Negative	
011990000163	fx2	5	1:21:44 PM	17	Low RU	0	Negative	
011990000163	fx22	5	1:22:44 PM	31	0.17 kUA/l	0	Negative	
011990000163	gx2	5	1:23:44 PM	22	0.06 kUA/l	0	Negative	
011990000163	wx1	5	1:24:43 PM	32	0.18 kUA/l	0	Negative	
011990000163	t20	5	1:25:43 PM	15	Low RU	0		
011990000165	fx5	5	1:27:43 PM	20	0.04 kUA/l	0	Negative	
011990000165	fx1	5	1:28:43 PM	15	Low RU	0	Negative	
011990000165	fx2	5	1:29:43 PM	98	0.95 kUA/l	2	Positive	
011990000165	fx22	5	1:30:44 PM	29	0.15 kUA/l	0	Negative	
011990000165	gx2	5	1:31:43 PM	22	0.06 kUA/l	0	Negative	
011990000165	wx1	5	1:32:43 PM	28	0.13 kUA/l	0	Negative	
011990000165	t20	5	1:33:43 PM	16	Low RU	0		
011990000166	fx5	5	1:51:43 PM	22	0.06 kUA/l	0	Negative	
011990000166	fx1	5	1:52:43 PM	15	Low RU	0	Negative	
011990000166	fx2	5	1:53:42 PM	19	0.02 kUA/l	0	Negative	
011990000166	fx22	5	1:54:42 PM	30	0.16 kUA/l	0	Negative	
011990000166	gx2	5	1:55:42 PM	24	0.08 kUA/l	0	Negative	
011990000166	wx1	5	1:56:42 PM	28	0.14 kUA/l	0	Negative	
011990000166	t20	5	1:57:42 PM	17	0.01 kUA/l	0		
011990000169	fx5	5	11:43:47 AM	38	0.25 kUA/l	0	Negative	
011990000169	fx1	5	11:44:47 AM	19	0.02 kUA/l	0	Negative	
011990000169	fx2	5	11:45:47 AM	18	0.01 kUA/l	0	Negative	
011990000169	fx22	5	11:46:47 AM	29	0.15 kUA/l	0	Negative	
011990000169	gx2	5	11:47:47 AM	24	0.09 kUA/l	0	Negative	
011990000169	wx1	5	11:48:47 AM	29	0.14 kUA/l	0	Negative	
011990000169	t20	5	11:49:47 AM	17	Low RU	0		
011990000172	fx5	5	1:03:45 PM	18	0.01 kUA/l	0	Negative	
011990000172	fx1	5	1:04:44 PM	14	Low RU	0	Negative	
011990000172	fx2	5	1:05:44 PM	33	0.19 kUA/l	0	Negative	
011990000172	fx22	5	1:06:44 PM	34	0.21 kUA/l	0	Negative	
011990000172	gx2	5	1:07:44 PM	21	0.04 kUA/l	0	Negative	
011990000172	wx1	5	1:08:44 PM	31	0.17 kUA/l	0	Negative	
011990000172	t20	5	1:09:44 PM	17	Low RU	0		
011990000173	fx5	5	1:43:43 PM	22	0.06 kUA/l	0	Negative	
011990000173	fx1	5	1:44:43 PM	16	Low RU	0	Negative	

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000173	fx2	5	1:45:43 PM	16	Low RU	0	Negative	
011990000173	fx22	5	1:46:43 PM	29	0.14 kUA/I	0	Negative	
011990000173	gx2	5	1:47:42 PM	19	0.03 kUA/I	0	Negative	
011990000173	wx1	5	1:48:43 PM	28	0.14 kUA/I	0	Negative	
011990000173	t20	5	1:49:43 PM	18	0.01 kUA/I	0		
011990000174	fx5	5	1:35:43 PM	19	0.02 kUA/I	0	Negative	
011990000174	fx1	5	1:36:43 PM	14	Low RU	0	Negative	
011990000174	fx2	5	1:37:43 PM	14	Low RU	0	Negative	
011990000174	fx22	5	1:38:43 PM	29	0.14 kUA/I	0	Negative	
011990000174	gx2	5	1:39:43 PM	20	0.04 kUA/I	0	Negative	
011990000174	wx1	5	1:40:43 PM	31	0.16 kUA/I	0	Negative	
011990000174	t20	5	1:41:43 PM	17	Low RU	0		
011990000177	fx5	5	12:55:45 PM	20	0.04 kUA/I	0	Negative	
011990000177	fx1	5	12:56:45 PM	16	Low RU	0	Negative	
011990000177	fx2	5	12:57:44 PM	14	Low RU	0	Negative	
011990000177	fx22	5	12:58:44 PM	32	0.18 kUA/I	0	Negative	
011990000177	gx2	5	12:59:44 PM	26	0.11 kUA/I	0	Negative	
011990000177	wx1	5	1:00:44 PM	29	0.15 kUA/I	0	Negative	
011990000177	t20	5	1:01:44 PM	16	Low RU	0		
011990000178	fx5	5	10:47:49 AM	21	0.05 kUA/I	0	Negative	
011990000178	fx1	5	10:48:49 AM	21	0.05 kUA/I	0	Negative	
011990000178	fx2	5	10:49:49 AM	15	Low RU	0	Negative	
011990000178	fx22	5	10:50:49 AM	30	0.16 kUA/I	0	Negative	
011990000178	gx2	5	10:51:49 AM	23	0.08 kUA/I	0	Negative	
011990000178	wx1	5	10:52:49 AM	319	3.51 kUA/I	3	Positive	
011990000178	t20	5	10:53:49 AM	17	Low RU	0		
011990000180	fx5	5	11:51:47 AM	41	0.29 kUA/I	0	Negative	
011990000180	fx1	5	11:52:47 AM	18	0.01 kUA/I	0	Negative	
011990000180	fx2	5	11:53:47 AM	27	0.12 kUA/I	0	Negative	
011990000180	fx22	5	11:54:47 AM	34	0.20 kUA/I	0	Negative	
011990000180	gx2	5	11:55:47 AM	53	0.42 kUA/I	1	Positive	
011990000180	wx1	5	11:56:47 AM	34	0.20 kUA/I	0	Negative	
011990000180	t20	5	11:57:47 AM	17	Low RU	0		
011990000183	fx5	5	1:59:42 PM	27	0.12 kUA/I	0	Negative	
011990000183	fx1	5	2:00:42 PM	19	0.02 kUA/I	0	Negative	
011990000183	fx2	5	2:01:42 PM	386	4.28 kUA/I	3	Positive	
011990000183	fx22	5	2:02:42 PM	31	0.16 kUA/I	0	Negative	
011990000183	gx2	5	2:03:42 PM	27	0.12 kUA/I	0	Negative	
011990000183	wx1	5	2:04:42 PM	81	0.76 kUA/I	2	Positive	
011990000183	t20	5	2:05:42 PM	41	0.29 kUA/I	0		
011990000184	fx5	5	12:23:46 PM	19	0.02 kUA/I	0	Negative	
011990000184	fx1	5	12:24:46 PM	14	Low RU	0	Negative	
011990000184	fx2	5	12:25:46 PM	13	Low RU	0	Negative	
011990000184	fx22	5	12:26:46 PM	29	0.15 kUA/I	0	Negative	
011990000184	gx2	5	12:27:46 PM	21	0.05 kUA/I	0	Negative	
011990000184	wx1	5	12:28:46 PM	26	0.11 kUA/I	0	Negative	
011990000184	t20	5	12:29:46 PM	17	0.01 kUA/I	0		
011990000187	fx5	5	12:39:45 PM	18	0.02 kUA/I	0	Negative	
011990000187	fx1	5	12:40:45 PM	15	Low RU	0	Negative	
011990000187	fx2	5	12:41:45 PM	12	Low RU	0	Negative	
011990000187	fx22	5	12:42:45 PM	30	0.15 kUA/I	0	Negative	
011990000187	gx2	5	12:43:45 PM	24	0.08 kUA/I	0	Negative	
011990000187	wx1	5	12:44:45 PM	29	0.14 kUA/I	0	Negative	
011990000187	t20	5	12:45:45 PM	17	Low RU	0		

Laboratory Report

Identity	Test	Dil	Measured	Resp	Conc	Class	Cut-Off	Remarks
011990000190	fx5	5	11:59:47 AM	18	0.01 kUA/l	0	Negative	
011990000190	fx1	5	12:00:47 PM	17	Low RU	0	Negative	
011990000190	fx2	5	12:01:46 PM	14	Low RU	0	Negative	
011990000190	fx22	5	12:02:46 PM	34	0.20 kUA/l	0	Negative	
011990000190	gx2	5	12:03:46 PM	29	0.14 kUA/l	0	Negative	
011990000190	wx1	5	12:04:46 PM	30	0.15 kUA/l	0	Negative	
011990000190	t20	5	12:05:47 PM	16	Low RU	0		
011990000191	fx5	5	12:31:45 PM	21	0.05 kUA/l	0	Negative	
011990000191	fx1	5	12:32:45 PM	23	0.08 kUA/l	0	Negative	
011990000191	fx2	5	12:33:46 PM	64	0.55 kUA/l	1	Positive	
011990000191	fx22	5	12:34:46 PM	28	0.13 kUA/l	0	Negative	
011990000191	gx2	5	12:35:46 PM	21	0.05 kUA/l	0	Negative	
011990000191	wx1	5	12:36:45 PM	31	0.17 kUA/l	0	Negative	
011990000191	t20	5	12:37:45 PM	16	Low RU	0		
011990000192	fx5	5	12:15:46 PM	24	0.08 kUA/l	0	Negative	
011990000192	fx1	5	12:16:46 PM	14	Low RU	0	Negative	
011990000192	fx2	5	12:17:46 PM	19	0.03 kUA/l	0	Negative	
011990000192	fx22	5	12:18:46 PM	29	0.14 kUA/l	0	Negative	
011990000192	gx2	5	12:19:46 PM	29	0.15 kUA/l	0	Negative	
011990000192	wx1	5	12:20:46 PM	30	0.15 kUA/l	0	Negative	
011990000192	t20	5	12:21:46 PM	18	0.01 kUA/l	0		
011990000194	fx5	5	11:19:48 AM	25	0.10 kUA/l	0	Negative	
011990000194	fx1	5	11:20:48 AM	17	Low RU	0	Negative	
011990000194	fx2	5	11:21:48 AM	15	Low RU	0	Negative	
011990000194	fx22	5	11:22:48 AM	32	0.18 kUA/l	0	Negative	
011990000194	gx2	5	11:23:48 AM	22	0.06 kUA/l	0	Negative	
011990000194	wx1	5	11:24:48 AM	101	0.98 kUA/l	2	Positive	
011990000194	t20	5	11:25:48 AM	41	0.29 kUA/l	0		
011990000195	fx5	5	11:11:48 AM	36	0.23 kUA/l	0	Negative	
011990000195	fx1	5	11:12:48 AM	20	0.04 kUA/l	0	Negative	
011990000195	fx2	5	11:13:48 AM	14	Low RU	0	Negative	
011990000195	fx22	5	11:14:48 AM	28	0.13 kUA/l	0	Negative	
011990000195	gx2	5	11:15:48 AM	21	0.05 kUA/l	0	Negative	
011990000195	wx1	5	11:16:48 AM	26	0.11 kUA/l	0	Negative	
011990000195	t20	5	11:17:48 AM	17	Low RU	0		
011990000196	fx5	5	11:03:48 AM	18	0.01 kUA/l	0	Negative	
011990000196	fx1	5	11:04:48 AM	16	Low RU	0	Negative	
011990000196	fx2	5	11:05:48 AM	11	Low RU	0	Negative	
011990000196	fx22	5	11:06:48 AM	27	0.12 kUA/l	0	Negative	
011990000196	gx2	5	11:07:48 AM	20	0.03 kUA/l	0	Negative	
011990000196	wx1	5	11:08:48 AM	27	0.13 kUA/l	0	Negative	
011990000196	t20	5	11:09:49 AM	16	Low RU	0		

Lot Numbers

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-0.00	a_IgE	1:26:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.00	a_IgE	1:27:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:28:02 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.35	a_IgE	1:29:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:30:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-0.70	a_IgE	1:31:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:32:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-3.50	a_IgE	1:33:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Laboratory Report

Calibrator	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Calibrator	Remarks
CAL-17.5	a_lgE	1:34:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-17.5	a_lgE	1:35:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:36:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	
CAL-100	a_lgE	1:37:01 PM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BUADN 20/04	

Curve control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Curve ctrl	Remarks
CC-1	a_lgE	10:34:50 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	
CC-2	a_lgE	10:35:50 AM	865FX 20/02	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		A0VH7 21/04	

Quality control	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Qual ctrl	Diluent	Remarks
slgE M	t3	10:36:49 AM	352D0 19/12!	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ1AR 21/08		Exp
slgE L	e1	10:37:49 AM	359EK 20/10	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		CZ4AN 19/06!		Exp
slgE Neg	d1	10:38:49 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		BL1CP 19/12!		Exp
IQC271119	d1	10:39:49 AM	357J6 21/01	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		<lab> 81/11		

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000178	phad	10:46:49 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000178	fx5	10:47:49 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000178	fx1	10:48:49 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000178	fx2	10:49:49 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000178	fx22	10:50:49 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000178	gx2	10:51:49 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000178	wx1	10:52:49 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000178	t20	10:53:49 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	phad	10:54:49 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	fx5	10:55:49 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	fx1	10:56:49 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	fx2	10:57:49 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	fx22	10:58:49 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	gx2	10:59:49 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	wx1	11:00:49 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000157	t20	11:01:49 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	phad	11:02:49 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	fx5	11:03:48 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	fx1	11:04:48 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	fx2	11:05:48 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	fx22	11:06:48 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	gx2	11:07:48 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	wx1	11:08:48 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000196	t20	11:09:49 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	phad	11:10:48 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	fx5	11:11:48 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	fx1	11:12:48 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	fx2	11:13:48 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	fx22	11:14:48 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	gx2	11:15:48 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	wx1	11:16:48 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000195	t20	11:17:48 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	phad	11:18:48 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	fx5	11:19:48 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	fx1	11:20:48 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	fx2	11:21:48 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	fx22	11:22:48 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	gx2	11:23:48 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	wx1	11:24:48 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000194	t20	11:25:48 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	phad	11:26:48 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	fx5	11:27:48 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	fx1	11:28:48 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	fx2	11:29:48 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	fx22	11:30:48 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	gx2	11:31:47 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	wx1	11:32:48 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000156	t20	11:33:47 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000159	phad	11:34:47 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000159	fx5	11:35:47 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000159	fx1	11:36:47 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000159	fx2	11:37:48 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000159	fx22	11:38:47 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000159	gx2	11:39:47 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000159	wx1	11:40:47 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000159	t20	11:41:47 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	phad	11:42:47 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	fx5	11:43:47 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	fx1	11:44:47 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	fx2	11:45:47 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	fx22	11:46:47 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	gx2	11:47:47 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	wx1	11:48:47 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000169	t20	11:49:47 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	phad	11:50:47 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	fx5	11:51:47 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	fx1	11:52:47 AM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	fx2	11:53:47 AM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	fx22	11:54:47 AM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	gx2	11:55:47 AM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	wx1	11:56:47 AM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000180	t20	11:57:47 AM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	phad	11:58:47 AM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	fx5	11:59:47 AM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	fx1	12:00:47 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	fx2	12:01:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	fx22	12:02:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	gx2	12:03:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	wx1	12:04:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000190	t20	12:05:47 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	phad	12:06:47 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	fx5	12:07:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	fx1	12:08:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	fx2	12:09:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	fx22	12:10:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	gx2	12:11:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	wx1	12:12:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000160	t20	12:13:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	phad	12:14:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	fx5	12:15:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	fx1	12:16:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	fx2	12:17:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	fx22	12:18:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	gx2	12:19:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	wx1	12:20:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000192	t20	12:21:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	phad	12:22:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	fx5	12:23:46 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	fx1	12:24:46 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	fx2	12:25:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	fx22	12:26:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	gx2	12:27:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	wx1	12:28:46 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000184	t20	12:29:46 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	phad	12:30:46 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	fx5	12:31:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	fx1	12:32:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	fx2	12:33:46 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	fx22	12:34:46 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	gx2	12:35:46 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	wx1	12:36:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000191	t20	12:37:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	
011990000187	phad	12:38:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJOBP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000187	fx5	12:39:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000187	fx1	12:40:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000187	fx2	12:41:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000187	fx22	12:42:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000187	gx2	12:43:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000187	wx1	12:44:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000187	t20	12:45:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	phad	12:46:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	fx5	12:47:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	fx1	12:48:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	fx2	12:49:45 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	fx22	12:50:45 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	gx2	12:51:45 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	wx1	12:52:45 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000155	t20	12:53:45 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	phad	12:54:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	fx5	12:55:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	fx1	12:56:45 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	fx2	12:57:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	fx22	12:58:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	gx2	12:59:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	wx1	1:00:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000177	t20	1:01:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	phad	1:02:45 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	fx5	1:03:45 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	fx1	1:04:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	fx2	1:05:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	fx22	1:06:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	gx2	1:07:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	wx1	1:08:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000172	t20	1:09:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	phad	1:10:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	fx5	1:11:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	fx1	1:12:44 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	fx2	1:13:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	fx22	1:14:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	gx2	1:15:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	wx1	1:16:44 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000162	t20	1:17:44 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	phad	1:18:44 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	fx5	1:19:44 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	fx1	1:20:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	fx2	1:21:44 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	fx22	1:22:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	gx2	1:23:44 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	wx1	1:24:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000163	t20	1:25:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	phad	1:26:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	fx5	1:27:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	fx1	1:28:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	fx2	1:29:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	fx22	1:30:44 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	gx2	1:31:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	wx1	1:32:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000165	t20	1:33:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	phad	1:34:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	fx5	1:35:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	fx1	1:36:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	fx2	1:37:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	fx22	1:38:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	gx2	1:39:43 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	wx1	1:40:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000174	t20	1:41:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000173	phad	1:42:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000173	fx5	1:43:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Laboratory Report

Sample	Test	Measured	Cap/Well	Conjugate	Dev sol	Stop sol	Wash sol	Diluent	Remarks
011990000173	fx1	1:44:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000173	fx2	1:45:43 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000173	fx22	1:46:43 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000173	gx2	1:47:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000173	wx1	1:48:43 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000173	t20	1:49:43 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	phad	1:50:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	fx5	1:51:43 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	fx1	1:52:43 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	fx2	1:53:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	fx22	1:54:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	gx2	1:55:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	wx1	1:56:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000166	t20	1:57:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	phad	1:58:43 PM	610EL 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	fx5	1:59:42 PM	513EX 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	fx1	2:00:42 PM	447C1 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	fx2	2:01:42 PM	448BT 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	fx22	2:02:42 PM	BPHGE 21/06	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	gx2	2:03:42 PM	433BL 21/08	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	wx1	2:04:42 PM	436BB 21/05	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	
011990000183	t20	2:05:42 PM	475GG 21/07	A1HGR 21/06	BW1SS 20/02	BNDMB 21/02		AJ0BP 20/11	

Errors

Expired lot numbers used for Quality Control

352D0	ImmunoCAP
CZ4AN	Quality control vial
BL1CP	Quality control vial