

Type Evaluation / Verification Report

Instrument Under Test		Test Conditions	
Manufacturer	Accura Scales Pvt. Ltd.	Test Location	Lab Bay 2
Model	AX-500	Tested On	2026-09-25 11:00
Serial No.	AX500-2291	Tested By	Rahul Mehta
Accuracy Class	III	Temperature	24.5 °C
Max Capacity	5000 g	Humidity	55 %RH
Min Capacity	20 g	Pressure	1004 hPa
Scale Interval (e)	1 g		
Verification Scale Interval (d)	1 g		
Max Tare	2000 g		

⚠ **Rounding-trap detected.** The naive calculation and the R 76-1 changeover-point method disagree on the verdict for at least one reading below. The R 76-1 result is authoritative.

Marginal result(s). The following readings used more than 80% of the allowed MPE.

Test Results

Test	Load (g)	Error (g)	MPE (±g)	MPE Used	Naive rounding	Result	Clause	Explanation
weighing	1000	-3.0	2.5	120%	2.0 → PASS	FAIL	R76-1 3.6.1	Error of -3.0 g exceeds MPE of ±2.5 g for loads between 500e and 2000e.
eccentricity	2500	1.0	5.0	20%	1.0 → PASS	PASS	R76-1 3.6.2	Corner-load error within MPE.
repeatability ⚠	5000	2.1	2.5	84%	2.1 → PASS	PASS	R76-1 3.6.3	Spread of repeat readings uses 84% of MPE (marginal).

Overall Verdict: FAIL

Dr. Anita Rao
Chief Metrology Officer
Approved On: 2026-09-26 14:32



Verify online

Scan the QR code or visit the link below to verify this report is genuine and unmodified. <https://scalesaathi.example.org/verify/3f9a7c1e2b8d4f0a6c5e9b1d7a2f4c8e0b6d3a1f9c7e5b2d4a6f8c0e2b4d6f81>
This report is system-generated and cryptographically hashed at approval. Any change after approval will be flagged at the verification link above.