












Plan Quality Algorithm: NSABP B-51_RTOG 1304 Arm 2 Group 2A Right Side 64Gy Photons_Electon composite [26 Max Possible] [26 Metrics] (Page 1 of 3)

1	 V (%) D(Gy)	Volume (%) of the PTV EVAL covered by 47.5 (Gy)	1	
2	 V (%) D(Gy)	Volume (%) of the PTV EVAL covered by 45 (Gy)	1	
3	 V (%) D(Gy)	Dose (Gy) covering 50 (%) of the PTV EVAL	1	
4	 V (%) D(Gy)	Volume (%) of the PTV EVAL covered by 64 (Gy)	1	
5	 ROI Max (Gy)	Maximum dose (Gy) inside the PTV EVAL	1	
6	 V (cc) D(Gy)	Volume (cc) of the PTV EVAL covered by 62.5 (Gy)	1	
7	 V (%) D(Gy)	Volume (%) of the LUMPECTOMY PTV EVAL_64GY covered by 60.8 (Gy)	1	
8	 V (%) D(Gy)	Volume (%) of the LUMPECTOMY PTV EVAL_64GY covered by 57.6 (Gy)	1	
9	 V (%) D(Gy)	Volume (%) of the LUMPECTOMY PTV EVAL_64GY covered by 70.4 (Gy)	1	
10	 ROI Max (Gy)	Maximum dose (Gy) inside the LUMPECTOMY PTV EVAL_64GY	1	

Plan Quality Algorithm: NSABP B-51_RTOG 1304 Arm 2 Group 2A Right Side 64Gy Photons_Electon composite [26 Max Possible] [26 Metrics] (Page 2 of 3)



Plan Quality Algorithm: NSABP B-51_RTOG 1304 Arm 2 Group 2A Right Side 64Gy Photons_Electon composite [26 Max Possible] [26 Metrics] (Page 3 of 3)

21	 V (%) D(Gy)	Volume (%) of the IPSILATERAL LUNG covered by 10 (Gy)	1	
22	 V (%) D(Gy)	Volume (%) of the IPSILATERAL LUNG covered by 5 (Gy)	1	
23	 V (%) D(Gy)	Volume (%) of the CONTRALATERAL LUNG covered by 5 (Gy)	1	
24	 ROI Max (Gy)	Maximum dose (Gy) inside the HEART_RT SIDE	1	
25	 V (%) D(Gy)	Volume (%) of the HEART_RT SIDE covered by 15 (Gy)	1	
26	 ROI Mean (Gy)	Mean dose (Gy) inside the HEART	1	