

Indiana Standards Laboratory

2919 SHELBY STREET
INDIANAPOLIS, INDIANA 46203-5236

REPORT of CALIBRATION

REISSUED

| | | | |
|-----------------------|---|---------------|----------------|
| DATE | : 02/27/12 | TEST NUMBER | : 114000 |
| MANUFACTURER | : BW Technologies | MODEL NUMBER | : MC-XWHM-Y-NA |
| NOMENCLATURE | : Detector, Gas | SERIAL NUMBER | : KA208-013104 |
| TOLERANCE | : Manufacturer Specifications | | |
| | | | |
| SUBMITTED BY | : Dresser-Roots 900 W. Mount Street GE Energy Connersville, IN 47331 | P.O. NUMBER | : Credit Card |
| PROCEDURE | : Manufacturer | TEMPERATURE | : 20 °C |
| CAL INTERVAL | : 6 EOM | HUMIDITY | : 35 %RH |
| PREVIOUS TEST NO. | : 110017 | NEXT CAL DUE | : 08/29/12 |
| | | TECHNICIAN | : MWC |
| RCV'D CONDITION | : In-tolerance | | |
| RETURNED CONDITION | : In-tolerance | | |
| COMMENTS | : Oxygen replaced. Replaced Flammable Sensor | | |
| | | | |
| CALIBRATION SOURCE(S) | : 110914 | | |
| UNCERTAINTY: | As per the customer's request, measurement uncertainty is not reported. | | |

CERTIFICATION

After careful inspection of the above-described instrument, tests were performed on-site to establish calibration accuracy. All equipment utilized in the testing is maintained by this laboratory and checked frequently against standards that have been certified traceable to SI Units through NIST or another ILAC recognized NMI.

It is hereby certified that the test instrument meets all the original manufacturer's performance specifications unless otherwise noted above. Test documentation is on file and may be viewed in our office. Testing was completed per ISL Quality Manual: Doc-001-Rev12 dated 8-11-2009.

Richard F. Chance



Chief, Metrology

This report shall not be reproduced except in full, without the written approval of Indiana Standards Laboratory.

DATE : 02/27/12

TEST NUMBER : 114000

MANUFACTURER : BW Technologies

MODEL NUMBER : MC-XWHM-Y-NA

NOMENCLATURE : Detector, Gas

SERIAL NUMBER : KA208-013104

TOLERANCE : Manufacturer Specifications

CALIBRATION SOURCES

| Source | Manufacturer | Nomenclature | Model | Serial | Cal Due |
|--------|--------------|-----------------|------------|----------|----------|
| 110914 | NorLab | Calibration Gas | F105340PM3 | 1-237-21 | 09/30/12 |

CALIBRATION DATA

| Gas | Standard | Indicated |
|----------------|----------|-----------|
| H2S (ppm) | 40 | 40 |
| CO (ppm) | 100 | 100 |
| Methane (%LEL) | 50 | 49 |
| Oxygen (%) | 15 | 14.8 |

*Out-of-tolerance

**Indeterminate

Ø = Fixed Zero

Note: Uncertainty is not taken into account in pass / fail function checks.