

System Configuration

Software Product Name:	QuantStudio™ Real-Time PCR Software v1.7.2
Report Generation Date:	10-18-2023 13:32:16 CDT
Instrument Name:	278871950
Instrument Type:	QuantStudio™ 7 Flex System
Instrument Serial Number:	278871950
Instrument Firmware Version:	1.1.2
Block Type:	96-Well Block (0.2mL)
Block Serial Number:	278023333
Block Firmware Version:	20
Heated Cover Serial Number:	278123247
Heated Cover Firmware Version:	10
Heated Cover Type:	96 Wells
Lamp Life:	205 hours

Run Information

Run Operator ID:	GUEST
Run Operator Name:	GUEST
Calibration Run Date:	05-30-2023 15:15:07 CDT
Analysis Status:	PASSED
Comments:	

Reagent Information


Plate Barcode:	C920AOUH
Reagent Name:	Cal Plate 2
Reagent Part Number:	A26332
Reagent Lot Number:	2208215
Reagent Expiration Date:	2023-08-24



Dependent Calibration

ROI Calibration Date: 05-30-2023 14:08:23 CDT
Background Calibration Date: 05-30-2023 14:32:11 CDT
Uniformity Calibration Date: 05-30-2023 14:42:45 CDT

Customer is responsible for any validation of assays and compliance with regulatory requirements that pertain to their procedures and uses of reagents and/or instruments.
© 2021 Thermo Fisher Scientific. All rights reserved. The trademarks mentioned herein are the property of Thermo Fisher Scientific or their respective owners.



For Research Use Only. Not for use in diagnostic procedures.

Printed:

10-18-2023 13:32:16 CDT