

Field reference notes: calibration record sheet

Before you start

- Note the reference instrument and its certificate number.
- Choose steady weather: no frontal passage, no rain for pressure checks.

For each sensor, record

Field	Example
Sensor and serial number	Barometer, serial as on label
Station reading	Value shown by the station sensor
Reference reading	Value shown by the reference at the same time
Difference	Station minus reference
Action	Within tolerance / corrected / reported for adjustment

Exposure checks

- Temperature sensor inside a clean, ventilated radiation screen, 1.25 to 2 m above natural ground.
- Rain gauge level, funnel and filter clear, no nearby obstruction.

Always

- Take several paired readings, never just one.
- Write the entry in the maintenance log the same day, with your name.