Kinematic Motion Gating and Multi-Attribute Signal Trust Calibration for Resilient Ambulatory Heart-Rate Alerting: An Empirical Benchmark

Student 1
Dept. of Computer Science & Eng.
Vel Tech R&D Institute
Chennai, India

Student 2
Dept. of Computer Science & Eng.
Vel Tech R&D Institute
Chennai, India

Student 3
Dept. of Computer Science & Eng.
Vel Tech R&D Institute
Chennai, India

Faculty Supervisor
Dept. of Computer Science & Eng.
Vel Tech R&D Institute
Chennai, India

***Abstract*—Wearable cardiovascular monitors frequently combine signal quality heuristics with alerting thresholds to detect resting tachycardia. However, the incremental utility of multi-signal software trust scoring over simple kinematic motion gating is rarely benchmarked under pre-registered protocols. We present a reproducible, pre-registered fault-injection benchmark evaluating five alerting arms across fifteen ambulatory subjects from the public PPG-DaLiA dataset. To address natural event sparsity in healthy individuals, 112 labelled synthetic resting tachycardia episodes were injected into derived pulse-rate streams alongside two seeded telemetry fault regimes and raw optical noise. Evaluated in a strict leave-one-subject-out design with standard persistence and refractory hold-off, our primary hypothesis H1 was not supported. Kinematic motion gating drove nearly the entire reduction in false alerts per hour (reducing false alerts by roughly two-thirds to three-quarters relative to ungated personal baselines), while multi-attribute software trust scoring yielded no statistically significant improvement across any evaluated condition (paired differences were negligible and all ninety-five percent bootstrap confidence intervals spanned zero). Furthermore, aggressive software trust scoring degraded event recall under intense optical noise. Fixed population thresholding with persistence proved to be an unexpectedly robust baseline. Because events were synthetically injected and evaluated in healthy volunteers, these findings are exploratory and carry no clinical standing.**

***Index Terms*—wearable sensors, photoplethysmography, heart-rate alerting, data quality, fault injection, negative result, reproducibility**

I. Introduction

Ambulatory continuous monitoring with consumer and medical-grade wearables offers substantial promise for early arrhythmia detection. However, wrist photoplethysmography (PPG) is acutely vulnerable to motion artifacts, optical decoupling, transmission dropouts, and flatline sensor lockups. When raw optical or derived telemetry signals degrade, alerting systems frequently emit spurious clinical notifications. In home monitoring and clinical telemetry alike, high false-alert frequencies precipitate severe alarm fatigue, inducing desensitization and compromising user adherence [10].

To prevent spurious alarms, modern digital health frameworks deploy complex signal quality indices (SQIs) and multi-component trust heuristics that penalize missing samples, flag frozen sensor values, and bound rate-of-change dynamics. Despite their architectural prevalence, the isolated incremental utility of multi-component software trust scores over straightforward kinematic stillness gating (e.g., Euclidean Norm Minus One, ENMO) has not been demonstrated in pre-registered, cross-validated benchmarks. Many prior evaluations suffer from train-test leakage, post-hoc hyperparameter selection, or lack of minimal baseline comparators.

This paper reports a rigorous, pre-registered fault-injection benchmark for wearable heart-rate alerting. Our contributions are:

1) **Open Benchmarking Framework:** A fully reproducible pipeline formalizing an eight-type fault taxonomy (clean, gaps, stale values, duplicates, spikes, drift, scaling errors, and raw optical noise) evaluated across 31.25 hours of multi-activity ambulatory recordings.
2) **Causal Leave-One-Subject-Out (LOSO) Evaluation:** Five distinct alerting arms evaluated with strict data quarantine, separating training folds from held-out evaluation subjects.
3) **Rigorous Negative Finding:** Empirical refutation of pre-registered hypothesis H_1, showing that kinematic motion gating delivers 67.9%–74.2% false alert reduction, whereas multi-attribute software trust scoring provides no statistically detectable incremental benefit ($p \geq 0.28$).

II. Literature Survey

Wrist-based pulse rate monitoring relies on optical absorption changes in microvascular tissue beds [1], [3]. Relative motion between sensor photodiodes and cutaneous capillary beds introduces severe motion artifacts that overlap spectrally with true cardiac fundamental frequencies [1], [9].

To isolate motion artifact and evaluate waveform integrity, numerous signal quality indices (SQIs) have been proposed, spanning statistical moments (skewness, kurtosis), spectral purity, and template cross-correlation [2]. While SQIs achieve strong classification performance on stationary laboratory datasets, ambulatory evaluations reveal severe performance drops during daily locomotion [1], [3]. In clinical deployments, uncertainty quantification and selective prediction frameworks, such as conformal prediction, provide distribution-free finite-sample guarantees that offer structured paths for principled abstention [4]. Table I contrasts existing wearable alerting paradigms with our pre-registered framework.

TABLE I
COMPARATIVE OVERVIEW OF WEARABLE ALERTING PARADIGMS

Study / Paradigm	Pre-Reg.	Ablation	LOSO	Fault Inject.
Deep PPG [1]	No	No	Yes	No
Optimal PPG SQI [2]	No	No	No	No
PPG Roadmap [3]	Review	N/A	N/A	N/A
Conformal Clinical UQ [4]	Review	N/A	N/A	N/A
This Benchmark	**Yes**	**Yes**	**Yes**	**Yes**

III. DATASET AND PROTOCOL

A. PPG-DaLiA Dataset

We evaluate on the public PPG-DaLiA benchmark (UCI ID 495, CC BY 4.0) [1]. The dataset comprises 15 healthy subjects undergoing 2.5 hours of authentic daily activities (Table II). Reference heart rates are derived from gold-standard chest ECG. Raw wrist PPG (64 Hz) and triaxial accelerometry (32 Hz) were captured with an Empatica E4 wristband. Strict quarantine was maintained: chest ECG reference was isolated solely for ground truth evaluation and was inaccessible to alerting arms.

TABLE II
PPG-DALIA BENCHMARK DATASET SUMMARY [1]

Attribute	Specification
Subjects	$N = 15$ healthy individuals (7M, 8F, age 21–55, mean 30.6)
Protocol Duration	$\approx$ 2.5 h per subject (31.25 total ambulatory hours)
Activities	Sitting, Walking, Stairs, Cycling, Driving, Lunch, Soccer
Wrist Sensors	Empatica E4 (BVP: 64 Hz, 3D ACC: 32 Hz, GSR/Temp: 4 Hz)
Reference Standard	Chest ECG (RespiBAN Professional, 700 Hz)
Quarantine	Chest ECG strictly quarantined from alerting arms

B. Labelled Synthetic Tachycardia Injection

Because natural resting tachycardia was virtually absent in this healthy volunteer cohort (manifesting in only 3 of 15 subjects), evaluating sensitivity on raw data is statistically uninformative. Per Amendment 2, we injected 112 labelled synthetic resting tachycardia episodes into derived pulse-rate streams during verified low-motion windows ($\text{ENMO}_{10\text{s}} \leq 0.05$ g). Peak heart rates were sampled between 105 and 135 bpm, with linear ramp durations of 15 s and minimum inter-episode gaps of 60 s. To respect participant baseline heart rates HR_{base}, lower and upper ramp bounds were adaptively adjusted: $r_{\text{lo}} = \max(2.0, (105 - HR_{\text{base}})/15)$, $r_{\text{hi}} = \min(4.0, (135 - HR_{\text{base}})/15)$, with plateau peak heart rate clamped to $[105.0, 135.0]$ bpm.

C. Seeded Fault Regimes

Sensor telemetry faults were simulated under two pre-specified schedules alongside an optical corruption condition (Table III):

- **Condition 0 (No-Fault):** Clean ambulatory background.
- **Regime A (Low Duty, $\approx 10\%$ faulted time):** Sporadic gaps, stale values, duplicates, and slow drift (seed 42).
- **Regime B (Nominal Duty, $\approx$ 25–30% faulted time):** Frequent dropouts, prolonged flatlines, and scale errors (seed 42).
- **F7 (High-Motion Optical Noise):** Additive Gaussian noise continuously modulated by wrist 3D acceleration injected into raw 64 Hz BVP at SNR $+5.0$ dB prior to Step 3a spectral estimation.

TABLE III
FAULT TAXONOMY AND SEEDED REGIMES (PROTOCOL AMENDMENT 3)

Fault ID	Mechanism & Parameter	Reg. A	Reg. B
F0_CLEAN	Ambient background	28.09 h	21.67 h
F1_GAPS	Packet dropout (5–15 s NaN)	0.94 h	3.15 h
F2_STALE	Sensor freeze (10–30 s constant)	0.62 h	1.58 h
F3_DUPL	Repeated packets (burst 2–5)	0.62 h	1.56 h
F4_SPIKE	Sudden jumps ($\pm$30–60 bpm)	0.07 h	0.21 h
F5_DRIFT	Slow ramp ($+0.2$ bpm/s trend)	0.45 h	1.55 h
F6_UNIT	Scale error ($0.5\times$ or $2.0\times$)	0.48 h	1.56 h
F7_OPTICAL	Motion noise (SNR $= +5$ dB, 64 Hz)	Standalone	Standalone

IV. SYSTEM ARCHITECTURE

The experimental harness, telemetry platform, and symptom screening engine were implemented as a modular full-stack application (Fig. 1). The front-end is a React 18 Progressive Web Application (PWA) aligned with Figma interactive specifications, featuring live optical telemetry monitoring, threshold calibration, and an interactive LOSO benchmark dashboard. Per product requirements (PRD), all 13 core functional specifications are operational: Requirement FR-04 enforces automated data-quality exclusion of corrupted windows from downstream inference; Module FR-06 implements a guided symptom assessment with mandatory answer review before screening; and Module FR-07 provides transparent local feature attribution with calibrated prediction uncertainty intervals ($\pm 8\%$). The back-end is powered by FastAPI and SQLite, incorporating the core `vitasense.quality` multi-attribute trust engine, precomputed LOSO caches, and safety-aligned routing.




Fig. 1. Visual Architecture: VitaSense full-stack telemetry and benchmark platform.

V. Methods and Algorithms

A. Evaluated Alerting Arms and Pipeline

The research pipeline processes raw sensor streams into calibrated, causal alerts without future peeking (Fig. 2). We benchmark five causal alerting arms (Fig. 3):

1) **Arm 1 (Fixed Population Cutoff):** Standard clinical threshold $\tau_{\text{pop}} \in \{100, 110, 120, 125, 130\}$ bpm.
2) **Arm 2 (Personal Baseline with Shrinkage):** Dynamic cutoff $\tau_i(t) = \max(\tau_{\text{floor}}, \hat{\mu}_i(t) + k \cdot \hat{\sigma}_i(t))$ regularized toward training-population priors $(\mu_{\text{pop}}, \sigma_{\text{pop}})$ via empirical Bayes shrinkage κ (Algorithm 2).
3) **Arm 3a (Kinematic Motion Gating):** Arm 2 personal baseline paired with a stillness gate: alert triggers if and only if $HR(t) > \tau_i(t)$ and $\text{ENMO}_{10\text{s}}(t) \leq \theta_{\text{rest}}$.
4) **Arm 3b (Multi-Attribute Trust Score Gating):** Arm 2 baseline gated by software trust score: alert triggers if and only if $HR(t) > \tau_i(t)$ and $S_{\text{trust}}(t) \geq \theta_{\text{trust}}$.
5) **Arm 3b-noMotion (Trust Ablation):** Arm 3b with kinematic motion term ablated ($w_{\text{mot}} = 0$).




Fig. 2. Visual Pipeline: Pre-registered fault injection and LOSO benchmarking.




Fig. 3. Visual Decision Flow: Causal alerting logic and dual gating pathways.

B. Shared Standard Alert Emitter

To ensure rigorous cross-arm comparability, all arms feed instantaneous binary triggers into a shared `StandardAlertEmitter` (Algorithm 3). Alerts fire if and only if triggering conditions hold continuously for $W_p = P/\text{step}$ windows ($P \in \{10, 20, 30\}$ s). Emitted alerts enforce a mandatory refractory hold-off of $W_r = R/\text{step}$ windows ($R = 60$ s).

VI. Experimental Setup

A. Leave-One-Subject-Out (LOSO) Cross-Validation

Hyperparameters were tuned strictly within 15 LOSO training folds (Algorithm 4). For each test subject $s \in \{1 \ldots 15\}$, candidate configurations were evaluated across the 14 training

Algorithm 1 Synthetic Tachycardia Injection & Fault Schedule

Require: Pulse rate HR, motion ENMO, seed $S = 42$, regime $\mathcal{R}$
Identify resting blocks $\mathcal{B} = \{[t_s, t_e] \mid \text{ENMO}[t] \leq 0.05\,\text{g}, \text{dur} \geq 80\,\text{s}\}$
for episode $j = 1$ to N_{tachy} (sample without overlap, gap ≥ 60 s) **do**
  Sample target plateau $HR_{\text{peak}} \sim \mathcal{U}(105, 135)$ bpm
  Compute adaptive ramp rates: $r \leftarrow (HR_{\text{peak}} - HR_{\text{base}})/15$ s
  Inject 15 s ramp, plateau (45–120 s), 20 s exponential decay
end for
if $\mathcal{R} \in \{\text{Regime A}, \text{Regime B}\}$ **then**
  Sample non-overlapping fault intervals outside synthetic events
  Inject F1 (gaps), F2 (stale), F3 (dupl), F4 (spikes), F5 (drift), F6 (unit scale)
else if $\mathcal{R} ==$ F7 **then**
  Add optical Gaussian noise modulated by $\|\mathbf{a}(t)\|$ at $+5.0$ dB SNR
end if
return Fault-injected stream HR_{faulted} and episode ground truth

Algorithm 2 Causal Personal Baseline & Trust Score Gating

Require: HR, ENMO, priors $\mu_{\text{pop}}, \sigma_{\text{pop}}$, grid $k, \kappa, \theta_{\text{rest}}, \theta_{\text{trust}}$
for window index $t = 0$ to $N - 1$ **do**
  $c \leftarrow \sum_{\tau=0}^{t-1} \mathbb{I}(\text{ENMO}[\tau] \leq 0.05 \wedge \text{finite}(HR[\tau]))$
  if $c > 1$ **then**
    $\bar{x} \leftarrow \frac{1}{c} \sum_{\tau<t} HR[\tau], \quad s \leftarrow \text{std}_{\tau<t}(HR[\tau])$
  else
    $\bar{x} \leftarrow \mu_{\text{pop}}, \quad s \leftarrow \sigma_{\text{pop}}$
  end if
  $w \leftarrow \frac{c \cdot \Delta t/3600}{c \cdot \Delta t/3600 + \kappa}; \quad \hat{\mu} \leftarrow w\bar{x} + (1-w)\mu_{\text{pop}}; \quad \hat{\sigma} \leftarrow ws + (1-w)\sigma_{\text{pop}}$
  $\tau_i[t] \leftarrow \max(100.0, \hat{\mu} + k \cdot \hat{\sigma})$
  $S_{\text{gap}} \leftarrow 1 - 2 \cdot \text{nan_ratio}; \quad S_{\text{stale}} \leftarrow \mathbb{I}(\text{var} \geq 10^{-6})$
  $S_{\text{phys}} \leftarrow \mathbb{I}(35 \leq HR \leq 220 \wedge |\Delta HR| \leq 40)$
  $S_{\text{mot}} \leftarrow \text{clip}(1.0 - 0.60 \cdot (\text{ENMO} - 0.05)/0.25, 0.2, 1.0)$
  $S_{\text{trust}}[t] \leftarrow \text{clip}(S_{\text{gap}} \cdot S_{\text{stale}} \cdot S_{\text{phys}} \cdot S_{\text{mot}}, 0.0, 1.0)$
  $Cond_{\text{3a}}[t] \leftarrow (HR[t] > \tau_i[t]) \wedge (\text{ENMO}[t] \leq \theta_{\text{rest}})$
  $Cond_{\text{3b}}[t] \leftarrow (HR[t] > \tau_i[t]) \wedge (S_{\text{trust}}[t] \geq \theta_{\text{trust}})$
  $Cond_{\text{3b-noMot}}[t] \leftarrow (HR[t] > \tau_i[t]) \wedge (S_{\text{gap}} \cdot S_{\text{stale}} \cdot S_{\text{phys}} \geq \theta_{\text{trust}})$
end for

Algorithm 3 Standard Alert Emitter with Refractory Hold-off

Require: Raw condition mask C, persistence $W_p = P/\Delta t$, refractory $W_r = R/\Delta t$
$run_len \leftarrow 0, \quad ref_cnt \leftarrow 0, \quad Alert \leftarrow \mathbf{0}$
for $t = 0$ to $N - 1$ **do**
  if $ref_cnt > 0$ **then**
    $ref_cnt \leftarrow ref_cnt - 1$
  end if
  if $C[t]$ is TRUE **then**
    $run_len \leftarrow run_len + 1$
    if $run_len \geq W_p$ **and** $ref_cnt == 0$ **then**
      $Alert[t] \leftarrow$ TRUE; $\quad ref_cnt \leftarrow W_r; \quad run_len \leftarrow 0$
    end if
  else
    $run_len \leftarrow 0$
  end if
end for
return $Alert$

TABLE IV
Evaluated Parameter Grids (Amendment 4)

Arm	Tuned Parameters	Refractory R
Arm 1	$\tau_{\text{pop}} \in \{100, 110, 120, 125, 130\}$ bpm; $P \in \{10, 20, 30\}$ s	60 s
Arm 2	$k \in \{2.0, 2.5, 3.0, 3.5, 4.0\}$; $\kappa \in \{0.5, 1.0, 2.0\}$ h; $P \in \{10, 20, 30\}$ s	60 s
Arm 3a	$\theta_{\text{rest}} \in \{0.03, 0.05, 0.08\}$ g; $P \in \{10, 20, 30\}$ s	60 s
Arm 3b	$\theta_{\text{trust}} \in \{0.60, 0.70, 0.80\}$; $P \in \{10, 20, 30\}$ s	60 s
Arm 3b-noMot	$\theta_{\text{trust}} \in \{0.60, 0.70, 0.80\}$; $P \in \{10, 20, 30\}$ s	60 s

TABLE V
LEAVE-ONE-SUBJECT-OUT COMBINED ALERTING BENCHMARK RESULTS ACROSS CONDITIONS

Condition	Alerting Arm	Recall (%)	FA/hr	True Alerts	False Alerts	Precision (%)	F_1 Score
Condition 0 (No-Fault)	Arm 1 (Fixed)	91.1	3.20	132	100	56.9	0.700
	Arm 2 (Personal)	80.4	5.69	133	178	42.8	0.558
	Arm 3a (+Kinematic)	80.4	1.47	133	46	74.3	0.772
	Arm 3b (+Trust Score)	79.5	1.69	130	53	71.0	0.750
	Arm 3b-noMotion	79.5	2.91	130	91	58.8	0.676
Regime A (Low Duty 10%)	Arm 1 (Fixed)	91.1	3.58	132	112	54.1	0.679
	Arm 2 (Personal)	80.4	6.08	132	190	41.0	0.543
	Arm 3a (+Kinematic)	80.4	1.63	132	51	72.1	0.760
	Arm 3b (+Trust Score)	79.5	1.82	129	57	69.4	0.741
	Arm 3b-noMotion	79.5	3.07	129	96	57.3	0.666
Regime B (Nominal 25-30%)	Arm 1 (Fixed)	91.1	4.51	132	141	48.4	0.632
	Arm 2 (Personal)	75.9	7.16	125	224	35.8	0.487
	Arm 3a (+Kinematic)	75.9	2.30	125	72	63.5	0.691
	Arm 3b (+Trust Score)	75.0	2.37	122	74	62.2	0.680
	Arm 3b-noMotion	75.0	3.61	122	113	51.9	0.614
F7 (High-Motion BVP Noise)	Arm 1 (Fixed)	92.0	3.07	131	96	57.7	0.709
	Arm 2 (Personal)	64.3	4.61	100	144	41.0	0.501
	Arm 3a (+Kinematic)	64.3	1.57	100	49	67.1	0.657
	Arm 3b (+Trust Score)	58.0	1.28	90	40	69.2	0.631
	Arm 3b-noMotion	58.0	2.05	90	64	58.4	0.582

subjects pooled. The pre-specified selection rule chooses the parameter vector achieving $\geq 85\%$ training recall while minimizing false alerts per hour (FA/hr). If no parameter setting reaches 85% recall, the pre-registered fallback rule selects the setting maximizing training recall.

Algorithm 4 LOSO Grid Tuning, Fallback Rule, and Dual Scoring

Require: Training folds $\mathcal{S}_{\text{train}}$, test subject s_{test}, parameter grid Θ
for $\theta \in \Theta$ **do**
  Evaluate pooled training recall $\text{Rec}_{\text{train}}(\theta)$ and $\text{FA/hr}_{\text{train}}(\theta)$
end for
$\mathcal{V} \leftarrow \{\theta \in \Theta \mid \text{Rec}_{\text{train}}(\theta) \geq 0.85\}$
if $\mathcal{V} \neq \emptyset$ **then**
  $\theta^* \leftarrow \arg\min_{\theta \in \mathcal{V}} \text{FA/hr}_{\text{train}}(\theta)$ {Primary selection rule}
else
  $\theta^* \leftarrow \arg\max_{\theta \in \Theta} \text{Rec}_{\text{train}}(\theta)$ {Fallback selection rule}
end if
Run chosen arm with θ^* on s_{test} to yield alert mask A
$\text{TP} \leftarrow \sum_j \mathbb{I}(\exists t \in [t_{s,j}, t_{e,j}] : A[t] == \text{TRUE})$
$\text{Recall} \leftarrow \text{TP}/N_{\text{injected}}$
$\text{Clean FA} \leftarrow \sum_t \mathbb{I}(A[t] == \text{TRUE} \wedge t \notin \mathcal{T}_{\text{injected}} \wedge t \notin \mathcal{T}_{\text{natural}\pm 30\text{s}})$
$\text{FA/hr} \leftarrow \text{Clean FA}/T_{\text{non-event hours}}$
return Test performance (Recall, FA/hr)

B. Evaluation Metrics and Paired Statistics

Scoring evaluates event-level recall (detecting injected episodes within $[t_{\text{start}}, t_{\text{end}}]$) alongside alert-level clean false alerts per hour (FA/hr), precision on non-event periods, and F_1 score. The primary statistical hypothesis test compares paired per-subject differences (Δ = Arm 3b$-$Arm 3a) using the two-sided Wilcoxon signed-rank test ($\alpha = 0.05$), Hodges-Lehmann median difference estimator, and 95% subject-bootstrap confidence intervals (10,000 resamples).




Fig. 4. Visual Timeline: Pre-registered protocol amendments and git provenance.

VII. RESULTS AND DISCUSSION

A. Primary Hypothesis H_1 Outcome: NOT SUPPORTED

Pre-registered hypothesis H_1 postulated that incorporating software trust scores would reduce FA/hr by $\geq 60\%$ over kinematic gating and personalized baselines while preserving $\geq 85\%$ recall.

Empirical finding: H_1 is not supported. In paired per-subject comparisons (Table VII and Fig. 5b), Arm 3b demonstrated no statistically detectable false-alert reduction over Arm 3a across any condition:

- **Condition 0 (Clean):** Mean difference $+0.13$ FA/h (95% CI $[-0.54, +0.89]$, $W = 31.0$, $p = 0.8589$).
- **Regime A (10% faults):** Mean difference $+0.05$ FA/h (95% CI $[-0.66, +0.82]$, $W = 40.0$, $p = 0.7007$).
- **Regime B (25–30% faults):** Mean difference -0.04 FA/h (95% CI $[-0.78, +0.69]$, $W = 52.0$, $p = 0.9750$).
- **F7 (Optical Noise):** Mean difference -0.49 FA/h (95% CI $[-1.47, +0.38]$, $W = 17.0$, $p = 0.2845$).

In every condition, 95% bootstrap confidence intervals spanned zero, and Wilcoxon tests yielded $p \geq 0.28$.


B. Secondary Benchmark Findings

Kinematic motion gating drives alert accuracy: Relative to ungated Arm 2, Arm 3a reduced false alerts per hour by 74.2% in clean data (1.47 vs 5.69 FA/h) and by 67.9% in Regime B (2.30 vs 7.16 FA/h) (Table V).


Recall degradation under optical corruption (F7): Under intense motion-coupled optical noise, Arm 3b recall degraded to 58.0% compared to 64.3% in Arm 3a, indicating that multi-component software abstention can suppress genuine emergencies during noisy episodes.

C. Edge-of-Grid Analysis and Fallback Rule Invocation


Hyperparameter tuning consistently selected parameter boundary values (Table IX). Most notably, **Arm 2 never attained the $\geq 85\%$ recall target on any training fold in any condition** (maximum training recall achieved was $\approx 80.4\%$).




(a) Event Recall vs Clean False Alerts per Hour across Arms and Regimes



(b) Paired Differences (Δ = Arm 3b − Arm 3a) with 95% Bootstrap CIs



(c) Per-Subject False Alerts per Hour in Clean Data (Condition 0)



(d) Arm 1 Threshold Sensitivity across Cutoffs τ_{pop} and Persistence P

Fig. 5. Benchmark results across arms and conditions: (a) Event recall vs clean false alerts per hour; (b) Forest plot of paired differences (Δ = Arm 3b − Arm 3a) with 95% bootstrap CIs; (c) Per-subject FA/hr in Condition 0; (d) Arm 1 sensitivity across τ_{pop} and persistence P.

TABLE VI
STRATIFIED ALERTING METRICS (RECALL % / MEDIAN DELAY (S) / FA PER HOUR)

Condition	Stratum	Arm 1 (Fixed)	Arm 2 (Personal)	Arm 3a (+Kinematic)	Arm 3b (+Trust)	Arm 3b-noMotion
Condition 0 (No-Fault)	Pooled All-15	91.1% / 40s / 3.20	80.4% / 22s / 5.69	80.4% / 22s / 1.47	79.5% / 22s / 1.69	79.5% / 22s / 2.91
	Without S5 (14 subj)	90.7% / 40s / 2.90	79.6% / 22s / 5.26	79.6% / 22s / 1.35	78.7% / 22s / 1.72	78.7% / 22s / 2.93
	S5/S6/S11 (3 subj)	100.0% / 37s / 9.39	94.4% / 20s / 16.13	94.4% / 20s / 5.10	94.4% / 22s / 3.88	94.4% / 22s / 6.74
Regime A (Low Duty 10%)	Pooled All-15	91.1% / 40s / 3.58	80.4% / 22s / 6.08	80.4% / 22s / 1.63	79.5% / 22s / 1.82	79.5% / 22s / 3.07
	Without S5 (14 subj)	90.7% / 40s / 3.27	79.6% / 22s / 5.52	79.6% / 22s / 1.45	78.7% / 24s / 1.75	78.7% / 24s / 3.00
	S5/S6/S11 (3 subj)	100.0% / 37s / 9.80	94.4% / 20s / 17.15	94.4% / 20s / 5.51	94.4% / 22s / 4.08	94.4% / 22s / 7.15
Regime B (Nominal 25-30%)	Pooled All-15	91.1% / 40s / 4.51	75.9% / 22s / 7.16	75.9% / 22s / 2.30	75.0% / 24s / 2.37	75.0% / 24s / 3.61
	Without S5 (14 subj)	90.7% / 40s / 4.14	75.0% / 24s / 6.40	75.0% / 24s / 2.19	74.1% / 24s / 2.29	74.1% / 24s / 3.54
	S5/S6/S11 (3 subj)	100.0% / 37s / 10.21	94.4% / 20s / 18.78	94.4% / 20s / 5.92	94.4% / 22s / 4.29	94.4% / 22s / 7.15
F7 (High-Motion BVP Noise)	Pooled All-15	92.0% / 40s / 3.07	64.3% / 24s / 4.61	64.3% / 24s / 1.57	58.0% / 24s / 1.28	58.0% / 24s / 2.05
	Without S5 (14 subj)	91.7% / 40s / 2.70	63.0% / 25s / 4.01	63.0% / 25s / 1.18	56.5% / 24s / 1.15	56.5% / 24s / 1.92
	S5/S6/S11 (3 subj)	100.0% / 38s / 8.98	88.9% / 21s / 14.90	88.9% / 21s / 5.92	88.9% / 22s / 3.47	88.9% / 22s / 5.51

TABLE VII
PAIRED STATISTICAL COMPARISON: ARM 3B VS ARM 3A

Condition	Mean Δ	95% Boot. CI	Wilcoxon W	p-val
Condition 0 (Clean)	+0.13	[-0.54, +0.89]	31.0	0.8589
Regime A (10%)	+0.05	[-0.66, +0.82]	40.0	0.7007
Regime B (25–30%)	-0.04	[-0.78, +0.69]	52.0	0.9750
F7 (Optical Noise)	-0.49	[-1.47, +0.38]	17.0	0.2845

TABLE VIII
DISAGGREGATED FALSE ALERTS PER FAULT TYPE (REGIMES A & B)

Reg.	Fault	Exp.	Arm 1	Arm 2	Arm 3a	Arm 3b	3b-nM
Reg A	F0_CLEAN	28.1h	3.2 (90)	5.9 (165)	1.5 (43)	1.8 (50)	3.0 (83)
	F1_GAPS	0.9h	0.0 (0)*	0.0 (0)*	0.0 (0)*	0.0 (0)*	0.0 (0)*
	F2_STALE	0.6h	21.1 (13)*	8.1 (5)*	3.2 (2)*	1.6 (1)*	1.6 (1)*
	F3_DUPL	0.6h	3.2 (2)*	8.0 (5)*	3.2 (2)*	0.0 (0)*	4.8 (3)*
	F4_SPIKE	0.1h	0.0 (0)*	14.8 (1)*	0.0 (0)*	0.0 (0)*	14.8 (1)*
	F5_DRIFT	0.5h	0.0 (0)*	2.2 (1)*	0.0 (0)*	0.0 (0)*	0.0 (0)*
	F6_UNIT	0.5h	14.5 (7)*	26.9 (13)*	8.3 (4)*	12.4 (6)*	16.6 (8)*
Reg B	F0_CLEAN	21.7h	3.5 (76)	5.8 (125)	1.6 (35)	1.9 (40)	2.9 (63)
	F1_GAPS	3.1h	0.0 (0)*	0.0 (0)*	0.0 (0)*	0.0 (0)*	0.0 (0)*
	F2_STALE	1.6h	20.2 (32)	10.8 (17)	3.2 (5)*	4.4 (7)*	7.6 (12)
	F3_DUPL	1.6h	3.2 (5)*	9.6 (15)	1.9 (3)*	1.3 (2)*	1.9 (3)*
	F4_SPIKE	0.2h	0.0 (0)*	0.0 (0)*	0.0 (0)*	4.8 (1)*	4.8 (1)*
	F5_DRIFT	1.6h	2.6 (4)*	6.5 (10)	1.9 (3)*	0.7 (1)*	1.9 (3)*
	F6_UNIT	1.6h	15.4 (24)	36.6 (57)	16.7 (26)	14.8 (23)	19.9 (31)

* Insufficient data ($<$ 10 alerts or $<$ 1.0 h). Rate shown as FA/h (count).

Consequently, the fallback selection rule (maximizing training recall) was invoked in 100% (15/15) of folds for Arm 2 across all regimes, selecting $k = 2.0$ and $P = 10\,\text{s}$.

TABLE IX
EDGE-OF-GRID PARAMETER BOUNDARY SELECTION ACROSS 15 LOSO FOLDS

Arm	Parameter Boundary	Cond 0	Reg A	Reg B	F7
Arm 1 (Fixed)	$\tau_{\text{pop}} = 100\,\text{bpm}_{\text{lo}}$	100%	100%	100%	100%
	$P = 30\,\text{s}_{\text{hi}}$	100%	100%	100%	100%
Arm 2 (Personal)	$k = 2.0_{\text{lo}}$	100%	100%	100%	100%
	$\kappa = 2.0\,\text{h}_{\text{hi}}$	93.3%	93.3%	100%	100%
	$P = 10\,\text{s}_{\text{lo}}$	100%	100%	100%	100%
Arm 3a (+Kin.)	$\theta_{\text{rest}} = 0.08\,\text{g}_{\text{hi}}$	93.3%	93.3%	93.3%	100%
	$\theta_{\text{rest}} = 0.03\,\text{g}_{\text{lo}}$	6.7%	6.7%	6.7%	0.0%
Arm 3b (+Trust)	$\theta_{\text{trust}} = 0.60_{\text{lo}}$	93.3%	93.3%	93.3%	100%
	$\theta_{\text{trust}} = 0.80_{\text{hi}}$	6.7%	6.7%	6.7%	0.0%

VIII. LIMITATIONS

This benchmark is subject to several methodological constraints (Table X): (1) Events were synthetically injected due to natural sparsity; (2) The cohort is limited to 15 healthy

participants; (3) Algorithmic errors in front-end PPG tracking couple into downstream logic; (4) Post-result protocol amendments (Table XI) document adjustments made after preliminary inspection; (5) Arm 4 (conformal prediction with selective abstention) was never built or evaluated; (6) PhysioNet Wearable 1.0.1 and WESAD [8] datasets were not used for benchmark evaluation, maintaining focus strictly on the ratified PPG-DaLiA protocol.

TABLE X
THREATS TO VALIDITY AND SCIENTIFIC LIMITATIONS

Threat Domain	Empirical Limitation and Mitigation
Synthetic Events	True resting tachycardia absent in cohort; 112 episodes injected into derived stream (105–135 bpm).
Cohort Size	$N = 15$ healthy young adults limits power for small paired differences (reported with 95% bootstrap CIs).
Noise Coupling	Estimator tracking errors and motion artifacts interact directly with downstream thresholding logic.
Grid Boundaries	Tuning selected boundaries ($\tau_{\text{pop}} = 100$, $k = 2.0$, $P = 10\,\text{s} / 30\,\text{s}$), reflecting trade-offs.
Scope & Datasets	PhysioNet 1.0.1 and WESAD unanalyzed; Arm 4 conformal abstention deferred. Academic benchmark only.

TABLE XI
PROTOCOL EVOLUTION AND AMENDMENTS (FROM GIT LOG)

Milestone [Commit]	Ratified Scope and Description
Protocol v1.0.0 [41eff1e]	Frozen protocol: H_1–H_3, faults F1–F7, LOSO cross-validation.
Amendment 1 [3b1f2f6]	Resting alert ground truth; non-penalization of motion; corrected accel scale [3a5ccc2].
Amendment 2 [696fd00]	Labelled synthetic tachycardia injection; adaptive ramps (r_{lo}, r_{hi}); S5 outlier handling rule.
Amendment 3 [a0d682a]	Restated H_1 as estimation; 15-subject LOSO; Arm 3b-noMotion ablation; $\geq$ 60 s gap [0bf4ca4].
Amendment 4 [a9ce657]	Standard emitter: $P \in \{10, 20, 30\}$ s, $R = 60$ s; Arm 2 fallback rule invoked in 100% of folds [f92ca8d].

IX. FUTURE SCOPE

Future efforts will extend this benchmark along five technical avenues:

1) **Arm 4 Conformal Abstention:** Building and integrating inductive conformal prediction to yield distribution-free validity sets with user-calibrated error bounds [4].
2) **External Dataset Generalization:** Evaluating on external public wearable corpora, including the PhysioNet Wearable Device Dataset [7] and WESAD [8].
3) **Clinical Patient Cohorts:** Validating on holter-monitored inpatient recordings with spontaneous atrial fibrillation and supraventricular tachycardia.
4) **On-Device Hardware Integration:** Deploying causal baseline and kinematic gating engines onto low-power embedded microcontrollers.
5) **Prospective Clinical Validation:** Translating the operational FR-06 guided symptom questionnaire and FR-07 transparent explanation modules into multicenter clinical cohorts.

X. CONCLUSION

In this pre-registered fault-injection benchmark, kinematic motion gating accounted for virtually all false-alert suppression in wearable heart-rate monitoring, while multi-component software trust scoring offered no statistically detectable benefit. Fixed thresholding with persistence provided a remarkably resilient baseline. All findings are derived from synthetic tachycardia episodes in healthy individuals and carry no clinical standing.

REGULATORY AND SAFETY STATEMENT

Not a Medical Device: The VitaSense AI software, alerting algorithms, and benchmark suite are academic research artifacts. They are neither cleared nor approved by any regulatory authority and must not be used for medical diagnosis, clinical treatment, or life-critical monitoring.

DATA AND CODE AVAILABILITY

The PPG-DaLiA dataset is openly accessible via the UCI Machine Learning Repository [1], [5]. The PhysioNet Wearable Device Dataset is available under ODC-By v1.0 [7]. Benchmark code, execution scripts, and numerical provenance logs are provided in the project repository.

REFERENCES

[1] A. Reiss, I. Indlekofer, P. Schmidt, and K. Van Laerhoven, "Deep PPG: Large-scale heart rate estimation with convolutional neural networks," *Sensors*, vol. 19, no. 14, p. 3079, 2019.

[2] M. Elgendi, "Optimal signal quality index for photoplethysmogram signals," *Bioengineering*, vol. 3, no. 4, p. 21, 2016.

[3] P. H. Charlton *et al.*, "The 2023 wearable photoplethysmography roadmap," *Journal of Physics: Photonics*, vol. 5, no. 4, p. 042501, 2023.

[4] J. Vazquez and J. C. Facelli, "Conformal prediction in clinical medical sciences," *Journal of Healthcare Informatics Research*, vol. 6, no. 3, pp. 257–285, 2022.

[5] A. Reiss, I. Indlekofer, P. Schmidt, and K. Van Laerhoven, "PPG-DaLiA," UCI Machine Learning Repository, 2019, DOI: 10.24432/C53890.

[6] R. Detrano, A. Jánosi, W. Steinbrunn, K. Pfisterer, and J. Schmid, "Heart Disease," UCI Machine Learning Repository, 1988, DOI: 10.24432/C52P4X.

[7] A. Kos, Y. Tomažič, and A. Umek, "Wearable Device Dataset from Induced Stress and Structured Exercise Sessions," PhysioNet, Version 1.0.1, 2022, DOI: 10.13026/8fca-2v93.

[8] P. Schmidt, A. Reiss, R. Duerichen, C. Marberger, and K. Van Laerhoven, "Introducing WESAD, a multimodal dataset for wearable stress and affect detection," in *Proc. 20th ACM Int. Conf. Multimodal Interaction (ICMI)*, 2018, pp. 400–408.

[9] Z. Zhang, Z. Pi, and B. Liu, "TROIKA: A general framework for heart rate monitoring using wrist-type photoplethysmographic signals during intensive physical exercise," *IEEE Trans. Biomed. Eng.*, vol. 62, no. 2, pp. 522–531, 2015.

[10] M. Cvach, "Monitor alarm fatigue: An integrative review," *Biomed. Instrum. Technol.*, vol. 46, no. 4, pp. 268–277, 2012.