Mettler/Paar DMA45 Densitometer Calibration and Density Measurements

Teng-Jui Lin

Department of Chemical Engineering, University of Washington

Surface and Colloid Science

Densitometer measures liquid densities

• Glass oscillator measures vibration period (direct) that relates to density (indirect)



Densitometer calibration constants require vibration period and reference density

Calibration constants

$$egin{aligned} A &= rac{ au_{ ext{H}_2 ext{O}}^2 - au_{ ext{air}}^2}{
ho_{ ext{H}_2 ext{O}} -
ho_{ ext{air}}} \ B &= au_{ ext{air}}^2 - A
ho_{ ext{air}} \end{aligned}$$

Water density (ITS-90 correlation)

$$ho_{
m H_2O}(T) = A + BT + CT^2 + DT^3$$

Air density

$$ho_{
m air}(T) = rac{P_{
m air}}{R_{
m air}T} + rac{P_{
m H_2O}}{R_{
m H_2O}T} \ R_i = rac{R}{\mathcal{M}_i}$$

Water partial pressure (Relative humidity)

$$P_{\mathrm{H_2O}} = P_{\mathrm{H_2O}}^* \mathrm{RH}$$


Water saturation pressure (Antoine's equation)


$$P^*_{
m H_2O}(T) = \exp\left(A - rac{B}{C+T}
ight)$$

Dry air partial pressure

$$P_{
m air} = P - P_{
m H_2O}$$

Top panel allows adjustment of calibration constants





1. Calibration

1. Use a screwdriver to switch densitometer to period mode (T)

Measure period of air for calibration constants

- 1. Calibration (cont.)
 - 2. Turn on the densitometer
 - 3. Turn on the light
 - 4. Record the period of air when value stabilizes



Measure period of water for calibration constants

- 1. Calibration (cont.)
 - 5. Use a plastic syringe to draw >1 mL deionized water
 - 6. Remove all bubbles from the syringe
 - 7. Inject deionized water into the U-tube (avoid air!)
 - 8. Record the period of water when value stabilizes or after 2 min



Calculate calibration constants using period of water and air

- 1. Calibration (cont.)
 - 9. Use the provided spreadsheet to calculate the calibration constants
 - 10. Record the calibration constants

Input the calibration constants into the densitometer and verify calibration

- 1. Calibration (cont.)
 - 11. Use a screwdriver to change the calibration constants on the densitometer
 - 12. Calibration is now complete!
- 2. Verify the calibration
 - 1. Use a screwdriver to switch densitometer to density mode (ρ)
 - 2. Read the density measurement of water and compare to literature
 - 3. Draw out the water
 - 4. Turn on the pump and put the airline in the U-tube until the U-tube is dry
 - 5. Read the density measurement of air and compare to literature

Measure density of unknown sample

- 3. Measure density for samples of interest
 - 1. Follow the same procedure for getting samples into densitometer as calibration
 - 2. Draw samples, avoid air bubbles, and inject into U-tube
 - 3. Record density measurement when value stabilizes
 - 4. Draw out samples and rinse the U-tube with acetone
 - 5. Turn on the pump and put the airline in the U-tube until the U-tube is dry
 - 6. Note: rinse the syringe with deionized water between samples

Shutting down the densitometer

- 4. Densitometer shutdown
 - 1. When the U-tube is dry, turn off the pump
 - 2. Turn off the light
 - 3. Turn off the densitometer
 - 4. Place back the calibration panel lid, screwdriver, and syringe

References

- Jones, F. E., & Harris, G. L. (1992). ITS-90 Density of Water Formulation for Volumetric Standards Calibration. Journal of Research of the National Institute of Standards and Technology, 97(3), 335—340. https://doi.org/10.6028/jres.097.013
- Koretsky, M. D. (2012). Engineering and Chemical Thermodynamics (2nd edition). Wiley.
- McVay, C. W. (1989). Determination of Density with the Paar Digital Density Meter. https://www.nrc.gov/docs/ML0329/ML032960458.pdf
- Richard Shelquist. (2023). Equations—Air Density and Density Altitude. https://wahiduddin.net/calc/density_altitude.htm