ECG Arrhythmia Beat Classification Using a Convolutional Bidirectional LSTM Deployed as a Tool for a Retrieval-Augmented LLM Agent

Student Name : PRIYATOSH SATPATI
Student Registration Number : 25113170096

Approved Guide Name : MR. G. SWAMINATHAN

Name of the Company with address : [COMPANY NAME AND ADDRESS IN CAPITAL LETTERS]

Abstract / Problem Statement

Cardiac arrhythmias are usually identified from long ambulatory electrocardiogram (ECG) recordings, where manual beat-by-beat review is slow and error-prone. Many published deep-learning heartbeat classifiers report near-perfect accuracy, but they often split beats randomly, so beats from the same patient appear in both training and test data and the reported performance does not carry over to new patients. This project builds a heartbeat classifier that is evaluated on unseen patients and makes it usable through a conversational assistant.

Single-lead recordings from the MIT-BIH Arrhythmia Database are band-pass filtered, segmented around each R-peak and grouped into the AAMI classes normal (N), supraventricular ectopic (S) and ventricular ectopic (V). A compact convolutional bidirectional long short-term memory (Conv-BiLSTM) network of about 50,000 parameters combines beat morphology with rhythm features expressed relative to each patient's own heart rate. Model selection uses patients held out from training, and final testing uses 22 patients never seen during development, reported with patient-level bootstrap confidence intervals and compared against baselines.

On the unseen patients the model reaches 94.5% accuracy and a macro-F1 score of 0.65, against 0.53 for a rhythm-only baseline. It detects 80% of ventricular beats with 85% precision, and its probabilities are well calibrated. Supraventricular beats remain difficult, and this limitation is reported explicitly.

The model is deployed in a FastAPI web service with an ECG signal-quality check and is exposed as a callable tool to a large language model agent. The agent combines the classifier's output with retrieval-augmented generation over a persistent Chroma vector database of reference documents, so its answers about an uploaded ECG rest on both the measurement and cited sources. The system is verified by 95 automated tests covering preprocessing, APIs, integration, accuracy and performance.

Signature of the Student
(PRIYATOSH SATPATI)

Signature of the Reporting Manager
(MR. G. SWAMINATHAN)