



Medical Device Details

Sponsor: Tere Health Limited
41 Adams Lane
Springlands
Blenheim
New Zealand

WAND Reference: 260729-WAND-786DQ9

Sponsors Own Reference: TERE-VITALS-001

GMDN: Multiple physiological parameter spot-check analysis system, clinical [57960]

Class: IIa

Intended Purpose: Tere Vitals is a Software as a Medical Device (SaMD) that uses the front-facing camera of a consumer smartphone or laptop to non-contact estimate and display heart rate, peripheral oxygen saturation (SpO₂), and respiratory rate using remote photoplethysmography (rPPG). It is intended as an adjunct to remote clinical triage during telehealth consultations in adults (18 years and over), providing indicative physiological measurements for a registered clinician to consider alongside clinical history and other assessment findings. It is not intended as a diagnostic device, not for continuous monitoring, and not for use in paediatric or neonatal populations, where a clinical decision requires laboratory-grade accuracy, or where physiological instability requires urgent in-person assessment.

Manufacturer: Tere Health Limited 41 Adams Lane
Springlands
Blenheim
New Zealand

ARTG ID:

- › The device is supplied unsterilised.
- › The device is not intended to be invasive.
- › The device is not intended for single use.
- › The device is an active device.
- › The device does not contain material or ingredients of microbial origin.
- › The device does not contain material or ingredients of recombinant origin.
- › The device does not contain material or ingredients manufactured or formulated using a genetically modified organism.
- › The device does not contain material or ingredients of human origin.
- › The device does not contain human blood or its components.
- › The device consists of: Single product
- › The device does not contain material or ingredients of animal origin.
- › The device is not medicated.
- › The device is not formulated.
- › The product does not contain a medicine that has consent for marketing in New Zealand.
- › The product does not contain a medical device which incorporates a medicine as an integral part and that has an action ancillary to the device.

Status: Active