

Trainer Debrief — Lab 7

Stakeholder Analysis, Resistance Management and Change Leadership

Case study: Meridian Medical Devices — seal weld burst pressure

Objective: Analyse stakeholders and plan change leadership for your capstone (A1, K1).

***TRAINER ONLY.** Do not hand this out before the lab is complete.*

Expected analysis

lab-07-stakeholder-analysis.xlsx

Three stakeholders are both HIGH influence and OPPOSED — the Suzhou Plant Manager, the Manufacturing Engineering Lead and the Suzhou night shift supervisor. They are exactly the people whose process the data has just criticised, so their resistance is predictable and usually political rather than technical: the Lab 15 and Lab 19 findings say their cell is the problem. Treating that as a communication task ('they need to understand the data') is the classic Black Belt failure. The action that moves them is co-ownership — put the Engineering Lead on the DOE team so the fixture fix is HIS, and let the Suzhou manager present the improvement to the steering committee. The Finance Controller must move from Neutral to Supportive before Lab 30, or the benefit will not be signed.

Every dataset is generated from a fixed seed, so these figures are reproducible — re-run the analysis on the shipped workbook and you will get these numbers.

What good looks like

Every stakeholder is plotted on the influence/support grid, every high-influence resistor has a named action matched to a diagnosed resistance type, and your ADKAR plan covers all five elements.

Where this sits in the capstone

Black Belt projects fail on people far more often than on statistics. Map influence against support, diagnose resistance by type, and build an ADKAR-based change plan. This is Black-Belt-only material — the Green Belt course stops at RACI. CAPSTONE BUILDING BLOCK — the output of this lab goes directly into your capstone project pack, which you present to the steering committee on Day 5.

The running process — figures that must reconcile

Checkpoint	Value	Established in
Baseline defect rate	2.9%	Lab 3
Measurement system	FAILS Gage R&R (~40% study variation)	Lab 11
Baseline capability	Pp 0.62 / Ppk 0.53	Lab 14
Proven Xs	clamp pressure, dwell time, electrode tip age	Labs 18-21
DOE optimum	clamp 47.4 bar, dwell 2.76 s	Labs 23-25
Improved capability	Ppk 1.40	Lab 27

Annual hard benefit	SGD 487,000 vs SGD 812,000 COPQ	Labs 1, 30
---------------------	------------------------------------	------------

If a learner's number disagrees with the column above, they have either used their own workplace data (fine — say so) or made an error (find it).

Certified Lean Six Sigma Black Belt (CLSSBB) Training · TGS-2024051900 · Version v4 · © 2026 Tertiary Infotech Academy Pte Ltd · TRAINER ONLY