

Dr. James Scott Head, Division of Occupational & Environmental Health Director, Gage Laboratories e-m: aerosol.testing.dlsph@utoronto.ca

JOB NUMBER: 3653-001

FILTRATION EFFICIENCY REPORT SUMMARY

Attn: Dr. William Ng, UHN, Toronto

From: Dr. James Scott, Division of Occupational & Environmental Health, DLSPH, U of Toronto

Method: MODIFIED NIOSH 42 CFR Part 84 (30 L/min)

No. of tests: 2

Report date: 2020-06-19

RESULTS SUMMARY

SAMPLE #	DESCRIPTION	*FE _{mass} (%)
1 [†]	Air Guard Clear – attached to face piece	99.7
2	Hydro Guard Mini – filter only	99.9

NOTES: *Filtration efficiency calculated as a function of particle mass

DESCRIPTION OF PROCEDURE

Masks were subjected to a modified NIOSH 42 CFR Part 84 test regimen. In brief, filtration efficiency was measured under conditions of unidirectional airflow at 30 L/min through masks fitted to a test mannequin and exposed in an atmosphere of salt aerosol of median diameter ~80 nm. Particles were measured by size category using a TSI SMPS model 3034 and expressed as counts or mass (standard protocol assesses filtration efficiency by mass).

DISCLAIMER

The Governing Council of the University of Toronto and the author(s) of this Report expressly disclaim any warranty, expressed or implied, that this Report or any information in connection therewith, including but not limited to warranties on design, performance, and materials described herein, meet the requirements of any laws, rules, standards, or governmental specifications; and use or reliance on this Report for any purpose is entirely the responsibility of the user of this Report. This Report is not to be reproduced without permission and is for use only of its intended recipients who commissioned it



[†]Pressure drop: -21.3 mm H₂O