

***** Flatness Measurement *****

////NIIGATASEIKI////

Saved Date : ----/--/-- --:--:--

Work Name : Kiem tra ban 1000x1500

Work Number : 002

Operator : Tony

Comments : Test2

Total crosswise measurement length = 1000[mm] (200[mm] * 5)

Total lengthwise measurement length = 1500[mm] (300[mm] * 5)

Maximum Offset = 0.90[um]

***** Measured Value [mm/M] *****

	[01]	[02]	[03]	[04]	[05]
X[01]	-0.003	0.000	-0.001	0.000	0.000
X[02]	-0.002	0.000	0.000	0.000	0.001
X[03]	-0.001	0.000	-0.001	-0.001	0.001
X[04]	0.001	0.001	0.001	0.005	0.001
X[05]	0.003	0.004	0.005	0.004	0.006
X[06]	0.005	0.004	0.004	0.004	0.005
Y[01]	0.005	0.006	0.007	0.007	0.006
Y[02]	0.005	0.006	0.006	0.006	0.006

***** Result (3 Corner Zero) [um] (*:MAX @:MIN) *****

	[00]	[01]	[02]	[03]	[04]	[05]
Y[00]	0.00	-0.44	-0.28	-0.32	-0.16	0.00
Y[01]	-0.36	-0.72	-0.68	-0.64	-0.60	-0.36
Y[02]	-0.42	-0.54	-0.46	-0.58	-0.70	-0.42
Y[03]	-0.18	-0.40	-0.62	@-0.84	-0.26	-0.48
Y[04]	*0.06	-0.34	-0.54	-0.54	-0.74	-0.54
Y[05]	0.00	0.00	-0.20	-0.40	-0.60	-0.60