

Hearing Aid Consultation

Consultation Date: 04-06-2026

Doctor: James Hargreaves

Audiologist: Daisy Song

Chart No.: P1-007

Patient Name: Sophia Turner

Ear Side: Right

Demo Model: -

Record Type: Repair Completed

Sophia brought in her right Signia Styletto X as it had stopped charging. Had a look at the charging port and found some moisture damage and a bit of debris built up inside. Cleaned it out and tested it and it is charging normally again. Talked to her about looking after the aids, particularly using a drying kit overnight and keeping them away from moisture. Gave the device back to her in working order. Reminded her of her next scheduled appointment.

Christchurch ENT Clinic