



PESHAWAR INSTITUTE OF CARDIOLOGY

MEDICAL TEACHING INSTITUTION, PIC - MTI

5-A, Sector B-3, Phase-V, Hayatabad, Peshawar, KP, Pakistan

☎ 0092 919219641-5 ✉ info@pic.edu.pk 🌐 www.pic.edu.pk

Date: 24-06-2021

ADDENDUM PIC-033

Following Point are revised for PIC-033: After Pre-Bid

(High End Echocardiography Machine)

Description	Previous Specs	After Addendum
Processing channel	System must be offered with a minimum of 550,000 digital processed channels or better	System must be offered with a minimum of 650,000 digital processed channels or better
Frame Rate	Frame rate with 2000 or more in 2D Mode	Frame rate with 2500 or more in 2D Mode
Mode	3d wall motion tracking	To Be removed


HOD Cardiology
PIC-MTI


AP Paeds Cardiology
PIC-MTI


Bio-Medical Engr
PIC-MTI


Hospital Director
PIC-MTI