



# PESHAWAR INSTITUTE OF CARDIOLOGY

MEDICAL TEACHING INSTITUTION, PIC - MTI

5-A, Sector B-3, Phase-V, Hayatabad, Peshawar, KP, Pakistan

☎0092 919219641-5 ✉info@pic.edu.pk 🌐www.pic.edu.pk

Date: 22-06-2021

## ADDENDUM PIC-032

Following Point are revised for PIC-032: (MRI) After Pre-Bid

Description after addendum	Previous Specs	After Addendum
Magnet Length	150 cm or less	165 cm or Less
2D Minimum Slice Thickness	0.1 mm or less	0.5 mm or less
Hard Disk Capacity for Raw Data	1,000,000 images with a matrix size of 256 x 256 or more	As Per OEM recommendation

HOD Radiology

PIC-MTI

Bio-Medical Eng

PIC-MTI

Hospital Director  
Peshawar Institute  
of Cardiology-MTI



## **PESHAWAR INSTITUTE OF CARDIOLOGY**

**MEDICAL TEACHING INSTITUTION, PIC - MTI**

**5-A, Sector B-3, Phase-V, Hayatabad, Peshawar, KP, Pakistan**

### **CORRIGENDUM (Ref: PIC 32)**

This is with reference to the Advertisement (Invitation for Bids) Ref: PIC 32, published in dawn & Mashriq dated 18.05.21 for the procurement of Magnetic Resonance Imaging (MRI). The Pre-bid meeting with the interested bidders will be now held on 31 May 2021 at 10:00 am in the office of Manager Material Management Office . All the other terms and conditions remain the same.

**HOSPITAL DIRECTOR  
PESHAWAR INSTITUTE OF CARDIOLOGY-MIT**



## **PESHAWAR INSTITUTE OF CARDIOLOGY**

**MEDICAL TEACHING INSTITUTION-  
KHYBER PAKHTUNKHWA, PAKISTAN**

### **CORRIGENDUM-II PIC-032**

This is with reference to the Advertisement (Invitation for Bids) Ref: PIC 32, published in Dawn & Mashriq dated 18.05.21 for the procurement of Magnetic Resonance Imaging (MRI).

The Tender opening and last submission date have been revised, now Tender opening and last submission date will be 29-06-2021 instead of 17-06-2021

Note: All other terms and condition will remain the same as per previous advertisement.

**Hospital Director  
Peshawar Institute of Cardiology-MTI**