

बिड दस्तावेज़ / Bid Document

बिड विवरण / Bid Details	
बिड बंद होने की तारीख/समय /Bid End Date/Time	30-07-2026 13:00:00
बिड खुलने की तारीख/समय /Bid Opening Date/Time	30-07-2026 13:30:00
बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date)	180 (Days)
मंत्रालय/राज्य का नाम/Ministry/State Name	Ministry Of Health And Family Welfare
विभाग का नाम/Department Name	Department Of Health And Family Welfare
संगठन का नाम/Organisation Name	Jawaharlal Institute Of Postgraduate Medical Education (Jipmer)
कार्यालय का नाम/Office Name	Jawaharlal Institute Of Postgraduate Medical Education
शिकायत निवारण के संपर्क विवरण/ Contact details of Grievance redressal	HOD Email id :jd1176@jipmer.edu.in Buyer Email id: dilipkumar.m.jir@nic.in
कुल मात्रा/Total Quantity	1
वस्तु श्रेणी /Item Category	Ultrasound Machine (V2) (Q2)
बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years)	25 Lakh (s)
मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years)	200 Lakh (s)
उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service	3 Year (s)
एमएसएमई के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है/MSE Relaxation for Years of Experience and Turnover	No
स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Relaxation for Years of Experience and Turnover	No
विक्रेता से मांगे गए दस्तावेज़/Document required from seller	Experience Criteria,Past Performance,Bidder Turnover in ATC),OEM Authorization Certificate,OEM Annual (Requested in ATC),Compliance of BoQ specific documents *In case any bidder is seeking exemption from Experience Criteria, the supporting documents to prove his eligibility be uploaded for evaluation by the buyer
क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid?	Yes (Documents submitted as part of a clarification the tender/bid process will also be displayed to others after log in)

बिड विवरण/Bid Details	
बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension	3
दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended	7
ऑटो एक्सटेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count	1
विगत प्रदर्शन / Past Performance	30 %
बिड से रिवर्स नीलामी सक्रिय किया/Bid to RA enabled	No
व्यापक रखरखाव शुल्क आवश्यक / Comprehensive Maintenance Charges Required	Yes
बिड का प्रकार/ Type of Bid	Two Packet Bid
तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय / Time allowed for Technical Clarifications during technical evaluation	2 Days
निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/ Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM)	No
अनुमानित बिड मूल्य / Estimated Bid Value	5000000
मूल्यांकन पद्धति/ Evaluation Method	Total value wise evaluation
मध्यस्थता खंड/ Arbitration Clause	No
सुलह खंड/ Mediation Clause	No

ईएमडी विवरण/EMD Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईएमडी राशि/EMD Amount	250000

ईपीबीजी विवरण /ePBG Detail

एडवाइजरी बैंक/Advisory Bank	State Bank of India
ईपीबीजी प्रतिशत (%) /ePBG Percentage(%)	5.00
ईपीबीजी की आवश्यक अवधि (माह) /Duration of ePBG required (Months).	63

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित कैटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है।/EMD EXEMPTION: The exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded Policy.

(b).ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in wherever it is applicable.

(c).ईएमडी और संपादन जमानत राशि लाभार्थी के पक्ष में होनी चाहिए। / Earnest Money Deposit (EMD) shall also be accepted by the surety bond.

लाभार्थी /Beneficiary :

Director , JIPMER, Puducherry

Jawaharlal Institute of Postgraduate Medical Education and Research, Department of Health and Family Welfare, Jawahar Postgraduate Medical Education and Research (JIPMER), Ministry of Health and Family Welfare (Vir Singh Negi)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता / MII Purchase Preference

एमआईआई खरीद वरीयता / MII Purchase Preference	Yes
मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में है / Purchase Preference to MII sellers available upto price within $L1+X\%$	20
मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference	50
सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time)	Yes, in compliance with the MII ORDER : DPIIT Order applicable time to time)

एमएसई खरीद वरीयता/MSE Purchase Preference

एमएसई खरीद वरीयता/MSE Purchase Preference	Yes
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं/सेवा प्रदाता को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ / Purchase Preference to MSE OEMs/ Service Provider available upto price within $L1+X\%$	15
सूक्ष्म और लघु उद्यम मूल उपकरण निर्माता/सेवा प्रदाता को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Percentage of Bid quantity/amount for MSE OEMs/ Service Provider Purchase preference	25

1. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheet or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

2. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Government Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criteria.

3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

4. Preference to Make In India products (For bids < 200 Crore): Preference shall be given to Class 1 local supplier as per Public Procurement (Preference to Make in India) Order, 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier shall be as indicated in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM or the local content and the details of locations at which the local value addition is made along with their bid, failing which the preference shall not be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be audited by a statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to bid. However, eligible micro and small enterprises will be allowed to participate. The buyers are advised to refer the OM No. 14/2021-PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017 and its subsequent Order/Notifications issued by concerned Ministry. Benefits of MSE will be allowed only if seller/service provider is validated on-line in GeM profile as approved by Buyer after evaluation of documents submitted.

5. Purchase preference to Micro and Small Enterprises (MSEs): Purchase preference will be given to MSEs having valid Certificate and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the Service provider or manufacturer / OEM of the offered product on GeM. In respect of bid for Services, the bidder must be the Service provider. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers of products manufactured by some other OEM are not eligible for any purchase preference. Relevant documentary evidence in the form of Certificate of Registration as MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference as per relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (Selected by Buyer) of total quantity. The buyers are advised to refer the OM No. 14/2021-PPD dated 18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

6. Estimated Bid Value indicated above is being declared solely for the purpose of guidance on EMD amount and for the purpose of shortlisting. It has no relevance or bearing on the bidders and is also not going to have any impact on bid participation. Also this is not going to be used as a criterion for the reasonableness of quoted prices which would be determined by the buyer based on its own assessment of reasonable and competitive prices received in Bid / RA process.

7. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products to any Government Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date to any Central / State Government Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of having supplied each of the Financial year. In case of bunch bids, the category related to primary product having highest value should meet this criterion.

Ultrasound Machine (V2) (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम कैटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement
Types of Probes & Features	Convex Probe	2.5 to 5 MHz (± 2)
	Microconvex probe	Not Provided
	Linear array probe	3 to 12 MHz (± 2)
	TVS Probe	4 to 9 MHz (± 2)
	Cardiac Probe	1 to 5 MHz (± 2)
	Hockey Stick Probe	Not Provided
	Volume Convex Probe	2.5 to 5 MHz (± 2), Not Provided
	Volume Linear Probe	3 to 12 MHz (± 2), 6 to 15 MHz (± 2), 6 to 18 MHz (± 2), Not Provided
	Volume TVS Probe	Not Provided
	Endorectal Probe	Not Provided
	E-Breast Elastography technique	Yes
	E-Thyroid Elastography technique	Yes
	Liver Elastography technique	Yes
	Gel Warmer	Yes, No
	Dedicated 3-D/4-D	Yes, No
Scan Modes	Contrast harmonic imaging with angio mode	Yes
	Provision of Harmonic imaging in Linear Probes	Yes
	Color Doppler imaging (CDI)	Yes
	Color Doppler 3-D/4-D option	Yes, No
	Power Doppler imaging (PDI)	Yes
	Power Doppler 3-D/4-D option	Yes, No
	Continuous wave	Yes
	Triplex mode	Yes
	Tissue Doppler imaging	Yes
	Contrast-enhanced ultrasound (CEUS) option	Yes
	Tissue synchronization, tissue velocity	Yes, No
	Provision for higher sensitivity for low frequency doppler in all probes	Yes
	Number of channels in color doppler system	4000000 or more (if channels are being designed)
	Live 3 D imaging in color doppler system	Yes, No

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Required
	Colour Doppler system with all application packages ,Quad loop for serial studies with high frame rate reviews, harmonic imaging capability in all modes and integrated stress echo package digital storage and retrieval	Yes
	Provision of Tissue Colorization (B-Colour) for improved contrast resolution	Yes
	Cine Loop Features	Yes
	Frame grabber facility for post analysis	Yes
	Number of transducer/probe ports	4 or More
	System should be capable of generating real time live 3- D images	Yes, No
	Contrast agent imaging	Yes, No
Image Display and Processing	DICOM 3 Compliant	Yes
	Automated B-mode (2-D) image	Yes
	Automated CDI(Color Doppler Imaging) image optimization	Yes
	Automated PW Doppler image optimization	Yes
	Touch Screen	Yes
	Monitor display size (inch)	21 or more Or higher
	Split screen	Twin View and Live Quad Screen
	Single/dual monitors	Dual (image display and touch panel)
	Max number frames	>=1000
	Digital storage hard drive, GB	>=500
	DVD/CD Writer/ USB	Yes
	Type of processor (in PC)	i7 or latest
	RAM capacity of PC in GB (hint: In case PC not provided put zero)	? 4
	Processor speed in GHz (PC)	? 3
	Frame Grabber incorporated(PC)	Yes
	Connectivity between USG and PC	Yes
	PC with all software inclusive interfaced with USG machine	Yes, No
	Laser Color printer provided	Yes, No
	Inkjet/Thermal printer provided	Yes

विवरण/Specification	विशिष्टि का नाम /Specification Name	बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement
	CD/DVD/USB Produced should be playable on any system	Yes, No
	Glossy Colour Print Paper supplied	500 sheets, 1000 sheets, NA
	Dicom Interfacability Dicom to PACS	Yes
	Interfacability Dicom to RIS	Yes
Power Requirements	Resettable over current breaker shall be fitted for protection	Yes
	Power backup	UPS
	Back up time in minutes (minimum)	30.0 (minute)
	Suitable Servo Controlled Stabilizer / CVT should be provided	Yes, No
Warranty and Maintenance	Warranty in Years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue)	5 Or higher (year)

व्यापक रखरखाव / Comprehensive Maintenance	
Comprehensive Maintenance (Minimum Percentage)	3 %
Comprehensive Maintenance (Maximum Percentage)	25 %
Warranty of required product	5 Year
Comprehensive Maintenance Duration (Post Warranty)	5 Year

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

Additional Specification Parameters - Ultrasound Machine (V2) (1 pieces)

Specification Parameter Name	Bid Requirement (Allowed Values)
Additional terms	All the specifications in Buyer added ATC must be complied. Physical delivery of machine at WCH block, JIPMER Puducherry is mandatory
Probes	Single Crystal Convex Abdominal probe with frequency range from 1 to 10 MHZ approx. with more than 180 Degree Angle Single Crystal Linear and Breast Imaging 2-14 Mhz Single Crystal Adult Cardiac Probe with 1-10 MHz Echo Studies
Warranty and CAMC	5 year comprehensive warranty on the whole machine and all the accessories 5 year CAMC on USG machine. CAMC should cover replacement of at least any manufacturing defect, poor visualization of image quality or non-functioning of any part

* Bidders offering must also comply with the additional specification parameters mentioned above.

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

क्र.सं./S.No.	परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer	पता/Address	मात्रा /Quantity	डिलीवरी
1	Anand Raju Mittisila	533464,Multispecialty Consulting Unit, JIPMER, Government Road, Yanam - 533464	1	30

Special terms and conditions-Version:1 effective from 15-04-2024 for category Ultrasound Machine (V2

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals and Fertilizers, Government of India, from time to time in this regard.
2. The sellers are registered on GeM based on self declaration of valid Medical Device License, product certification. However, buyers must check and validate the details at their end for all applicable licenses and certifications, authenticity/genuineness of Medical Device license, product certification, manufacturer certification/licence.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer. Authorized resellers/distributors are operating in compliance with all relevant laws and regulations and are properly handling the manufacturer's products, including verifying the validity and authenticity of Medical Device license held by the manufacturer.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by the Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price fixed by the Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC shall be rejected by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from the seller with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence. The above terms shall supersede General Terms and Conditions (GTC), whenever there are any.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including service /operational manual of the manufacturer, service charges and spares. During the warranty period, after the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be applicable for consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) warranty period. In case of downtime, penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to the buyer. It shall be the responsibility of the seller that authorized service centres are available to cater to the areas where supplies are made within reasonable distance so that the service calls can be handled. Details of toll-free numbers for service call and online registration of service calls shall be provided by the seller/bidder to the buyer/consignee at the time of supplies.
8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipment. Seller shall provide Performa invoice or any other documents to establish that the products supplied are manufactured by the seller/bidder from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods shall be provided strong enough to withstand transit including transshipment (if any), rough handling, open storage etc. without damage to the goods etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, mode of transport and handling facilities at all points during transit up to final destination,. Quality of packing shall be ensured within & outside the packages and provision of accompanying documentation shall take into consideration the nature of equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked with indelible paint of proper quality: indicating contract number and date, brief description of goods, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured or to be manufactured. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supply. In case of reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee. Such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM or reseller shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee. The price for most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning of the goods. Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for installation and Training of Consignee's representatives for operating and maintaining the equipment and providing a number of operation & maintenance manual for the goods. In case the category parameters are specified in the STC, they shall be followed.

regarding the installations, training and manuals the same shall also be applicable.

12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical Safety Standards or equivalent .In case they do not have required equipment for such testing should er checked for electrical safety compliance through labs with facilities for such checking during every prev
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/**Buyer Added Bid Specific Terms and Conditions**

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.

4. **Generic**

Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following ac

Dr. Anand Raju Mittisila
Multispecialty Consulting Unit, JIPMER,
Government Road
Yanam - 533464
East Godavari ANDHRA PRADESH - 533464

5. **Generic**

Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibilit
reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum

6. **Generic**

Data Sheet of the product(s) offered in the bid, are to be uploaded along with the bid documents. Buyers can
Sheet with the product specifications offered. In case of any unexplained mismatch of technical parameters, t

7. **Generic**

Installation, Commissioning, Testing, Configuration, Training (if any - which ever is applicable as per scope of
by OEM / OEM Certified resource or OEM authorised Reseller.

8. **Generic**

Scope of supply includes Training: Number of employees to be trained

5
, Place for Training
Multispecialty Consulting Unit, JIPMER, Government Road, Yanam - 533464
and Duration of training
2
days.

9. **Generic**

The successful bidder has to supply all essential accessories required for the successful installation and comn
supplied. Besides standard accessories as per normal industry practice, following accessories must be part of
included in bid price:

Yes

10. **Generic**

While generating invoice in GeM portal, the seller must upload scanned copy of GST invoice and the screenshot of payment of GST.

11. Scope of Supply

Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of

12. Warranty

Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance for 5 years for the offered products at the rate not more than 5 % of contract price per annum. Buyer reserves the right to terminate the agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and charges would be made Annually after rendering of the CMC Services of the relevant CMC period. Performance Security of bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include

500

(Upload the undertaking). The original Performance Security of contract will be returned only after submission of Performance Security for 5% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).

13. Warranty

Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service request during warranty period, Seller has to complete the required Service / Rectification within 5 days time limit. If the Seller fails to complete the service / rectification within defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover the penalty from the Performance Security (PBG). Cumulative Penalty cannot exceed more than 10% of the total contract value. Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller. Seller shall be liable to reimburse the cost of such service / rectification to the Buyer.

14. Certificates

Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid Document. If any, the Bidder must upload the same in the Bid Corrigendum.

15. Buyer Added Bid Specific ATC

Buyer Added text based ATC clauses

TECHNICAL SPECIFICATIONS OF Mid-End USG machine

TECHNICAL SPECIFICATIONS OF Mid-End USG machine	
S.NO.	TECHNICAL SPECIFICATIONS
1	It should be robust state of art, fully digital high end latest Color Doppler Ultrasound System with crystal-live / similar architecture capable of precision beam forming ,capable of performing abdominal, obstetric/gynaecology, Fetal Heart ,musculoskeletal, small parts, Urology, Breast, Pediatric, Cardiac
2	System should have Windows 11 Operating System for strengthened cybersecurity and ensure regular continuous updates
3	System should have broad band beam former capable of processing signals from 1-20 MHz
4	System should have latest state of the art Hybrid Beam forming technology to ensure no compromise in axial and Lateral resolution

5	System should have more than 23" LED Flat panel Monitor
6	System should have more than 14" wide LED tilting Touch Screen Control.
7	System processing channels must be more than 80,00,000.
8	Frame rates more than 5000 frames/sec preferred.
9	System should have Digital TGC and LGC controls
10	System should incorporate facility for high resolution 2D, M-mode, PW, Color Flow Imaging, g, Power Pulse Inversion Harmonics, Directional Color Power angio imaging modes and Corr
11	System should have Full Spectrum Imaging, Tissue Harmonic Imaging, Spatial Compound I armonic Imaging, Trapezoidal Imaging, Quad Imaging, Dual Imaging in Horizontal Split, 2D/ PW Doppler Adjustment and Auto 2D Adjustment.
12	The system should have capability to automatically adjust brightness and uniformity of th me to provide optimal image quality for each organ, helping to improve diagnosis and work
13	System should have Advanced Image Processing algorithm to analyze between targets and target anatomy, reduce the sparkle & artifacts to improve image quality.
14	System should have scan depth of 2 to 50 cm or more. Please specify through data sheet.
15	System should have 256 shades of gray display.
16	System should have algorithm to provides clearer images by mitigating the characteristics are slightly blurred than the actual vision
17	The system should have a feature to enhance hidden structures in shadowed regions by se uency and low-frequency of the ultrasound to identify shadowy areas such as fetal head or ccurs.
18	The should have the feature to get a three-dimensional visualization of blood flow by 2D pr stand the structure of blood flow and small vessels intuitively
19	The system should have a feature to visualize microcirculatory and slow blood flow to to de cularization and to provide a detailed view of blood flow in relation to surrounding tissue or spatial resolution. System should have to detect very low intensity vascularization.
20	The system should have a feature which optimize Color and PW images with one click of a l quickly acquire the optimal images for vascular structures , enhancing the workflow during
21	System should have feature for automatic biometric measurement. It enables users to mea al growth parameters with one click while maintaining exam consistency.
22	System should have a feature to provides automatic measurement and Z score required fc
23	The system should have feature to automatically extracts the centerline and thickness of th d provides a coronal view in 3D.

24	The system to have a feature which allows the user to rotate, zoom, crop, and move 3D chscreen, zoom, crop, and move 3D images right from the touchscreen
25	System should have Fetoscopic view technology that displays detailed volume rendering, e ntify subtle anatomical structures with change in position of light source. Anatomies look re
26	The Endocavity probe should have viewing angle of more than 180 Degree, to visualize ent undus.
27	Console height should be adjustable for user's comfort.
28	Convex Probe with Single Crystal will be accepted for higher frame rate and deep penetrati Real time Shearwave liver elastography with quantification.
29	System should be DICOM ready.
30	System should have built in Image Management Software, for off line application when pati ation, such as Image Manipulation, Multi Planner reformatting, surface & volume rendering sk memory of 512 GB or more.
31	System should have facility for real time or frozen, pan or point zoom.
32	System should have cine loop review minimum 50000 frames. Please specify through data
33	System should have upgradable automatic tool to derive 9 planes of fetal heart (as per AIL one volume sweep.
34	System should have an upgradable automatic tool for deriving 9 planes for Fetal Central Ne
35	System should have 2D real time Shear wave Imaging with Convex and Linear Probe.
36	The quoted model should be US FDA/CEA/BIS/CDSCO certification.
37	Please respond to each specification in the same format and order and support it with Prod
38	System should have 2D real time Shear wave Imaging with Convex and Linear Probe.
39	Please respond to each specification in the same format and order and support it with Prod
40	All the relevant measurement software should be included for the probes supplied.
41	OEM authorisation letter to be submitted.
	System should be provided with following transducer:
A	Single Crystal Convex Abdominal probe with frequency range from 1 to 7 MHz. (Single Crys or higher frame rate and deep penetration, also capable of doing Shear wave Elastography eptable

B	Endocavity (TVS) 2-10 MHZ approx. with more than 180 Degree Angle. +/-1 MHz Frequency biopsy guide along with probe.
C	Single Crystal Linear probe for vascular studies and Breast Imaging 2-14 Mhz. (Single Cryst or higher frame rate and deep penetration, also capable of doing Shear wave Elastography) ptable
D	Single Crystal Adult Cardiac Probe with 1-5 Mhz. approx. for 2D Echo Studies +/-1 MHz Frec
	Accessories:
	UPS with 30 mins backup-1 No
	Color ink jet Printer with ink tank (refillable)-1 No
	Additional Specification Parameters:
A	The Vendor must bring the equipment at own expenditure for demonstration on a given day. It is liable to be rejected. Place of demonstration WCH block, JIPMER Puducherry 605009
B	The user manual and service manual in English must be provided. Power input to be 220-240V Indian Plug.
C	Warranty: 5 year comprehensive warranty on the whole machine and all the accessories included.
D	CAMC: 5 year CAMC on USG machine. CAMC should cover replacement of at least one probe in the event of a manufacturing defect, poor visualization of image quality or non-functioning of probe.
E	Warranty of the product shall be Five years from the date of installation & CAMC cost for the product to be submitted. The user manual and service manual in English must be provided. The Bidder must submit the document related to the Past performance with 30% of the tender value. Cost break up details of the quoted product should be submitted along with the price bid. The performance security paid for the purchase of the equipment will be refunded only after the CAMC is paid as per the GeM guidelines. Performance bank guarantee during CAMC period shall be on total CAMC value.
D	Timely Servicing/ rectification of defects during warranty period: After having been notified of a service requirement during warranty period. Seller has to complete the required Service/ Rectification within the defined time limit. If the Seller fails to complete service/ Rectification within defined time limit, a penalty of the product shall be charged as penalty for each week of delay from the seller. Seller cannot shift the responsibility to the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Seller / Bank guarantee (PBG) Cumulative Penalty cannot exceed more than 5 % of the total contract value. Buyer shall have the right to get the service/ Rectification done from alternate sources at their cost besides forfeiture of PBG. Seller shall be liable to reimburse the cost of such service/ Rectification.

Buyer Added text based ATC clauses

EMD payment process:

1. Go to <https://jipmer.edu.in/>
2. Select online payment option
3. Redirect to SBI collect window
4. Click proceed
5. Select payment category : tender fee
6. Make the payment
7. Upload the SBI collect receipt in the GeM portal under additional documents (1 or 2 or more)

For further details kindly call JIPMER Purchase Section -0413-2296047

16. **Buyer Added Bid Specific ATC**

Buyer uploaded ATC document [Click here to view the file.](#)

अस्वीकरण/Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority, who is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof in the bidding process due to these ATCs and including the modification of technical specifications and / or terms governing the bid. All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization with GeM. If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice.

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ is the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Source Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension is not approved by GeM.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, is permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyer / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause of the Buyer's / Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be dealt with by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specific technical specifications, Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM rules.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional technical specifications or other document. If buyer needs more items along with the main item, the same must be added through bunching or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be added separately.

Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this representation against the same by using the Representation window provided in the bid details field in Seller dashboard. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply.

Additional Clause For Comprehensive Maintenance Charges

1. CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the equipment and charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC. Full uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by period.

2. CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.

3. GST shall be included in the CMC Charges quoted.

4. Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula. A 10% discounting rate per year. Charges for price evaluation on Net Present Value.

5. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by the seller.

6. While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges for number of years for CMC required after the warranty period shall be available. Under this option up to 15 years can be selected beyond warranty period.

7. In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into full warranty including preventive maintenance and calibration as per technical/ service /operational manual of the equipment and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty (excluding consumables) also in product Cost.

8. The CMC functionality shall be available in bid only and no direct RA shall be applicable. In case of bid to R/A decision shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC.

8.1. Buyer shall indicate number of years of warranty by selecting different options available in the field dependent on parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall be same as period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fee for CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable for CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A15 are to be indicated for 6th to 15th year of CMC if applicable.

8.2. The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC period while creating bid.

8.3. In the price evaluation, the system shall provide function to calculate the cost of each equipment by form including CMC and then show the inter-se-ranking of the bidders. The following are the variables to be considered:

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation =

$C + C * \{ (A1/100)/(1.10^n) + (A2/100)/(1.10^{n+1}) + (A3/100)/(1.10^{n+2}) + (A4/100)/(1.10^{n+3}) + (A5/100)/(1.10^{n+4}) \}$
C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall be percentage of CMC charges for 1st, 2nd, 3rd, 4th & 5th years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4. CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5. The CMC charges shall be offered within range of 3 to up to 50% of cost of equipment as defined by buyer.

9. Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount for CMC contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10. Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable for the entire CMC period as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11. In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by L1 bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.
12. The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.
13. CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tool for the work.

All GeM Sellers/Service Providers shall ensure full compliance with all applicable labour laws, including schemes and guidelines under the four Labour Codes i.e. the Code on Wages, 2019; the Industrial Relations Act, 1947; the Occupational Safety, Health and Working Conditions Code, 2020; and the Code on Social Security, 2020 and brought into force by the Government of India.

For all provisions of the Labour Codes that are pending operationalisation through rules, schemes or notifications, the corresponding provisions of the pre-existing labour enactments (such as The Minimum Wages Act, 1948; The Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 and relevant State Rules) shall continue to remain applicable.

The Seller/ Service Providers shall, therefore, be responsible for ensuring compliance under:

- All notified and enforceable provisions of the new Labour Codes as mentioned hereinabove; and
- All operative provisions of the erstwhile Labour Laws until their complete substitution.

All obligations relating to wages, social security, safety, working conditions, industrial relations etc. and other requirements shall be strictly met by the Seller/ Service Provider. Any non-compliance shall constitute a breach of contract and shall entitle the Buyer to take appropriate action in accordance with the contract and applicable law.

This Bid is governed by the General Terms and Conditions, conditions stipulated in Bid and Service Level Agreement for the Service, as the case may be, as provided in the Marketplace.

However, in case of Service, if any condition specified in General Terms and Conditions is contradicted or contradicted by the conditions stipulated in Service Level Agreement specific to said Service, then it will over-ride the conditions in the General Terms and Conditions.

[यह बिड सामान्य शर्तों के अंतर्गत भी शासित है /This Bid is also governed by the General Terms and Conditions](#)

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ सीमा साझा करने वाले देश का कोई भी बिडर इस निविदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिडर को बिड देने का अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने पर इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के तहत कार्रवाई का आधार होगा।/In terms of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of the same. Non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---