


CENTRE FOR BARIATRIC SURGERY
Gastrointestinal Physiology Laboratory
VIDEO MANOMETRY REPORT



Name: Sue McKenzie

Date: 3/4/08

Summary of findings: Despite adding up to 5 mls into the LAP-BAND, no restriction was able to be produced to either liquid or semi-solid (porridge) swallows. Consistent with erosion or mechanical failure of LAP-BAND mechanism. Gastroscopy pending. AP and Lateral images in this show the appearance with 5 mls.

Number of images in this report: 2

