

KNEE MR ARTHROGRAM

FOV = 16 x 16

Ax T2 FSE fat sat	3.5/0.7
Sag PD FSE	3.5/0.7
Cor PD FSE fat sat	3.5/0.7
Sag T1 fat sat	3.5/0.7
Cor T1 fat sat	3.5/0.7
Cor T1 FSE	3.5/0.7

<b>Ax PD fat sat (thin as possible)</b>	<b>1/0</b>
---	------------