

#### 4. Routine Liver with MRCP with EOVISt (FNH/ biliary leak/ Mets)

<b>Routine Liver with MRCP with Contrast</b>
3 Plane Localizer free breathing – 3 Plane Localizer breath hold T2 Haste Coronal T2 Haste axial T1 ViBE Dixon breath hold /Fat quantification T1 VIBE fs axial breath hold
<b>Post contrast</b>
T1 VIBE fs axial 30s T1 VIBE fs axial 70s T2 Haste axial fs T1 VIBE fs axial 3 min T2 TrueFISP axial bh T1 VIBE fs axial 5 min T1 Vibe coronal Diffusion 50, 800 T2 Haste fs coronal thick slab breath hold T1 VIBE fs axial 10 min T1 VIBE fs axial 20 min (Check with radiologist) T1 VIBE fs axial 30 min (Check with radiologist) T1 VIBE fs axial 40 min (Check with radiologist)

Note: after the 20 min please check with Radiologist if the exam needs to be extend.