6. Routine/Benign Female Pelvic

Routine/Benign Female Pelvic

- 3 PL localizer free breathing
- T2 Haste sagittal
- T2 Haste coronal
- T2 Haste axial
- T2 TSE fs Axial
- T1 Dixon fs axial
- T1 VIBE fs axial breath hold
- T1 VIBE fs sagittal breath hold
- T1 VIBE fs coronal breath hold
- T2 TrueFISP axial breath hold
- T2 coronal TrueFISP (cover from top of the kidney to pubic symphysis)

Contrast

- T1 VIBE fs axial +c (50s)
- T1 VIBE fs axial +c (90s)
- T1 VIBE fs sagittal+c (2 min)
- T1 VIBE fs coronal+c (3 min)
- T1 tse axial +c (4 min)

If Congenital female pelvic has been ordered by the staff/Fellow radiologist please add;

T2 SPACE axial iso (3 mm)