Lecture Module - Data Acquisition

ME3023 - Measurements in Mechanical Systems

Mechanical Engineering

Tennessee Technological University

Module 8 - Data Acquisition

Module 8 - Data Acquisition

- Topic 1 - Analog to Digitial Conversion
- Topic 2 - DAQ Hardware and Applications
- Topic 3 - Sampling and Aliasing

Topic 1 - Analog to Digitial Conversion

- DAQ and Computer Storage
- Number Types
- Analog to Digital Conversion and DAQ
- Activity: ADC Resolution Calculation

DAQ and Computer Storage

Types of Signals:

- ___________ - magnitude is continuous in time
- ___________ ______ - magnitude at points in time

  - sampling at repeated time intervals
- ___________ - exists in discrete points in time

  - magnitude is also discrete

DAQ and Computer Storage

A data acquisition system is the portion of a measurement system that quantifies and stores data. - Theory and Design of Mechanical Measurements



Image: T.Hill

Number Types

- Integers
  - Binary
  - Decimal
  - Hexadecimal
- Fixed Point
- Floating Point

Number Types

Binary	Decimal	Hexadecimal
0	0	0
1	1	1
10	2	2
11	3	3
100	4	4
	5	5
	6	6
	7	7
	8	8
	9	9
	10	A
	11	B

Binary	Decimal	Hexadecimal
	12	C
	13	D
	14	E
	15	F
	16	
	17	
	18	
	19	
	20	
	21	
	22	
	23	

some reference

Number Types

Binary	Decimal	Hex.
0	0	0
1	1	1
10	2	2
11	3	3
100	4	4

Binary	Decimal	Hex.
0	0	0
1	1	1
10	2	2
11	3	3
100	4	4

some reference

Number Types

Standard storage of a floating point value in memory



Wikipedia: IEEE Float Wikipedia: Fixed Point

Number Types

Integer	Floating Point	Fixed Point
Pros:	Pros:	Pros:
Cons:	Cons:	Cons:
Examples:	Examples:	Examples:

Text: Theory and Design for Mechanical Measurements

Analog to Digital Conversion and DAQ

In electronics, an ______________ (ADC, A/D, or A-to-D) is a system that converts an analog signal, such as a sound picked up by a microphone or light entering a digital camera, into a digital signal. An ADC may also provide an isolated measurement such as an electronic device that converts an analog input voltage or current to a digital number representing the magnitude of the voltage or current. Typically the digital output is a two's complement binary number that is proportional to the input, but there are other possibilities.



wikipedia, image

Analog to Digital Conversion and DAQ

Activity: ADC Resolution Calculation

It is important to realize the potential for data loss resulting in a reduced quality measurement based on the parameters of the analog to digital conversion process. This issue can occur when designing systems around a low-level analog to digital converter as well as when using high-end DAQ equippment.



Activity: ADC Resolution Calculation

Activity: - Consider setting up a data acquisition system to record pressure measurements in a vehicle system. Multiple sensors are available, and the DAQ device has different operating modes. Each sensor and DAQ mode has a different input and output signal ranges and different sampling frequencies.

Signal

- Measure Variable: Pressure (psi) in automobile tire
- Expected Range: 0-100 psi

Sensor+Transducer	Input Range (psi)	Output Range (volts)
A	0-200	0-3.0
B	0-120	0-0.50

DAQ Mode	Input Range (volts)	ADC Resolution
1	0 to 3.3	10bit
2	-10 to 10	12bit
3	0 to 10	12bit

Activity: ADC Resolution Calculation

Activity (continued):

1. Choose a **sensor+transducer** pair and an appropriate **DAQ mode** to record the signal shown with best (lowest) possible resolution. Support your choices with a resolution calculation for smallest detectable voltage (*volts*) and smallest detectable pressure (*psi*)

2. Approximate the sensivity of the measuement system in units of $\frac{psi}{volts}$.

NI - Smallest Detectable Change

Topic 2 - DAQ Hardware and Applications

- Signal Types and Wiring Configurations
- EMI Considerations
- Available Hardware
- Software Integration

Signal Types and Wiring Configurations

Most data acquisition devices and systems ______________ and ______________ analog voltage signals and possibly additional signal types. Signal ______________ may also be a feature on some systems.

A voltage signal requires a **common** reference or ______________.

Signal Sources:

- Grounded or Ground-Referenced
- Ungrounded or Floating

Measurement (DAQ) Systems:

- Common Ground
- Common Mode Voltage
- Isolated Ground

NI, Digilent
Text: Theory and Design for Mechanical Measurements

Signal Types and Wiring Configurations

Most data acquisition devices and systems measure and record analog voltage signals and possibly additional signal types. Signal generation may also be a feature on some systems.

2 Major Configurations:

- Single-Ended Signals
  The signal is measured as a voltage between a ______________ conductor and the ______________ which must be carried on a separate conductor or wire.

- Double-Ended (Differential) Signals
  The signal is measured as the difference between ______________ voltages (double) carried on separate conductors, or wires. Typically a ______________ is shared between the two devices requiring a third conductor.

Read more here: MCCDAQ

Signal Types and Wiring Configurations

Ground Referenced Source Differential:





Images: NI myDAQ User Guide and Specifications

Signal Types and Wiring Configurations

Floating Source Differential:





Images: NI myDAQ User Guide and Specifications

Signal Types and Wiring Configurations

Floating Source Differential with Resistors:





Images: NI myDAQ User Guide and Specifications

Signal Types and Wiring Configurations

Multiple Signal Sources:



Figure 4. USB-6211 OEM Connector Pinout

Images: NI UBS 6211 User Guide and Specifications

Signal Types and Wiring Configurations

Single-Ended Signals

Pros:

Cons:

Examples:

Double-Ended Signals

Pros:

Cons:

Examples:

Text: Theory and Design for Mechanical Measurements

EMI Considerations

______________, also called radio-frequency interference (RFI) when in the radio frequency spectrum, is a disturbance generated by an external source that affects an electrical circuit by electromagnetic induction, electrostatic coupling, or conduction.

A *combination* of naturally occuring and human made sources of interference is always present. The total EMI affecting a system is determined by the local conditions as well as global environmental influences.

Sources of EMI:

-
-
-
-
-

EMI Considerations

In data acquisition, electromagnetic interference (EMI) can cause ______________ of signal quality and data ______________ in the form of ______________ and or ______________.

Consider the case of an analog signal transmitted from a sensor to a DAQ device. What can be done to avoid issues associated with EMI?

Methods of reducing EMI affects:

- Proximity -
- Differential signal -
- Noise rejection cables/wires -

wikipedia: twisted pair, Mouser-Alphawire

Available Hardware

- National Instruments
- Measurement Computing
- dSPACE
- Arduino or other

Available Hardware

- National Instruments
- Measurement Computing
- dSPACE
- Arduino or other

Available Hardware

Software Integration

Software Integration

Topic 3 - Sampling and Aliasing

- Sampling
- The Aliasing Phenomenon
- Example by Hand
- MATLAB Example

Sampling

... A discrete time signal usually results from the __________ of a continuous variable at __________ finite time intervals. ...



Discrete time signal

$\{y(r\delta t)\}$

r	Discrete data
0	0
1	5.9
2	9.5
3	9.5
4	5.9
5	0
6	−5.9
7	−9.5
8	−9.5
9	−5.9
10	0

Text,Figure: Theory and Design for Mechanical Measurements Ch. 7

Sampling

The Aliasing Phenomenon



Figure: Theory and Design for Mechanical Measurements Ch. 7

Example by Hand



Image: T.Hill

Example by Hand

MATLAB Example

```
% ME3023 - Tennessee Technological University
% Tristan Hill - October 10, 2019 - April 14,
    2021
% Data Acquisition Topic 3 - Sampling and
    Aliasing

clear variables; close all; clc

% simulate a continuous signal
A1=5; f1=3;
w1=2*pi*f1;

dt_sim=0.001; t_stop=6;
t_sim=0:dt_sim:t_stop;
y_sim=A1*sin(w1*t_sim);
```

MATLAB Example

```
% simulate sampling the signal
dt_sam = 0.3;
t_sam=0:dt_sam:t_stop;
y_sam=A1*sin(w1*t_sam);

% show the figure
figure(1); hold on
plot(t_sim,y_sim,'-',t_sam,y_sam,'o')
axis([0 t_stop -1.2*A1 1.2*A1])
grid on
```

MATLAB code: T. Hill