



PASS

Ready for use

Readiness score 96/100 · All safety checks passed · Philips HeartStart HS1

INSPECTED

28 Sept 2026, 10:14 IST

Took 3 min 12 s

NEXT STEPS

- Order replacement pads — they expire in Nov 2026 (2 months left).
- Next routine check due by 28 Oct 2026.

aedsmartx by Think Health supplies replacement pads, batteries and accessories for the Philips HeartStart HS1 — aedsmartx.com

DEVICE

AED MODEL

Philips HeartStart HS1

SERIAL NUMBER

A18A-06336

PADS EXPIRY

Nov 2026

2 months left

BATTERY EXPIRY

Dec 2028

2+ years left

READINESS INDICATOR

Ready

INSPECTION

INSPECTED BY

Priya Sharma

EMAIL

priya.sharma@example.com

PHONE

+91 98765 43210

REPORT

7C1E4A92

CHECKS

Consumables & identification



Serial number

A18A-06336

Serial number read clearly from the SN label on the back panel.

PASS

Confidence 97%



Pads expiry

Nov 2026 · 2 months left

Pads cartridge expiry date is 2026/11, beside the hourglass symbol.

PASS

Confidence 95%



Battery expiry
Dec 2028 · 2+ years left

Battery label shows an install-before date of 2028-12-31.

PASS

Confidence 93%

Physical status



Battery attached

Battery is pushed fully in and sits flush with the case.

PASS

Confidence 94%




Pads connected

Pads cartridge is fitted in the front well, green PULL handle down.

PASS

Confidence 92%



Readiness indicator

Ready

Green Ready light seen blinking at a steady interval; no fault indicators.

PASS

Confidence 90%

Accessories & signage



AED cabinet OPTIONAL

Cabinet door closed and AED signage clearly visible.

PASS

Confidence 90%



Fast response kit OPTIONAL

Kit holds gloves, a razor, scissors and a CPR face shield.

PASS

Confidence 88%

An AI-assisted visual inspection, made from photos and video taken on site. It supports the manufacturer’s maintenance schedule and does not replace it.