THINK HEALTHCARE AND SAFETY

Inspection Summary — ETV Parcel 9, Bengaluru

Period: 01 Aug 2026 – 30 Oct 2026 · Generated 30 Oct 2026, 05:22 by Technical Reviewer

Filters: City: Bengaluru · Site: ETV Parcel 9

56	1.8%	0%	1	2
Assets in scope	Compliance (in date)	Pass rate (period)	Inspections (period)	Open defects

1	0	0	55	56
Out of service	Overdue	Due in 14 days	Never inspected	Incomplete records

Breakdown by floor

Floor	Assets	Inspected	Pass rate	Defects	Out of svc	Overdue	Due 14d	Never	Compliance
Unassigned floor — ETV Parcel 9, Bengaluru	56	1	0%	2	1	0	0	55	1.8%
Total	**56**	**1**	**0%**	**2**	**1**	**0**	**0**	**55**	**1.8%**

Open defects (2)

Asset	Site / floor	Check	Remark	Raised
JLL-EP9-WC-001	ETV Parcel 9	**#3 Wheel locks: engagement, holding function and release (critical)**	Left wheel lock does not hold on ramp	30 Oct 2026
JLL-EP9-WC-001	ETV Parcel 9	**Load test Functional / load test (critical)**	Chair rolled with left lock engaged	30 Oct 2026

Assets overdue or due within 14 days (0)

Nothing due in the next 14 days.