

TANAFS Logo

TANAFS

Patient Summary & Clinical Review Report

MEDICAL SUMMARY REPORT

Generated on 21 Apr 2026, 02:12 PM

Patient Information

Patient Name:

Ali AlHarbi

Patient ID:

1001

Gender:

Male

Date of Birth:

14 Mar 1990

Phone:

0551234567

Status:

Active

Generated By

Healthcare Professional:

Njood Turki

Role:

Respiratory therapist

Email:

njood_tmo@hotmail.com

Phone:

0537335500

Report Date:

21 Apr 2026

Report Time:

02:12 PM

Clinical Summary

This report summarizes the patient's waveform analysis history and documented healthcare professionals' comments. A total of 4 waveform analysis record(s) and 1 comment(s) were identified for this patient. Among these analyses, 4 anomaly detection(s) were documented, with 'Ineffective Effort Flow' appearing as the most frequent recorded pattern when applicable. The most recent analysis available in the system was recorded on 18 Mar 2026, 08:30 AM.

4

Total
Analyses

4

Detected
Anomalies

1

Clinical
Comments

**Ineffective
Effort Flow**

Most Frequent Pattern

All Previous Waveform Analyses

Premature Cycling Flow

Recorded on 18 Mar 2026, 08:30 AM

ANOMALY

Performed by:

Njood Turki (Respiratory therapist)

Anomaly Type:

Premature Cycling Flow

Doctor Action:

Increased inspiratory time and monitored waveform ..

Double Triggering Flow

Recorded on 12 Feb 2026, 05:32 PM

ANOMALY

Performed by:

Njood Turki (Respiratory therapist)

Anomaly Type:

Double Triggering Flow

Doctor Action:

Reviewed sedation level and optimized ventilator support.

Ineffective Effort Flow

Recorded on 11 Feb 2026, 12:11 PM

ANOMALY

Performed by:

Abdullah AlQahtani (Doctor)

Anomaly Type:

Ineffective Effort Flow

Doctor Action:

No additional Doctor Action were documented for this analysis.

Ineffective Effort Flow

Recorded on 11 Feb 2026, 12:11 PM

ANOMALY

Performed by:

Abdullah AlQahtani (Doctor)

Anomaly Type:

Ineffective Effort Flow

Doctor Action:

Increased trigger sensitivity to detect patient effort.

Healthcare Professionals' Comments

Njood Turki - Respiratory therapist | 12 Feb 2026, 04:05 PM

Feeling better

Key Recommendations

- Review cycling criteria and inspiratory timing to improve synchrony and avoid early cycle termination.
- Review ventilator sensitivity, inspiratory time, and patient-ventilator synchrony to reduce repeated triggering.
- Assess patient effort and consider adjusting trigger settings, sedation level, or ventilator support parameters.