

0.1% Formic Acid in Water - Mobile Phase

Ready-to-Use Solutions | Application-Specific Mobile Phase Kits

1. Specification scope

Product	0.1% Formic Acid in Water - Mobile Phase
CAS number	64-18-6
Grade scope	LC-MS Grade
Document status	Reference specification summary

The values below are representative catalog specifications. Where several grades are listed, acceptance limits may differ by grade. The current controlled grade-specific specification agreed at quotation or qualification is the governing document.

2. Test requirements

Test item	Specification / reference value	Test method
Formic Acid Concentration	0.1% v/v ($\pm 2\%$)	Titration
Water Quality	TOC < 50 ppb, Resistivity > 18 M Ω ·cm	Approved method
Filtration	0.2 μ m membrane filtered	Approved method
Packaging Atmosphere	Nitrogen-filled	Approved method

3. Packaging and release documentation

Package size	Container / format	Supply status
5L	FLEX BAG	Available
10L	FLEX BAG	Available
20L	FLEX BAG	Available

Each released batch is supported by a batch-specific Certificate of Analysis. Packaging sizes are consolidated in this document; actual availability, closure system and minimum order quantity are confirmed with the quotation.

4. Document control notice

This downloadable file is a reference summary generated from the LANCHROM product catalog. Request the current approved specification, SDS and sample COA before method validation, supplier qualification, regulatory submission or

purchase release.