

# pH Buffer Standard Solutions

Reference Standards & Analytical Reagents | Physicochemical Calibration Standards

## 1. Specification scope

|                 |                                 |
|-----------------|---------------------------------|
| Product         | pH Buffer Standard Solutions    |
| CAS number      | -                               |
| Grade scope     | HPLC Grade                      |
| Document status | Reference specification summary |

The values below are representative catalog specifications. Where several grades are listed, acceptance limits may differ by grade. The current controlled grade-specific specification agreed at quotation or qualification is the governing document.

## 2. Test requirements

| Test item       | Specification / reference value | Test method     |
|-----------------|---------------------------------|-----------------|
| pH 4.00 Buffer  | 4.00 ± 0.01 @ 25°C              | Approved method |
| pH 7.00 Buffer  | 7.00 ± 0.01 @ 25°C              | Approved method |
| pH 10.01 Buffer | 10.01 ± 0.01 @ 25°C             | Approved method |
| Traceability    | NIST-traceable reference        | Approved method |

## 3. Packaging and release documentation

| Package size | Container / format | Supply status |
|--------------|--------------------|---------------|
| 250mL        | HDPE               | Available     |
| 500mL        | HDPE               | Available     |

Each released batch is supported by a batch-specific Certificate of Analysis. Packaging sizes are consolidated in this document; actual availability, closure system and minimum order quantity are confirmed with the quotation.

## 4. Document control notice

This downloadable file is a reference summary generated from the LANCHROM product catalog. Request the current approved specification, SDS and sample COA before method validation, supplier qualification, regulatory submission or purchase release.