An alternative proposal for dynamic ISF in open-source AID systems

The Diabeloop glucose curve, tested on other cohorts and set against the logarithmic and sigmoid models

Tim Street, Diabettech

12 November 2026

Summary

The dynamic ISF models in AndroidAPS and Trio have two origins. The logarithmic model began with Chris Wilson, who described the work in a talk to Loop and Learn in February 2022 [2]. Through an online form he asked volunteers from the diabetes online community what correction dose they would give at a stated glucose, with glucose steady and no insulin on board. About 50 people answered, from a two-year-old using 8 U a day to people in their seventies using more than 100 U, and the usable answers gave about 300 data points. The doses and their effects were reported by the respondents from experience, and neither was measured. He trained a neural network on those points and gave the surface it found across glucose and TDD as roughly 277,700 ÷ (glucose × TDD). AndroidAPS first carried that equation, from February 2022, and the glucose term was replaced by a logarithm in June 2022, which is the form in use now [3]. The talk predates that change and does not describe it.

The sigmoid model offered in Trio came from its author's observation of glucose response, and no analysis of data has been published in support of it.

Diabeloop has since examined the same idea on the records of its own closed-loop system and presented the result at the American Diabetes Association's Scientific Sessions in 2025 [1]. From 11,201 adults it reported a sensitivity of 79 mg/dL per unit at a glucose of 100 mg/dL, 35 at 180 and below 20 above 300, with a fitted equation.

This document builds on that work. It sets out the equation and how it was derived, and reports what the same measurement shows on other cohorts. It then describes how the equation could be offered in AndroidAPS and Trio as an alternative to both the logarithmic and the sigmoid models.

The measurement was repeated on 1,790 people in eight cohorts held in this series. Within person, the value at 100 mg/dL was 2.2 to 4.0 times the value at 200, against 2.61 in the published equation, so the steep fall towards 200 mg/dL is reproducible. On records where no algorithm was dosing, the published equation fits the measured values from 100 to 200 mg/dL at a single scale. On Loop records no single scale fits. Refitting the equation's five coefficients to each cohort agreed with both results, and returned a flatter curve than the published one above 200 mg/dL.

The measurement was also run on simulated people whose true sensitivity was known, and those runs are reported in full below. They describe how the measurement behaves. This document takes the published equation as given and asks how it should be scaled to a person and offered in the two systems.

The published equation has no term for the person, and it needs one before it can dose. Two are described here: the profile ISF, and total daily dose (TDD) as the logarithmic model uses. They were compared on 3,433 corrections given by 150 adults in a trial where no algorithm was dosing. The fall in glucose per unit followed the person's entered ISF, with a coefficient of 0.95 (95% interval 0.55 to 1.32) where the profile link predicts 1. Daily dose added nothing once the entered ISF was known, at -0.07 (-0.43 to 0.26) where the TDD link predicts -1. The link to the profile is therefore the candidate, and the link to TDD is kept as an arm of the evaluation.

The curve is fitted to a person by taking their profile ISF to be valid at one glucose and scaling the whole published curve to pass through it. That glucose is 140 mg/dL in this proposal, which is a modelling choice and open to revision. The publication's own reference is 180 mg/dL, and anchoring there would make every dose 1.44 times smaller. For an adult on 40 U a day with a profile ISF of 50 mg/dL per unit, the candidate gives 79.3 mg/dL per unit at 100 mg/dL, 50.0 at 140 and 30.3 at 200. AndroidAPS at an adjustment factor of 100% gives that person 43.4, 35.6 and 29.3. Both links are modelling choices made in this document and appear nowhere in the publication.

Multiplying gives every person the published shape, and two checks reported below support doing so. Attaching the curve by adding a constant in place of multiplying gave a sensitivity of zero or below at 299 of the 3,433 corrections.

A curve measured separately for each person had a shape that predicted that person's other events no better than the shape shared by everyone, in each of six cohorts.

The ends of the curve sit outside Trio's default limits. Trio holds a dynamic ISF between the profile ISF divided by 1.2 and the profile ISF divided by 0.7. The candidate stays inside that band only between 109 and 158 mg/dL. Below 109 mg/dL it gives a higher sensitivity than Trio allows, reaching 1.59 times the profile ISF at 100 mg/dL against a ceiling of 1.43. Above 158 mg/dL it gives a lower one, reaching 0.30 times the profile ISF at 380 mg/dL against a floor of 0.83. With the limits unchanged, Trio would dose from the curve over 49 of the 280 mg/dL it covers. The limits need a decision from Trio's maintainers before the model can run there as published. AndroidAPS applies no such limit in its dynamic ISF calculation.

The candidate has since been run in simulation (INV-019). Forty simulated adults were run for 35 days in TimSim with shipped oref0 dosing, and only the sensitivity rule differed between arms. Against the same people on their unchanged profile, the candidate lowered mean glucose by 1.8 mg/dL (95% interval 0.7 to 2.9) and raised time below 54 mg/dL by 0.43 percentage points (0.24 to 0.61), which is about what a profile ISF entered 30% too low did. About seven tenths of that extra time followed a glucose peak above 180 mg/dL, where the rule gives more insulin than the profile does. A harm rule written before the run was read on two simulated months together. On the low side it read harm against oref0 with autosens, at +0.37 points of time below 54 mg/dL (0.24 to 0.51). Against a logarithmic model matched to the profile it read unresolved, at +0.17 points (0.04 to 0.32). It read cleared on the high side against both. Held inside Trio's default limits, the candidate raised time below 54 mg/dL by 0.13 points (-0.02 to 0.27) and time in range by 0.90 points (0.11 to 1.69). The simulated body is built the opposite way to the equation's premise, since a unit of insulin does more there as glucose rises. The run therefore ranks the rules in that simulator, and its sizes are not estimates for people. The harm reading is reported and looked into, and it does not stop the candidate from moving to shadow running. Code for that stage has been written for AndroidAPS and for Trio. It calculates the candidate each cycle and stores it, and nothing doses from it. It has passed each project's unit tests and has not yet run on a device.

Two things are unsettled. The first is how far the comparison of links reaches. It covers adults on a pump who gave their own corrections, and it compares people with each other, so it does not show whether one person's value should also move with their recent insulin use, as the TDD link does. People

using a loop were not tested, and nor were children. The second is whether a loop dosing to this equation does better in people than one dosing to either existing model. The simulation is the only test so far, and nothing has been tested in people.

This is a draft for discussion. The equation has not run on a device, and the candidate uses the published coefficients unchanged. Readers are assumed to know how dynamic ISF works in both systems. Glucose is in mg/dL throughout because the code is.

What is the Diabeloop equation?

The source [1] is a conference abstract and poster from Diabeloop, the maker of the DBLG1 closed-loop system. It used the records of 11,201 adults with type 1 diabetes from March 2021 onwards. No full paper was found.

The abstract reports a sensitivity that rises from 65 to 79 mg/dL per unit between 50 and 100 mg/dL. Above that it falls, to 35 at 180 mg/dL and below 20 above 300. Relative to a dose sized at 180 mg/dL, the authors suggest the dose could be halved at 120 mg/dL or below and may need doubling above 340.

The poster's results figure gives the fit in two pieces, for glucose G in mg/dL. Up to 100 mg/dL:


```
ISF = 98.03 − 1.077 G + 0.008868 G^2
```

Above 100 and up to 380 mg/dL:

```
ISF = 272 − 3.121 G + 0.01511 G^2 − 3.305e−05 G^3 + 2.69e−08 G^4
```

The coefficients in the code (`inv008.dynisf.quartic_isf`) were checked against that figure on 11 November 2026 and match digit for digit. The quartic gives 80.6 mg/dL per unit at 100 mg/dL, 35.3 at 180, 21.1 at 300 and 15.3 at 380. It implies dose multiples of 0.56 at 120 mg/dL and 1.95 at 340, in line with the abstract's half and double. The quadratic gives 79.0 at 100 mg/dL and 65.3 at 60, so the two pieces differ by 2% where they meet.

The figure below redraws the poster's results figure. The poster gives its bin medians and quartiles only as a drawing, so they were measured from the image, and the measured medians sit within 2.6 mg/dL per unit of the published equations (`results/inv013_poster_figure.md`).



The published results figure, redrawn. Bars are bin medians and whiskers are quartiles, both measured from the poster's image. The curves are the published equations.

The equation is steeper than the logarithmic model. Between 100 and 200 mg/dL its value falls by a factor of 2.61, where the logarithmic model with a divisor of 55 falls by 1.48. From 200 to 300 mg/dL its value is multiplied by 0.69 and the logarithmic model's by 0.82.

It has no term for the person. Everyone gets 35.3 mg/dL per unit at 180 mg/dL, whatever their daily dose or their profile.

How was it derived?

The unit of analysis is a fall in glucose. On days with at least 75% of time in closed loop, the authors took each glucose peak between 50 and 400 mg/dL that came at least an hour after a meal, with less than 2.5 g of carbohydrate still on board, and followed it to the next minimum more than 20 minutes later. The sensitivity for the event is the fall divided by the insulin used over it. Insulin used is the drop in insulin on board plus whatever was delivered above the reference basal. The curve is the fit through those event values against glucose.

Each event ends where glucose stopped falling, so events are selected on what glucose did. The figure is therefore a ratio of two outcomes, the fall and the insulin used during it. The events are how the variation of sensitivity with glucose was measured. The proposal uses the equation fitted through them, applied at the current glucose.

Four things the derivation depends on are absent from the publication. It does not define the reference basal or say which insulin action model produced the insulin on board. The figure's horizontal axis is labelled G(t) with no statement of which glucose in an event that is. It gives no daily dose for the people measured. The rest of the poster could not be retrieved, and the authors have not been asked.

What does it show on our data?

The same construction was run in this series as INV-012, on 1,790 people in eight cohorts, and is reported in section 7 of document 5. It was built before the abstract was read, and two of its event rules differ from the published ones. It required three hours clear of a meal where the publication requires one hour with under 2.5 g on board, and a peak standing 15 mg/dL clear of its surroundings where the publication requires a 20 minute interval. Each event was filed under the glucose halfway between its peak and its minimum, and multiplied by the person's daily dose.

The shape between 100 and 200 mg/dL

The steep fall is reproducible. Within person, the value at 100 mg/dL was 2.2 to 4.0 times the value at 200 mg/dL across the eight cohorts, against 2.61 in the published equation. In REPLACE-BG, where no algorithm was dosing, the ratio was 2.55 (95% interval 2.25 to 2.82), which includes the published figure. Under closed loop it was steeper: 3.26 (3.11 to 3.39) for Loop, 3.86 (2.72 to 6.04) for oref0 and 3.21 (2.24 to 5.62) for Trio. The first two intervals lie above 2.61 and Trio's includes it.

Fitting the equation to the measured values

The table divides the measured value in each glucose bin by the published value at that glucose. The result is the daily dose, in U, at which the published equation would return what was measured. A row that stays level means the published shape fits that cohort at one scale. Only the four cohorts with carbohydrate records are shown, with the 95% interval over people.

Cohort	People	100 mg/dL	150	200	250	300
No algorithm (REPLACE-BG)	196	103 (89 to 115)	100 (94 to 108)	106 (94 to 125)	120 (103 to 137)	174 (132 to 233)
Loop	842	88 (84 to 91)	58 (56 to 60)	71 (67 to 73)	82 (77 to 89)	108 (95 to 119)

Cohort	People	100 mg/dL	150	200	250	300
Trio	22	90 (63 to 111)	52 (40 to 64)	74 (51 to 104)		
oref0	86	109 (90 to 136)	70 (59 to 86)	78 (58 to 188)		

With no algorithm dosing, the published equation fits from 100 to 200 mg/dL at one scale: 103, 100 and 106 U, each inside the other two intervals. Above 200 mg/dL the row rises to 120 and then 174 U, so the published equation falls faster there than these records do. Under Loop no single scale fits. The figure is 88 U at 100 mg/dL, 58 at 150 and 71 at 200, and those three intervals do not overlap. Trio and oref0 show the same dip at 150 mg/dL with wider intervals.

Two cautions apply to the table. The columns hold different people, since fewer reach the higher bins, and the within-person ratios in this section are the firmer comparison. The published values were read at the centre of each bin on the assumption that the publication files an event under its midpoint glucose, which it does not state.

The scale itself is not a daily dose anyone was using. It is about 100 U in REPLACE-BG because the measurement reads high where no algorithm doses, as the simulations below show.

Refitting the equation

The equation was also refitted. The quartic, with all five coefficients free, was fitted to each cohort's values in 10 mg/dL bins, from 100 mg/dL to the highest bin holding ten people. Each bin was weighted by the square root of the people in it, which is a modelling choice, since the poster does not state its weighting. For comparison the published fit was laid over the same bins with one free number, its scale in U. The miss is the typical distance between a fit and the bin values, as a percentage of each value.

Cohort	People	Bins fitted, mg/dL	Best scale of the published fit, U	Miss, published fit	Miss, refit
No algorithm (REPLACE-BG)	196	100 to 370	103 (96 to 113)	14%	7%
Loop	842	100 to 380	71 (69 to 74)	17%	6%
Trio	22	100 to 200	65 (48 to 80)	18%	7%

Cohort	People	Bins fitted, mg/dL	Best scale of the published fit, U	Miss, published fit	Miss, refit
oref0	86	100 to 240	82 (71 to 99)	23%	14%

The refit misses by less in every cohort, as a fit with four more free numbers must. The second table shows where the two differ. It gives each fit relative to its own value at 180 mg/dL.

Fit	People	100 mg/dL	150	200	250	300
Published		2.29	1.30	0.87	0.70	0.60
Refit, no algorithm (REPLACE-BG)	196	2.17 (1.91 to 2.40)	1.28 (1.21 to 1.33)	0.89 (0.86 to 0.92)	0.81 (0.72 to 0.87)	0.90 (0.77 to 1.02)
Refit, Loop	842	3.19 (3.02 to 3.34)	1.25 (1.22 to 1.27)	1.00 (0.98 to 1.01)	1.07 (1.03 to 1.11)	0.94 (0.88 to 1.01)
Refit, Trio	22	3.29 (2.15 to 4.76)	1.19 (0.85 to 1.53)	1.12 (0.66 to 1.61)		
Refit, oref0	86	2.96 (2.22 to 3.83)	1.00 (0.78 to 1.24)	1.02 (0.82 to 1.38)		

With no algorithm dosing, the refit returns the published shape from 100 to 200 mg/dL, and all three intervals there include the published figure. At 250 and 300 mg/dL the refit sits at 0.81 and 0.90 of its value at 180, where the published fit is at 0.70 and 0.60. Both of those intervals exclude the published figure. Under Loop the refit differs from the published fit at every glucose in the table. It is higher at 100 mg/dL, 3.19 against 2.29, and from 200 mg/dL upward it stays close to its value at 180. Trio's intervals, from 22 people, include the published figure at each of its three points. The oref0 values resemble Loop's, with an interval that excludes the published figure only at 150 mg/dL.

Two limits apply. A quartic suits these records poorly, because the values fall steeply to about 150 mg/dL and then level off. Under Loop the refit rises from 1.00 to 1.07 between 200 and 250 mg/dL, while the bin values themselves go from 2,182 to 2,025. Above 200 mg/dL the bins and the within-person ratios reported below are the better guide. The second limit is that the refit describes the same outcome ratio as the published curve, so the simulations in the next subsection bear on it equally.

The refit gives no single equation to offer in place of the published one. Its coefficients differ between cohorts, and its value at 180 mg/dL is 3,657 (3,431 to 4,055) with no algorithm dosing and 2,123 (2,048 to 2,214) under Loop. It

confirms the published shape between 100 and 200 mg/dL on records that no loop shaped. It also agrees with the within-person ratios that, by this measure, the published fall above 200 mg/dL is not reproduced here.

What the method returns when the truth is known

The method has been run twice on simulated people whose true sensitivity was known.

In the first run (INV-012) the sensitivity was constant at every glucose, 2,000 when multiplied by daily dose, and 45% of meals went unlogged. Two arms were dosed by a simple reactive loop, at a tenth and at a fifth of its gain. They returned values 1.4 to 3.4 times too high, declining as glucose rose. Their ratio of the value at 250 to the value at 200 mg/dL was 0.87 (0.83 to 0.90) in one arm and 0.92 (0.89 to 0.95) in the other. The same loop at full gain returned 1,816 and 1,520 at 150 and 200 mg/dL, below the true value, and reached no higher bin. An arm with no loop returned 4,560, 4,509 and 6,091 at 150, 200 and 250 mg/dL. Its ratio was 1.31 (1.00 to 1.94), from 25 people, so it read high and did not fall. A constant sensitivity therefore returned a falling curve in the arms where a loop was dosing, and in the arm with none it did not.

The six Jaeb cohorts gave 0.84 to 1.05 for the same ratio, a range that contains both reduced-gain arms, so the decline in those cohorts between 150 and 250 mg/dL needs no change in what a unit of insulin does to explain it. The published equation gives 0.80, which is steeper than either arm by an amount this series cannot account for. At 100 mg/dL between 59 and 93% of the simulated events had no net insulin used, and no arm gave a value at 300 mg/dL, so the first run says nothing about either end.

The second run (INV-014) was made to cover those ends. It used TimSim, which runs the UVA/Padova model of glucose and insulin [4, 5] with added layers for day-to-day physiology, behaviour, the sensor and the infusion site. That model has no single sensitivity, so the truth was measured before the method was read. Each of 40 simulated adults was held fasting at twelve steady glucose levels, and at each a small bolus was given in one of two otherwise identical runs. The true sensitivity at that glucose is the largest gap that opened between the two, per unit given. A unit lowered glucose by 11.1 mg/dL (10.1 to 13.7) at 100 mg/dL, from 39 people, by 23.0 (16.8 to 25.9) at 200, from 39, and by 34.6 (32.0 to 51.0) at 300, from 24. Within person the value at 100 mg/dL was 0.52 (0.49 to 0.54) of the value at 200, and the value

at 300 was 1.30 (1.25 to 1.34) of it. In this simulator a unit of insulin does more at higher glucose. That is a property of the model in a fasting steady state and is no evidence about people.

The same people then lived 56 simulated days under each of five dosing regimes. Two ran oref0 with SMB, one with the simulator's infusion-site faults and one without them. Two had no algorithm and gave corrections by hand: scheduled basal and a bolus for each announced meal, with a full correction towards 140 mg/dL whenever the sensor read above 180 mg/dL, at most once in three hours. Those two differed over site faults in the same way. The fifth had scheduled basal and meal boluses, with no corrections between meals. The correction rule is a modelling choice and was not fitted to anything.

The method was run on what a real record would hold, so unannounced meals and rescue carbohydrate were absent from it. Events were filed in bins 50 mg/dL wide, where the cohorts above used 10, because 56 days gave too few events for narrower ones. Of the 40 people, 31 completed every regime, and the table uses those 31. It gives the method's value divided by the truth, within person, and then the ratio of the value at 100 mg/dL to the value at 200 by the method and in truth.

Regime	Method ÷ truth at 100 mg/dL	At 150	At 200	100:200 by the method	100:200 in truth	People with both bins
oref0 with SMB	4.29 (3.83 to 4.50)	1.28 (1.19 to 1.40)	0.62 (0.53 to 0.73)	3.62 (3.06 to 4.30)	0.53 (0.49 to 0.55)	28
oref0 with SMB, no site faults	4.27 (4.02 to 4.61)	1.40 (1.27 to 1.51)	1.15 (0.84 to 1.24)	1.98 (1.61 to 3.02)	0.50 (0.47 to 0.61)	11
Corrections by hand	15.55 (13.47 to 20.00)	4.71 (4.00 to 5.21)	1.69 (1.35 to 1.85)	4.85 (3.92 to 6.67)	0.52 (0.49 to 0.54)	29
Corrections by hand, no site faults	17.94 (16.67 to 20.06)	5.00 (4.26 to 5.76)	1.88 (1.72 to 2.28)	4.64 (4.33 to 5.70)	0.52 (0.49 to 0.54)	29
Meal boluses only	14.90 (11.62 to 16.70)	16.00 (12.72 to 19.59)	28.18 (17.33 to 70.38)	0.24 (0.10 to 0.45)	0.52 (0.49 to 0.54)	29

In the four regimes that gave corrections, the method's value at 100 mg/dL was 2.0 to 4.9 times its value at 200, and every interval lies above 1. The truth ran the other way, at about half. The real cohorts gave 2.2 to 4.0 and the published equation gives 2.61, so the simulator returned the published shape

between 100 and 200 mg/dL from a sensitivity that rose with glucose. The arithmetic behind it is the same as in the real records, described under “Near target” below. Under oref0 the median fall was 32 mg/dL for events at 100 mg/dL and 41 for events at 200, while the median insulin used rose from 0.63 to 3.88 U.

The table’s values are medians over events with positive insulin used, as on the cohorts. Under oref0, 6 to 7% of events at 100 mg/dL had no net insulin used and none did at 150 or 200. With corrections by hand the share was 32 to 40% at 100 and 150 mg/dL and nil at 200.

With meal boluses only, the method’s curve rose with glucose. From 150 mg/dL upward the median event in that regime had no net insulin used, and 66 to 77% of events had none. Its values rest on the remaining events, whose doses were near zero, and they ran from 15 times the truth at 100 mg/dL to several thousand times it at 300. The arm with no loop in the first run behaved in the same way. Across both runs the curve fell where insulin was given in response to high glucose and did not fall where it was not, whatever the true sensitivity did.

The level was wrong in both directions. Under oref0 the method read 4.29 times the truth at 100 mg/dL and 0.62 of it at 200.

Above 200 mg/dL the check is thin. In the regimes with corrections the middle 90% of a typical person’s sensor readings lay between 78 and 241 mg/dL. One of them, corrections by hand with site faults, held ten people at 250 mg/dL. There the method’s value was 0.83 (0.75 to 1.07) of its value at 200, from 17 people, where the truth was 1.16 (1.13 to 1.22) of it. No regime with corrections filled a bin at 300 mg/dL.

Four limits apply to the second run. TimSim’s own audit records that its validation figures predate a rebuild, so it is not described here as validated. The run shows what the method returns from a known truth and says nothing about what the truth is in people. The truth was measured fasting with the added layers off, and a free-living high usually has carbohydrate still absorbing and more insulin on board. A truth measured inside the free-living runs was not attempted. The bins are five times wider than those used on the cohorts. The nine people left out each had a run that ended early, so the 31 are selected on having completed. Nine runs ended through hypoglycaemia, eight of them under oref0, and four through hyperglycaemia, all with meal boluses only.

Near target

Near target the decline comes from the insulin, with the fall almost unchanged. Across cohorts the median fall was 30 to 43 mg/dL for events at 100 mg/dL and 39 to 64 for events at 200. The median insulin used was 0.02 to 0.2 U at 100 mg/dL and 0.3 to 1.5 U at 200. A similar fall divided by much less insulin gives a much higher figure. At 100 mg/dL between 9 and 58% of events had no net insulin used at all, and a line fitted through each person's events put the fall at zero net insulin at 27 to 44 mg/dL. In type 1 diabetes a raised glucose does not fall without insulin, so that fall is basal and earlier doses acting, together with the selection of events that end at a minimum. The steep part of the curve below 150 mg/dL therefore rests on very small net doses. The second simulation showed the same pattern, and there the value at 100 mg/dL was 4.3 to 17.9 times the truth.

Above 200 mg/dL

The two measures disagree here. By the ratio of fall to insulin used, the value at 300 against 200 mg/dL within person was 0.90 (0.77 to 1.26) in REPLACE-BG, from 60 people, and 0.96 (0.84 to 1.08) under Loop, from 137 people. Both intervals exclude the published 0.69. Measured as the extra fall per extra unit, the same comparison gave 0.66 (0.49 to 0.73) in REPLACE-BG, from 24 people, and 0.73 (0.64 to 0.93) under Loop, from 49 people, and both of those include 0.69. No simulated regime with corrections reached 300 mg/dL, so neither measure has a known-truth check there. The extra fall per extra unit was not computed in the second simulation.

As a predictor

The shape has been tried once as a predictor on oref records, where it did worse than the static profile. On 80 people and 62,344 overnight windows, with each shape anchored to the profile at 100 mg/dL, the typical miss was 24.2 mg/dL for the static profile and 29.8 for this shape. At 205 to 260 mg/dL it was 37.8 against 73.7. No interval was reported for those figures. Two things limit the test. Its windows were selected on the later outcome, as the audit of 16 September 2026 recorded, and it anchored the shape at 100 mg/dL where this document anchors at 140.

Where that leaves the equation

The publication is a conference abstract from the system's manufacturer, drawn from the records of 11,201 adults, and this document takes its equation as given. The same measurement on other people's records

reproduces it between 100 and 200 mg/dL. Above 200 mg/dL the records here are flatter than the published fit by one measure and agree with it by the other.

The simulations describe how the measurement behaves on constructed records. They say nothing about what the truth is in people, and they are reported so that a reader can weigh them.

Two questions remain for this series. One is the scale for each person, taken up in the next section. The other is whether a loop that doses to the equation does better than one dosing to either existing model or to a flat profile. The third stage of the evaluation below ran that comparison in simulation, and it has not been tested in people.

A proposal from June 2026 that used this shape (document 3) was withdrawn for live dosing on 16 September 2026. The audit of that date left the way open for such candidates to be simulated and replayed in shadow, and to be studied prospectively. The evaluation proposed below stays inside that.

How could it be implemented?

What it would replace

AndroidAPS and Trio run the same logarithmic equation:

sensitivity = 1800 ÷ (adjustment factor × TDD × ln(glucose ÷ divisor + 1))

Trio computes a ratio, profile ISF × adjustment factor × TDD × ln(glucose ÷ divisor + 1) ÷ 1800, and divides the profile ISF by it, so the profile value cancels. Given the same inputs the two systems return the same number. This was checked on 11 November 2026 against AndroidAPS dev 0c593a4b09 and Trio dev 902be75e8, both of 10 November 2026. Neither calculation differs from the released versions (AndroidAPS master f77abd9b82, Trio main 0b29cd534), and Trio's `DynamicISF.swift` was last changed on 24 August 2026.

What separates the two in use is the defaults around the equation. Trio's adjustment factor defaults to 0.8 where AndroidAPS defaults to 100%, which puts Trio's value 25% higher. Trio then holds its ratio inside the autosens limits, 0.7 and 1.2 by default, and `calculateRawDynIsf` in AndroidAPS applies no limit. For the worked person Trio's upper limit binds from 157 mg/dL. Above that glucose its value stays at 41.7 mg/dL per unit while AndroidAPS continues down to 29.3 at 200 mg/dL. The daily dose is blended

differently in each system, and the divisors differ for faster insulins: 55, 65 or 75 by peak time in AndroidAPS, and 55 or 70 in Trio, or 120 less a custom peak time.

Trio's sigmoid mode returns the profile ISF at the profile target and moves it within the same limits. oref0 as published has autosens and no dynamic ISF based on daily dose. Loop and iAPS were not read.

Linked to the profile ISF, or to TDD?

The published values need a scale for each person. At 140 mg/dL the equation gives 50.9 mg/dL per unit to a person on 20 U a day and to a person on 80 U a day. Among 422 adults with entered pump settings analysed earlier in this series, sensitivity multiplied by daily dose was 1915 (95% CI 1828 to 1986), which puts the first of those people near 96 mg/dL per unit and the second near 24.

The approach is to titrate the curve to the person. Their entered setting is taken to be valid at one glucose, the anchor, and the whole published curve is multiplied by the one number that makes it pass through that value. The shape is left as published. The curve moves up or down for each person, and the ratio between its values at any two glucose levels stays as published.

The anchor in this proposal is 140 mg/dL. That is a modelling choice, made as the initial proposal, and nothing in the publication or in the analyses here selects it. The publication's own reference point is 180 mg/dL, the glucose against which it states its halving and doubling of doses, and earlier drafts of this document anchored there. The published value at 140 mg/dL is 1.44 times the value at 180. Anchoring at 140 therefore gives every person a sensitivity 1.44 times lower at every glucose than anchoring at 180 would, and a dose 1.44 times larger.

Linked to the profile, the entered ISF is taken to hold at 140 mg/dL:

sensitivity = profile ISF × (published value at this glucose ÷ published value at 140 mg/dL)

This fits Trio's structure, which already expresses dynamic ISF as a ratio of the profile. The schedule through the day is kept and no level has to be chosen. The value does not respond to recent dosing. It is also only as good as the entered number. It suits a profile ISF the person has evaluated, and it carries the whole of any error in one they have not.

Linked to TDD, a person’s sensitivity at 140 mg/dL is taken to be 1915 ÷ TDD:

sensitivity = (1915 ÷ TDD) × (published value at this glucose ÷ published value at 140 mg/dL)

This is the logarithmic model with its glucose term swapped. The daily dose is each system’s existing blend and the adjustment factor multiplies it as now. The divisor for insulin type has no counterpart and drops out. The value keeps moving with recent insulin use by the same multiplier as today: 1.39 for a person whose blended dose is 0.72 of their own average and 0.88 at 1.14 of it, which are the tenth and ninetieth percentiles among 960 people in document 6. The profile ISF and its schedule through the day play no part, as in AndroidAPS now. The adult figure of 1915 is a modelling choice. The figure pooled across all ages in the same analysis is higher, and 51% of that group are children, so it was not used.

The two links agree for anyone whose profile ISF multiplied by daily dose is 1915. They differ for everyone else, and that difference was used to test them.

Which link corrections followed

The test (INV-015) used REPLACE-BG, the one cohort here in which a correction was the person’s own dose and no algorithm added to it. A correction counted when it was given at 150 to 350 mg/dL and was at least 0.5 U, with no carbohydrate entered for six hours either side and no further bolus for six hours. That left 3,433 corrections from 150 adults who each had five or more. Among them the entered ISF multiplied by daily dose had a median of 1929. A tenth of people sat below 1345 and a tenth above 2997, and that spread is what separates the links.

For each person the fall in glucose over six hours, per unit given, was divided by the published shape at the starting glucose, and the median was taken. That figure is a ratio of outcomes and only its ordering of people was used. It was related to daily dose and to the entered ISF together, in logarithms. The expectation written before the run was that the result would lean towards the profile.

	Coefficient on daily dose	Coefficient on entered ISF
The link to TDD predicts	-1	0
The link to the profile predicts	0	1
Measured (95% interval)	-0.07 (-0.43 to 0.26)	0.95 (0.55 to 1.32)

Both measured intervals include what the link to the profile predicts and exclude what the link to TDD predicts.

The second table puts the same people in four groups of equal size by their entered ISF multiplied by daily dose. The first group entered an ISF well below what their daily dose would suggest, and the last group one well above it. The last column is the fall achieved per unit as a multiple of the entered ISF.


Entered ISF × daily dose	People	Start, mg/dL	Target	At six hours	Achieved ÷ entered ISF (95% interval)
900–1665	38	209	106	135	1.38 (1.23 to 1.76)
1665–1929	37	240	110	130	1.76 (1.61 to 1.92)
1929–2322	37	234	110	142	1.50 (1.29 to 1.68)
2322–5211	38	229	107	138	1.47 (1.40 to 1.63)

All four groups stood at 130 to 142 mg/dL six hours after a correction. The last column is 1.38 in the first group and 1.47 in the last, with the second group highest. Under the link to TDD it would fall as the product rises.

The comparison between groups does not depend on the anchor, and nor do the coefficients above. The level of the column does. With the anchor at 140 mg/dL all four intervals lie above 1, so the six-hour fall per unit was 1.38 to 1.76 times the sensitivity the candidate assigns at the starting glucose. With the anchor at 180 mg/dL the same column read 0.96 to 1.22. The fall includes basal and earlier doses acting, so the column overstates what the correction did and cannot place the anchor. It is the one figure in this document in which the anchor meets a measured fall, and it is reported for that reason.

The test has a known lean. A correction is sized from the entered ISF, so any part of the fall that does not depend on the dose makes the entered ISF look like the better scale. Falls were therefore constructed for the same people and doses with each link as the truth, and the test was run on them. The two cases in which the fall follows the distance from target were added after the first run had returned a coefficient near 1 on the entered ISF. With the link to TDD as the truth and nothing else acting it returned -0.97 on daily dose and -0.05 on the entered ISF, close to that link's prediction. With 46% of the fall following the distance from target whatever the dose, it returned -0.54 (-0.68 to -0.42) and 0.37 (0.22 to 0.51), and both values lie outside the measured intervals. The 46% is the fall on nights with no bolus as a share of the fall per

unit on correction nights in this cohort (INV-009). At 75%, which is a modelling choice with no measurement behind it, the constructed values came inside the measured intervals. With the link to the profile as the truth, the test returned 0.88 to 0.97 on the entered ISF in every case. The link to TDD is therefore excluded as the scale between people unless about three quarters of the six-hour fall was independent of the dose.




The measured coefficients against what each link predicts. The small points are the constructed cases, in which the true link is known.

The result has limits. It comes from 150 adults in one trial. It compares people with each other, so it does not test whether one person's sensitivity moves with their recent dosing, which is what the link to TDD adds within a person. Document 6 measured that movement at an exponent of -0.645 (standard error 0.033) across 1,313 people. People using a loop were not tested, because in their records the algorithm doses alongside the correction, and children were not tested. The 46% was measured overnight and the corrections were given at any hour.

Why the curve is multiplied, and why one shape is used for everyone


The curve could be attached to the profile ISF by addition. Every published value would then be raised or lowered by the difference between the profile ISF and the published 50.9 mg/dL per unit at 140 mg/dL. Adding keeps the

published drop in mg/dL per unit between any two glucose levels, where multiplying keeps the published ratio. The two forms were compared on the 3,433 REPLACE-BG corrections of the previous test (INV-017). The added form gave a sensitivity of zero or below at 299 of them, in 32 of the 150 people, and it reaches zero inside the published range for any profile ISF below 35.6 mg/dL per unit. People were then put in four groups by entered ISF, and the fall achieved per unit was divided by what each form predicted. A form that attaches the curve correctly gives the same figure in every group. Between the lowest and the highest group the figure differed by a factor of 1.53 (95% interval 0.81 to 2.89) for the multiplied form and 6.88 (3.05 to 17.78) for the added one. The first interval includes 1 and the second lies above it. Falls constructed with the added form as the truth came inside the measured intervals only when three quarters of the six-hour fall was taken to be independent of the dose, which is the same reservation as in the test of the links.




The published curve attached to four profile ISFs, by multiplying and by adding a constant.

A further possibility is a curve measured for each person from their own events. The publication’s measurement was run person by person (INV-018), first on the 196 adults of REPLACE-BG, who had a median of 448 events each. In thirds of entered ISF, their curves kept 0.58, 0.56 and 0.61 of the reading at 140 mg/dL by 250 mg/dL, and the share in the highest third was 1.06 (0.70 to 1.45) times the share in the lowest. The drop in mg/dL per unit over the same span was 2.32 (1.23 to 5.51) times as large in the highest third as in the lowest, where curves that differ by an added constant give 1. People with different entered settings therefore had curves of one shape at different heights.



The publication's measurement run on each of 196 adults in REPLACE-BG, as measured and with each curve divided by its own reading at 140 mg/dL.

Each person's events were then divided into two halves and a curve was built from each half. The height repeated: in REPLACE-BG the reading at 140 mg/dL had a rank correlation of 0.74 (0.66 to 0.82) between the halves, and in the seven other cohorts it was 0.78 to 0.95. The shape was tested by prediction. A person's share kept at 190 and at 250 mg/dL in one half was predicted from their own other half and from the median person's, and the second miss was divided by the first, so that a figure above 1 favours the person's own shape. Across the six cohorts with enough people it was 0.87 to 1.01 at 190 mg/dL and 0.96 to 1.03 at 250, and no interval lay above 1. Longer records did not change that. Among 38 people using Loop with at least 1,600 events each, the figure was 1.06 (0.93 to 1.20) at 190 mg/dL, the highest in the run. Trio and oref0 had too few people for the comparison of shapes.



Each REPLACE-BG adult's curve built from two halves of their events. The height at 140 mg/dL repeats between the halves, and the share kept at 250 mg/dL repeats weakly.


The height of a person’s measured curve is not used either. Its reading at 140 mg/dL was 2.5 to 2.9 times the entered ISF in every third, because the measurement credits the whole fall to insulin above scheduled basal and leaves out the events with no net insulin used, which were 30% of the median person’s events at that glucose. The third of people with the lowest settings entered a median of 30 mg/dL per unit and read 69 by the measurement. A correction from 200 to 100 mg/dL sized on 69 would be 1.4 U, against 3.3 U on the entered setting. The candidate therefore takes its height from the profile ISF and its shape from the publication.

One person under each model

The worked person has a profile ISF of 50 mg/dL per unit, a daily dose of 40 U, rapid-acting insulin and recent dosing equal to their average. Their product is 2,000, which is why the two links give them values about 4% apart. The table gives the sensitivity each method assigns them, in mg/dL per unit, with both systems at their defaults.

Glucose, mg/dL	AndroidAPS, divisor 55	Trio logarithmic	Trio sigmoid	Equation linked to TDD	Equation linked to profile
80	50.1	62.6	53.7	75.9	79.3
100	43.4	54.3	50.0	75.9	79.3
120	38.9	48.6	47.1	59.8	62.5
140	35.6	44.4	45.0	47.9	50.0
150	34.2	42.8	44.3	43.2	45.1
180	31.0	41.7	42.9	33.2	34.7
200	29.3	41.7	42.4	29.0	30.3
250	26.3	41.7	41.8	23.3	24.4
300	24.1	41.7	41.7	19.9	20.8
380	21.8	41.7	41.7	14.4	15.0

Against AndroidAPS at 100% the equation linked to the profile gives this person less insulin below 211 mg/dL with a divisor of 55, and below 168 mg/dL with a divisor of 75, and more above. The ratio between the two sensitivities is 1.83 at 100 mg/dL, 1.32 at 150, 1.03 at 200 and 0.86 at 300. Against Trio at its defaults it gives less insulin below 158 mg/dL in logarithmic mode and below 152 in sigmoid mode, and more above. Against the person’s own profile it gives less insulin below 140 mg/dL and more above, which holds for everyone. The comparison with AndroidAPS holds for this person only, since AndroidAPS scales by daily dose and the candidate by the profile.



The candidate for three profile ISFs, and the worked person under each model with the band Trio's default limits allow.

The published fit stops at both ends, and the reference implementation (`inv013/candidate.py`) makes a choice at each. Below 100 mg/dL it holds the value reached at 100. The published quadratic falls there, towards 65 mg/dL per unit at 60 mg/dL, and following it would give more insulin as glucose drops below 100. Work in this series in March 2026 declined the quadratic for dosing on that ground. Above 380 mg/dL, where the fit ends, the value at 380 is held.

This document carries the link to the profile as the candidate, on the result above, and the link to TDD as an arm of the evaluation.

The ends of the curve and Trio's limits

Trio holds the ratio of the profile ISF to the dynamic value inside the autosens limits, 0.7 and 1.2 by default. The candidate is the profile ISF multiplied by the curve's shape, so the limits act on the shape itself: it may run from 0.83 to 1.43 of its value at 140 mg/dL. The published shape runs from 0.30 to 1.59. It lies inside the limits between 109 and 158 mg/dL and outside them over the rest of its range. Those two glucose values are the same for every person on the profile link, whatever their profile ISF.

The table gives the worked person's sensitivity from the curve, and the value Trio's default limits would pass on to the dose, in mg/dL per unit.

Glucose, mg/dL	Curve linked to profile	With Trio's default limits	Held by
100	79.3	71.4	Lower limit, 0.7
120	62.5	62.5	
140	50.0	50.0	

Glucose, mg/dL	Curve linked to profile	With Trio's default limits	Held by
150	45.1	45.1	
180	34.7	41.7	Upper limit, 1.2
200	30.3	41.7	Upper limit, 1.2
250	24.4	41.7	Upper limit, 1.2
300	20.8	41.7	Upper limit, 1.2
380	15.0	41.7	Upper limit, 1.2

The two ends are affected unequally. Near target the curve withholds slightly more insulin than the limits allow: at 100 mg/dL it asks for 0.63 of what the profile would, and the lower limit holds that at 0.7. At high glucose the gap is wide. The curve asks for 2.41 times the profile's insulin at 300 mg/dL and 3.33 times at 380, where the upper limit holds 1.2. A correction from 300 mg/dL to 100 is 9.62 U from the curve and 4.80 U inside the limits.

With the limits as they are, the model in Trio would follow the curve between 109 and 158 mg/dL and give 1.2 times the profile's insulin at every glucose above that, so the published curve would not be what doses over most of its range. Running the curve as published needs limits for this mode of about 0.63 and 3.33, or no limits on it, and the upper figure is the one that carries risk, since it more than triples a correction at the top of the range. Which of these Trio should do is a question for its maintainers. The third stage of the evaluation ran the candidate with and without the default limits, in simulation. Against the unchanged profile, the full curve raised time below 54 mg/dL by 0.43 percentage points (0.24 to 0.61) and the curve inside the limits by 0.13 (-0.02 to 0.27). Their changes in time above 250 mg/dL were -0.35 points (-0.72 to 0.02) and -0.14 (-0.40 to 0.16). `calculateRawDynIsf` in AndroidAPS applies no limit, so the whole curve would reach the dose there, and the same question about the top of the range applies to it.

The level of the TDD arm

The link to the profile needs no level, because the entered ISF supplies it. The link to TDD needs one, and the publication gives nothing to set it from. One way to compare the possible levels is the daily dose at which the published values would apply unchanged, called the reference dose here.

Basis for the level	People	Sensitivity × daily dose at 200 mg/dL (95% interval)	Reference dose, U
Adult entered settings, taken to hold at 140 mg/dL (the TDD arm)	422	1,161 (1,108 to 1,204)	37.7
Adult entered settings, taken to hold at 180 mg/dL (the publication's reference)	422	1,674 (1,598 to 1,736)	54.3
The 1800 rule, taken to hold at 105 mg/dL (this series, March 2026)		731	23.7
Measured under Loop	842	2,182 (2,065 to 2,266)	70.8
Measured under Trio	22	2,273 (1,569 to 3,218)	73.7
Measured under oref0	86	2,398 (1,779 to 5,809)	77.8
Measured with no algorithm dosing (REPLACE-BG)	196	3,283 (2,898 to 3,854)	106.5

The first row takes a person's entered sensitivity to be right at 140 mg/dL, and the second at 180. The third takes the conventional rule to be right near target, and it puts the whole curve lower again. The last four rows are the measured values from the previous section, read at 200 mg/dL. All four intervals lie above the arm's 1,161. Trio's, from 22 people, includes the 1,674 of the second row, and the other three lie above that as well. How far the measured rows sit from a true level is not known. At 200 mg/dL the same measurement on simulated people read between 0.62 and 2.4 times the truth across the seven regimes that dosed in response to glucose, and daily dose was not defined the same way in each source.

No default can be chosen from these figures. The reference doses run from 24 to 107 U. The measurement cannot narrow that range alone, given how far it sat from the truth in simulation, so the arm's level is left to the fourth stage of the evaluation, which gives each person a starting adjustment factor. For

any one person the adjustment factor moves the level in either direction, as it does now. Its guidance has to be worked out again for this arm, because at 100% the arm's value at 100 mg/dL is 1.75 times the existing one.

Correction doses at the current glucose

The equation is applied at the current glucose. The existing code applies the logarithmic model the same way, at the current glucose or at a glucose chosen by one of three branches in `DetermineBasalSMB.kt`, and the candidate takes whichever glucose that code already passes. The publication's interval from peak to minimum is how the variation of sensitivity with glucose was measured. It plays no part in how the equation is used.

The table gives the units needed to bring the worked person to 100 mg/dL, with the equation linked to the profile. It is the arithmetic of the sensitivity alone, before insulin on board and the loop's predictions change it.

Current glucose, mg/dL	AndroidAPS, divisor 55	Trio logarithmic	Profile	Equation, profile link
150	1.46	1.17	1.00	1.11
200	3.41	2.40	2.00	3.30
250	5.71	3.60	3.00	6.15
300	8.29	4.80	4.00	9.62

At 150 mg/dL the equation asks for 1.11 U, close to the 1.00 U of the flat profile. At 300 mg/dL it asks for 9.62 U, against 4.00 U from the profile and 8.29 U from AndroidAPS at 100%.

The anchor sets how much insulin the candidate asks for overall. Each arm was replayed on the recorded loop decisions of 138 people (INV-016), of whom 22 used Trio, 19 AndroidAPS and 97 oref0. The insulin asked for at every reading above 100 mg/dL was summed and set against the same sum under the profile ISF alone. The candidate asked for 1.39 (95% interval 1.28 to 1.42), 1.38 (1.29 to 1.47) and 1.44 (1.39 to 1.51) times the profile's correction insulin in those cohorts. By glucose at the reading, it asked for 0.82 to 0.83 times the profile's insulin between 100 and 150 mg/dL, 1.33 times between 150 and 200, 1.83 to 1.84 times between 200 and 260, and 2.31 to 2.37 times above 260. An anchor of 180 mg/dL lowers every one of those figures by the factor of 1.44. The replay shows what each arm would have asked for at readings recorded under another algorithm, and says nothing about where glucose would then have gone.

What changes in the code

The glucose function is the smallest part. One function per system returns the published quartic divided by its value at 140 mg/dL, with test vectors shared between the two codebases so that both return the same values.

Trio's limits need a decision from its maintainers, as set out above: with the defaults unchanged the curve reaches the dose only between 109 and 158 mg/dL.

The forecasts are affected as well as the dose. Both systems calculate a sensitivity at each predicted glucose, and the call sites can stay as they are. The equation's value rises 2.61 times as predicted glucose falls from 200 to 100 mg/dL, against 1.48 for the logarithmic model, so a forecast heading towards target is computed with a higher sensitivity than now. The effect on minimum predicted glucose and on time spent at zero basal has not been measured and is part of the replay.

The remaining additions are smaller. The loop result already stores the sensitivity it used (`variableSens` in AndroidAPS), and during evaluation both the existing and the new value are stored while the existing one doses. Settings gain a mode selector and, in a first version, no new number for the person to set. The first version also leaves out any adjustment for recent dosing. The code is written against the standard AndroidAPS and Trio repositories, so that builds derived from them inherit it. Each system gets unit tests on the shared reference vectors and replay assets for the new mode, alongside its existing tests (`ReplayApsResultsTest` in AndroidAPS, `DynamicISFTests` in Trio). A guard test confirms that the existing modes' output is unchanged.

The shadow part was written on 13 November 2026, on the development branches of nightscout/AndroidAPS and nightscout/Trio. In each system one function returns the candidate from the current glucose and the profile ISF, and the value is stored with the loop's decision. Nothing reads it back, so the insulin delivered and the forecasts are as they were. AndroidAPS stores it as a new field of the loop result (`curve_sens`), which the SMB plugin fills and the other algorithm plugins leave empty. Trio writes it into the reason text of each determination, in the form `Curve ISF: 55→49.6`. The reason text already reaches Trio's database and Nightscout and is shown on the loop status screen, so nothing in the stored schema changed. For a person who uses mmol/L, Trio converts the new tag as it converts the existing ISF tag, and the

value is then kept to one decimal place. The profile ISF is the profile's own value in both systems, with the profile percentage applied in AndroidAPS and any active override applied in Trio.

The two implementations and the Python reference are tested against one table of ten glucose values. Each device test was confirmed to fail when the upper end of the range was moved from 380 to 400 mg/dL. AndroidAPS's unit tests passed with the change (14,562 tests) and so did Trio's (998 tests). Trio keeps 14 recorded decisions that its tests compare character by character. The 13 that carry a full reason were recorded again, and each differs from the earlier file only by the new tag. A fresh-context review of both changes found no route from the new value to a dose. The code has no mode selector, and it has not been built for or run on a device. It has not been published or offered to either project.

In AndroidAPS the code touched is `calculateRawDynIsf` and the preference screen, together with the per-prediction lines and the dose-sensitivity branches in `DetermineBasalSMB.kt`. In Trio it is `DynamicISF.swift` and the forecast adjustment used in logarithmic mode. The link to the profile brings the profile ISF into the AndroidAPS calculation, which does not use it now.

Evaluation

Seven arms are compared throughout. Three are references. The logarithmic model as shipped and Trio's sigmoid are the two existing models, and the profile alone is the baseline for both. Three arms are forms of the equation, each applied at the current glucose. One is linked to the profile and is the candidate. The other two are linked to TDD and to a plain average daily dose. The seventh is the first draft of this document, a sensitivity anchored to the profile and flat above 100 mg/dL. It is the arm the equation has to beat for its shape to matter.

The first stage replays each arm's value on recorded histories using the INV-008 pipeline. Results are split by glucose and by time since the last meal, and reported separately for people who announce meals and people who do not. The stage shows how much insulin each arm asks for and in whom. It is descriptive and has no pass rule. It was run on 138 people as INV-016, and its totals for the candidate are quoted above.

The second stage runs the publication's measurement (`inv012/peak_trough.py`) on AndroidAPS and Trio records under the published event rules. An earlier draft had this stage decide the default level of the link to TDD. The simulation with a measured truth has since covered 100 to 250 mg/dL,

and the measurement sat too far from the truth there to set one. What remains for the simulation is 300 mg/dL and a truth measured inside the free-living runs.

The third stage is simulation in TimSim, with runs paired by seed. It was run as INV-019 and is reported at the end of this section. The candidate was run with Trio's default limits and without them, and with the anchor at 140 and at 180 mg/dL. It was also run on a mis-set profile ISF. A harm rule was written before the runs and has two sides, time below 54 mg/dL and time above 250 mg/dL. Both are needed. For the worked person this model gives less insulin than AndroidAPS at 100% below 211 mg/dL and more above it, and at 380 mg/dL it asks for 3.33 times the profile's insulin. The arm for unannounced meals was left for a second run, as were the TDD arm with a mis-set adjustment factor and the arm on a plain average daily dose.

A harm reading in the simulation is reported and looked into, and it does not by itself stop the candidate. The simulated body differs from the equation's premise, since its true sensitivity rises with glucose between 100 and 300 mg/dL, and its titrated profiles sit well below the level that adults enter.

The fourth stage is shadow running. Both values are calculated on the device each cycle and stored, and the existing one doses. That gives the distribution of the ratio between them for each person, and for the TDD arm a starting adjustment factor. The code for this stage has been written, as described under "What changes in the code", and no shadow records exist yet.

The fifth stage is use, on the developer's own device first and on released lines after that. It follows only when the third stage has been reported and the fourth has passed, and it stops on the harm rule.

The first two stages are analysis and carry the analysis checklist. The last two change code that can change a dose. They carry unit and guard tests, the repository's review checklist, a replay against real history and a fresh-context review of the diff.

The way back at every stage is the mode selector. The existing modes stay in the code unchanged and remain the default until the fifth stage has run, and any build that includes the new mode names the build before it.

The third stage in simulation

The third stage was run on 12 November 2026 (INV-019). Each arm was the shipped oref0 algorithm run by TimSim's adapter, with the sensitivity it is handed every five minutes replaced by the arm's rule at the current sensor

glucose. The 40 simulated adults, their meals, the sensor noise and their titrated settings were the same in every arm, so a difference between two arms is the rule. Every arm with a rule ran with autosens held at 1.0, which is what AndroidAPS does in dynamic mode at its default settings. A rule switched on once seven days of delivery were held, as AndroidAPS requires, and outcomes were read on true glucose from day 8 to day 35. The arms and the margins of the harm rule were committed before the run, with the expected sign of each result (`inv019/outcome_test.py`). There were sixteen arms and 640 runs, and none raised an error. Twelve runs ended early through hypoglycaemia and none through hyperglycaemia.

The checks were read before any result. Week one, before any rule had switched on, differed from the reference by 0.000 mg/dL of mean glucose in every arm that shares its profile. Each rule reached its runs, as the second column of the table below shows. The null arm was a profile ISF multiplied by 1.01, which shows how far two runs drift apart when nothing of size is changed. Every one of its intervals included zero and it read cleared on both sides of the harm rule, so both sides could be judged with a panel of this size. Median glucose across the arms ran from 123 to 138 mg/dL, with 5th percentiles of 75 to 91.


The panel differs in two ways from the people the rule is written for. Its titrated profile ISF multiplied by daily dose has a median of 988, about half of the 1,915 that adults enter. The arms scaled by daily dose therefore dosed these people more weakly than their profiles did, and an arm was added with the logarithmic model's factor set per person so that it returns the profile ISF at 140 mg/dL. That arm removes the difference of level and leaves the difference of shape. The other difference is in the simulated body. Its true sensitivity rises with glucose between 100 and 300 mg/dL (INV-014), and the equation's falls.

The table gives each arm against the same people on their unchanged profile ISF. Differences are means over the people who completed both runs, with 95% intervals from resampling people, and those in time are percentage points.

Arm	Sensitivity applied ÷ profile ISF	Time below 54 mg/dL	Time above 250 mg/dL	Time in range	Mean glucose, mg/dL
Candidate (profile ISF at 140 mg/dL)	1.12	+0.43 (+0.24 to +0.61)	-0.35 (-0.72 to +0.02)	+0.15 (-0.74 to +1.07)	-1.77 (-2.86 to -0.70)

Arm	Sensitivity applied ÷ profile ISF	Time below 54 mg/dL	Time above 250 mg/dL	Time in range	Mean glucose, mg/dL
Candidate inside Trio's limits, 0.7 to 1.2	1.12	+0.13 (-0.02 to +0.27)	-0.14 (-0.40 to +0.16)	+0.90 (+0.11 to +1.69)	-1.16 (-1.93 to -0.37)
Candidate, profile ISF at 180 mg/dL	1.53	-0.01 (-0.13 to +0.12)	-0.14 (-0.41 to +0.15)	-0.33 (-1.03 to +0.34)	+3.06 (+2.40 to +3.72)
Equation linked to TDD	2.26	-0.09 (-0.29 to +0.09)	+0.05 (-0.32 to +0.40)	-1.89 (-2.94 to -0.91)	+8.87 (+6.77 to +10.96)
First draft: profile above 100 mg/dL, curve below	1.00	+0.03 (-0.04 to +0.11)	+0.05 (-0.16 to +0.27)	-0.20 (-0.74 to +0.34)	+0.57 (+0.08 to +1.15)
oref0 with autosens, no dynamic rule	1.00	+0.02 (-0.09 to +0.11)	-0.03 (-0.22 to +0.18)	+0.38 (-0.28 to +1.01)	-0.75 (-1.43 to -0.02)
Logarithmic, AndroidAPS at 100%	1.62	-0.02 (-0.22 to +0.18)	-0.31 (-0.62 to +0.02)	-0.26 (-1.05 to +0.55)	+4.85 (+2.78 to +6.89)
Logarithmic, matched to the profile at 140	0.98	+0.23 (+0.09 to +0.37)	-0.37 (-0.63 to -0.12)	+0.72 (-0.11 to +1.63)	-1.36 (-2.39 to -0.31)
Logarithmic, Trio (factor 0.8, limits)	1.43	-0.17 (-0.32 to -0.03)	-0.06 (-0.31 to +0.21)	-0.18 (-1.03 to +0.69)	+4.35 (+3.39 to +5.35)
Trio sigmoid, defaults	0.95	+0.16 (+0.05 to +0.27)	-0.19 (-0.42 to +0.06)	+0.33 (-0.40 to +1.04)	-1.55 (-2.31 to -0.71)
Placebo: profile ISF x 1.01	1.00	+0.03 (-0.08 to +0.13)	+0.09 (-0.10 to +0.29)	-0.07 (-0.79 to +0.61)	+0.42 (-0.16 to +1.06)
Profile ISF entered 30% too low	1.00	+0.39 (+0.22 to +0.56)	-0.36 (-0.69 to -0.04)	+1.00 (+0.09 to +1.91)	-5.95 (-6.90 to -4.95)
Profile ISF entered 40% too high	1.00	-0.17 (-0.31 to -0.05)	+0.32 (-0.03 to +0.69)	-1.45 (-2.53 to -0.36)	+5.81 (+4.89 to +6.72)


The candidate lowered mean glucose by 1.77 mg/dL (0.70 to 2.86) and time above 180 mg/dL by 1.03 points (0.20 to 1.89). Its changes in time above 250 mg/dL and in time in range had intervals that include zero. Time below 54 mg/dL rose by 0.43 points (0.24 to 0.61). The median person spent 0.22% of time there on the unchanged profile and 0.81% under the candidate. A profile ISF entered 30% too low raised it by 0.39 points (0.22 to 0.56), so on the low side the candidate cost about what that mis-setting did.



Each rule against the same simulated people on their unchanged profile ISF, with 95% intervals. Results from TimSim are relative.

The arms that change one part of the candidate indicate where the extra time below 54 mg/dL came from. The first draft's form follows the curve below 100 mg/dL and the profile above it, and it changed time below 54 mg/dL by +0.03 points (-0.04 to +0.11), so the easing near target was not the source. Inside Trio's limits the candidate stops at 1.2 times the profile's insulin, and the rise was 0.13 points (-0.02 to 0.27). Anchored at 180 mg/dL the rule asks for more insulin than the profile only above 180, and the change was -0.01 (-0.13 to +0.12) with mean glucose 3.06 mg/dL higher (2.40 to 3.72). The rise therefore went with how much more insulin than the profile the rule asked for above target.

The lows were then traced in the raw runs (`inv019/low_trace.py`), with the expected result written down first. Each period below 54 mg/dL was filed under the highest true glucose in the four hours before it began, and four hours is a choice. The candidate's runs held 5.4 such periods per person, against 2.8 on the unchanged profile. Of the 0.43 extra points, 0.24 (0.14 to 0.35) followed a peak between 180 and 250 mg/dL and 0.06 (-0.04 to 0.18) a peak above 250 mg/dL, which together is about seven tenths. Lows that followed a peak below 140 mg/dL added 0.02 points (0.00 to 0.05). The placebo's interval included zero in every band.



Time below 54 mg/dL in three arms, split by the highest glucose in the four hours before each low began. Each arm is compared with the same simulated people on their unchanged profile ISF.

In the four hours before a low that followed a peak of 180 to 250 mg/dL, the candidate's run had delivered a median of 11.7 U. The run on the unchanged profile delivered 9.0 U to the same person over the same hours. Across people the candidate delivered 1.90 U more (1.02 to 2.95) before its lows. The matched logarithmic model showed the same pattern at a smaller size, with 0.16 points (0.06 to 0.27) after a peak of 180 to 250 mg/dL. The extra time below 54 mg/dL was therefore the tail of corrections given at high glucose. The trace covers the first simulated month only.

Among the rules run on the unchanged profile, only the candidate inside Trio's limits had an interval for time in range that lay above zero, at +0.90 points (0.11 to 1.69). The profile ISF entered 30% too low did the same, at +1.00 (0.09 to 1.91).

The existing models moved the same way where they dosed at the profile's level. The matched logarithmic model raised time below 54 mg/dL by 0.23 points (0.09 to 0.37) and Trio's sigmoid by 0.16 (0.05 to 0.27). The logarithmic model at AndroidAPS's 100% and at Trio's defaults dosed more weakly than the profile on this panel, and mean glucose rose by 4.85 mg/dL (2.78 to 6.89) and 4.35 (3.39 to 5.35). The equation linked to TDD applied 2.26 times the profile's sensitivity and raised mean glucose by 8.87 mg/dL (6.77 to 10.96). That figure describes the level of this panel, and it does not test the link in adults whose product is near 1,915.

On a profile ISF entered 30% too low, the candidate added 0.20 points of time below 54 mg/dL (-0.01 to 0.43) to what the mis-setting did alone, and four of its 40 runs ended early through hypoglycaemia against one for the mis-set profile alone. On a profile entered 40% too high it added 0.19 points (0.11 to 0.28) and lowered time above 250 mg/dL by 0.60 points (0.29 to 0.92).

The harm rule compares the candidate with two arms. The matched logarithmic model has the candidate's level at 140 mg/dL and differs in shape. oref0 with autosens is what a person has with dynamic sensitivity switched off. The margins are 0.25 points of time below 54 mg/dL and 1.25 points above 250, a quarter of the consensus limits of 1% and 5% [6], and the fraction is a choice. A side reads harm when the interval lies wholly above zero and the mean is above the margin. It reads cleared when the upper bound is below the margin, and unresolved otherwise. The logarithmic model at 100% is shown beside these and is outside the rule, because on this panel it doses at a different level.

Arm	Against	Side, mg/dL	People	Difference, points of time	Margin	Reading
Placebo	Unchanged profile	Below 54	39	+0.03 (-0.08 to +0.13)	0.25	Cleared
Placebo	Unchanged profile	Above 250	39	+0.09 (-0.10 to +0.29)	1.25	Cleared
Candidate	Logarithmic, matched at 140	Below 54	39	+0.25 (+0.09 to +0.42)	0.25	Unresolved
Candidate	Logarithmic, matched at 140	Above 250	39	+0.03 (-0.24 to +0.31)	1.25	Cleared
Candidate	oref0 with autosens	Below 54	40	+0.41 (+0.26 to +0.59)	0.25	Harm
Candidate	oref0 with autosens	Above 250	40	-0.36 (-0.80 to +0.04)	1.25	Cleared
Candidate	Logarithmic, AndroidAPS at 100%	Below 54	40	+0.46 (+0.29 to +0.63)	0.25	Harm
Candidate	Logarithmic, AndroidAPS at 100%	Above 250	40	-0.04 (-0.27 to +0.21)	1.25	Cleared

On the low side the candidate read harm against oref0 with autosens, at +0.41 points (0.26 to 0.59). Against the matched logarithmic model it read unresolved at +0.25 (0.09 to 0.42), an interval that lies above zero with its

mean at the margin. On the high side it read cleared against both. No comparison met the separate criterion on runs that ended early. The largest rise in one person’s time below 54 mg/dL was 1.97 points.

Two of the signs written down before the run did not hold. Time above 250 mg/dL was expected to be lower under the candidate than under the matched logarithmic model, and the difference was +0.03 points (-0.24 to +0.31). It was expected to be higher under the logarithmic model at 100% than on the profile, and it was -0.31 (-0.62 to +0.02), with mean glucose higher as expected. The second is unexplained. The low side had been expected to read harm or unresolved.

Two properties of the simulator could produce the low-side reading without the same holding in people. The simulated body responds more to a unit of insulin as glucose rises, so a rule that gives more insulin at high glucose gives it where each unit already does most. The equation was measured in people in whom, by its own account, a unit does less there. The titrated profiles also sit at about half the level adults enter for their daily dose, and how that moves the result was not tested. Neither property was shown to be the cause.

The run has further limits. The vendored oref0 takes one sensitivity per cycle. AndroidAPS in dynamic mode also uses the value along its forecast, and that use was not exercised. The sixteen arms were run on one simulated month, and five of them on a second, as described below. TimSim orders options and finds failures, and the sizes above are not estimates for people.

What follows was fixed before the run. The harm reading is looked into, and the candidate carries on to the fourth stage. The unresolved reading against the matched model called for a larger panel or a second seed. Shadow running, in which the device calculates the value and doses nothing from it, is unaffected by either reading.

A second simulated month

The second seed was run on 13 November 2026. Its plan was committed before it ran, with the first run’s results in hand. Five arms were rerun on TimSim’s second measured month with the same people and settings: the candidate, its two comparators, the unchanged profile and the placebo. The meals, the sensor noise, the infusion-site failures and the day-to-day sensitivity differ between the two months, and the people do not. A person’s difference was taken as the mean over the months in which both runs completed, and the harm rule was applied to those means with its margins

unchanged. Combining the months in that way is a choice. None of the 200 runs raised an error. Week one differed by 0.000 mg/dL between the arms that share a profile, and the placebo read cleared on both sides.

Arm	Against	Side, mg/dL	People	Difference, points of time	Margin	Reading
Placebo	Unchanged profile	Below 54	40	-0.02 (-0.11 to +0.06)	0.25	Cleared
Placebo	Unchanged profile	Above 250	40	-0.02 (-0.15 to +0.11)	1.25	Cleared
Candidate	Logarithmic, matched at 140	Below 54	39	+0.17 (+0.04 to +0.32)	0.25	Unresolved
Candidate	Logarithmic, matched at 140	Above 250	39	+0.09 (-0.12 to +0.29)	1.25	Cleared
Candidate	oref0 with autosens	Below 54	40	+0.37 (+0.24 to +0.51)	0.25	Harm
Candidate	oref0 with autosens	Above 250	40	-0.31 (-0.59 to -0.05)	1.25	Cleared

Every reading was the one the first month gave, and each had been written down as expected. Against oref0 with autosens the candidate's time below 54 mg/dL was 0.37 points higher (0.24 to 0.51), which reads harm. Against the matched logarithmic model it was 0.17 points higher (0.04 to 0.32). That reads unresolved, because the interval lies above zero and its upper end is past the margin. On the high side the candidate read cleared against both, and its time above 250 mg/dL was 0.31 points lower than under oref0 with autosens (0.05 to 0.59). One more of the candidate's runs than of the autosens runs ended early through hypoglycaemia, where the separate criterion needs three. The largest rise in one person's time below 54 mg/dL was 1.69 points. Against the unchanged profile over both months, the candidate raised time below 54 mg/dL by 0.35 points (0.20 to 0.50) and lowered mean glucose by 2.12 mg/dL (1.22 to 3.02).

The second month read alone gave smaller differences on the low side. They were +0.25 points (0.09 to 0.43) against oref0 with autosens and +0.04 (-0.15 to 0.20) against the matched logarithmic model, where the first month gave +0.41 and +0.25. Reading one month alone was not in the plan, and the figures are given to show how far two months differ. In that month the placebo's time in range differed from the unchanged profile by +0.98 points (0.13 to 1.94). A difference of about one point of time in range within a single month can therefore arise with nothing of size changed.

No third month is planned. The low side against the matched logarithmic model stays unresolved, with a mean below the margin. Against oref0 with autosens the candidate cost about a third of a percentage point of time below 54 mg/dL in this simulator, and the trace places most of that after corrections given at high glucose.

What is not settled

Some of what is missing can only come from the authors. The level of the TDD arm and the comparison with the measurements in this series depend on four things the publication does not give: which glucose an event is filed under, how the reference basal was defined, which insulin action model gave the insulin on board, and what daily doses the people measured were using. The authors have not been asked. They have also not been asked whether an equation derived by a manufacturer from its own users' records may be built into open-source code.

The link is settled only as far as INV-015 reaches, which is between adults who gave their own corrections with no algorithm dosing. Whether the profile ISF is as good a scale for a person whose loop does most of the correcting has not been tested. The same test on loop records would need the algorithm's insulin counted with the correction. The first version of the candidate carries no multiplier for recent dose against the person's own average, which is what the link to TDD adds within a person. Whether a later version should carry one is open.

The anchor is open. The 140 mg/dL used here is an initial proposal, and the publication's reference is 180 mg/dL. The two differ by a factor of 1.44 in every dose. In the simulation the anchor at 180 mg/dL left time below 54 mg/dL where the profile had it and raised mean glucose by 3.06 mg/dL (2.40 to 3.72). The anchor at 140 lowered mean glucose by 1.77 mg/dL (0.70 to 2.86) and raised time below 54 mg/dL. That is the ranking in one simulator and it does not settle the choice.

The level of the TDD arm is open, as described above. A level set from entered settings at 140 mg/dL may not apply to open-source systems, since the measurement on Loop, Trio and oref0 records read higher at 200 mg/dL.

It is an open question whether the steep part between 100 and 150 mg/dL should be used for dosing at all, given how many events there involved no net insulin. The question of whether to use the published quadratic below 100 mg/dL belongs with it.

The fall above 200 mg/dL is open in the same way. By the ratio of fall to insulin used, neither REPLACE-BG nor Loop reproduces it, whichever way that ratio is summarised. By the extra fall per extra unit, both do.

Parts of the existing systems were not read. The bolus calculator's sensitivity in dynamic mode is one. Autosens in that mode was read in AndroidAPS only, where the default setting holds it at 1.0 (`ApsDynIsfAdjustSensitivity` off, nightscout/AndroidAPS dev f66398c9d0). The third stage did not run AndroidAPS's or Trio's own code for the existing models. It ran each model's equation from `inv013/candidate.py` inside the vendored oref0, with TDD blended as AndroidAPS blends it, and Trio's own blend was not reproduced.

The third stage has a further limit. In TimSim a unit of insulin does more at higher glucose, and the equation assumes it does less. A result from that stage describes the equation in a body built the other way, and whether people resemble the simulator in this respect is not known. The low-side reading reported above is subject to that limit.

The last question is for the maintainers of AndroidAPS and Trio, and it is whether they want one shared definition. The two systems run the same equation and, at their defaults, return values 42% apart for the worked person at 200 mg/dL. A reference implementation with shared test vectors would be new to both.

Reproducing the numbers

`python3 -m inv013.candidate` writes `results/inv013_comparator.md`, which holds the numbers in this document's worked examples and in every table except the refit. `python3 -m inv013.refit` writes `results/inv013_refit.md`, with the refit tables and the refitted coefficients. It reads the per-person bins that `python3 -m inv012.peak_trough` writes, a file too large to keep in git, and checks them against the committed INV-012 table before fitting. `python -m inv014.known_truth --timsim <TimSim checkout>` writes `results/inv014_known_truth.md`, the simulation with a measured truth. It needs an interpreter with simglucose and took about three hours on seven workers. Its per-person tables stay out of git.
`python3 -m inv015.link` writes `results/inv015_link.md`, the comparison of the two links on REPLACE-BG corrections, and `inv015/tests/test_link.py` checks that the regression returns a link it was given.
`python3 -m inv016.replay --replay <INV-008 replay cache> --cohort <canonical_cohort.json>` writes `results/inv016_replay.md`, the replay on recorded histories. `python3 -m inv017.shift_or_scale` writes `results/`

`inv017_shift_or_scale.md`, the comparison of multiplying with adding. `python3 -m inv018.person_curves` writes `results/inv018_person_curves.md`, the curve of each person in REPLACE-BG. `python3 -m inv018.split_half` and `python3 -m inv018.split_half_cohorts` write the comparison of halves, on REPLACE-BG and on all eight cohorts. `python -m inv019.outcome_test --timsim <TimSim checkout>` writes `results/inv019_outcome_test.md`, the outcome test in simulation. It needs the same interpreter as INV-014 and took about seven hours on seven workers, and its raw runs stay out of git. The same command with `--seed 101 1101` and the five arms named in the script's docstring writes `results/inv019_outcome_test_seeds_101_1101.md`, the two months together, and took about two and a half hours. With `--seed 1101 --stage report` it writes the second month alone. `python -m inv019.low_trace` writes `results/inv019_low_trace.md`, the trace of the lows, from the raw runs of the first month. `python3 -m inv013.figures` draws the standalone figures in `charts/proposal8/` and writes `results/inv013_poster_figure.md`, the values measured from the poster's figure. `inv013/tests/test_candidate.py` checks the logarithmic equation against the replay code of INV-008, Trio's form against AndroidAPS's, and the published fit against the values in the abstract and on the poster's figure. It also checks that the refit returns a quartic it was given, and it holds the table of ten values that the device code is tested against.

Figures from other analyses are quoted from where they were produced. The measurement on other cohorts is in `results/inv012_peak_trough.md` and section 7 of document 5. The adult figure of 1915 is from `docs/preprint-dosing-constants.md`, and the overnight test is in `oref-investigations/results/bridge_diabeloop.md`.

Earlier drafts of this document are in git. The first (commit fd3662e) proposed the profile-anchored, flat form now kept as an arm. The second (commit b324528) opened on the comparison with AndroidAPS and described both systems for a general reader. The third (commit 1667e8f) carried the link to TDD as the candidate and judged the equation against the simulations. The fourth (commit b19e3f1) anchored both links at 180 mg/dL. The fifth (commit e856095) moved the anchor to 140 mg/dL and set a reading of the curve at the midpoint of a correction beside the reading at the current glucose. The sixth (commit 1468fc2) applied the equation at the current glucose only. The seventh (commit f1672de) added the evidence for multiplying the curve and

for one shape shared by everyone. The eighth (commit 24b0706) added the outcome test in simulation and the figures. This version adds the second simulated month, the trace of the lows and the shadow code.

References

1. Romero-Ugalde HM, Adenis A, Cartier-Michaud T, Gauthier P, Desir C, Le Roux-Mallouf T, Huneker E, Benhamou PY. 920-P: The Relationship between Glycemia and Insulin Sensitivity Factor. Diabetes 2025;74(Supplement 1):920-P. doi:10.2337/db25-920-P
2. Wilson C. Chris Wilson on Insulin Sensitivity (Correction Factor) with Loop and Learn, 2/6/2022. Loop and Learn, YouTube. https://www.youtube.com/watch?v=oL49FhOts3c. The survey is described from 10 to 16 minutes and the equation is given at 28 minutes.
3. AndroidAPS source history, github.com/nightscout/AndroidAPS. Commit f98175755e (12 February 2022) introduced 277700 / (TDD * bg), and commit c233f9ba12 (13 June 2022) replaced it with 1800 / (TDD * ln(bg / divisor + 1)).
4. Kovatchev BP, Breton M, Dalla Man C, Cobelli C. In silico preclinical trials: a proof of concept in closed-loop control of type 1 diabetes. J Diabetes Sci Technol 2009;3(1):44-55.
5. Xie J. Simglucose v0.2.1. 2018. github.com/jxx123/simglucose
6. Battelino T, Danne T, Bergenstal RM, et al. Clinical targets for continuous glucose monitoring data interpretation: recommendations from the International Consensus on Time in Range. Diabetes Care 2019;42(8):1593-1603. doi:10.2337/dci19-0028