

Department of Electronic and Telecommunication
Engineering
University of Moratuwa



BM3110 – Electronic Instrumentation
ECG Specifications

200003P-T.L Abeygunathilaka

200119C-D.K.D Dewagiri

200287L – P.B.K.S.G kapukotuwa

1. Specifications

Parameter	Value
CMRR	100 dB
Gain	50 dB
Bandwidth	150Hz
Input Impedance	50M Ω
Input voltage	100 μ V – 5mV
Slew rate	10V/ μ s
Drift	0.5 μ V/ $^{\circ}$ C max