

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

\\USER\UserProtocols\Merriam\VASO_playground\26_slices_assym_TR5s_Version_129_DNM

TA: 25:41

PAT: 3

Voxel size: 0.8x0.8x0.8 mm

Rel. SNR: 1.00

USER: VASO_124

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

Routine

Slab group 1	
Slabs	1
Dist. factor	50 %
Position	R7.9 A25.9 F5.7
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
Slice oversampling	7.7 %
Slices per slab	26
FoV read	133.0 mm
FoV phase	133.3 %
Slice thickness	0.82 mm
TR	2736.60 ms
TE	24 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A32

Contrast

Perfusion mode	SS-SI VASO
TI2	650 ms
TI1	50 ms
TI1s	50 ms
Flip angle	26 deg
Fat suppr.	Fat sat.
Fat sat. mode	Strong

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	563
Delay in TR	0 ms
Multiple series	Off

Perfusion mode	PICORE Q2T
Inversion time 1	50 ms
Saturation stop time	50 ms
Inversion time 2	650.0 ms
Flow limit	100 cm/s

Resolution

Base resolution	162
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

PAT mode	GRAPPA
Accel. factor PE	3
Ref. lines PE	45
Accel. factor 3D	1
Ref. lines 3D	22
Reference scan mode	Separate

Prescan Normalize	Off
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Ascending
Special sat.	Parallel F
Gap	25.0 mm
Thickness	100 mm
Table position	H
Table position	0 mm
Inline Composing	Off

System

V32	Off
A32	On
Positioning mode	REF
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Save uncombined	Off
Coil Combine Mode	Sum of Squares
AutoAlign	---
Auto Coil Select	Default

Shim mode	Standard
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
! Ref. amplitude 1H	220.000 V
Adjustment Tolerance	Auto
Adjust volume	
! Position	L0.0 A30.0 F1.4
! Orientation	T > C0.2
! Rotation	90.00 deg
! A >> P	178 mm
! R >> L	133 mm
! F >> H	39 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

Motion correction	Off
Spatial filter	Off

Sequence

Introduction	On
Dimension	3D
Reordering	Linear
Contrasts	1
Bandwidth	1188 Hz/Px
Free echo spacing	Off
Echo spacing	0.97 ms
EPI factor	216

SIEMENS MAGNETOM Investigational_Device_7T syngo MR B17

RF pulse type	Normal
Gradient mode	Normal
Excitation	Slab-sel.
RF spoiling	On
<hr/>	
Ampl	150
BWDTH	150 3.1kHz
ph.skip 4 Robert (the one)	1
use Ernst angle	Off
Maxwell Correction	Off
log physio files	Off
FFT scale	3.00
dummy prepscan time	3 s
z shim	0.00 mT/m*ms
RF duration	2000 us
RF BWTP	25.0
Renzo: Delta TI	72 ms
EFFECTIVE TR	76624 ms
PatPartitions	28
EPI phase correction	local
PAT refscan mode	Flash
FlashRef BaseRes	162
FlashRef BW	100 Hz/px
FlashRef TE	6500 us
FlashRef FA	5 deg
use CAIPI	Off