Sample Name :69-88
Sample ID :U230SGAJG0-7
Time Processed :9:02:21
Month-Day-Year Processed :03/25/2024

Pump A : 0.065% trifluoroacetic in 100% water (v/v)
Pump B : 0.05% trifluoroacetic in 100% acetonitrile (v/v)
Total Flow:1 ml/min
Wavelength:220 nm
<<LC Time Program>>

Time	Module	Command	Value
0.01	Pumps	Pump A B.Conc	5
25.00	Pumps	Pump A B.Conc	65
25.01	Pumps	Pump A B.Conc	95
27.00	Pumps	Pump A B.Conc	95
27.01	Pumps	Pump A B.Conc	5
35.00	Pumps	Pump A B.Conc	5
35.01	Controller	Stop	

<<Column Performance>>
<Detector A>
Column :Inertsil ODS-3 4.6 x 250 mm
Equipment: ZJ23010112

<Chromatogram>



<Peak Table>

Detector A Channel 1 220nm

Peak#	Ret. Time	Area	Height	Area%
1	22.336	59407	7918	0.149
2	22.700	92168	9849	0.232
3	22.958	288290	25389	0.724
4	23.502	39075071	2890324	98.169
5	24.054	288912	24618	0.726
Total		39803848	2958099	100.000