

TI Stimulation Virtual Bench

Waveform Visualiser

Parameter settings:

I1	
Frequency 1 (Hz)	2000
Phase 1 (Degrees)	0
Amplitude 1 (mA)	1
I2	
Frequency 2 (Hz)	2010
Phase 2 (Degrees)	180
Amplitude 2 (mA)	1

Inspect the waveform of the applied currents and voltages.

Is the amplitude of the current or voltage modulated? Why or Why Not?

	Current	Voltage
Modulated?		

If modulation is observed – why?

Inspect the spectrum of the sum current.

Does it have a frequency component at Δf ? Why or Why Not?

Inspect the waveform and spectrum of the sum current.

What is the maximum amplitude?

What is the Amplitude Modulation (AM) Index?

Decrease the amplitude of I1 (Amplitude 1) to 0.5 mA.

Increase the amplitude of I2 (Amplitude 2) to 1.5 mA.

Inspect the waveform and spectrum of the sum current.

Is the maximum current different?

Is the AM Index different? Why?

Field Distribution Visualiser

Switch to the Field Distribution Visualiser tab, and apply the following settings:

Parameter Settings:

Pair A Electrodes	
A+	E11

A-	E8
f1 (Hz)	2000
I1 (mA)	1
Pair B Electrodes	
B+	E1
B-	E4
f2 (Hz)	2010
I2 (mA)	1

Inspect the locus of the TI stimulation (envelope modulation amplitude) in the x and y directions.

Which one is stronger?

Change A+ to E10 and B+ to E2

How does it affect the TI stimulation locus?

Change the Pair A current I2 amplitude to 0.5mA

Change the Pair B current I2 amplitude to 1.5mA

What is current sum and current amplitude ratio?

How does it affect the location of the TI locus relative to the electrodes?

Change the Pair A current I1 amplitude to 0.2 mA

Change the Pair B current I2 amplitude to 1.8 mA

What is current sum and current ratio?

How does it affect the location of the TI locus relative to the electrodes?

Switch back to the Waveform Visualiser tab.

FSK TI stimulation

Change to TI stimulation with frequency shift keying (FSK) waveform and set the following parameters:

FSK Parameters	
PRF (Hz)	10
Pulse Duration (ms)	10
I1 Parameters	
Frequency 1 (Hz)	2000
Phase 1 (Degrees)	0
Amplitude I1 (mA)	1
I2 Parameters	
Freq 2 Outside Pulse (Hz)	2000

Freq 2 Inside Pulse (Hz)	2100
Phase 2 (Degrees)	180
Amplitude I2 (mA)	1

Inspect the waveform of the applied **currents**.

The simulation speed can be slowed down with the 'Time Speed' slider in the Playback controls. The simulation can also be paused with the pause button in the Playback controls.

Is the frequency shift evident in I2?

Inspect the spectrum of the applied currents.

How does it differ from normal TI stimulation?

Inspect the spectrum of the sum current.

How does it differ from the individual applied currents?

Does it have a frequency component at Δf ?

How is the pulse repetition frequency evident in the spectrum?

Change Phase 2 to various values (0, 90, 180, 270)

How does the phase delay between the currents affect the sum current?

Change Phase 2 to 180 degrees

Change Freq 2 inside pulse to 2020Hz

Change PRF to 5Hz

Change Δf duration to 50ms

How many pulses are there?

Change Δf duration to 150

How many pulses are there? Why?

Set the parameters to deliver 5 pulses of a 100Hz Δf every 200ms.

PWM TI stimulation

Change to TI stimulation with PWM waveform and set the following parameters:

Parameter settings:

I1	
Frequency 1 (Hz)	2000
Phase 1 (Degrees)	0
Amplitude 1 (mA)	1
I2	
Frequency 2 (Hz)	2010
Phase 2 (Degrees)	180
Amplitude 2 (mA)	1

Inspect the waveform and spectrum of the sum current.

How does it differ from normal TI stimulation?

Does it have a frequency component at Δf ?