



- NOTES:
- 1. PASSIVATE PER ASTM-A967
NITRIC ACID QQ-P-35C TYPE VIII
CITRIC ACID TREATMENT PERMISSABLE
 - 2. ▽ INDICATES KEY FEATURES TO BE VERIFIED

| REV | REVISION DESCRIPTION | BY | DATE | ECO # |
|-----|----------------------|----|----------|-----------|
| A | ORIGINAL RELEASE | SD | 08/11/06 | BCT017484 |



1-6-11 408

| | | | | | |
|---|-------|--|---------------------------|--|----------|
| THESE DRAWINGS AND SPECIFICATIONS ARE THE PROPERTY OF GAMBRO BCT INC. AND SHALL NOT BE REPRODUCED, COPIED, OR USED FOR MANUFACTURE OR SALE OF APPARATUS WITHOUT WRITTEN PERMISSION OF GAMBRO BCT INC. | | | |  10811 Collins Avenue, Lakewood, Colorado 80215-4440 USA GAMBRO BCT | |
| TOLERANCES AND SYMBOLS PER ANSI Y14.5M-1982 DIMENSIONS SPECIFIED ARE AFTER ADDITIVE FINISH. | | DIMENSIONS ARE IN INCHES UNLESS OTHERWISE SPECIFIED | | TITLE | |
| .X ± 0.04 .XX ± 0.010 .XXX ± 0.005 .XXXX ± 0.0020 | | MATERIAL 303 STAINLESS STEEL | | NUT, JAM SMALL VIEW PORT | |
| ANGLES ± 1 | | SURF. FINISH 125 | | ORIGINATOR | |
| DATE | | DATE | | DATE | |
| SIZE | SCALE | DD NOT SCALE DWG | THIRD ANGLE PROJECTION | DRAWN | DATE |
| B | 1:1 | | | S DURKEE | 09/21/05 |
| PART NO. | | | | REV | SHT / OF |
| 777094-216 | | | | A | 1 / 1 |

UNIGRAPHICS