



REV	REVISION DESCRIPTION	BY	DATE	ECO #
D	ADDED INSPECTION CRITERIA	MQ	9/16/91	57071
E	ADDED TOLERANCE TO .X PLACE	RH	02/08/96	BCT 000922



- NOTES:
1. BREAK ALL SHARP EDGES & CORNERS .02 R MAX.
 2. FINISH: MATTE MULTI-DIRECTIONAL LAY IE AS FROM VIBRATORY DEBURRING.
 3. ♦ INDICATE KEY DIM TO BE CHECKED BY INCOMING INSPECTION.



AUTOCAD DWG
ORIG DISK:
FILE: 066246D.DWG

THESE DRAWINGS AND SPECIFICATIONS ARE THE PROPERTY OF COBE BCT INC. AND SHALL NOT BE REPRODUCED, COPIED, OR USED FOR MANUFACTURE OR SALE OF APPARATUS WITHOUT WRITTEN PERMISSION OF COBE BCT INC.				COBE BCT COBE BCT INC. Blood Component Technology a gambro company 1201 Oak Lakewood, Colorado 80215-4407 USA	
TOLERANCES AND SYMBOLS PER ANSI Y14.5M-1982 DIMENSIONS SPECIFIED ARE AFTER ADDITIVE FINISH. .X ± .1 .XX ± .01 .XXX ± .005 .XXXX ±		DIMENSIONS ARE IN INCHES UNLESS OTHERWISE SPECIFIED MATERIAL BLACK DELRIN		TITLE BUTTON	
ANGLES ±		SURF. FINISH 125 RMS		ORIGINATOR SCANNED	
SIZE B		SCALE 2X		DATE	
DO NOT SCALE DWG		THIRD ANGLE PROJECTION		DATE	
PART NO. 777066-246				REV E SHT 1 OF 1	

Dated Print
3-1-96

408