


Test Procedure Reference	Calibration	Job No.	Calibration
Unique No.	Calibration	Test Description	Calibration
R&D Reference	Calibration	Test Date	26/09/2025
Valve Description	Calibration	Valve Drawing No.	Calibration



Calibration Downstream



Applied (µA)	4000	8000	12000	16000	20000
Counts (avg)	19	1952836	3906037	5858527	7811778
Converted (µA)	4000.04	7999.41	11999.56	15998.26	19998.52
Abs Error (µA) - ±3.6 mV	0.04	0.59	0.44	1.74	1.48

Data Recording Equipment Used

Data Logger	DLS Prototype
Serial No.	N/A
Transducers	

Gauges

Torque Transducer N/A

Cycle Count	1
-------------	---

Test Pressure	0 psi
---------------	-------

Breakout Torque	N/A
Running Torque	N/A

3rd Party Stamp and Date

Operative: Jason Bradley