


Test Procedure Reference	Test Procedure Reference	Job No.	Job Number
Unique No.	Unique Number	Test Description	Calibration
R&D Reference	R&D Reference	Test Date	25/09/2025
Valve Description	Valve Description	Valve Drawing No.	Valve Drawing Number



Calibration Downstream



Applied (mA)	4 mA	8 mA	12 mA	16 mA	20 mA
Counts (avg)	93	1953181	3906786	5858989	7812730
Converted (µA)	4000.19	8000.12	12001.1	15999.21	20000.47
Abs Error (µA)	0.19	0.12	1.1	0.79	0.47

Data Recording Equipment Used	
Data Logger	DLS Prototype
Serial No.	N/A
Transducers	
Gauges	
Torque Transducer N/A	
Cycle Count	1
Test Pressure	0 psi
Breakout Torque	N/A
Running Torque	N/A
3rd Party Stamp and Date	
Operative: Toby Lelean	