


Test Procedure Reference	N/A	Job No.	N/A
Unique No.	N/A	Test Description	Calibration
R&D Reference	N/A	Test Date	26/09/2025
Valve Description	N/A	Valve Drawing No.	N/A



Calibration Upstream



Applied (µA)	4000	8000	12000	16000	20000
Counts (avg)	4395881	4883871	5371962	5859216	6346601
Converted (µA)	13002.77	14002.17	15001.78	15999.68	16997.84
Abs Error (µA) - ±3.6 mV	9002.77	6002.17	3001.78	0.32	3002.16

Data Recording Equipment Used

Data Logger	DLS Prototype
Serial No.	N/A
Transducers	

Gauges

Torque Transducer N/A

Cycle Count	1
-------------	---

Test Pressure	0 psi
---------------	-------

Breakout Torque	N/A
-----------------	-----

Running Torque	N/A
----------------	-----

3rd Party Stamp and Date

Operative: Jason Bradley