


Test Procedure Reference	Test Procedure Reference	Job No.	Job Number
Unique No.	Unique Number	Test Description	Calibration
R&D Reference	R&D Reference	Test Date	11/09/2025
Valve Description	Valve Description	Valve Drawing No.	Valve Drawing Number



Calibration Upstream



	1	2	3	4	5
	nan	nan	nan	nan	nan
Measured	-94.0	1953092.0	3909902.0	5862926.0	7816124.0
Expected	0.0	1953125.0	3906250.0	5859375.0	7812500.0
Error %	nan	-0.002	0.093	0.061	0.046

Data Recording Equipment Used

Data Logger	DLS Prototype
Serial No.	N/A
Transducers	

Gauges

Torque Transducer N/A

Cycle Count	1
Test Pressure	0 psi
Breakout Torque	N/A
Running Torque	N/A

3rd Party Stamp and Date

Operative: Toby Lelean