


Test Procedure Reference	Test Procedure Reference	Job No.	Job Number
Unique No.	Unique Number	Test Description	Calibration
R&D Reference	PI-11	Test Date	08/09/2025
Valve Description	Valve Description	Valve Drawing No.	Valve Drawing Number



Calibration Upstream



	1	2	3	4	5
	nan	nan	nan	nan	nan
Measured	0.0	4998.0	9996.0	14991.0	19990.0
Expected	0.0	5000.0	10000.0	15000.0	20000.0
Error %	0.0	-0.04	-0.04	-0.06	-0.05

Data Recording Equipment Used

Data LoggerDLS Prototype
Serial No.N/A
Transducers

Gauges

Torque TransducerN/A

Cycle Count1

Test Pressure0 psi

Breakout TorqueN/A
Running TorqueN/A

3rd Party Stamp and Date

Operative: Jason Bradley