


Test Procedure Reference	Test Procedure Reference	Job No.	Job Number
Unique No.	Unique Number	Test Description	Calibration
R&D Reference	R&D Reference	Test Date	25/09/2025
Valve Description	Valve Description	Valve Drawing No.	Valve Drawing Number



Calibration BX Port



Applied (mA)	4 mA	8 mA	12 mA	16 mA	20 mA
Counts (avg)	325	1953536	3907244	5859488	7813089
Converted (µA)	4000.67	8000.84	12002.04	16000.23	20001.21
Abs Error (µA)	0.67	0.84	2.04	0.23	1.21

Data Recording Equipment Used

Data Logger	DLS Prototype
Serial No.	N/A
Transducers	

Gauges

Torque Transducer N/A

Cycle Count	1
-------------	---

Test Pressure	0 psi
---------------	-------

Breakout Torque	N/A
Running Torque	N/A

3rd Party Stamp and Date

Operative: Toby Lelean