


Test Procedure Reference	Test Procedure Reference	Job No.	Job Number
Unique No.	Unique Number	Test Description	Calibration
R&D Reference	R&D Reference	Test Date	26/09/2025
Valve Description	Valve Description	Valve Drawing No.	Valve Drawing Number



Calibration Ambient Temperature



Applied (mV)	-5.89	9.28	24.46	39.64	54.81
Counts (avg)	-198	193	585	978	1370
Converted (mV)	-13.55	1.57	16.72	31.88	47.04
Abs Error (mV) [±0.12 mV]	7.66	7.71	7.74	7.76	7.77

Data Recording Equipment Used

Data Logger	DLS Prototype
Serial No.	N/A
Transducers	

Gauges

Torque Transducer N/A

Cycle Count	1
-------------	---

Test Pressure	0 psi
---------------	-------

Breakout Torque	N/A
Running Torque	N/A

3rd Party Stamp and Date

Operative: Toby Lelean