


Test Procedure Reference	Test Procedure Reference	Job No.	Job Number
Unique No.	Unique Number	Test Description	Calibration
R&D Reference	R&D Reference	Test Date	10/09/2025
Valve Description	Valve Description	Valve Drawing No.	Valve Drawing Number



Calibration Upstream



	1	2	3	4	5
	nan	nan	nan	nan	nan
Measured	3602.0	1956436.0	3905718.0	5857605.0	7811856.0
Expected	0.0	1953125.0	3906250.0	5859375.0	7812500.0
Error %	nan	0.17	-0.014	-0.03	-0.008

Data Recording Equipment Used	
Data Logger	DLS Prototype
Serial No.	N/A
Transducers	
Gauges	
Torque Transducer	N/A
Cycle Count	1
Test Pressure	0 psi
Breakout Torque	N/A
Running Torque	N/A

3rd Party Stamp and Date

Operative: Toby Lelean