



Eastern Research Group
601 Keystone Park Drive
Suite 700
Morrisville, NC 27560

June 13, 2026

Mr. Bryan Sims
Oklahoma Dept. of Environmental Quality
707 N. Robinson, PO Box 1677
Oklahoma City, OK 73101-1677
Project Name: OCOK

Dear Mr. Bryan Sims,

This report contains the analytical results for the sample(s) received under chain(s) of custody by Eastern Research Group on 04/08/26 10:46 through 05/06/26 12:20.

Values below the MDL for QC results in this report are recorded as ND, however the actual values are reported in the accompanying Excel report with a "U" flag (Under the detection limit). The actual values are reported in AQS.

This test is accredited under the 2016 TNI Standard for Environmental Laboratories (FL DOH Certification # E87673). All analyses were performed as described in the US EPA-approved QAPP, under the contract for National Hazardous Air Pollutant Support (US EPA Contract No. 68HERH22D0002). This cover page is an integral part of this report, and any exceptions or comments are noted on the last page.

Release of the data contained in this data package and in the data submitted in the electronic data deliverable, has been authorized by the Program Manager, or the Program Manager's designee as verified by the following signature.

The issuance of the final Certificate of Analysis takes precedence over any previous Report. If you have any questions, please contact me at 919-468-7872.

Sincerely,

Randolph Bower
Program Manager
randy.bower@erg.com

The information contained in this report and its attachment(s) are intended only for the use of the individual to whom it is addressed and may contain information that is privileged, confidential, or exempt from disclosure. If the reader of this message is not the intended recipient, you are hereby notified that any dissemination, distribution, or copying of this report is strictly prohibited. If you have received this report in error, please notify randy.bower@erg.com and delete the report without retaining any copies.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

ANALYTICAL REPORT FOR SAMPLES

<u>SampleName</u>	<u>LabNumber</u>	<u>Matrix</u>	<u>Sampled</u>	<u>Received</u>
OCOK	6040829-01	Air	04/02/26 23:59	04/08/26 10:46
OCOK	6041472-01	Air	04/08/26 23:59	04/14/26 10:41
OCOK	6042310-01	Air	04/14/26 23:59	04/23/26 10:22
OCOK	6042310-02	Air	04/20/26 23:59	04/23/26 10:22
OCOK	6050635-01	Air	04/26/26 23:59	05/06/26 12:20
OCOK FB	6050635-02	Air	04/26/26 23:59	05/06/26 12:20
OCOK D1	6050635-03	Air	05/02/26 23:59	05/06/26 12:20
OCOK D2	6050635-04	Air	05/02/26 23:59	05/06/26 12:20



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6040829-01

Sampled: 04/02/26 23:59

Temp Rec'd (°C): 10.1

Sample Volume: 1143.1 L

Received: 04/08/26 10:46

Cartridge Pouch #: 26G2649

Analyzed: 04/16/26 16:43

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.95	DNPH, TT	0.0554
Acetaldehyde	1.58	DNPH, TT	0.0229
Acetone	1.85	DNPH, TT	0.0849
Propionaldehyde	0.535	DNPH, QB-01, TT	0.181
Crotonaldehyde	0.0419	DNPH, TT	0.0166
Butyraldehyde	0.292	DNPH, TT	0.00980
2-Butanone	0.716	DNPH, TT	0.0369
Benzaldehyde	0.115	DNPH, TT	0.0379
Valeraldehyde	0.210	DNPH, TT	0.0106
Hexaldehyde	0.184	DNPH, TT	0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6041472-01

Sampled: 04/08/26 23:59

Temp Rec'd (°C): 2.1

Sample Volume: 1142.8 L

Received: 04/14/26 10:41

Cartridge Pouch #: 26G2650

Analyzed: 04/16/26 18:15

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	2.87	LB	0.0554
Acetaldehyde	1.82		0.0229
Acetone	4.48		0.0849
Propionaldehyde	0.736	QB-01	0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.286		0.00980
2-Butanone	0.707		0.0369
Benzaldehyde	0.0962		0.0379
Valeraldehyde	0.141		0.0106
Hexaldehyde	0.104		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6042310-01

Sampled: 04/14/26 23:59

Temp Rec'd (°C): 1.5

Sample Volume: 1143.5 L

Received: 04/23/26 10:22

Cartridge Pouch #: 26G2651

Analyzed: 04/27/26 22:06

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.97		0.0554
Acetaldehyde	1.32		0.0229
Acetone	0.674		0.0848
Propionaldehyde	0.421	QB-01	0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.201		0.00979
2-Butanone	0.243		0.0369
Benzaldehyde	0.0820		0.0379
Valeraldehyde	0.149		0.0106
Hexaldehyde	0.169		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6042310-02

Sampled: 04/20/26 23:59

Temp Rec'd (°C): 1.5

Sample Volume: 1143.8 L

Received: 04/23/26 10:22

Cartridge Pouch #: 26G2652

Analyzed: 04/30/26 16:25

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.01		0.0553
Acetaldehyde	1.69		0.0229
Acetone	4.12		0.0848
Propionaldehyde	0.647	QB-01	0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.231		0.00979
2-Butanone	0.615		0.0369
Benzaldehyde	ND	U	0.0379
Valeraldehyde	0.116		0.0106
Hexaldehyde	0.0742		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK

Lab ID: 6050635-01

Sampled: 04/26/26 23:59

Temp Rec'd (°C): 0.5

Sample Volume: 1146.1 L

Received: 05/06/26 12:20

Cartridge Pouch #: 26G2653

Analyzed: 05/07/26 22:55

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	1.76	DNPH	0.0552
Acetaldehyde	0.947	DNPH	0.0229
Acetone	0.462	DNPH	0.0846
Propionaldehyde	0.340	DNPH	0.181
Crotonaldehyde	0.0710	DNPH	0.0166
Butyraldehyde	0.164	DNPH	0.00977
2-Butanone	1.56	DNPH	0.0368
Benzaldehyde	0.131	DNPH	0.0378
Valeraldehyde	0.103	DNPH	0.0106
Hexaldehyde	0.111	DNPH	0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK FB

Lab ID: 6050635-02

Sampled: 04/26/26 23:59

Temp Rec'd (°C): 0.5

Sample Volume: 1146.1 L

Received: 05/06/26 12:20

Cartridge Pouch #: 26G2654

Analyzed: 05/07/26 23:12

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	0.0361	U	0.0552
Acetaldehyde	0.0325		0.0229
Acetone	0.0560	U	0.0846
Propionaldehyde	0.0680	U	0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	ND	U	0.00977
2-Butanone	ND	U	0.0368
Benzaldehyde	0.0419		0.0378
Valeraldehyde	ND	U	0.0106
Hexaldehyde	ND	U	0.0106



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK D1

Lab ID: 6050635-03

Sampled: 05/02/26 23:59

Temp Rec'd (°C): 0.5

Sample Volume: 1142.2 L

Received: 05/06/26 12:20

Cartridge Pouch #: 26G2655

Analyzed: 05/11/26 15:16

Comments:

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.32		0.0554
Acetaldehyde	1.82		0.0229
Acetone	4.70		0.0849
Propionaldehyde	0.686	QB-01	0.181
Crotonaldehyde	ND	U	0.0166
Butyraldehyde	0.248		0.00981
2-Butanone	0.691		0.0369
Benzaldehyde	0.107		0.0379
Valeraldehyde	0.122		0.0106
Hexaldehyde	0.127		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Description: OCOK D2	Lab ID: 6050635-04	Sampled: 05/02/26 23:59
Temp Rec'd (°C): 0.5	Sample Volume: 1141.1 L	Received: 05/06/26 12:20
	Cartridge Pouch #: 26M0033	Analyzed: 05/11/26 16:05
Comments: Duplicate Makeup from 3-15-26		

Air Toxics by EPA Compendium Method TO-11A

<u>Analyte</u>	<u>Results</u>	<u>Flag</u>	<u>MDL</u>
	<u>ug/m³ Air</u>		<u>ug/m³ Air</u>
Formaldehyde	3.30		0.0555
Acetaldehyde	1.81		0.0230
Acetone	4.60		0.0850
Propionaldehyde	0.670	QB-01	0.181
Crotonaldehyde	ND	U	0.0167
Butyraldehyde	0.249		0.00982
2-Butanone	0.691		0.0370
Benzaldehyde	0.102		0.0379
Valeraldehyde	0.127		0.0106
Hexaldehyde	0.130		0.0107



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604052 - B6D1410

Calibration Blank (2604052-CCB1)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Blank (2604052-CCB2)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.00	ug/ml							U
Acetaldehyde	0.00	ug/ml							U
Acetone	0.00	ug/ml							U
Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604052-CCV1)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.144	ug/ml	0.15000		95.9	84.9-115.1			
Acetaldehyde	0.152	ug/ml	0.15000		101	84.9-115.1			
Acetone	0.150	ug/ml	0.15000		100	84.9-115.1			
Propionaldehyde	0.150	ug/ml	0.15000		100	84.9-115.1			
Crotonaldehyde	0.150	ug/ml	0.15000		99.7	84.9-115.1			
Butyraldehyde	0.147	ug/ml	0.15000		98.0	84.9-115.1			
2-Butanone	0.151	ug/ml	0.15000		101	84.9-115.1			
Benzaldehyde	0.157	ug/ml	0.15000		105	84.9-115.1			
Valeraldehyde	0.148	ug/ml	0.15000		98.8	84.9-115.1			
Hexaldehyde	0.148	ug/ml	0.15000		98.8	84.9-115.1			

Calibration Check (2604052-CCV2)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.144	ug/ml	0.15000		96.0	84.9-115.1			
Acetaldehyde	0.150	ug/ml	0.15000		99.8	84.9-115.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604052 - B6D1410

Calibration Check (2604052-CCV2) Contin

Prepared & Analyzed: 04/16/26

Acetone	0.150	ug/ml	0.15000	100	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.152	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Hexaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1

Calibration Check (2604052-CCV3)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.6	84.9-115.1
Propionaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Crotonaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Butyraldehyde	0.147	ug/ml	0.15000	97.9	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	101	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1

Batch 2604053 - B6D0706

Calibration Blank (2604053-CCB1)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2604053-CCB2)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604053 - B6D0706

Calibration Blank (2604053-CCB2) Contin

Prepared & Analyzed: 04/16/26

Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604053-CCV1)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Acetone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.149	ug/ml	0.15000	99.0	84.9-115.1
2-Butanone	0.151	ug/ml	0.15000	100	84.9-115.1
Benzaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1

Calibration Check (2604053-CCV2)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Acetone	0.151	ug/ml	0.15000	100	84.9-115.1
Propionaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.149	ug/ml	0.15000	99.0	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.0	84.9-115.1
Benzaldehyde	0.160	ug/ml	0.15000	107	84.9-115.1
Valeraldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.3	84.9-115.1

Calibration Check (2604053-CCV3)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.147	ug/ml	0.15000	98.2	84.9-115.1
Acetaldehyde	0.150	ug/ml	0.15000	99.8	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.7	84.9-115.1
Propionaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Crotonaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.5	84.9-115.1
Benzaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1
Valeraldehyde	0.158	ug/ml	0.15000	105	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604053 - B6D0706

Calibration Check (2604053-CCV3) Contin

Prepared & Analyzed: 04/16/26

Hexaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
-------------	-------	-------	---------	-----	------------

Batch 2604088 - B6D2709

Calibration Blank (2604088-CCB1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2604088-CCB2)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2604088-CCV1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.143	ug/ml	0.15000	95.0	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	97.9	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.6	84.9-115.1
Propionaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.1	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1

Calibration Check (2604088-CCV2)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.143	ug/ml	0.15000	95.2	84.9-115.1
--------------	-------	-------	---------	------	------------

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604088 - B6D2709

Calibration Check (2604088-CCV2) Contin

Prepared & Analyzed: 04/27/26

Acetaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.2	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	99.0	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Hexaldehyde	0.142	ug/ml	0.15000	94.5	84.9-115.1

Calibration Check (2604088-CCV3)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.0	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.0	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1

Batch 2604099 - B6D3002

Calibration Blank (2604099-CCB1)

Prepared & Analyzed: 04/30/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2604099-CCB2)

Prepared & Analyzed: 04/30/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604099 - B6D3002

Calibration Blank (2604099-CCB2) Contin

Prepared & Analyzed: 04/30/26

Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2604099-CCV1)

Prepared & Analyzed: 04/30/26

Formaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Acetone	0.149	ug/ml	0.15000	99.0	84.9-115.1
Propionaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Crotonaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Butyraldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
2-Butanone	0.150	ug/ml	0.15000	99.9	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.3	84.9-115.1

Calibration Check (2604099-CCV2)

Prepared & Analyzed: 04/30/26

Formaldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Propionaldehyde	0.154	ug/ml	0.15000	103	84.9-115.1
Crotonaldehyde	0.147	ug/ml	0.15000	97.9	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	95.7	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	97.0	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	99.0	84.9-115.1

Calibration Check (2604099-CCV3)

Prepared & Analyzed: 04/30/26

Formaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
Acetaldehyde	0.149	ug/ml	0.15000	99.1	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Propionaldehyde	0.151	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.1	84.9-115.1
Benzaldehyde	0.155	ug/ml	0.15000	103	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2604099 - B6D3002

Calibration Check (2604099-CCV3) Contin

Prepared & Analyzed: 04/30/26

Valeraldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1

Batch 2605024 - B6E0702

Calibration Blank (2605024-CCB1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Blank (2605024-CCB2)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.00	ug/ml				U
Acetaldehyde	0.00	ug/ml				U
Acetone	0.00	ug/ml				U
Propionaldehyde	0.00	ug/ml				U
Crotonaldehyde	0.00	ug/ml				U
Butyraldehyde	0.00	ug/ml				U
2-Butanone	0.00	ug/ml				U
Benzaldehyde	0.00	ug/ml				U
Valeraldehyde	0.00	ug/ml				U
Hexaldehyde	0.00	ug/ml				U

Calibration Check (2605024-CCV1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.4	84.9-115.1
Acetone	0.148	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.3	84.9-115.1
2-Butanone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Benzaldehyde	0.151	ug/ml	0.15000	101	84.9-115.1
Valeraldehyde	0.147	ug/ml	0.15000	97.7	84.9-115.1
Hexaldehyde	0.148	ug/ml	0.15000	98.7	84.9-115.1

Calibration Check (2605024-CCV2)

Prepared & Analyzed: 05/07/26

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605024 - B6E0702

Calibration Check (2605024-CCV2) Contin

Prepared & Analyzed: 05/07/26

Formaldehyde	0.142	ug/ml	0.15000	94.7	84.9-115.1
Acetaldehyde	0.148	ug/ml	0.15000	98.8	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.1	84.9-115.1
Propionaldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	99.9	84.9-115.1
Butyraldehyde	0.143	ug/ml	0.15000	95.5	84.9-115.1
2-Butanone	0.148	ug/ml	0.15000	98.7	84.9-115.1
Benzaldehyde	0.153	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.5	84.9-115.1
Hexaldehyde	0.147	ug/ml	0.15000	98.1	84.9-115.1

Calibration Check (2605024-CCV3)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.6	84.9-115.1
Acetone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.2	84.9-115.1
Butyraldehyde	0.141	ug/ml	0.15000	94.2	84.9-115.1
2-Butanone	0.149	ug/ml	0.15000	99.1	84.9-115.1
Benzaldehyde	0.152	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Hexaldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1

Batch 2605031 - B6E1108

Calibration Blank (2605031-CCB1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U
Propionaldehyde	0.00	ug/ml			U
Crotonaldehyde	0.00	ug/ml			U
Butyraldehyde	0.00	ug/ml			U
2-Butanone	0.00	ug/ml			U
Benzaldehyde	0.00	ug/ml			U
Valeraldehyde	0.00	ug/ml			U
Hexaldehyde	0.00	ug/ml			U

Calibration Blank (2605031-CCB2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.00	ug/ml			U
Acetaldehyde	0.00	ug/ml			U
Acetone	0.00	ug/ml			U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605031 - B6E1108

Calibration Blank (2605031-CCB2) Contin

Prepared & Analyzed: 05/11/26

Propionaldehyde	0.00	ug/ml							U
Crotonaldehyde	0.00	ug/ml							U
Butyraldehyde	0.00	ug/ml							U
2-Butanone	0.00	ug/ml							U
Benzaldehyde	0.00	ug/ml							U
Valeraldehyde	0.00	ug/ml							U
Hexaldehyde	0.00	ug/ml							U

Calibration Check (2605031-CCV1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.143	ug/ml	0.15000	95.4	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.3	84.9-115.1
Acetone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Propionaldehyde	0.149	ug/ml	0.15000	99.4	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.6	84.9-115.1
Butyraldehyde	0.140	ug/ml	0.15000	93.4	84.9-115.1
2-Butanone	0.145	ug/ml	0.15000	96.7	84.9-115.1
Benzaldehyde	0.157	ug/ml	0.15000	105	84.9-115.1
Valeraldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.5	84.9-115.1

Calibration Check (2605031-CCV2)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000	96.1	84.9-115.1
Acetaldehyde	0.146	ug/ml	0.15000	97.5	84.9-115.1
Acetone	0.147	ug/ml	0.15000	97.7	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Crotonaldehyde	0.149	ug/ml	0.15000	99.3	84.9-115.1
Butyraldehyde	0.144	ug/ml	0.15000	96.2	84.9-115.1
2-Butanone	0.146	ug/ml	0.15000	97.5	84.9-115.1
Benzaldehyde	0.156	ug/ml	0.15000	104	84.9-115.1
Valeraldehyde	0.148	ug/ml	0.15000	98.6	84.9-115.1
Hexaldehyde	0.146	ug/ml	0.15000	97.1	84.9-115.1

Calibration Check (2605031-CCV3)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.144	ug/ml	0.15000	95.9	84.9-115.1
Acetaldehyde	0.147	ug/ml	0.15000	98.0	84.9-115.1
Acetone	0.147	ug/ml	0.15000	98.3	84.9-115.1
Propionaldehyde	0.150	ug/ml	0.15000	99.7	84.9-115.1
Crotonaldehyde	0.150	ug/ml	0.15000	100	84.9-115.1
Butyraldehyde	0.142	ug/ml	0.15000	94.6	84.9-115.1
2-Butanone	0.144	ug/ml	0.15000	95.7	84.9-115.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch 2605031 - B6E1108

Calibration Check (2605031-CCV3) Contin

Prepared & Analyzed: 05/11/26

Benzaldehyde	0.154	ug/ml	0.15000	102	84.9-115.1
Valeraldehyde	0.143	ug/ml	0.15000	95.3	84.9-115.1
Hexaldehyde	0.145	ug/ml	0.15000	96.4	84.9-115.1

Batch B6D1410 - TO-11A extraction

Blank (B6D1410-BLK1)

Prepared: 04/14/26 Analyzed: 04/16/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							QB-01, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0456	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6D1410-BLK3)

Prepared: 04/14/26 Analyzed: 04/16/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D1410-BS1)

Prepared: 04/14/26 Analyzed: 04/16/26

Formaldehyde	0.278	ug/ml	0.30000	92.7	79.9-120.1
Acetaldehyde	0.285	ug/ml	0.30000	95.1	79.9-120.1
Acetone	0.292	ug/ml	0.30000	97.2	69.9-130.1
Propionaldehyde	0.329	ug/ml	0.30000	110	69.9-130.1
Crotonaldehyde	0.272	ug/ml	0.30000	90.8	69.9-130.1
Butyraldehyde	0.244	ug/ml	0.30000	81.3	69.9-130.1
2-Butanone	0.237	ug/ml	0.30000	79.1	69.9-130.1
Benzaldehyde	0.306	ug/ml	0.30000	102	69.9-130.1
Valeraldehyde	0.251	ug/ml	0.30000	83.7	69.9-130.1
Hexaldehyde	0.247	ug/ml	0.30000	82.5	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D1410 - TO-11A extraction

LCS Dup (B6D1410-BSD1)

Prepared: 04/14/26 Analyzed: 04/16/26

Formaldehyde	0.282	ug/ml	0.30000		93.9	79.9-120.1	1.27	20	
Acetaldehyde	0.284	ug/ml	0.30000		94.7	79.9-120.1	0.411	20	
Acetone	0.293	ug/ml	0.30000		97.6	69.9-130.1	0.390	20	
Propionaldehyde	0.340	ug/ml	0.30000		113	69.9-130.1	3.22	20	
Crotonaldehyde	0.268	ug/ml	0.30000		89.2	69.9-130.1	1.75	20	
Butyraldehyde	0.249	ug/ml	0.30000		83.1	69.9-130.1	2.23	20	
2-Butanone	0.237	ug/ml	0.30000		79.2	69.9-130.1	0.0940	20	
Benzaldehyde	0.307	ug/ml	0.30000		102	69.9-130.1	0.230	20	
Valeraldehyde	0.249	ug/ml	0.30000		83.0	69.9-130.1	0.798	20	
Hexaldehyde	0.244	ug/ml	0.30000		81.5	69.9-130.1	1.19	20	

Batch B6D1603 - TO-11A extraction

Blank (B6D1603-BLK1)

Prepared & Analyzed: 04/16/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							QB-01, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0449	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6D1603-BLK3)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.110	ug/m ³ Air							LB
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D1603-BS1)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.289	ug/ml	0.30000		96.2	79.9-120.1			
Acetaldehyde	0.290	ug/ml	0.30000		96.7	79.9-120.1			
Acetone	0.291	ug/ml	0.30000		97.1	69.9-130.1			

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D1603 - TO-11A extraction

LCS (B6D1603-BS1) Continued

Prepared & Analyzed: 04/16/26

Propionaldehyde	0.353	ug/ml	0.30000	118	69.9-130.1
Crotonaldehyde	0.285	ug/ml	0.30000	94.9	69.9-130.1
Butyraldehyde	0.257	ug/ml	0.30000	85.7	69.9-130.1
2-Butanone	0.235	ug/ml	0.30000	78.3	69.9-130.1
Benzaldehyde	0.314	ug/ml	0.30000	105	69.9-130.1
Valeraldehyde	0.262	ug/ml	0.30000	87.3	69.9-130.1
Hexaldehyde	0.250	ug/ml	0.30000	83.2	69.9-130.1

LCS Dup (B6D1603-BSD1)

Prepared & Analyzed: 04/16/26

Formaldehyde	0.295	ug/ml	0.30000	98.5	79.9-120.1	2.32	20
Acetaldehyde	0.291	ug/ml	0.30000	97.1	79.9-120.1	0.429	20
Acetone	0.296	ug/ml	0.30000	98.6	69.9-130.1	1.51	20
Propionaldehyde	0.351	ug/ml	0.30000	117	69.9-130.1	0.633	20
Crotonaldehyde	0.288	ug/ml	0.30000	96.0	69.9-130.1	1.17	20
Butyraldehyde	0.258	ug/ml	0.30000	86.1	69.9-130.1	0.371	20
2-Butanone	0.239	ug/ml	0.30000	79.6	69.9-130.1	1.56	20
Benzaldehyde	0.318	ug/ml	0.30000	106	69.9-130.1	1.15	20
Valeraldehyde	0.261	ug/ml	0.30000	87.1	69.9-130.1	0.263	20
Hexaldehyde	0.248	ug/ml	0.30000	82.6	69.9-130.1	0.708	20

Batch B6D2709 - TO-11A extraction

Blank (B6D2709-BLK1)

Prepared & Analyzed: 04/27/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0322	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					QB-01, U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

Blank (B6D2709-BLK3)

Prepared & Analyzed: 04/27/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	ND	ug/m ³ Air					U
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D2709 - TO-11A extraction

Blank (B6D2709-BLK3) Continued

Prepared & Analyzed: 04/27/26

2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D2709-BS1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.279	ug/ml	0.30000		93.0	79.9-120.1			
Acetaldehyde	0.289	ug/ml	0.30000		96.2	79.9-120.1			
Acetone	0.287	ug/ml	0.30000		95.7	69.9-130.1			
Propionaldehyde	0.332	ug/ml	0.30000		111	69.9-130.1			
Crotonaldehyde	0.271	ug/ml	0.30000		90.3	69.9-130.1			
Butyraldehyde	0.255	ug/ml	0.30000		85.0	69.9-130.1			
2-Butanone	0.234	ug/ml	0.30000		77.9	69.9-130.1			
Benzaldehyde	0.306	ug/ml	0.30000		102	69.9-130.1			
Valeraldehyde	0.256	ug/ml	0.30000		85.2	69.9-130.1			
Hexaldehyde	0.247	ug/ml	0.30000		82.2	69.9-130.1			

LCS Dup (B6D2709-BSD1)

Prepared & Analyzed: 04/27/26

Formaldehyde	0.277	ug/ml	0.30000		92.3	79.9-120.1	0.699	20	
Acetaldehyde	0.289	ug/ml	0.30000		96.5	79.9-120.1	0.298	20	
Acetone	0.288	ug/ml	0.30000		95.9	69.9-130.1	0.165	20	
Propionaldehyde	0.331	ug/ml	0.30000		110	69.9-130.1	0.398	20	
Crotonaldehyde	0.274	ug/ml	0.30000		91.2	69.9-130.1	1.08	20	
Butyraldehyde	0.252	ug/ml	0.30000		84.0	69.9-130.1	1.22	20	
2-Butanone	0.229	ug/ml	0.30000		76.4	69.9-130.1	2.02	20	
Benzaldehyde	0.308	ug/ml	0.30000		103	69.9-130.1	0.824	20	
Valeraldehyde	0.253	ug/ml	0.30000		84.3	69.9-130.1	1.01	20	
Hexaldehyde	0.248	ug/ml	0.30000		82.6	69.9-130.1	0.420	20	

Batch B6D3002 - TO-11A extraction

Blank (B6D3002-BLK1)

Prepared & Analyzed: 04/30/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	0.0320	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							QB-01, U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6D3002 - TO-11A extraction

Blank (B6D3002-BLK1) Continued

Prepared & Analyzed: 04/30/26

Hexaldehyde	ND	ug/m ³ Air							U
-------------	----	-----------------------	--	--	--	--	--	--	---

Blank (B6D3002-BLK3)

Prepared & Analyzed: 04/30/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6D3002-BS1)

Prepared & Analyzed: 04/30/26

Formaldehyde	0.286	ug/ml	0.30000	95.4	79.9-120.1		
Acetaldehyde	0.289	ug/ml	0.30000	96.4	79.9-120.1		
Acetone	0.289	ug/ml	0.30000	96.3	69.9-130.1		
Propionaldehyde	0.388	ug/ml	0.30000	129	69.9-130.1		
Crotonaldehyde	0.277	ug/ml	0.30000	92.2	69.9-130.1		
Butyraldehyde	0.258	ug/ml	0.30000	86.0	69.9-130.1		
2-Butanone	0.232	ug/ml	0.30000	77.5	69.9-130.1		
Benzaldehyde	0.315	ug/ml	0.30000	105	69.9-130.1		
Valeraldehyde	0.260	ug/ml	0.30000	86.8	69.9-130.1		
Hexaldehyde	0.253	ug/ml	0.30000	84.2	69.9-130.1		

LCS Dup (B6D3002-BSD1)

Prepared & Analyzed: 04/30/26

Formaldehyde	0.290	ug/ml	0.30000	96.7	79.9-120.1	1.30	20
Acetaldehyde	0.289	ug/ml	0.30000	96.4	79.9-120.1	0.0206	20
Acetone	0.291	ug/ml	0.30000	97.2	69.9-130.1	0.862	20
Propionaldehyde	0.390	ug/ml	0.30000	130	69.9-130.1	0.526	20
Crotonaldehyde	0.278	ug/ml	0.30000	92.8	69.9-130.1	0.608	20
Butyraldehyde	0.258	ug/ml	0.30000	86.1	69.9-130.1	0.0932	20
2-Butanone	0.232	ug/ml	0.30000	77.4	69.9-130.1	0.144	20
Benzaldehyde	0.308	ug/ml	0.30000	103	69.9-130.1	2.03	20
Valeraldehyde	0.261	ug/ml	0.30000	86.9	69.9-130.1	0.137	20
Hexaldehyde	0.250	ug/ml	0.30000	83.4	69.9-130.1	1.01	20

Batch B6E0702 - TO-11A extraction

Blank (B6E0702-BLK1)

Prepared & Analyzed: 05/07/26

Formaldehyde	ND	ug/m ³ Air							U
--------------	----	-----------------------	--	--	--	--	--	--	---

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E0702 - TO-11A extraction

Blank (B6E0702-BLK1) Continued

Prepared & Analyzed: 05/07/26

Acetaldehyde	0.0303	ug/m ³ Air							
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	0.0517	ug/m ³ Air							
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

Blank (B6E0702-BLK3)

Prepared & Analyzed: 05/07/26

Formaldehyde	ND	ug/m ³ Air							U
Acetaldehyde	ND	ug/m ³ Air							U
Acetone	ND	ug/m ³ Air							U
Propionaldehyde	ND	ug/m ³ Air							U
Crotonaldehyde	ND	ug/m ³ Air							U
Butyraldehyde	ND	ug/m ³ Air							U
2-Butanone	ND	ug/m ³ Air							U
Benzaldehyde	ND	ug/m ³ Air							U
Valeraldehyde	ND	ug/m ³ Air							U
Hexaldehyde	ND	ug/m ³ Air							U

LCS (B6E0702-BS1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.288	ug/ml	0.30000		96.0	79.9-120.1			
Acetaldehyde	0.294	ug/ml	0.30000		98.1	79.9-120.1			
Acetone	0.294	ug/ml	0.30000		98.1	69.9-130.1			
Propionaldehyde	0.335	ug/ml	0.30000		112	69.9-130.1			
Crotonaldehyde	0.277	ug/ml	0.30000		92.2	69.9-130.1			
Butyraldehyde	0.260	ug/ml	0.30000		86.8	69.9-130.1			
2-Butanone	0.242	ug/ml	0.30000		80.7	69.9-130.1			
Benzaldehyde	0.310	ug/ml	0.30000		103	69.9-130.1			
Valeraldehyde	0.259	ug/ml	0.30000		86.4	69.9-130.1			
Hexaldehyde	0.251	ug/ml	0.30000		83.6	69.9-130.1			

LCS Dup (B6E0702-BSD1)

Prepared & Analyzed: 05/07/26

Formaldehyde	0.288	ug/ml	0.30000		95.9	79.9-120.1	0.125	20	
Acetaldehyde	0.294	ug/ml	0.30000		97.9	79.9-120.1	0.145	20	
Acetone	0.293	ug/ml	0.30000		97.6	69.9-130.1	0.495	20	
Propionaldehyde	0.334	ug/ml	0.30000		111	69.9-130.1	0.412	20	
Crotonaldehyde	0.277	ug/ml	0.30000		92.4	69.9-130.1	0.162	20	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E0702 - TO-11A extraction

LCS Dup (B6E0702-BSD1) Continued

Prepared & Analyzed: 05/07/26

Butyraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1	0.315	20
2-Butanone	0.244	ug/ml	0.30000	81.4	69.9-130.1	0.906	20
Benzaldehyde	0.314	ug/ml	0.30000	105	69.9-130.1	1.30	20
Valeraldehyde	0.259	ug/ml	0.30000	86.5	69.9-130.1	0.0858	20
Hexaldehyde	0.251	ug/ml	0.30000	83.6	69.9-130.1	0.102	20

Batch B6E1108 - TO-11A extraction

Blank (B6E1108-BLK1)

Prepared & Analyzed: 05/11/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	0.0325	ug/m ³ Air					
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					QB-01, U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	0.0448	ug/m ³ Air					
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

Blank (B6E1108-BLK3)

Prepared & Analyzed: 05/11/26

Formaldehyde	ND	ug/m ³ Air					U
Acetaldehyde	ND	ug/m ³ Air					U
Acetone	ND	ug/m ³ Air					U
Propionaldehyde	ND	ug/m ³ Air					U
Crotonaldehyde	ND	ug/m ³ Air					U
Butyraldehyde	ND	ug/m ³ Air					U
2-Butanone	ND	ug/m ³ Air					U
Benzaldehyde	ND	ug/m ³ Air					U
Valeraldehyde	ND	ug/m ³ Air					U
Hexaldehyde	ND	ug/m ³ Air					U

LCS (B6E1108-BS1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.288	ug/ml	0.30000	95.9	79.9-120.1
Acetaldehyde	0.293	ug/ml	0.30000	97.6	79.9-120.1
Acetone	0.294	ug/ml	0.30000	98.0	69.9-130.1
Propionaldehyde	0.351	ug/ml	0.30000	117	69.9-130.1
Crotonaldehyde	0.276	ug/ml	0.30000	92.0	69.9-130.1
Butyraldehyde	0.263	ug/ml	0.30000	87.6	69.9-130.1
2-Butanone	0.238	ug/ml	0.30000	79.4	69.9-130.1
Benzaldehyde	0.314	ug/ml	0.30000	105	69.9-130.1

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Analyte	Result	Units	Spike Level	Source Result	%REC	%REC Limits	RPD	RPD Limit	Notes
---------	--------	-------	-------------	---------------	------	-------------	-----	-----------	-------

Air Toxics by EPA Compendium Method TO-11A - Quality Control

Batch B6E1108 - TO-11A extraction

LCS (B6E1108-BS1) Continued

Prepared & Analyzed: 05/11/26

Valeraldehyde	0.261	ug/ml	0.30000	87.0	69.9-130.1
Hexaldehyde	0.254	ug/ml	0.30000	84.8	69.9-130.1

LCS Dup (B6E1108-BSD1)

Prepared & Analyzed: 05/11/26

Formaldehyde	0.290	ug/ml	0.30000	96.6	79.9-120.1	0.660	20
Acetaldehyde	0.294	ug/ml	0.30000	97.9	79.9-120.1	0.305	20
Acetone	0.293	ug/ml	0.30000	97.5	69.9-130.1	0.439	20
Propionaldehyde	0.348	ug/ml	0.30000	116	69.9-130.1	0.877	20
Crotonaldehyde	0.271	ug/ml	0.30000	90.4	69.9-130.1	1.74	20
Butyraldehyde	0.259	ug/ml	0.30000	86.2	69.9-130.1	1.60	20
2-Butanone	0.240	ug/ml	0.30000	80.1	69.9-130.1	0.778	20
Benzaldehyde	0.315	ug/ml	0.30000	105	69.9-130.1	0.358	20
Valeraldehyde	0.255	ug/ml	0.30000	84.9	69.9-130.1	2.43	20
Hexaldehyde	0.249	ug/ml	0.30000	83.1	69.9-130.1	2.06	20

Duplicate (B6E1108-DUP1)

Source: 6050635-03 Prepared & Analyzed: 05/11/26

Formaldehyde	3.32	ug/m ³ Air	3.32	0.0115	10.1	
Acetaldehyde	1.82	ug/m ³ Air	1.82	0.289	10.1	
Acetone	4.69	ug/m ³ Air	4.70	0.113	10.1	
Propionaldehyde	0.693	ug/m ³ Air	0.686	0.942	10.1	QB-01
Crotonaldehyde	ND	ug/m ³ Air	ND		10.1	U
Butyraldehyde	0.250	ug/m ³ Air	0.248	1.03	10.1	
2-Butanone	0.698	ug/m ³ Air	0.691	0.955	10.1	
Benzaldehyde	0.109	ug/m ³ Air	0.107	2.36	10.1	
Valeraldehyde	0.126	ug/m ³ Air	0.122	2.82	10.1	
Hexaldehyde	0.120	ug/m ³ Air	0.127	5.82	10.1	

Duplicate (B6E1108-DUP2)

Source: 6050635-04 Prepared & Analyzed: 05/11/26

Formaldehyde	3.29	ug/m ³ Air	3.30	0.451	10.1	
Acetaldehyde	1.80	ug/m ³ Air	1.81	0.303	10.1	
Acetone	4.60	ug/m ³ Air	4.60	0.0396	10.1	
Propionaldehyde	0.681	ug/m ³ Air	0.670	1.62	10.1	QB-01
Crotonaldehyde	ND	ug/m ³ Air	ND		10.1	U
Butyraldehyde	0.249	ug/m ³ Air	0.249	0.119	10.1	
2-Butanone	0.691	ug/m ³ Air	0.691	0.00172	10.1	
Benzaldehyde	0.111	ug/m ³ Air	0.102	8.51	10.1	
Valeraldehyde	0.117	ug/m ³ Air	0.127	7.57	10.1	
Hexaldehyde	0.119	ug/m ³ Air	0.130	8.82	10.1	

Eastern Research Group

The results in this report apply only to the samples analyzed in accordance with the chain of custody document. This analytical report must be reproduced in its entirety.



CERTIFICATE OF ANALYSIS

U.S. Environmental Protection Agency, Region 6

707 N. Robinson, PO Box 1677

Oklahoma City, OK 73101-1677

ATTN: Mr. Bryan Sims

PHONE: (405) 702-0100

FILE #: 3508.00

REPORTED: 06/13/26 10:10

SUBMITTED: 04/08/26 to 05/06/26

AQS SITE CODE: 40 109 1037

SITE CODE: OCOK

Notes and Definitions

U	Under Detection Limit
TT	Transport temperature is out of specs.
QB-01	Analyte exceeds method blank criteria
O-05	This sample was extracted outside of the EPA recommended holding time.
LB	Lab blank value above acceptable limit.
DNPH	DNPH fails QAPP criteria. Results for this sample should be considered an estimate and may be biased low.
D-F	Duplicate exceeds DQO criteria.
D	This result obtained by dilution.
CE	Not reportable due to a co-eluting compound.
ND	Analyte NOT DETECTED
NR	Not Reported
MDL	Method Detection Limit
RPD	Relative Percent Difference

Note: Breakthrough cartridges were not submitted or analyzed as specified by Compendium Method TO-11A. This deviation from the method has been approved by the US EPA, under the contract for UATMP, NATTS, CSATAM, PAMS and NMOC Support (US EPA Contract No. EP-D-14-030).

Note: This test is accredited under the 2016 TNI Standard.